

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 1 of 7

OOS Number: OOS- 261242

Phase I

Section A – Test Details					
Initiator:	Simin Mohammad	Test Date:	14-May-2026	Date Initiated:	26-May-2026
Date of Incident:	26-May-2026	Test Name:	Environmental Monitoring	Sample ID:	ETX-260526-0395
Name of Analyst who Performed the Test:	Simin Mohammad (Weekly Active air Sampling Plate Setup)	Sample / Active Name	EM SMO 115B Air 14May2026	Lot #:	EM SMO 115B Air 14May2026
	Sophia Santamaria (Weekly Active air Sampling Plate Readers)	Dosage Form	plate		
Description of Incident:	The CFU count for the environmental monitoring plate exceeded the action level.		☐ Client Care notified of OOS ☐ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS		
	SOP / Test Method # MICRO-SOP-2	Effective Date: 05-AUG-2025	SOP / Test Method Rev: 15	Limits/Specification: Action level: >10CFU/plate	
Manager notified	Manager/designee name: Kathan Parikh	Date: 26-May-2026	Quality Assurance Department notified:	Yes ☐ No ☐	
Section B – Investigation and Root Cause Analysis					
Personnel Discrepancy			Comments		
Analyst interviewed?	☒ Yes ☐ No ☐ N/A		Yes, analysts Simin Mohammad, & Sophia Santamaria were comprehensively interviewed.		
Correct samples analyzed?	☐ Yes ☐ No ☒ N/A		N/A		
Correct SOP / Test Method used/followed?	☒ Yes ☐ No ☐ N/A		Yes, as per MICRO-SOP-2		
Correct sample collection/technique?	☒ Yes ☐ No ☐ N/A		Yes, as per MICRO-SOP-2		
Are the Test Result Records Attached & Calculations Verified and Correct?	☒ Yes ☐ No ☐ N/A		Yes, Information is available in EagleTrax under ETX-260526-0395		
Is the analyst qualified to perform the test?	☒ Yes ☐ No ☐ N/A		Yes, the analysts are trained and qualified by quality to perform the test		
Is formula worksheet provided and verified?	☐ Yes ☐ No ☒ N/A		N/A		
Does the processing method/method set conform to the required specification?	☐ Yes ☐ No ☒ N/A		N/A		
Was the sample stored appropriately?	☒ Yes ☐ No ☐ N/A		Yes, as per MICRO-SOP-2		
Material Discrepancy			Lot Number	Expiry Date	
Is the integrity of the sample(s) questionable?	☐ Yes ☐ No ☒ N/A				
Was the sample transported appropriately?	☐ Yes ☐ No ☒ N/A				
Are the other test results from this run under investigation?	☐ Yes ☐ No ☒ N/A				

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 2 of 7

OOS Number: OOS- 261242

Appropriate/correct glassware used?	☐ Yes ☐ No ☒ N/A		
Are reagents stored in proper condition and not expired?	☒ Yes ☐ No ☐ N/A	TSA Plate: 1011772520	TSA Plate: 11-Feb-2027
Standard Used	N/A	N/A	N/A
Solvent Used	N/A	N/A	N/A
Is the standard the source of the root cause?	☐ Yes ☐ No ☒ N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐ Yes ☐ No ☒ N/A		
Equipment Discrepancy		Equipment ID	Calibration Due Date
Was the instrument set up correctly?	☒ Yes ☐ No ☐ N/A	Please see below	Please see below
Is the clean room facility certified?	☒ Yes ☐ No ☐ N/A	CR115 (E001737)	DEC-2026
System Performance Check (if applicable)?	☐ Yes ☐ No ☒ N/A	N/A	N/A
System Suitability RSD < 2% (PDA)?	☐ Yes ☐ No ☒ N/A	N/A	N/A
System Suitability RSD < 10% (QDA, CAD)?	☐ Yes ☐ No ☒ N/A	N/A	N/A
Do the environmental monitoring results conform to required standards?	☐ Yes ☐ No ☒ N/A	N/A	N/A
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☐ Yes ☐ No ☒ N/A	N/A	N/A
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☒ Yes ☐ No ☐ N/A	Incubator E001034 (Sensor E001501)	Aug-2026 Feb-2027
		Incubator E001031 (Sensor E001505)	Aug-2026 Feb-2027

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 3 of 7

OOS Number: OOS- 261242

Are chromatogram/spectra/raw data correct?		☐ Yes ☐ No ☐ N/A	
Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)			
☒ N/A ☐ Re-injection of sample vial ☐ Re-injection of original standard vial ☐ Re-vialing of original working standard solution		☐ Re-vialing of original working sample solution ☐ Re-dilution of a working sample solution from the original stock ☐ Re-dilution of a working standard solution from the original stock standard ☐ Other: <u>N/A SMO 18-Aug-2026</u>	
Section D – Phase I Summary			
Phase I Summary: The analyst involved in the Active air Sampling plate setup, Simin Mohammad, and the analysts involved in reading the plate Sophia Sanatamaria, were interviewed comprehensively. Their answers are recorded throughout this document. The EM plates were stored in compliance with the supplier's recommendations, and their integrity was visually inspected prior to use. Furthermore, the plates were confirmed to be within their valid expiration dates. All the supplies were thoroughly disinfected according to MICRO-SOP-9 Cleaning and Disinfecting Procedure for Microbiology. The functionality of both incubators were verified through a review of data obtained from our comprehensive in-house continuous monitoring system. Active air Sampling was performed by analyst Simin Mohammad during weekly environmental monitoring processing, in ISO 8 115, and ISO 7 115A, and ISO 7 115B on 14 May 2026 as per MICRO-SOP-2 Environmental Monitoring of the Cleanroom Facility. The plates were initially incubated at a temperature of 30–35°C in incubator E001031 for a minimum duration of 48 hours, commencing on 14 May 2026. Following completion of minimum of 48 hours of incubation on 18 May 2026, the plates were further incubated for minimum of 5 days, with the incubation concluding on 26 May 2026. Please see Table 1 for detailed information on the observations during respective incubations. Based on the observations in Table 1, since the CFU count exceeded the action level for the one of the Active air Sampling plates (115B), the plate was submitted for Microbial Identification under ETX-260526-0395. The colonies were identified as Candida orthopsilosis (Budding yeast) To observe if the organisms identified, were transient in nature or recurring, weekly monitoring environmental monitoring plates for the clean room, were bracketed to include the week before testing, the week of testing and week after testing as detailed in Table 2 (please see attached).			

OOS Number: OOS- 261242

Phase I Summary (continued):

All H2O2 indicators passed.

OOS Number: OOS- 261242

Phase I Summary (continued):

N/A SMO 18-AUG-2026

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 7 of 7

VERSION HISTORY

Laboratory OOS Investigation Form
(v11.1)

Approved: 10-JUL-2026 | Effective: 10-JUL-2026 | Retired: N/A
Minor up version to fillable PDF

Laboratory OOS Investigation Form
(v11.0)

Approved: 30-SEP-2025 | Effective: 30-SEP-2025 | Retired: 10-JUL-2026
Refer to original approved version on ETX

ELECTRONIC SIGNATURES

Approvals

Preeti Purwaha (DPP) Director of Quality Assurance ppurwaha@eagleanalytical.com <i>Comment: Author review completed.</i>	Document Author I am the author of this document. Signed 05:05:09 PM UTC 09-JUL-2026
Minh-Anh (Alison) Eng (AE) Quality Site Manager meng@eagleanalytical.com <i>Comment: approved. AE10JUL2026</i>	Technical Manager Approval I have reviewed and approve this Minor Revision. Signed 05:32:28 PM UTC 10-JUL-2026
Jacqueline Esqueda (JE) VP of Quality and Regulatory Comp jesqueda@eagleanalytical.com <i>Comment: Approved</i>	Quality Approval I have reviewed and approve this Minor Revision. Signed 05:41:18 PM UTC 10-JUL-2026

CURRENT AUTHORS

Preeti Purwaha (DPP)	Director of Quality Assurance
--------------------------------------	-------------------------------

CUSTOM FIELDS

Location	Houston (Corporate)
ETX ID	3.100.019.F01

LINKS/ATTACHMENTS

Document CORP-SOP-13: Out-of-Specification (OOS) - Test Failure Investigation (10.0) [effective]