

Table 1: Read Dates & Incubation Observation

ETX Submission ID	Set-up Analyst & Date	Sampling Site & Location	Plate Reading Analyst ($\geq$ 48H)	CFUs Observed after 48 Hour Incubation at 30–35°C (E001031)	Plate Reading Analyst (NLT 5 days)	CFUs Observed after NLT 5-day Incubation at 20–25°C (E001034)	Microbial Identification
ETX-260526-0461	SMO 08MAY 2026	Active air sampling plate	MC	0 CFU	SAS	11 CFU in 115B ISO 7 sampling plate	Talaromyces purpurogenus

Table 2: Environmental Monitoring for Analyst & Cleanroom

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMM YYYY)	Analyst (Initials)	Day /Week(s)	Observation*	Environmental Monitoring Plate ETX ID	Microbial ID	Notes
Weekly Active Air Sampling Bracketing 115								
Active Air Sampling of Cleanrooms	Weekly	29APR 2026	SMO	Week (Before Testing Date)	1 CFU for 115 ISO 8	ETX-260506-0280	Micrococcus luteus	None
Active Air Sampling of Cleanrooms	Weekly	08MAY 2026	SMO	Week (On of Testing Date)	1 CFU in 115A ISO 7 and 11 CFU in 115B ISO 7	ETX-260518-0249 ETX-260518-0250	Staphylococcus aureus Penicillium decumbens Cladosporium tenuissimum Cladosporium langeronii Cladosporium halotolerans	None
Active Air Sampling of Cleanrooms	Weekly	14MAY 2026	SMO	Week (On or After Testing Date)	1 CFU in 115 ISO 8 and 11 CFU in 115B ISO 7	ETX-260526-0391 ETX-260526-0395	Gram (+) cocci Candida orthopsilosis Budding yeast	None
Surface Sampling of Anteroom and Cleanroom Bracketing 115								
Active Air Sampling of Cleanrooms	Weekly	29APR 2026	SMO	Week (Before Testing Date)	No growth	N/A	N/A	None
Surface Sampling of Cleanrooms	Weekly	08MAY 2026	SMO	Week (On of Testing Date)	No growth	N/A	N/A	None
Surface Sampling of Cleanrooms	Weekly	14MAY 2026	SMO	Week (On or After Testing Date)	1 CFU on table in ISO 8 115 and 1 CFU on the cart in 115 ISO 8	ETX-260526-0424 ETX-260526-0425	Gram (+) rods Staphylococcus capitis	None