

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			
Page 1 of 7			

OOS Number: OOS-

Phase I

Section A – Test Details					
Initiator:	Simin Mohammad	Test Date:	07-May-2026	Date Initiated:	18-May-2026
Date of Incident:	18-May-2026	Test Name:	Environmental Monitoring	Sample ID:	ETX-260518-0254
Name of Analyst who Performed the Test:	Simin Mohammad (Weekly Active Air Sampling Analyst)	Sample / Active Name	EM SMO 116A Air 07MAY2026	Lot #:	EM SMO 116A Air 07MAY2026
	Sophia Santamaria (Weekly Active Air Sampling Plate Reader)		Dosage Form		
Description of Incident:	The CFU count for environmental monitoring plate exceeded the action level.		☐ Client Care notified of OOS ☐ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS		
	SOP / Test Method # MICRO-SOP-2	Effective Date: 23-Jul-2026	SOP / Test Method Rev: 16	Limits/Specification: Action level: >= 10 CFU/Plate	
Manager notified	Manager/designee name: Kathan Parikh	Date: 18-May-2026	Quality Assurance Department notified:	Yes ☐ No ☐	
Section B – Investigation and Root Cause Analysis					
Personnel Discrepancy			Comments		
Analyst interviewed?	☒ Yes ☐ No ☐ N/A		Yes, analysts Simin Mohammad and Sophia		
Correct samples analyzed?	☐ Yes ☐ No ☒ N/A		N/A		
Correct SOP / Test Method used/followed?	☒ Yes ☐ No ☐ N/A		Yes, as per MICRO-SOP-2		
Correct sample collection/technique?	☒ Yes ☐ No ☐ N/A		Yes, as per MICRO-SOP-2		
Are the Test Result Records Attached & Calculations Verified and Correct?	☒ Yes ☐ No ☐ N/A		Yes, information is available in EagleTrax under ETX-260518-0254.		
Is the analyst qualified to perform the test?	☒ Yes ☐ No ☐ N/A		Yes, the analysts are trained and qualified by quality to perform the test		
Is formula worksheet provided and verified?	☐ Yes ☐ No ☒ N/A		N/A		
Does the processing method/method set conform to the required specification?	☐ Yes ☐ No ☒ N/A		N/A		
Was the sample stored appropriately?	☒ Yes ☐ No ☐ N/A		Yes, as per MICRO-SOP-2		
Material Discrepancy			Lot Number	Expiry Date	
Is the integrity of the sample(s) questionable?	☐ Yes ☐ No ☒ N/A				
Was the sample transported appropriately?	☐ Yes ☐ No ☒ N/A				
Are the other test results from this run under investigation?	☐ Yes ☐ No ☒ N/A				

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 2 of 7

OOS Number: OOS-

Appropriate/correct glassware used?	☐ Yes ☐ No ☒ N/A		
Are reagents stored in proper condition and not expired?	☒ Yes ☐ No ☐ N/A	TSA Plate: 1011543730	TSA Plate: 29-SEP-2026
Standard Used	N/A	N/A	N/A
Solvent Used	N/A	N/A	N/A
Is the standard the source of the root cause?	☐ Yes ☐ No ☒ N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐ Yes ☐ No ☒ N/A		
Equipment Discrepancy		Equipment ID	Calibration Due Date
Was the instrument set up correctly?	☒ Yes ☐ No ☐ N/A	Please see below	Please see below
Is the clean room facility certified?	☒ Yes ☐ No ☐ N/A	CR116 (E001738)	Dec-2026
System Performance Check (if applicable)?	☐ Yes ☐ No ☒ N/A	N/A	N/A
System Suitability RSD < 2% (PDA)?	☐ Yes ☐ No ☒ N/A	N/A	N/A
System Suitability RSD < 10% (QDA, CAD)?	☐ Yes ☐ No ☒ N/A	N/A	N/A
Do the environmental monitoring results conform to required standards?	☐ Yes ☐ No ☒ N/A	N/A	N/A
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☐ Yes ☐ No ☒ N/A	N/A	N/A
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☒ Yes ☐ No ☐ N/A	Incubator E001034 (Sensor E001501) Incubator E001031 (Sensor E001505)	Aug 2026 / Feb 2027 Aug 2026 / Feb 2027

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 3 of 7

OOS Number: OOS-

Are chromatogram/spectra/raw data correct?				☐ Yes	☐ No	☒ N/A	N/A	N/A
Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)								
☒ N/A		☐ Re-vialing of original working sample solution						
☐ Re-injection of sample vial		☐ Re-dilution of a working sample solution from the original stock						
☐ Re-injection of original standard vial		☐ Re-dilution of a working standard solution from the original stock standard						
☐ Re-vialing of original working standard solution		☐ Other: N/A SMO 18-Aug-2026						
Section D – Phase I Summary								
Phase I Summary:								
The analyst involved in the Active air Sampling plate setup, Simin Mohammad, and the analysts involved in reading the plate Sophia Santamaria, were interviewed comprehensively. Their answers are recorded throughout this document.								
The EM plates were stored in compliance with the supplier's recommendations, and their integrity was visually inspected prior to use. Furthermore, the plates were confirmed to be within their valid expiration dates. All the supplies were thoroughly disinfected according to MICRO-SOP-9 Cleaning and Disinfecting Procedure for Microbiology. The functionality of both incubators were verified through a review of data obtained from our comprehensive in-house continuous monitoring system.								
Active air Sampling was performed by analyst Simin Mohammad during weekly environmental monitoring processing, in ISO 8 116, and ISO 7 116A, and ISO 7 116B on 07 May 2026 as per MICRO-SOP-2 - Environmental Monitoring of the Cleanroom Facility.								
The plates were initially incubated at a temperature of 30–35°C in incubator E001031 for a minimum duration of 48 hours, commencing on 07 May 2026. Following completion of minimum of 48 hours of incubation on 11 May 2026, the plates were further incubated for minimum of 5 days, with the incubation concluding on 18 May 2026. Please see Table 1 for detailed information on the observations during respective incubations.								
Based on the observations in Table 1, since the CFU count exceeded the action level for the one of the Active air Sampling plates (116A), the plate was submitted for Microbial Identification under ETX-260518-0254. The colonies were identified as Kocuria palustris (Gram (+) cocci)								
To observe if the organisms identified, were transient in nature or recurring, weekly environmental monitoring plates for the clean room were bracketed to include the week before testing, the week of testing and week after testing as detailed in Table 2 (please see attached).								
Environmental Monitoring Summary: Weekly environmental sampling for the previous week showed no growth for the active air and weekly surface sample. Week of testing showed 748 CFUs in 116A ISO 7 and 1 CFU in 116B ISO 7 for active air identified as Kocuria palustris (Gram (+) cocci). Lastly, the following week of testing showed 6 CFUs in ISO 8 116 for active air, 1 CFU on cart in ISO 8 116, and 2 CFUs on floor in 116A ISO 7 identified as Gram (+) cocci, Gram-variable coccobacilli, Gram (+) cocci, Sporosarcina psychrophila, and Bacillus pumilus.								
During the interview with the analyst, they indicated that no obvious abnormalities or deviations in the testing procedure were observed. All the samples were thoroughly disinfected prior to testing. Moreover, The CR116 was thoroughly cleaned and prepared before initiating the testing as per MICRO-SOP-2 (Environmental Monitoring of the Clean Room Facility), and MICRO-SOP-9 (Cleaning and Disinfecting Procedure for Microbiology).								
Prior to the event, monthly cleaning and disinfection of the outermost ISO 8 Anteroom (Suite 116), the middle ISO 7 Buffer room (Suite 116A), the innermost ISO 7 clean room (Suite 116B), and its containing ISO 5 Biosafety Cabinets were performed on 26-Apr-2026 by analysts Rey Estrada and Tamiru Kotisso as per MICRO-SOP-9 (Cleaning and Disinfecting Procedure for Microbiology). It was documented that all H2O2 indicators passed, confirming the efficient pre-event cleaning and established state of control of Suite 116. Furthermore, subsequent monthly cleaning was performed on 31-May-2026 by analysts Tamiru Kotisso and Cuong Du with all H2O2 indicators passing, confirming ongoing facility control. Additionally, cleaning and disinfecting was performed both prior to and after the testing process as per MICRO-SOP-9.								

OOS Number: OOS-

Phase I Summary (continued):

It is also important to note that no samples processed in Suite 116 for the week of testing (03-May-2026 to 09-May-2026) failed testing.

Based on the available investigation findings, no specific analyst-related, procedural, equipment-related, or other laboratory-related cause was identified. The excursion appears to have been transient and non-recurring, and no assignable root cause was established. It is to be noted that the growth observed on the weekly active air and weekly surface sampling plates for the day of testing did not follow a trend, indicating that the contamination was transient in nature and that routine daily disinfection procedures were effective in eliminating the contamination. Furthermore, no trend was observed in the clean room's previous weekly EM data, therefore, no preventive and corrective actions are deemed necessary at this time.

OOS Number: OOS-

Phase I Summary (continued):

N/A SMO 18-AUG-2026

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 6 of 7

OOS Number: OOS-

Has root cause been identified as a possible laboratory error? <i>Note: Root cause must be conclusive, verifiable, and well-documented.</i> ☒ Yes ☐ If <u>retesting</u> is required , complete Phase II of the OOS (initiate Form 3.100.019.F02). ☐ No, the OOS result is valid.	
Reviewer comment(s):	
Prepared by (Print): Simin Mohammad	Signature/Date: Simin Mohammad Digitally signed by Simin Mohammad Date: 2026.09.22 16:03:12 -05'00'
Lab Manager/Designee (Print): Qiyue Chen	Signature/Date: Qiyue Chen Digitally signed by Qiyue Chen DN: OU=Eagle Analytical Services, O=Strerie Lab, CN=Qiyue Chen, E=qchen@eagleanalytical.com Location: Houston, TX Date: 2026.09.23 08:19:43-05'00' Foxit PDF Editor Version: 13.2.6
Section E – Quality Disposition	
Can investigation be closed? ☐ Yes ☐ Original result is valid. ☐ No, initiate Form 3.100.019.F02.	
Approved by QA Representative (Print):	Signature/Date:

	CORP-FORM-21 v11.1	Category: Form Corporate Name: Laboratory OOS Investigation Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.019.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 19-Aug-2026 at 8:54:35 PM UTC			Page 7 of 7

VERSION HISTORY

Laboratory OOS Investigation Form (v11.1)

Approved: 10-JUL-2026 | Effective: 10-JUL-2026 | Retired: N/A
Minor up version to fillable PDF

Laboratory OOS Investigation Form (v11.0)

Approved: 30-SEP-2025 | Effective: 30-SEP-2025 | Retired: 10-JUL-2026
Refer to original approved version on ETX

ELECTRONIC SIGNATURES

Approvals

Preeti Purwaha (DPP) Director of Quality Assurance ppurwaha@eagleanalytical.com <i>Comment: Author review completed.</i>	Document Author I am the author of this document. Signed 05:05:09 PM UTC 09-JUL-2026
Minh-Anh (Alison) Eng (AE) Quality Site Manager meng@eagleanalytical.com <i>Comment: approved. AE10JUL2026</i>	Technical Manager Approval I have reviewed and approve this Minor Revision. Signed 05:32:28 PM UTC 10-JUL-2026
Jacqueline Esqueda (JE) VP of Quality and Regulatory Comp jesqueda@eagleanalytical.com <i>Comment: Approved</i>	Quality Approval I have reviewed and approve this Minor Revision. Signed 05:41:18 PM UTC 10-JUL-2026

CURRENT AUTHORS

Preeti Purwaha (DPP)	Director of Quality Assurance
--------------------------------------	-------------------------------

CUSTOM FIELDS

Location	Houston (Corporate)
ETX ID	3.100.019.F01

LINKS/ATTACHMENTS

Document CORP-SOP-13: Out-of-Specification (OOS) - Test Failure Investigation (10.0) [effective]

Table 1: Read Dates & Incubation Observation

ETX Submission ID	Set-up Analyst & Date	Sampling Site & Location	Plate Reading Analyst (≥ 48H)	CFUs Observed after 48 Hour Incubation at 30–35°C (E001031)	Plate Reading Analyst (NLT 5 days)	CFUs Observed after NLT 5-day Incubation at 20–25°C (E001034)	Microbial Identification
ETX-260518-0254	SMO 07MAY 2026	Active air sampling plate	SAS	504 CFUs on active air sample plate	SAS	748 CFUs on active air sample plate	Kocuria palustris, Gram (+) cocci

Table 2: Environmental Monitoring Plates for Analyst and Cleanroom Bracketing

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMMM YYYY)	Analyst (Initials)	Day /Week(s)	Observation*	Environmental Monitoring Plate ETX ID	Microbial ID	Notes
Weekly Active Air Sampling Bracketing 116								
Active Air Sampling of Cleanrooms	Weekly	29APR 2026	SMO	Week Before Testing Date	No growth	N/A	N/A	None
Active Air Sampling of Cleanrooms	Weekly	07MAY 2026	SMO	Week of Testing Date	748 CFUs in ISO 7 116A and 1 CFU in ISO 7 116B	ETX-260518-0254 ETX-260518-0256	Kocuria palustris, Gram (+) cocci	None
Active Air Sampling of Cleanrooms	Weekly	15MAY 2026	SMO	Week After Testing Date	6 CFUs in ISO 8 116	ETX-260526-0428	Gram (+) cocci Gram-variable coccobacilli	None
Surface Sampling of Anteroom and Cleanroom Bracketing 116								
Surface Sampling of Cleanrooms	Weekly	29APR 2026	SMO	Week Before Testing Date	No growth	N/A	N/A	None
Surface Sampling of Cleanrooms	Weekly	07MAY 2026	SMO	Week of Testing Date	No growth	N/A	N/A	None
Surface Sampling of Cleanrooms	Weekly	15MAY 2026	SMO	Week After Testing Date	1 CFU on cart in ISO 8 116 and 2 CFUs on floor in 116A ISO 7	ETX-260526-0442 ETX-260526-0448	Gram (+) cocci Sporosarcina psychrophila Bacillus pumilus	None