

Table 1: Read Dates & Incubation Observation

| ETX Submission ID | Set-up Analyst & Date | Sampling Site & Location  | Plate Reading Analyst ( $\geq$ 48H) | CFUs Observed after 48 Hour Incubation at 30–35°C (E001031) | Plate Reading Analyst (NLT 5 days) | CFUs Observed after NLT 5-day Incubation at 20–25°C (E001034) | Microbial Identification |
|-------------------|-----------------------|---------------------------|-------------------------------------|-------------------------------------------------------------|------------------------------------|---------------------------------------------------------------|--------------------------|
| ETX-260526-0395   | SMO<br>14MAY<br>2026  | Active air sampling plate | SAS                                 | 11 CFUs on active air sample plate                          | SAS                                | 11 CFUs on active air sample plate                            | Candida orthopsilosis    |

Table 2: Environmental Monitoring for Analyst & Cleanroom

| Environmental Monitoring (EM) Sampling Site               | Frequency | Date (DDMMM YYYY) | Analyst (Initials) | Day /Week(s)             | Observation*                                               | Environmental Monitoring Plate ETX ID | Microbial ID                                                                                                                       | Notes |
|-----------------------------------------------------------|-----------|-------------------|--------------------|--------------------------|------------------------------------------------------------|---------------------------------------|------------------------------------------------------------------------------------------------------------------------------------|-------|
| Weekly Active Air Sampling Bracketing 115                 |           |                   |                    |                          |                                                            |                                       |                                                                                                                                    |       |
| Active Air Sampling of Cleanrooms                         | Weekly    | 08MAY 2026        | SMO                | Week Before Testing Date | 1 CFU in ISO 7 115A and 11 CFUs in ISO 7 115B              | ETX-260518-0249<br>ETX-260518-0250    | Staphylococcus aureus<br>Penicillium decumbens<br>Cladosporium tenuissimum<br>Cladosporium langeronii<br>Cladosporium halotolerans | None  |
| Active Air Sampling of Cleanrooms                         | Weekly    | 14MAY 2026        | SMO                | Week of Testing Date     | 1 CFU in ISO 8 115 and 11 CFUs in ISO 7 115B               | ETX-260526-0391<br>ETX-260526-0395    | Gram (+) cocci<br><br>Candida orthopsilosis                                                                                        | None  |
| Active Air Sampling of Cleanrooms                         | Weekly    | 22MAY 2026        | SMO                | Week After Testing Date  | No growth                                                  | N/A                                   | N/A                                                                                                                                | None  |
| Surface Sampling of Anteroom and Cleanroom Bracketing 115 |           |                   |                    |                          |                                                            |                                       |                                                                                                                                    |       |
| Surface Sampling of Cleanrooms                            | Weekly    | 08MAY 2026        | SMO                | Week Before Testing Date | No growth                                                  | N/A                                   | N/A                                                                                                                                | None  |
| Surface Sampling of Cleanrooms                            | Weekly    | 14MAY 2026        | SMO                | Week of Testing Date     | 1 CFU on table in 115 ISO 8 and 1 CFU on cart in 115 ISO 8 | ETX-260526-0424<br>ETX-260526-0425    | Gram (+) rods<br>Staphylococcus capitis                                                                                            | None  |
| Surface Sampling of Cleanrooms                            | Weekly    | 22MAY 2026        | SMO                | Week After Testing Date  | No growth                                                  | N/A                                   | N/A                                                                                                                                | None  |