**Table 1: Information for ETX-260902-0505 under investigation**

| Processing Analyst | **Reading Analyst** | Sample ID | Related Microbial ID | Media with microbial growth | Microbial ID |
|---|---|---|---|---|---|
| Alex Saravia | Elysse Nioupin | ETX-260902-0505 | ETX-260910-0290 | 1 x 100mL TSB bottle | *Microbacterium sp. PM5* (Gram (+) rods) |

Table 2: Environmental Monitoring from Processing Performed on 04Sep26

| **Environmental Monitoring (EM) Sampling Site** | **Frequency** | **Date (DDMMMYY)** | **Analyst (Initials)** | **Day /Week(s)** | **Observation*** | **Environmental Monitoring Plate ETX ID** | **Microbial ID** | **Notes** |
|---|---|---|---|---|---|---|---|---|
| **Personnel EM** | | | | | | | | |
| Personnel (Left Touch and Right Touch) | Daily | 03Sep26 | ES | Date Before Testing | No growth | N/A | N/A | None |
| Personnel (Left Touch and Right Touch) | Daily | 04Sep26 | ES | Date of Testing | No growth | N/A | N/A | None |
| Personnel (Left Touch and Right Touch) | Daily | 08Sep26 | ES | Date After Testing | No growth | N/A | | None |
| **Biological Safety Cabinet EM for E001316** | | | | | | | | |
| Surface Sampling of ISO 5 (4 locations) | Daily | 03Sep26 | ES | Date Before Testing | No growth | N/A | N/A | None |
| Surface Sampling of ISO 5 (4 locations) | Daily | 04Sep26 | ES | Date of Testing | No growth | N/A | N/A | None |
| Surface Sampling of ISO 5 (4 locations) | Daily | 08Sep26 | ES | Date After Testing | No growth | N/A | | None |
| Settling Sampling of ISO 5 (2 locations) | Daily | 03Sep26 | ES | Date Before Testing | No growth | N/A | N/A | None |
| Settling Sampling of ISO 5 (2 locations) | Daily | 04Sep26 | ES | Date of Testing | No growth | N/A | N/A | None |
| Settling Sampling of ISO 5 (2 locations) | Daily | 08Sep26 | ES | Date After Testing | No growth | N/A | | None |
| **Weekly Active Air Sampling of Cleanroom Suite 114 (E001736) with Processing BSC** | | | | | | | | |
| Active Air Sampling of Cleanrooms | Weekly | 04Sep26 | ISS | Week of Testing | 2 CFUs (ISO 8 114) | ETX-260914-0487 | *Corynebacterium ureicelerivorans* *Mycobacterium grossiae* | None |
| **Weekly Surface Sampling of Cleanroom Suite 114 (E001736) with Processing BSC** | | | | | | | | |
| Surface Sampling of Cleanrooms | Weekly | 04Sep26 | ISS | Week of Testing | No growth | N/A | N/A | None |