

Table 1: Read Dates & Incubation Observation

ETX Submission ID	Set-up Analyst & Date	Sampling Site & Location	Plate Reading Analyst (≥ 48H)	CFUs Observed after 48 Hour Incubation at 30–35°C (E001031)	Plate Reading Analyst (NLT 5 days)	CFUs Observed after NLT 5-day Incubation at 20–25°C (E001034)	Microbial Identification
ETX-260518-0273	CGS 11MAY 2026	Surface sampling plate	MC	0	SAS	1 CFU on surface #1 sampling plate	N/A

Table 2: Environmental Monitoring Plates for Analyst and Cleanroom Bracketing

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMM YYYY)	Analyst (Initials)	Day /Week(s)	Observation*	Environmental Monitoring Plate ETX ID	Microbial ID	Notes
Personnel EM Bracketing								
Personal (Left Touch and Right Touch)	Daily	08MAY 2026	CGS	Date Before Testing	No growth	N/A	N/A	None
Personal (Left Touch and Right Touch)	Daily	11MAY 2026	CGS	Date of Testing	No growth	N/A	N/A	None
Personal (Left Touch and Right Touch)	Daily	12MAY 2026	CGS	Date After Testing	No growth	N/A	N/A	None
Biological Safety Cabinet EM Bracketing Biological Safety Cabinet (BSC) 1309								
Surface Sampling of ISO 5 (4 locations)	Daily	09MAY 2026	VV	Date Before Testing	No growth			None
Surface Sampling of ISO 5 (4 locations)	Daily	11MAY 2026	CGS, DT, GA, VV	Date of Testing	1 CFU on S1 for CGS	ETX-260518-0273	N/A White, shiny, hardened, smooth area embeded on agar	None
Surface Sampling of ISO 5 (4 locations)	Daily	12MAY 2026	PG, DT, HS, ELB, GA, VV	Date After Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	09MAY 2026	VV	Date Before Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	11MAY 2026	CGS, DT, GA, VV	Date of Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	12MAY 2026	PG, DT, HS, ELB, GA, VV	Date After Testing	No growth	N/A	N/A	None
Weekly Active Air Sampling Bracketing 117								
Active Air Sampling of Cleanrooms	Weekly	29APR 2026	SMO	Week (On of Testing Date)	No growth	N/A	N/A	None
Active Air Sampling of Cleanrooms	Weekly	07MAY 2026	SMO	Week (On or After Testing Date)	2 CFU for 117 Air	ETX-260518-0263	Gram (+) cocci Gram (+) pleomorphic rods	None
Surface Sampling of Anteroom and Cleanroom Bracketing 117								
Surface Sampling of Cleanrooms	Weekly	29APR 2026	SMO	Week (On of Testing Date)	No growth	N/A	N/A	None
Surface Sampling of Cleanrooms	Weekly	07MAY 2026	SMO	Week (On or After Testing Date)	No growth	N/A	N/A	None