

	CORP-FORM-5 v9.1	Category: Form Corporate Name: Nonconformance Report Form	Effective 10-JUL-2026
Location Houston (Corporate)		ETX ID 3.100.007.F01	
Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 21-Aug-2026 at 3:28:56 PM UTC			Page 1 of 2

NCR#: N 26487

IDENTIFICATION	
Initiator: Qiyue Chen	Date: 18-Sep-2026
Department: Microbiology	Sample Name/Material Description:
Equipment ID #: E001230	Scan RDI Sterility Testing Samples (Session 17 Sep 2026 - 3)
Sample ID #: See Desc (3 Submissions)	Sample/Material Lot #: N/A (Multiple Client Lots)
NONCONFORMITY DESCRIPTION (Who, What, When, Where, Why, & How)	
Nonconformance Description: On 17-Sep-2026, during routine Scan RDI sterility testing operations, an unexpected software crash occurred on the Scan RDI instrument (Equipment ID: E001230) while running session '17 Sep 2026 - 3'. Following the crash, third-shift analysts and Microbiology Laboratory Supervisor Robin Seymour attempted to recover the session; however, the instrument displayed system error messages stating 'Unable to Create New Session Unknown Error' and 'Unable to Load Session Unknown Error', preventing completion of the verification and reading process. On 18-Sep-2026, Eagle IT Coordinator Kevin Torres conducted a technical assessment of instrument E001230 and confirmed that the session file was corrupted, exporting as only 7 KB, and the corresponding raw acquisition dataset within the D:\ directory was completely empty and unrecoverable. A total of three (3) sterility test sample submissions were actively processed in this corrupted session and impacted by data loss: ETX-260915-0653, ETX-260916-0044, and ETX-260916-0374. No valid test results could be generated or verified for these submissions from the corrupted run. Work Order WO-260470 was promptly issued to Engineering/IT to formally investigate the software failure and initiate technical remediation.	
CAUSE(S) ANALYSIS	
Description of cause(s): Scan RDI software crash and session file corruption on instrument E001230, resulting in unrecoverable raw acquisition data in the D:\ directory. Category (select all that apply): ☒ Equipment/System ☐ Process/Method ☐ Personnel ☐ External Phenomena ☐ Other	
INTERIM CORRECTIVE ACTION(S)	
Description of corrective action(s): - Work Order WO-260470 was issued to Engineering and IT to investigate and resolve the software crash, clear corrupted temporary files, and evaluate database integrity on instrument E001230. - Instrument E001230 was taken out of service for live sample testing pending completion and sign-off of WO-260470. - Sample re-testing was coordinated for the three affected submissions: ETX-260916-0374 was received and in-process with the third shift for re-testing on an alternate qualified Scan RDI unit; client notifications and rerun procedures were initiated for ETX-260915-0653 and ETX-260916-0044 per SOP.	
QUALITY EVALUATION	
Was action effective in addressing the root cause of the nonconformity? ☒ Yes ☐ No ☐ N/A CAR required: ☐ YES ☐ NO If yes, CAR #: _____ If yes, explain how effectiveness was determined: Additional Comments: Impact is isolated to the 3 sample submissions processed in the corrupted session (ETX-260915-0653, ETX-260916-0044, ETX-260916-0374). No impact to other testing sessions or alternate operational Scan RDI instruments (E002017, E002225). Rerun procedures were immediately initiated. Instrument E001230 remains under observation and repair under WO-260470.	
ACKNOWLEDGEMENT	
Prepared By (Print): Qiyue Chen	Signature Date
Reviewed By (Print): Robin Seymour	Signature Date



CORP-FORM-5
v9.1

Category: Form Corporate
Name: Nonconformance Report Form

Effective
10-JUL-2026

Location
Houston (Corporate)

ETX ID
3.100.007.F01

Printed by smohammad@eagleanalytical.com from https://app.zenqms.com on 21-Aug-2026 at 3:28:56 PM UTC

Page 2 of 2

NCR#: N 26487

CLOSING THE NONCONFORMANCE (*Closure by Quality*)

Assessment Impact:

Frequency: ²_____

<2 = 1 Between 2 to 10 = 2 >10 = 3

Severity: ²_____

Typo Corrections = 1
Changes in processes and/or systems with no retroactive implications = 2
Changes in processes and/or systems with retroactive implication = 3

Magnitude: ²_____

<2 Process/Material/Equipment Affected = 1
>2 to <10 Process/Material/Equipment Affected = 2
>10 Process/Material/Equipment = 3

1 to 2: Low
3 to 17: Major
18 to 27: Critical

3	18	27
2	17	18
1	2	3

Assessment Impact Grading: ⁸

Impact Level

☐ Low

☒ Major

☐ Critical

Quality Reviewed By (Print):

Robin Seymour

Signature

Date

Quality Approved By (Print):

Signature

Date