

Good morning @Simin Mohammad,

I have reviewed this OOS. Please find the summary of edits in the table below. All required updates, including the amended EM table and all form field revisions, have already been completed on your behalf and assembled into the attached finalized 7-page PDF package.

| Section                                | Unreviewed                                                                                                                                                                                                                                                                                                      | Reviewed                                                                                                                                                                                                                                                                                                                                                                                                                       | Impact                                                                              |
|----------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------|
| EM Table Attachment & Formatting       | The EM Table was omitted from the PDF submission. In the draft Word document, Table 1 listed ETX ID as <b>ETX-260526-0461</b> and Microbial ID as <b>Talaromyces purpurogenus</b> , Table 2 contained unclosed parentheses in timing descriptions, and multi-organism entries ran together without line breaks. | Amended both tables to reflect the correct ETX submission ID <b>ETX-260518-0250</b> , updated the microbial identification to <b>Penicillium decumbens</b> , <b>Cladosporium tenuissimum</b> , <b>Cladosporium langeronii</b> , <b>Cladosporium halotolerans</b> , corrected timing descriptions, formatted multi-organism line breaks, and attached the amended 1-page table as <b>Page 7</b> to complete the 7-page package. | Document Assembly & Technical Uniformity Correction. Completed on analyst's behalf. |
| Initiator Name                         | Listed as <b>"Simin Mohammad (written by Simin Mohammad)"</b>                                                                                                                                                                                                                                                   | Revised to <b>"Simin Mohammad"</b>                                                                                                                                                                                                                                                                                                                                                                                             | Clerical Correction.                                                                |
| Name of Analyst who Performed the Test | Listed as <b>"Simin Mohammad(weekly Active air Sampling Plate Setup)Maraya Chukwumerije &amp; Sophia Santamaria(Weekly active air Sampling Plate Readers)"</b>                                                                                                                                                  | Revised to separate entries with proper role titles and line breaks: <b>Simin Mohammad (Weekly Active Air Sampling Plate Setup)</b><br><b>Maraya Chukwumerije (Weekly Active Air Sampling Plate Reader)</b><br><b>Sophia Santamaria (Weekly Active Air Sampling Plate Reader)</b>                                                                                                                                              | Formatting & Administrative Correction.                                             |
| SOP Reference & Effective Date         | Listed as <b>"20600.002"</b> Rev <b>15</b> , Effective Date <b>05AUG2025</b>                                                                                                                                                                                                                                    | Revised to <b>"MICRO-SOP-2"</b> Rev <b>16</b> , Effective Date <b>23-Jul-2026</b>                                                                                                                                                                                                                                                                                                                                              | Compliance & Document Control Correction.                                           |
| Limits / Specification                 | Listed as <b>"Action level &gt;10"</b>                                                                                                                                                                                                                                                                          | Revised to <b>"Action level: &gt;= 10 CFU/Plate"</b>                                                                                                                                                                                                                                                                                                                                                                           | Technical Specification Correction.                                                 |
| Section B SOP Citations                | Comments listed <b>"SOP 2.600.00"</b> missing trailing digit                                                                                                                                                                                                                                                    | Revised to <b>"MICRO-SOP-2"</b>                                                                                                                                                                                                                                                                                                                                                                                                | SOP Citation Correction.                                                            |
| Analyst Interview Comment              | Listed as <b>"Yes, analysts Simin Mohammad, Maraya Chukwumerije, &amp; Sophia Santamaria were comprehensively interviewed."</b>                                                                                                                                                                                 | Revised to <b>"Yes, analysts Simin Mohammad, Maraya Chukwumerije, and Sophia Santamaria were comprehensively interviewed."</b>                                                                                                                                                                                                                                                                                                 | Grammatical Correction.                                                             |
| Equipment &                            | Equipment IDs and calibration                                                                                                                                                                                                                                                                                   | Revised with clean line breaks:                                                                                                                                                                                                                                                                                                                                                                                                | Equipment                                                                           |

|                                                             |                                                                                                                                                                                                                             |                                                                                                                                                                                                                                                                                                            |                                                 |
|-------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------|
| Calibration Details                                         | <p>dates running together without spacing:<br/> <b>"Incubator E001034(Sensor E001501)Incubator E001031 (Sensor E001505)"</b><br/> <b>"Aug 2026 Feb 2027 Aug 2026 Feb 2027"</b></p>                                          | <p><b>Incubator E001034 (Sensor E001501)</b><br/> <b>Incubator E001031 (Sensor E001505)</b></p> <p><b>Aug 2026 / Feb 2027</b></p>                                                                                                                                                                          | Traceability & Readability Correction.          |
| Phase I Summary – Setup & Incubation Narrative              | <p>Contained spelling error in reader's name (<b>"Sophia Sanatamaria"</b>), grammatical error (<b>"incubators were verified"</b>), and inaccurate reference to biosafety cabinet locations instead of cleanroom suites.</p> | <p>Corrected name to <b>Sophia Santamaria</b>, corrected grammar to <b>incubator was verified</b>, updated locations to <b>Suite 115 (ISO 8), Suite 115A (ISO 7), and Suite 115B (ISO 7)</b>, and cited <b>MICRO-SOP-2</b> and <b>MICRO-SOP-9</b>.</p>                                                     | Grammatical & Scope Correction.                 |
| Phase I Summary – Monthly Cleaning & Disinfection Narrative | <p>Contained contradictory monthly cleaning dates of <b>"22 Feb 2026"</b> and <b>"26APR2026"</b> within the same sentence, a missing period, and legacy SOP citations.</p>                                                  | <p>Reconciled cleaning date to <b>26-Apr-2026</b> performed by Tamiru Kotisso and Cuong Du, inserted missing period before the H2O2 indicators statement, and updated procedure reference to <b>MICRO-SOP-9</b>.</p>                                                                                       | Chronological Accuracy & Compliance Correction. |
| Phase I Summary – Root Cause Statement                      | <p>Stated that the OOS result observed for the Environmental Monitoring (EM) <b>"Settling Sampling plate"</b> may be attributed to a potential analyst error.</p>                                                           | <p>Revised plate type to <b>Active Air Sampling plate</b> to accurately reflect the test performed, while maintaining the determination that the result may be attributed to a potential analyst error, noting that the contamination was transient in nature and no corrective actions are necessary.</p> | Sampling Type & Technical Accuracy Correction.  |

@Simin Mohammad please review and sign the attached finalized 7-page PDF package.

Thanks,