Table 1: Information for ETX-261021-0286

Processing Analyst	Reading Analyst	Sample ID	Event Count	Confirmed Microbial Events	Morphology
Karla Silva	Sonal Uprety	ETX-261021-0286	145	84	***Rod-shaped morphology***

Table 2: Environmental Monitoring from Testing Performed on 22Oct26

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMMMYY)	Analyst	Day /Week(s)	Observation	EM Plate ETX Number	Microbial ID	Notes
Personnel EM								
Personal (Left Touch and Right Touch)	Daily	22Oct26	KSM	Date of Testing (Scan)	No Growth	Not Applicable	Not Applicable	None
Personal (Left Touch and Right Touch)	Daily	22Oct26	SU	Date of Testing (Scan C/O)	No Growth	Not Applicable	Not Applicable	None
Biological Safety Cabinet EM for E001938 and E001937								
Surface Sampling of ISO 5 BSC E001938 (4 Locations)	Daily	22Oct26	KSM	Date of Testing (Scan)	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of ISO 5 BSC E001937 (4 Locations)	Daily	22Oct26	SU	Date of Testing (Scan C/O)	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 BSC E001938	Daily	22Oct26	KSM	Date of Testing (Scan)	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 BSC E001937	Daily	22Oct26	SU	Date of Testing (Scan C/O)	No Growth	Not Applicable	Not Applicable	None
Weekly Active Air Sampling of Cleanroom 145, 144, 143, 142 (L-Suite) with Processing BSCs								
Active Air Sampling of Cleanrooms	Weekly	20Oct26	SMO	Week of Testing	2 CFUs (ISO 7 144 Sec I) 21 CFUs (ISO 8 143 Sec I) 17 CFUs (ISO 8 143 Sec II) 16 CFUs (ISO 8 142)	ETX-261027-0484 ETX-261027-0487 ETX-261027-0492 ETX-261027-0494	Pending Pending Pending Pending	None
Weekly Surface Sampling of Cleanroom 145, 144, 143, 142 (L-Suite) with Processing BSCs								
Surface Sampling of Cleanrooms	Weekly	20Oct26	SMO	Week of Testing	No Growth	Not Applicable	Not Applicable	None