

Table 1: Information for ETX-260902-0505 under investigation

Processing Analyst	Reading Analyst	Sample ID	Related Microbial ID	Media with microbial growth	Microbial ID
Alex Saravia	Elyse Nioupin	ETX-260902-0505	ETX-260910-0290	1 x 100mL TSB bottle	Microbacterium sp. PM5 (Gram (+) rods)

Table 2: Environmental Monitoring from Processing Performed on 04Sep26

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMMYY)	Analyst (Initials)	Day /Week(s)	Observation*	Environmental Monitoring Plate ETX ID	Microbial ID	Notes
Personnel EM Bracketing								
Personnel (Left Touch and Right Touch)	Daily	03Sep26	ES	Date Before Testing	No growth	N/A	N/A	None
Personnel (Left Touch and Right Touch)	Daily	04Sep26	ES	Date of Testing	No growth	N/A	N/A	None
Personnel (Left Touch and Right Touch)	Daily	08Sep26	ES	Date After Testing	No growth	N/A	N/A	None
Biological Safety Cabinet EM Bracketing Biological Safety Cabinet (BSC)								
Surface Sampling of ISO 5 (4 locations)	Daily	03Sep26	ES	Date Before Testing	No growth	N/A	N/A	None
Surface Sampling of ISO 5 (4 locations)	Daily	04Sep26	ES	Date of Testing	No growth	N/A	N/A	None
Surface Sampling of ISO 5 (4 locations)	Daily	08Sep26	ES	Date After Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	03Sep26	ES	Date Before Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	04Sep26	ES	Date of Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	08Sep26	ES	Date After Testing	No growth	N/A	N/A	None
Weekly Active Air Sampling Bracketing								
Active Air Sampling of Cleanrooms	Weekly	04Sep26	SMO	Week on Testing Date	1 CFU (ISO 8 114)	ETX-260914-0487	Pending	None
Surface Sampling of Anteroom and Cleanroom Bracketing								
Surface Sampling of Cleanrooms	Weekly	04Sep26	SMO	Week on Testing Date	No growth	N/A	N/A	None