Table 1: Information for ETX-260624-0251 (BUD-T90) under investigation

Processing Analyst	Reading Analyst	Sample ID	Related Microbial ID	Media with microbial growth	Microbial ID
Guanchen Li	Andrew Carrillo	ETX-260624-0251 (BUD-T90)	ETX-261012-0367	2 x 200mL TSB (Original and Subculture Bottles)	*Bacillus megaterium* (Gram (+) rods)

Table 2: Environmental Monitoring from Testing Performed on 23Sep26

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMMMYY)	Analyst	Day /Week(s)	Observation	EM Plate ETX Number	Microbial ID	Notes
Personnel EM								
Personnel (Left Touch and Right Touch)	Daily	22Sep26	GL	Date Before Testing	No Growth	Not Applicable	Not Applicable	None
Personnel (Left Touch and Right Touch)	Daily	23Sep26	GL	Date of Testing	No Growth	Not Applicable	Not Applicable	None
Personnel (Left Touch and Right Touch)	Daily	24Sep26	GL	Date After Testing	No Growth	Not Applicable	Not Applicable	None
Biological Safety Cabinet EM for E001314								
Surface Sampling of ISO 5 (4 locations)	Daily	22Sep26	AOD	Date Before Testing	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of ISO 5 (4 locations)	Daily	23Sep26	GL	Date of Testing	No Growth	Not Applicable	Not Applicable	None
Surface Sampling of ISO 5 (4 locations)	Daily	24Sep26	AOD	Date After Testing	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 (2 locations)	Daily	22Sep26	AOD	Date Before Testing	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 (2 locations)	Daily	23Sep26	GL	Date of Testing	No Growth	Not Applicable	Not Applicable	None
Settling Sampling of ISO 5 (2 locations)	Daily	24Sep26	AOD	Date After Testing	No Growth	Not Applicable	Not Applicable	None
Weekly Active Air Sampling of Cleanroom Suite 115 (E001737) with Processing BSC								
Active Air Sampling of Cleanrooms	Weekly	23Sep26	SMO	Week of Testing	6 CFUs (ISO 8 115)	ETX-261001-0625	*Micrococcus luteus* (4 CFUs) *Kocuria indica* and *Kocuria rhizophila* (1 CFU) Pending FunITS (1 CFU, Hyphae)	None
Weekly Surface Sampling of Cleanroom Suite 115 (E001737) with Processing BSC								
Surface Sampling of Cleanrooms	Weekly	23Sep26	SMO	Week of Testing	No Growth	Not Applicable	Not Applicable	None