

Table 1: Information for ETX-260804-0335 under investigation

Processing Analyst	Reading Analyst	Sample ID	Related Microbial ID	Media with microbial growth	Microbial ID
[Processing Analyst]	Qiyue Chen	ETX-260804-0335	ETX-260904-0331	1 x 100mL TSB bottle	Pending

Table 2: Environmental Monitoring for Analyst and Cleanroom

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMMM YYYY)	Analyst (Initials)	Day /Week(s)	Observation	Environmental Monitoring Plate ETX ID	Microbial ID	Notes
Personnel EM Bracketing 25 Aug 2026								
Personal (Left Touch and Right Touch)	Daily	24 Aug 2026	TBD	Date Before Testing	No growth	N/A	N/A	None
Personal (Left Touch and Right Touch)	Daily	25 Aug 2026	TBD	Date of Testing	No growth	N/A	N/A	None
Personal (Left Touch and Right Touch)	Daily	26 Aug 2026	TBD	Date After Testing	No growth	N/A	N/A	None
Biological Safety Cabinet EM Bracketing Biological Safety Cabinet (BSC) E001316								
Surface Sampling of ISO 5 (4 locations)	Daily	24 Aug 2026	TBD	Date Before Testing	No growth	N/A	N/A	None
Surface Sampling of ISO 5 (4 locations)	Daily	25 Aug 2026	TBD	Date of Testing	No growth	N/A	N/A	None
Surface Sampling of ISO 5 (4 locations)	Daily	26 Aug 2026	TBD	Date After Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	24 Aug 2026	TBD	Date Before Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	25 Aug 2026	TBD	Date of Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 locations)	Daily	26 Aug 2026	TBD	Date After Testing	No growth	N/A	N/A	None
Weekly Active Air Sampling of Cleanroom - CR114 (E001736)								
Active Air Sampling of Cleanrooms	Weekly	25 Aug 2026	SMO	Week of Testing	4 CFU in ISO 8 Room 114	ETX-260901-0112	3 Gram (+) cocci, 1 Hyphae	None
Surface Sampling of Anteroom and Cleanroom – CR114 (E001736)								
Surface Sampling of Cleanrooms	Weekly	25 Aug 2026	SMO	Week of Testing	No growth	N/A	N/A	None