



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqeda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 1 of 6

OOS Number: OOS-

Phase I

Section A – Test Details

Initiator:		Test Date:		Date Initiated:	
Date of Incident:		Test Name:		Sample ID:	
Name of Analyst who Performed the Test:		Sample / Active Name		Lot #:	
		Dosage Form			
Description of Incident:			☐ Client Care notified of OOS ☐ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS		
	SOP / Test Method #	Effective Date:	SOP / Test Method Rev:	Limits/Specification:	
Manager notified	Manager/designee name:	Date:	Quality Assurance Department notified:	Yes ☐ No ☐	

Section B – Investigation and Root Cause Analysis

Personnel Discrepancy		Comments
Analyst interviewed?	☐ Yes ☐ No ☐ N/A	
Correct samples analyzed?	☐ Yes ☐ No ☐ N/A	
Correct SOP / Test Method used/followed?	☐ Yes ☐ No ☐ N/A	
Correct sample collection/technique?	☐ Yes ☐ No ☐ N/A	
Are the Test Result Records Attached & Calculations Verified and Correct?	☐ Yes ☐ No ☐ N/A	
Is the analyst qualified to perform the test?	☐ Yes ☐ No ☐ N/A	
Is formula worksheet provided and verified?	☐ Yes ☐ No ☐ N/A	
Does the processing method/method set conform to the required specification?	☐ Yes ☐ No ☐ N/A	
Was the sample stored appropriately?	☐ Yes ☐ No ☐ N/A	
Material Discrepancy		Lot Number
Is the integrity of the sample(s) questionable?	☐ Yes ☐ No ☐ N/A	
Was the sample transported appropriately?	☐ Yes ☐ No ☐ N/A	
Are the other test results from this run under investigation?	☐ Yes ☐ No ☐ N/A	



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 2 of 6

OOS Number: OOS-

Appropriate/correct glassware used?	☐ Yes ☐ No ☐ N/A		
Are reagents stored in proper condition and not expired?	☐ Yes ☐ No ☐ N/A		
Standard Used			
Solvent Used			
Is the standard the source of the root cause?	☐ Yes ☐ No ☐ N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐ Yes ☐ No ☐ N/A		
Equipment Discrepancy		Equipment ID	Calibration Due Date
Was the instrument set up correctly?	☐ Yes ☐ No ☐ N/A		
Is the clean room facility certified?	☐ Yes ☐ No ☐ N/A		
System Performance Check (if applicable)?	☐ Yes ☐ No ☐ N/A		
System Suitability RSD < 2% (PDA)?	☐ Yes ☐ No ☐ N/A		
System Suitability RSD < 10% (QDA, CAD)?	☐ Yes ☐ No ☐ N/A		
Do the environmental monitoring results conform to required standards?	☐ Yes ☐ No ☐ N/A		
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☐ Yes ☐ No ☐ N/A		
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☐ Yes ☐ No ☐ N/A		



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 3 of 6

OOS Number: OOS-

Are chromatogram/spectra/raw data correct?

☐ Yes☐ No☐ N/A

Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)

☐ N/A☐ Re-injection of sample vial☐ Re-injection of original standard vial☐ Re-vialing of original working standard solution☐ Re-vialing of original working sample solution☐ Re-dilution of a working sample solution from the original stock☐ Re-dilution of a working standard solution from the original stock standard☐ Other: _____

Section D – Phase I Summary

Phase I Summary:



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 4 of 6

OOS Number: OOS-

Phase I Summary (continued):



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED
JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 5 of 6

OOS Number: OOS-

Phase I Summary (continued):



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 6 of 6

OOS Number: OOS-

Has root cause been identified as a possible laboratory error?*Note: Root cause must be conclusive, verifiable, and well-documented.*☐ Yes☐ If retesting is **required**, complete Phase II of the OOS (initiate Form 3.100.019.F02).☐ No, the OOS result is valid.**Reviewer comment(s):****Prepared by (Print):**

Signature/Date:

Lab Manager/Designee (Print):

Signature/Date:

Section E – Quality Disposition**Can investigation be closed?**☐ Yes☐ Original result is valid.☐ No, initiate Form 3.100.019.F02.**Approved by QA Representative (Print):**

Signature/Date:

Table 1: Read Dates and Incubation Observation

Sampling Location	Read Date (30-35°C, NLT 48h)	Read By	CFU Count / Observation	Read Date (20-25°C, NLT 5d)	Read By	CFU Count / Observation	Microbial Identification
Weekly Active Air Sampling Plate (Cleanroom Suite 116A)	06 Jun 2026	MC	No microbial growth was observed	11 Jun 2026	SMO	748 CFU on EM SMO 116A Air 07MAY2026	Kocuria palustris (Gram (+) cocci)

Table 2: Environmental Monitoring Plates for Analyst and Cleanroom Bracketing

Sampling Type	Sampling Location	Date Before Testing (06 May 2026)		Date of Testing (07-May-2026)		Date After Testing (08 May 2026)	
		Analyst / Result	ETX / Identification	Analyst / Result	ETX / Identification	Analyst / Result	ETX / Identification
Personnel Monitoring	Glove Touch (Left / Right)	SMO No growth	N/A N/A	SMO No growth	N/A N/A	SMO No growth	N/A N/A
Surface Sampling	ISO 5 Cleanroom Suite 116A	SMO No growth	N/A N/A	SMO No growth	N/A N/A	SMO No growth	N/A N/A
Settling Sampling	ISO 5 Cleanroom Suite 116A	SMO No growth	N/A N/A	SMO No growth	N/A N/A	SMO No growth	N/A N/A
Weekly Cleanroom Active Air	Cleanroom CR116 (E001738)	Rey Estrada No growth	N/A N/A	Rey Estrada No growth	N/A N/A	Rey Estrada No growth	N/A N/A
Weekly Cleanroom Surface	Anteroom & Buffer (CR116 (E001738))	Rey Estrada No growth	N/A N/A	Rey Estrada No growth	N/A N/A	Rey Estrada No growth	N/A N/A