



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 1 of 6

OOS Number: OOS- 261187

Phase I

Section A – Test Details

Initiator:	Clea S. Garza (written by: Qiyue Chen)	Test Date:	11May26	Date Initiated:	11May26
Date of Incident:	11May26	Test Name:	Environmental Monitoring	Sample ID:	ETX-260518-0273
Name of Analyst who Performed the Test:	Clea S. Garza (CGS) (Surface Sampling Plate Setup)	Sample / Active Name	ScanC/O CGS E001309 S1 11MAY2026	Lot #:	ScanC/O CGS E001309 S1 11MAY2026
	Simin Mohammad (Surface Sampling Plate Reader)	Dosage Form	Plate		
Description of Incident:	The CFU count for the environmental monitoring plate exceeded the action level.		☒ Client Care notified of OOS ☒ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS		
	SOP / Test Method # 2.600.002	Effective Date: 02Apr26 07Nov25	SOP / Test Method Rev: Rev: 14 Rev: 04	Limits/Specification: Action Level: >= 1 CFU/Plate	
Manager notified	Manager/designee name: Kathan Parikh	Date:	Quality Assurance Department notified:	Yes ☒ No ☐	

Section B – Investigation and Root Cause Analysis

Personnel Discrepancy		Comments
Analyst interviewed?	☒ Yes ☐ No ☐ N/A	Yes, Clea S. Garza and Simin Mohammad were interviewed comprehensively.
Correct samples analyzed?	☒ Yes ☐ No ☒ N/A	N/A
Correct SOP / Test Method used/followed?	☒ Yes ☐ No ☐ N/A	Yes, as per SOP 2.600.002
Correct sample collection/technique?	☒ Yes ☐ No ☐ N/A	Yes, as per SOP 2.600.002
Are the Test Result Records Attached & Calculations Verified and Correct?	☒ Yes ☐ No ☐ N/A	Yes, Information is available in EagleTrax under ETX-260518-0273.
Is the analyst qualified to perform the test?	☒ Yes ☐ No ☐ N/A	Yes, the analysts are trained and qualified by quality to perform the test.
Is formula worksheet provided and verified?	☐ Yes ☐ No ☒ N/A	Not Applicable
Does the processing method/method set conform to the required specification?	☐ Yes ☐ No ☒ N/A	Not Applicable
Was the sample stored appropriately?	☒ Yes ☐ No ☐ N/A	Yes, as per SOP 2.600.002
Material Discrepancy		Lot Number
Is the integrity of the sample(s) questionable?	☐ Yes ☒ No ☐ N/A	
Was the sample transported appropriately?	☒ Yes ☐ No ☐ N/A	
Are the other test results from this run under investigation?	☐ Yes ☒ No ☐ N/A	



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda, 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 2 of 6

OOS Number: OOS- 261187

Appropriate/correct glassware used?	☐ Yes ☐ No ☒ N/A		
Are reagents stored in proper condition and not expired?	☒ Yes ☐ No ☐ N/A	Scan Consumables: See the attached data packet Environmental Plates: TSA and Surface Plate: see attached environmental logs Monthly Cleaning: H2O2 strips and IPA: See attached monthly cleaning logs	Scan Consumables: See the attached data packet Environmental Plates: TSA and Surface Plate: see attached environmental logs Monthly Cleaning: H2O2 strips and IPA: See attached monthly cleaning logs
Standard Used	Not Applicable	Not Applicable	Not Applicable
Solvent Used	Not Applicable	Not Applicable	Not Applicable
Is the standard the source of the root cause?	☐ Yes ☐ No ☒ N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐ Yes ☐ No ☒ N/A		
Equipment Discrepancy		Equipment ID	Calibration Due Date
Was the instrument set up correctly?	☒ Yes ☐ No ☐ N/A	Please see below	Please see below
Is the clean room facility certified?	☒ Yes ☐ No ☐ N/A		Dec 2026
System Performance Check (if applicable)?	☒ Yes ☐ No ☒ N/A	Not Applicable	Not Applicable
System Suitability RSD < 2% (PDA)?	☐ Yes ☐ No ☒ N/A	Not Applicable	Not Applicable
System Suitability RSD < 10% (QDA, CAD)?	☐ Yes ☐ No ☒ N/A	Not Applicable	Not Applicable
Do the environmental monitoring results conform to required standards?	☒ Yes ☐ No ☒ N/A	Not Applicable	Not Applicable
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☒ Yes ☐ No ☒ N/A	Not Applicable	Not Applicable
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☒ Yes ☐ No ☐ N/A	Incubator E001034 (Sensor E001501) Incubator E001031 (Sensor E001505)	Aug 2026 Feb 2027 Aug 2026 Feb 2027



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqveda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 3 of 6

OOS Number: OOS- 261187

Are chromatogram/spectra/raw data correct?

☐ Yes

☐ No

☒ N/A

Not Applicable

Not Applicable

Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)

☒ N/A

☐ Re-injection of sample vial

☐ Re-injection of original standard vial

☐ Re-vialing of original working standard solution

☐ Re-vialing of original working sample solution

☐ Re-dilution of a working sample solution from the original stock

☐ Re-dilution of a working standard solution from the original stock standard

☐ Other: N/A QYC

Section D – Phase I Summary

Phase I Summary:



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 4 of 6

OOS Number: OOS- 261187

Phase I Summary (continued):



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 5 of 6

OOS Number: OOS- 261187

Phase I Summary (continued):

N/A QYC



QUALITY SYSTEM

FORM

DOCUMENT #: 3.100.019.F01

LEVEL: 3

Effective Date:

APPROVED

JEsqueda , 9/30/2025, 4:15:53 PM

Revision #: 11

TITLE:

LABORATORY OOS INVESTIGATION FORM

Page 6 of 6

OOS Number: OOS- 261187

Has root cause been identified as a possible laboratory error?

Note: Root cause must be conclusive, verifiable, and well-documented.☒ Yes☐ If retesting is **required**, complete Phase II of the OOS (initiate Form 3.100.019.F02).☒ No, the OOS result is valid.

Reviewer comment(s):

Prepared by (Print):

Maryam Naeem

Signature/Date:

Lab Manager/Designee (Print):

Kathan Parikh

Signature/Date:

Section E – Quality Disposition

Can investigation be closed?

☐ Yes☐ Original result is valid.☐ No, initiate Form 3.100.019.F02.

Approved by QA Representative (Print):

Signature/Date: