

Table 1: Read Dates & Incubation Observation

ETX Submission ID	Set-up Analyst & Date	Sampling Site & Location	Plate Reading Analyst (≥ 48H)	CFUs Observed after 48 Hour Incubation at 30–35°C (E001031)	Plate Reading Analyst (NLT 5 days)	CFUs Observed after NLT 5-day Incubation at 20–25°C (E001034)	Microbial Identification
ETX-260216-0348	GS 05 Feb 2026	Settling Sampling Plate (BSC 1314)	MC	10 CFU on Settling S2 sampling plate	SMO	10 CFU on Settling S2 sampling plate	Staphylococcus capitis (Gram (+) cocci), Staphylococcus hominis (Gram (+) cocci), Kocuria indica (Gram (+) cocci), Micrococcus luteus (Gram (+) cocci) and Staphylococcus epidermidis (Gram (+) cocci)

Table 2: Environmental Monitoring for Analyst & Cleanroom

Environmental Monitoring (EM) Sampling Site	Frequency	Date (DDMMYYYY Y)	Analyst (Initials)	Day /Week(s)	Observation	Plate ETX ID	Microbial ID	Notes
Personnel EM Bracketing								
Personal (Left & Right Touch)	Daily	04 Feb 2026	GS	Date Before Testing	No growth	N/A	N/A	None
Personal (Left & Right Touch)	Daily	05 Feb 2026	GS	Date of Testing	No growth	N/A	N/A	None
Personal (Left & Right Touch)	Daily	06 Feb 2026	GS	Date After Testing	No growth	N/A	N/A	None
Biological Safety Cabinet (BSC) EM Bracketing								
Surface Sampling of ISO 5 (4 loc)	Daily	04 Feb 2026	GS	Date Before Testing	No growth	N/A	N/A	None
Surface Sampling of ISO 5 (4 loc)	Daily	05 Feb 2026	GS	Date of Testing	No growth	N/A	N/A	None
Surface Sampling of ISO 5 (4 loc)	Daily	06 Feb 2026	GS	Date After Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 loc)	Daily	04 Feb 2026	GS	Date Before Testing	No growth	N/A	N/A	None
Settling Sampling of ISO 5 (2 loc)	Daily	05 Feb 2026	GS	Date of Testing	10 CFU (Settling S2)	ETX-260216-0348	Staphylococcus capitis (Gram (+) cocci), Staphylococcus hominis (Gram (+) cocci), Kocuria indica (Gram (+) cocci), Micrococcus luteus (Gram (+) cocci) and Staphylococcus epidermidis (Gram (+) cocci)	None
Settling Sampling of ISO 5 (2 loc)	Daily	06 Feb 2026	GS	Date After Testing	No growth	N/A	N/A	None
Weekly Active Air Sampling Bracketing								
Active Air Sampling of Cleanrooms	Weekly	06 Feb 2026	SMO	Week (On or After Date)	No growth	N/A	N/A	None
Surface Sampling of Anteroom & Cleanroom Bracketing								
Surface Sampling of Cleanrooms	Weekly	06 Feb 2026	SMO	Week (On or After Date)	No growth	N/A	N/A	None