OOS Number: OOS-

Phase I

Section A – Test Details					
Initiator:		**Test Date:**		**Date Initiated:**	
Date of Incident:		**Test Name:**		**Sample ID:**	
Name of Analyst who Performed the Test:		**Sample / Active Name**		**Lot #:**	
		Dosage Form			
Description of Incident:		☐ Client Care notified of OOS ☐ Sterility and Endotoxin test related OOS; Client must be notified within 24 hours of OOS			
	SOP / Test Method #	Effective Date:	SOP / Test Method Rev:	Limits/Specification:	
Manager notified	Manager/designee name:	Date:	Quality Assurance Department notified:	Yes ☐ No ☐	

Section B – Investigation and Root Cause Analysis			
Personnel Discrepancy		**Comments**	
Analyst interviewed?	☐Yes ☐ No ☐ N/A		
Correct samples analyzed?	☐Yes ☐ No ☐ N/A		
Correct SOP / Test Method used/followed?	☐Yes ☐ No ☐ N/A		
Correct sample collection/technique?	☐Yes ☐ No ☐ N/A		
Are the Test Result Records Attached & Calculations Verified and Correct?	☐Yes ☐ No ☐ N/A		
Is the analyst qualified to perform the test?	☐Yes ☐ No ☐ N/A		
Is formula worksheet provided and verified?	☐Yes ☐ No ☐ N/A		
Does the processing method/method set conform to the required specification?	☐Yes ☐ No ☐ N/A		
Was the sample stored appropriately?	☐Yes ☐ No ☐ N/A		
Material Discrepancy		**Lot Number**	**Expiry Date**
Is the integrity of the sample(s) questionable?	☐Yes ☐ No ☐N/A		
Was the sample transported appropriately?	☐Yes ☐ No ☐N/A		
Are the other test results from this run under investigation?	☐Yes ☐ No ☐N/A		

OOS Number: OOS-

Appropriate/correct glassware used?	☐Yes ☐ No ☐N/A		
Are reagents stored in proper condition and not expired?	☐Yes ☐ No ☐N/A		
Standard Used			
Solvent Used			
Is the standard the source of the root cause?	☐Yes ☐ No ☐N/A		
Has the analyst been instructed to keep the original test solution for possible retesting?	☐Yes ☐ No ☐N/A		
Equipment Discrepancy		**Equipment ID**	**Calibration Due Date**
Was the instrument set up correctly?	☐Yes ☐ No ☐ N/A		
Is the clean room facility certified?	☐Yes ☐ No ☐ N/A		
System Performance Check (if applicable)?	☐Yes ☐ No ☐ N/A		
System Suitability RSD < 2% (PDA)?	☐Yes ☐ No ☐ N/A		
System Suitability RSD < 10% (QDA, CAD)?	☐Yes ☐ No ☐ N/A		
Do the environmental monitoring results conform to required standards?	☐Yes ☐ No ☐ N/A		
Do the controls (i.e., positive, negative, etc.) conform to the required standards?	☐Yes ☐ No ☐ N/A		
Is equipment/instrument functional and calibrated/qualified/certified/PM?	☐Yes ☐ No ☐ N/A		

OOS Number: OOS-

Are chromatogram/spectra/raw data correct?	☐ Yes ☐ No ☐ N/A		

Section C – Evaluation of Laboratory Testing/Hypothesis Testing (if applicable)

- ☐ N/A
- ☐ Re-injection of sample vial
- ☐ Re-injection of original standard vial
- ☐ Re-vialing of original working standard solution
- ☐ Re-vialing of original working sample solution
- ☐ Re-dilution of a working sample solution from the original stock
- ☐ Re-dilution of a working standard solution from the original stock standard
- ☐ Other: _____

Section D – Phase I Summary

Phase I Summary:

OOS Number: OOS-

Phase I Summary (continued):

OOS Number: OOS-

Phase I Summary (continued):

OOS Number: OOS-

Has root cause been identified as a possible laboratory error?
Note: Root cause must be conclusive, verifiable, and well-documented.

☐ Yes

☐ If retesting is ***required***, complete Phase II of the OOS (initiate Form 3.100.019.F02).

☐ No, the OOS result is valid.

Reviewer comment(s):

Prepared by (*Print*)**:**	Signature/Date:
Lab Manager/Designee (*Print*)**:**	Signature/Date:

Section E – Quality Disposition

Can investigation be closed?

☐ Yes

☐ Original result is valid.

☐ No, initiate Form 3.100.019.F02.

Approved by QA Representative (*Print*)**:**	Signature/Date: