

IAEA-TERC-2022-01 World Wide Open Proficiency Test Exercise, Pie-charts, S-Shapes and Reported Results with Scores

Terrestrial Environmental Radiochemistry Laboratory, International Atomic Energy Agency

November 3, 2022

Contents

1 Introduction	1
1.1 Evaluation criteria	1
2 Data Evaluation	3
3 Proficiency Test Parameters	3
4 Intercomparison Parameters	73
5 Gamma spectrum analysis exercise results (Sample 7)	112
6 References	168
7 Contributors to evaluation and report	169

Abstract

This report describes the evaluation method for the Proficiency Test Exercise on the determination of anthropogenic and natural radionuclides in water, soil (gamma-ray spectrum analysis exercise), and simulated contaminated surface samples IAEA-TERC-2022-01. The evaluation is performed by the proficiency test team of the Terrestrial Environmental Radiochemistry Laboratory of the IAEA using their standard approach.

1 Introduction

The following figures show the results (displayed in pie charts, characteristic S-shape graphs and tables) for the World Wide Open Proficiency Test. In the S-shape graphs, the vertical axis displays the activity in [Bq/kg] and the horizontal axis shows the labcodes for the participating laboratories. The range on the vertical axis is limited to $\pm 2 \times$ Maximum acceptable relative bias (MARB). The red line is the target value for the exercise, whereas the green lines are the assigned uncertainty for each analyte. The results are grouped by samples, then analytes.

Disclaimer: This report has been generated automatically to provide information on-line to the participants. A more detailed evaluation will be published in the final report. If you find that any information provided in this report might be incorrect please contact us as soon as possible.

1.1 Evaluation criteria

The data is evaluated according to the following steps:

1. The relative bias between the reported and the target value (the best estimation of the true value) is expressed by the following equation:

$$Bias_{relative} = \frac{Value_{reported} - Value_{target}}{Value_{target}} * 100\%$$

2. The relative bias is compared to the Maximum Acceptable Relative Bias (MARB) which has been determined for each analyte, considering the radioanalytical methods, the level of radioactivity and the complexity of the analysis.

If the $|Bias_{relative}| \leq MARB$, the result will be "Accepted" for accuracy.

3. Based on fit-for-purpose and good laboratory practice principles, the expanded relative combined uncertainty should cover the relative bias:

$$P = \sqrt{\left(\frac{u_{target}}{A_{target}}\right)^2 + \left(\frac{u_{reported}}{A_{reported}}\right)^2} * 100$$

$$|Bias_{relative}| \leq k * P$$

where k is the coverage factor, for the 99% confidential level, $k = 2.58$.

4. If the result is between the $\pm MARB$ values, but it is not overlapping the target value within the uncertainty, this equation is used to decide if they are significantly different or not.

5. The P value is compared to the MARB also. If both the:

$$P \leq MARB$$

and

$$|Bias_{relative}| \leq k * P$$

are fulfilled, the reported results will be "Accepted" for precision. The result will be assigned "Not Accepted" for precision if either conditions are not fulfilled.

6. The final score is assigned according to the detailed evaluation described above. The possible scores are listed below:

- "Accepted" when both, accuracy and precision achieved "Accepted" status
- "Not Accepted" when the accuracy is "Not Accepted"
- "Warning" when accuracy is "Accepted" but the precision is "Not Accepted"

7. As additional information, a Z-Score parameter is shown in the evaluation tables that is calculated by using the robust standard deviation described in [2]:

$$z = \left| \frac{Value_{reported} - Value_{target}}{robust\ sd} \right|$$

8. If the analyte is included in the proficiency test evaluation scheme, the stated target value is used to calculate the Z-Score. For those analytes, which are subject of an intercomparison only, the robust mean of the values reported is used instead.

9. For analytes evaluated by intercomparison, the final scores (Z-Score evaluation) are assigned as follows:

- "Accepted" for Z-Scores less than 2
- "Not Accepted" Z-Scores above 3
- "Warning" for Z-Scores less than 3 and larger than 2

2 Data Evaluation

A total of 287 laboratories from 70 different countries successfully reported data in the IAEA-TERC-2022-01 proficiency test exercise.

For each of the analytes evaluated, the target value, target uncertainty and the MARB are listed. A pie chart showing the distribution of accepted, warning, not acceptable and not reported results, the s-shape graph, and a detailed table listing all evaluation results for the analyte (sorted by the laboratory code) are provided. The data points are coloured in blue, yellow and red, corresponding to a result of "Accepted", "Warning", and "Not Accepted", respectively.

For analytes evaluated by intercomparison (IC) parameters, the target value table lists the robust mean, robust standard deviation (SD) and the number of results evaluated (N). All results are combined from the World Wide Open and ALMERA exercises for the purposes of calculating the robust statistics. A detailed table listing all evaluation results for the analyte (sorted by the laboratory code) are provided, where the Z-Scores are calculated using the robust mean as the $Value_{target}$ in the equation listed in 1.1 Evaluation criteria, point 7.

3 Proficiency Test Parameters

Reported Results for Co-60 Determination in Sample 1

TABLE 1. Target values

Parameter	Value
Target Value, Bq/kg	17.7
UNC, Bq/kg	1.1
MARB, %	20




FIGURE 1. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 2. Reported results and evaluation parameters for Co-60. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 2. Evaluation Results for Co-60 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
1	18	0.7	1.69	7.33	A	A	A	0.27
4	16.92	0.24	-4.41	6.37	A	A	A	0.71
5	15.77	0.53	-10.90	7.07	A	A	A	1.75
6	17.39	1.15	-1.75	9.07	A	A	A	0.28
7	17.5	0.8	-1.13	7.71	A	A	A	0.18
8	17.7	1.4	0.00	10.06	A	A	A	0.00
9	26.88	0.55	51.86	6.54	N	N	N	8.35
11	17.5	1.1	-1.13	8.84	A	A	A	0.18
12	18.2	0.5	2.82	6.79	A	A	A	0.45
13	17	1	-3.95	8.56	A	A	A	0.64
14	18.06	0.90	2.03	7.97	A	A	A	0.33
16	16.37	0.73	-7.51	7.65	A	A	A	1.21
17	18.0	2.0	1.69	12.73	A	A	A	0.27
18	18.18	1.48	2.71	10.24	A	A	A	0.44
19	17.88	1.69	1.02	11.31	A	A	A	0.16
20	16	1	-9.60	8.81	A	A	A	1.55
21	17.6	0.6	-0.56	7.09	A	A	A	0.09
22	17.001	0.267	-3.95	6.41	A	A	A	0.64
23	17.56	0.51	-0.79	6.86	A	A	A	0.13
24	17.5	1.0	-1.13	8.44	A	A	A	0.18
25	16.49	0.81	-6.84	7.92	A	A	A	1.10
26	17.8	0.5	0.56	6.82	A	A	A	0.09
28	17.8	0.9	0.56	8.01	A	A	A	0.09
29	19.25	0.38	8.76	6.52	A	A	A	1.41
31	18.1	1.4	2.26	9.92	A	A	A	0.36
32	18.7	0.4	5.65	6.57	A	A	A	0.91
33	16.84	0.65	-4.86	7.32	A	A	A	0.78
34	17.40	1.03	-1.69	8.58	A	A	A	0.27
35	18.9	0.5	6.78	6.75	A	A	A	1.09
36	18	3	1.69	17.79	A	A	A	0.27

TABLE 2. Evaluation Results for Co-60 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
37	18.1	1.0	2.26	8.32	A	A	A	0.36
38	19	1.3	7.34	9.24	A	A	A	1.18
39	17.8	0.8	0.56	7.67	A	A	A	0.09
40	17.4	0.2	-1.69	6.32	A	A	A	0.27
41	17.6	1.06	-0.56	8.65	A	A	A	0.09
42	4.77	0.18	-73.05	7.27	N	N	N	11.75
43	17.4	0.7	-1.69	7.40	A	A	A	0.27
45	18.02	0.37	1.81	6.55	A	A	A	0.29
46	16.9	0.7	-4.52	7.47	A	A	A	0.73
47	19.5	2.1	10.17	12.43	A	A	A	1.64
49	18.6	0.6	5.08	7.00	A	A	A	0.82
50	18.0	0.3	1.69	6.43	A	A	A	0.27
51	17.4	0.3	-1.69	6.45	A	A	A	0.27
52	16.51	1.14	-6.72	9.29	A	A	A	1.08
53	21.02	0.55	18.76	6.74	A	N	W	3.02
55	16.7	0.56	-5.65	7.06	A	A	A	0.91
56	17.6	0.9	-0.56	8.05	A	A	A	0.09
57	18.52	0.88	4.63	7.82	A	A	A	0.75
58	17.65	0.39	-0.28	6.60	A	A	A	0.05
59	17.8	0.6	0.56	7.07	A	A	A	0.09
60	18.2	1.5	2.82	10.32	A	A	A	0.45
61	18.1	0.91	2.26	7.99	A	A	A	0.36
62	16.7	0.6	-5.65	7.18	A	A	A	0.91
63	16.94	0.56	-4.29	7.04	A	A	A	0.69
64	18.1	1	2.26	8.32	A	A	A	0.36
67	16.82	1.12	-4.97	9.11	A	A	A	0.80
68	17.5	0.9	-1.13	8.07	A	A	A	0.18
69	16.81	0.16	-5.03	6.29	A	A	A	0.81
70	17.8	0.5	0.56	6.82	A	A	A	0.09
71	16.63	0.45	-6.05	6.78	A	A	A	0.97
72	17	1	-3.95	8.56	A	A	A	0.64
74	17.95	0.99	1.41	8.31	A	A	A	0.23
75	16.8	0.8	-5.08	7.83	A	A	A	0.82
76	17.8	0.60	0.56	7.07	A	A	A	0.09
77	18.2	0.900	2.82	7.94	A	A	A	0.45
78	17.4	0.5	-1.69	6.85	A	A	A	0.27
79	18.6	1.1	5.08	8.58	A	A	A	0.82
80	16.43	0.49	-7.18	6.89	A	A	A	1.15
81	15.0	1.4	-15.25	11.21	A	A	A	2.45
83	16.98	0.77	-4.07	7.69	A	A	A	0.65
84	17.825	0.846	0.71	7.82	A	A	A	0.11
86	16.4	1.23	-7.34	9.74	A	A	A	1.18
87	17	0.7	-3.95	7.46	A	A	A	0.64
88	19.3	0.8	9.04	7.47	A	A	A	1.45
89	17.7	1.2	0.00	9.20	A	A	A	0.00
90	17.4	0.54	-1.69	6.95	A	A	A	0.27
91	16.9	1.2	-4.52	9.44	A	A	A	0.73
92	18.66	0.59	5.42	6.97	A	A	A	0.87
93	17.7	0.31	0.00	6.46	A	A	A	0.00
94	19.3	1.0	9.04	8.09	A	A	A	1.45
95	17.28	0.37	-2.37	6.57	A	A	A	0.38
96	17.45	0.31	-1.41	6.46	A	A	A	0.23
97	16.47	0.67	-6.95	7.43	A	A	A	1.12
98	17.9	1.1	1.13	8.74	A	A	A	0.18

TABLE 2. Evaluation Results for Co-60 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
101	17.8	1.5	0.56	10.47	A	A	A	0.09
102	18.4	0.757	3.95	7.45	A	A	A	0.64
103	18.3	1.0	3.39	8.28	A	A	A	0.55
105	17.55	0.999	-0.85	8.43	A	A	A	0.14
106	16.87	0.627	-4.69	7.24	A	A	A	0.75
107	17.8	0.56	0.56	6.97	A	A	A	0.09
109	18.40	0.70	3.95	7.29	A	A	A	0.64
110	18.3	1.09	3.39	8.61	A	A	A	0.55
111	18.2	1.11	2.82	8.71	A	A	A	0.45
113	15.98	0.16	-9.72	6.29	A	A	A	1.56
114	20.2	0.9	14.12	7.65	A	A	A	2.27
115	18.26	0.92	3.16	8.00	A	A	A	0.51
117	17.1	0.6	-3.39	7.14	A	A	A	0.55
118	18	1	1.69	8.34	A	A	A	0.27
119	17.252	1.811	-2.53	12.20	A	A	A	0.41
120	18.0	0.7	1.69	7.33	A	A	A	0.27
121	17.945	0.539	1.38	6.90	A	A	A	0.22
123	18.1	1.0	2.26	8.32	A	A	A	0.36
124	17.52	0.44	-1.02	6.70	A	A	A	0.16
125	17.15	1.20	-3.11	9.36	A	A	A	0.50
126	18.8	1.1	6.21	8.54	A	A	A	1.00
127	17.55	1.41	-0.85	10.16	A	A	A	0.14
128	17.9	0.6	1.13	7.06	A	A	A	0.18
129	16.7	0.9	-5.65	8.23	A	A	A	0.91
130	17.7	0.3	0.00	6.44	A	A	A	0.00
132	16.4	0.5	-7.34	6.92	A	A	A	1.18
133	17.2	1.8	-2.82	12.17	A	A	A	0.45
134	17.7	0.9	0.00	8.03	A	A	A	0.00
135	16.2	4.1	-8.47	26.06	A	N	W	1.36
136	17.8	1.4	0.56	10.02	A	A	A	0.09
137	16.90	0.27	-4.52	6.42	A	A	A	0.73
138	17.274	0.511	-2.41	6.88	A	A	A	0.39
139	17.9	0.25	1.13	6.37	A	A	A	0.18
140	17.8	1.1	0.56	8.76	A	A	A	0.09
143	17.8	0.82	0.56	7.74	A	A	A	0.09
145	16.0	0.9	-9.60	8.38	A	A	A	1.55
146	16.751	0.0416	-5.36	6.22	A	A	A	0.86
147	17.48	0.81	-1.24	7.75	A	A	A	0.20
148	17.05	0.81	-3.67	7.82	A	A	A	0.59
149	16.0	0.7	-9.60	7.60	A	A	A	1.55
150	17.167	0.978	-3.01	8.43	A	A	A	0.48
152	16.9	1.1	-4.52	9.00	A	A	A	0.73
153	14.9	0.3	-15.82	6.53	A	A	A	2.55
154	17.2	0.75	-2.82	7.59	A	A	A	0.45
155	16.03	0.95	-9.44	8.59	A	A	A	1.52
156	17.9	1.4	1.13	9.99	A	A	A	0.18
157	17.6	1.4	-0.56	10.09	A	A	A	0.09
158	19.240	0.770	8.70	7.39	A	A	A	1.40
159	18.0	2.0	1.69	12.73	A	A	A	0.27
160	17.2	0.3	-2.82	6.45	A	A	A	0.45
161	17.05	0.8	-3.67	7.79	A	A	A	0.59
162	16.6	0.8	-6.21	7.86	A	A	A	1.00
163	17.59	0.91	-0.62	8.09	A	A	A	0.10
164	17.68	0.72	-0.11	7.43	A	A	A	0.02

TABLE 2. Evaluation Results for Co-60 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
165	18.45	0.52	4.24	6.82	A	A	A	0.68
166	17.2	0.5	-2.82	6.86	A	A	A	0.45
167	17.6	1.0	-0.56	8.42	A	A	A	0.09
168	21.97	0.49	24.12	6.60	N	N	N	3.88
169	17.95	0.43	1.41	6.66	A	A	A	0.23
171	18.9	1.1	6.78	8.51	A	A	A	1.09
172	17.05	0.84	-3.67	7.93	A	A	A	0.59
173	18.41	0.35	4.01	6.50	A	A	A	0.65
175	18.1	0.9	2.26	7.96	A	A	A	0.36
177	17.8	0.8	0.56	7.67	A	A	A	0.09
178	17.4	0.6	-1.69	7.11	A	A	A	0.27
179	18.657	0.49252	5.41	6.75	A	A	A	0.87
180	19.0	0.9	7.34	7.81	A	A	A	1.18
182	19.1	0.5	7.91	6.74	A	A	A	1.27
183	17.27	0.46	-2.43	6.76	A	A	A	0.39
184	17.9	0.7	1.13	7.34	A	A	A	0.18
187	19.2	1.0	8.47	8.11	A	A	A	1.36
188	18.7	1.25	5.65	9.13	A	A	A	0.91
189	18.26	1.49	3.16	10.26	A	A	A	0.51
191	17.6	0.5	-0.56	6.83	A	A	A	0.09
192	18.10	0.73	2.26	7.41	A	A	A	0.36
193	2.059	0.131	-88.37	8.89	N	N	N	14.22
194	63	6	255.93	11.37	N	N	N	41.18
195	15.5	0.7	-12.43	7.68	A	A	A	2.00
196	17.5	0.587	-1.13	7.06	A	A	A	0.18
197	15.8	3.7	-10.73	24.23	A	N	W	1.73
198	17.89	0.71	1.07	7.37	A	A	A	0.17
201	18.0	0.5	1.69	6.81	A	A	A	0.27
202	17.7	0.91	0.00	8.07	A	A	A	0.00
203	17.9	0.6	1.13	7.06	A	A	A	0.18
204	17.5	1.1	-1.13	8.84	A	A	A	0.18
206	19.6	0.07	10.73	6.22	A	A	A	1.73
207	18.47	0.57	4.35	6.94	A	A	A	0.70
210	18.3	1.0	3.39	8.28	A	A	A	0.55
213	18.3	1.0	3.39	8.28	A	A	A	0.55
214	17.2	0.6	-2.82	7.13	A	A	A	0.45
215	17.2	0.92	-2.82	8.20	A	A	A	0.45
216	17.8	0.3	0.56	6.44	A	A	A	0.09
217	18.1	1.1	2.26	8.69	A	A	A	0.36
218	17.04	0.74	-3.73	7.58	A	A	A	0.60
219	16.0	0.43	-9.60	6.77	A	A	A	1.55
220	17.5	0.8	-1.13	7.71	A	A	A	0.18
221	18.13	0.48	2.43	6.76	A	A	A	0.39
222	17.8	0.8	0.56	7.67	A	A	A	0.09
223	17.11	0.11	-3.33	6.25	A	A	A	0.54
224	17.19	0.18	-2.88	6.30	A	A	A	0.46
226	19.2	0.448	8.47	6.64	A	A	A	1.36
227	17.15	0.9	-3.11	8.13	A	A	A	0.50
228	18.1	2.1	2.26	13.16	A	A	A	0.36
230	17.7	1.3	0.00	9.62	A	A	A	0.00
231	16.57	1.66	-6.38	11.79	A	A	A	1.03
233	13.2	0.8	-25.42	8.68	N	N	N	4.09
234	17.30	0.40	-2.26	6.63	A	A	A	0.36
235	17.2	0.28	-2.82	6.42	A	A	A	0.45

TABLE 2. Evaluation Results for Co-60 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
237	17.74	0.17	0.23	6.29	A	A	A	0.04
238	17.5	0.6	-1.13	7.10	A	A	A	0.18
239	18.16	2.63	2.60	15.76	A	A	A	0.42
241	18.48	0.77	4.41	7.48	A	A	A	0.71
243	18.62	0.95	5.20	8.04	A	A	A	0.84
244	18.9	1.7	6.78	10.93	A	A	A	1.09
246	18.0	0.4	1.69	6.60	A	A	A	0.27
247	18.56	0.92	4.86	7.95	A	A	A	0.78
248	18.74	0.59	5.88	6.97	A	A	A	0.95
249	11.2	4	-36.72	36.25	N	N	N	5.91
250	17.68	0.34	-0.11	6.51	A	A	A	0.02
251	16.24	0.88	-8.25	8.25	A	A	A	1.33
253	17.62	0.27	-0.45	6.40	A	A	A	0.07
255	17.7	0.4	0.00	6.61	A	A	A	0.00
256	17.8	0.68	0.56	7.29	A	A	A	0.09
258	18.3	0.34	3.39	6.49	A	A	A	0.55
259	18.21	1.26	2.88	9.30	A	A	A	0.46
262	18.6	2.41	5.08	14.37	A	A	A	0.82
263	17.962	1.193	1.48	9.10	A	A	A	0.24
264	17.7	1.7	0.00	11.44	A	A	A	0.00
265	15.53	0.40	-12.26	6.73	A	A	A	1.97
266	17.06	0.59	-3.62	7.11	A	A	A	0.58
269	17.3	1.02	-2.26	8.57	A	A	A	0.36
271	18.6	0.9	5.08	7.88	A	A	A	0.82
272	21.45	2.10	21.19	11.60	N	A	N	3.41
273	19.5	0.9	10.17	7.74	A	A	A	1.64
274	17.95	2.947	1.41	17.55	A	A	A	0.23
276	20.86	1.33	17.85	8.90	A	A	A	2.87
278	19.0	0.97	7.34	8.04	A	A	A	1.18
281	18.0	1.0	1.69	8.34	A	A	A	0.27
284	15.31	0.68	-13.50	7.64	A	A	A	2.17
285	17.32	1.14	-2.15	9.05	A	A	A	0.35
286	17.5	1.4	-1.13	10.13	A	A	A	0.18
287	17.185	0.614	-2.91	7.17	A	A	A	0.47
288	15.8	1.6	-10.73	11.88	A	A	A	1.73
289	17.5	1.0	-1.13	8.44	A	A	A	0.18
290	19.52	0.88	10.28	7.68	A	A	A	1.65
292	19.0	2.7	7.34	15.51	A	A	A	1.18
295	14.86	0.62	-16.05	7.49	A	A	A	2.58
296	16.6	0.6	-6.21	7.19	A	A	A	1.00
297	0.14	0.09	-99.21	64.59	N	N	N	15.96
298	17.77	0.25	0.40	6.37	A	A	A	0.06
299	14.36	0.07	-18.87	6.23	A	N	W	3.04
301	18.8	0.5	6.21	6.76	A	A	A	1.00
303	17.66	0.26	-0.23	6.39	A	A	A	0.04
305	11.611	1.346	-34.40	13.15	N	N	N	5.54
306	16.0	0.5	-9.60	6.96	A	A	A	1.55
307	112.00	25.1	532.77	23.26	N	N	N	85.73
309	18.38	1.71	3.84	11.19	A	A	A	0.62
310	17.5	1.9	-1.13	12.51	A	A	A	0.18
312	17.40	0.65	-1.69	7.25	A	A	A	0.27
313	18.1	1.2	2.26	9.09	A	A	A	0.36

Reported Results for Sr-90 Determination in Sample 1

TABLE 3. Target values

Parameter	Value
Target Value, Bq/kg	26.4
UNC, Bq/kg	1.6
MARB, %	30




FIGURE 3. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 4. Reported results and evaluation parameters for Sr-90. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 4. Evaluation Results for Sr-90 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
4	27.33	1.24	3.52	7.57	A	A	A	0.58
7	23.7	0.6	-10.23	6.57	A	A	A	1.69
8	27.0	1.4	2.27	7.98	A	A	A	0.38
9	7.15	2.18	-72.92	31.09	N	N	N	12.03
10	7.07	0.36	-73.22	7.92	N	N	N	12.08
13	29.4	1.6	11.36	8.15	A	A	A	1.88
14	26.33	0.56	-0.27	6.42	A	A	A	0.04
16	27.6	1.67	4.55	8.56	A	A	A	0.75
17	25.7	2.8	-2.65	12.47	A	A	A	0.44
23	26.88	1.24	1.82	7.62	A	A	A	0.30
24	26.3	1.1	-0.38	7.36	A	A	A	0.06
27	26.9	1.48	1.89	8.19	A	A	A	0.31
31	27.2	1.6	3.03	8.45	A	A	A	0.50
36	25	2	-5.30	10.04	A	A	A	0.87
41	26.4	1.5	0.00	8.31	A	A	A	0.00
42	25.28	2.836	-4.24	12.75	A	A	A	0.70
45	24.4	1.9	-7.58	9.87	A	A	A	1.25
47	28.1	2.0	6.44	9.35	A	A	A	1.06
49	25.4	1.0	-3.79	7.23	A	A	A	0.62
50	27.0	1.1	2.27	7.30	A	A	A	0.38
53	25.72	0.78	-2.58	6.78	A	A	A	0.42
57	26.14	0.63	-0.98	6.52	A	A	A	0.16
60	27.6	2.5	4.55	10.90	A	A	A	0.75
61	27.0	3.1	2.27	12.98	A	A	A	0.38
62	30.7	1.4	16.29	7.58	A	A	A	2.69
65	27.15	1.4	2.84	7.96	A	A	A	0.47
68	27.32	1.34	3.48	7.80	A	A	A	0.58
69	28.78	0.27	9.02	6.13	A	A	A	1.49
74	27.1	1.4	2.65	7.96	A	A	A	0.44
76	35.1	0.71	32.95	6.39	N	N	N	5.44
78	47.42	4.87	79.62	11.92	N	N	N	13.14
80	26.28	1.20	-0.45	7.59	A	A	A	0.07
81	29.0	0.2	9.85	6.10	A	A	A	1.63
83	20.8	0.7	-21.21	6.93	A	N	W	3.50
87	25.84	1.42	-2.12	8.18	A	A	A	0.35
88	27.0	3.0	2.27	12.66	A	A	A	0.38
90	26.5	0.87	0.38	6.89	A	A	A	0.06
92	26.4	1.4	0.00	8.05	A	A	A	0.00
93	24.92	0.55	-5.61	6.45	A	A	A	0.92
94	28.8	3.7	9.09	14.21	A	A	A	1.50
96	24.96	0.25	-5.45	6.14	A	A	A	0.90
98	26.8	3.3	1.52	13.72	A	A	A	0.25
99	26.47	2.65	0.27	11.70	A	A	A	0.04
103	30.1	2.24	14.02	9.60	A	A	A	2.31
110	24.3	0.756	-7.95	6.81	A	A	A	1.31
113	30.81	0.98	16.70	6.84	A	A	A	2.76
118	25	1	-5.30	7.26	A	A	A	0.87
119	31.560	5.523	19.55	18.52	A	A	A	3.22
121	13.37	0.06	-49.36	6.08	N	N	N	8.14
123	26.3	1.8	-0.38	9.14	A	A	A	0.06
128	27.6	2.5	4.55	10.90	A	A	A	0.75
133	31.6	5.5	19.70	18.43	A	A	A	3.25
134	40.5	2.2	53.41	8.14	N	N	N	8.81

TABLE 4. Evaluation Results for Sr-90 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
137	26.40	1.39	0.00	8.03	A	A	A	0.00
138	27.47	0.08	4.05	6.07	A	A	A	0.67
140	8.5	1.3	-67.80	16.45	N	N	N	11.19
143	24.3	0.67	-7.95	6.66	A	A	A	1.31
149	26.6	0.4	0.76	6.24	A	A	A	0.13
156	26.8	1.9	1.52	9.33	A	A	A	0.25
157	25.7	2.3	-2.65	10.81	A	A	A	0.44
159	22.5	2.0	-14.77	10.76	A	A	A	2.44
164	24.6	1.2	-6.82	7.78	A	A	A	1.12
171	27.3	0.9	3.41	6.90	A	A	A	0.56
177	23.4	2.5	-11.36	12.28	A	A	A	1.88
184	27.8	2.4	5.30	10.55	A	A	A	0.88
186	27.0	4.8	2.27	18.78	A	A	A	0.38
187	25.6	1.8	-3.03	9.28	A	A	A	0.50
188	26.3	1.0	-0.38	7.15	A	A	A	0.06
192	29.07	2.10	10.11	9.43	A	A	A	1.67
196	28.6	2.55	8.33	10.78	A	A	A	1.38
199	25.70	1.53	-2.65	8.50	A	A	A	0.44
203	28.8	2.6	9.09	10.87	A	A	A	1.50
204	28.3	1.9	7.20	9.04	A	A	A	1.19
211	27.0	0.7	2.27	6.59	A	A	A	0.38
214	28.57	0.5	8.22	6.31	A	A	A	1.36
216	23.0	0.9	-12.88	7.21	A	A	A	2.12
217	26.9	1.3	1.89	7.75	A	A	A	0.31
228	26.3	2.6	-0.38	11.60	A	A	A	0.06
230	26.5	1.06	0.38	7.26	A	A	A	0.06
234	23.60	3.00	-10.61	14.08	A	A	A	1.75
236	32.58	1.46	23.41	7.54	A	N	W	3.86
239	23.91	3.47	-9.43	15.73	A	A	A	1.56
240	26.9	1.60	1.89	8.49	A	A	A	0.31
241	27.24	0.12	3.18	6.08	A	A	A	0.52
247	25.0	1.2	-5.30	7.73	A	A	A	0.87
248	28.1	2.3	6.44	10.18	A	A	A	1.06
250	22.60	2.00	-14.39	10.73	A	A	A	2.37
258	17.30	7.06	-34.47	41.26	N	N	N	5.69
261	27.4	1.4	3.79	7.93	A	A	A	0.62
262	21.7	2.81	-17.80	14.30	A	A	A	2.94
272	20.74	0.48	-21.44	6.49	A	N	W	3.54
273	27.6	1.5	4.55	8.14	A	A	A	0.75
276	24.04	1.41	-8.94	8.43	A	A	A	1.47
281	30.0	2.0	13.64	9.01	A	A	A	2.25
284	24.15	1.77	-8.52	9.51	A	A	A	1.41
309	27	2.65	2.27	11.54	A	A	A	0.38

Reported Results for Cs-134 Determination in Sample 1

TABLE 5. Target values

Parameter	Value
Target Value, Bq/kg	15.9
UNC, Bq/kg	1
MARB, %	20




FIGURE 5. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 6. Reported results and evaluation parameters for Cs-134. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 6. Evaluation Results for Cs-134 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
1	14.5	0.52	-8.81	7.24	A	A	A	1.40
4	15.68	0.50	-1.38	7.05	A	A	A	0.22
6	15.27	1.04	-3.96	9.27	A	A	A	0.63
7	16.1	0.7	1.26	7.65	A	A	A	0.20
8	16.0	1.2	0.63	9.79	A	A	A	0.10
9	16.61	0.51	4.47	7.00	A	A	A	0.71
10	54.34	8.56	241.76	16.96	N	N	N	38.44
11	14.86	0.95	-6.54	8.97	A	A	A	1.04
12	16.8	0.6	5.66	7.23	A	A	A	0.90
13	16	1	0.63	8.87	A	A	A	0.10
14	16.05	0.60	0.94	7.32	A	A	A	0.15
16	15.29	0.29	-3.84	6.57	A	A	A	0.61
17	16.3	2.1	2.52	14.34	A	A	A	0.40
18	15.68	1.27	-1.38	10.25	A	A	A	0.22
19	16.17	1.52	1.70	11.31	A	A	A	0.27
20	16.7	0.8	5.03	7.91	A	A	A	0.80
21	15.4	0.7	-3.14	7.76	A	A	A	0.50
22	16.062	0.193	1.02	6.40	A	A	A	0.16
23	15.91	0.39	0.06	6.75	A	A	A	0.01
24	15.9	0.6	0.00	7.33	A	A	A	0.00
25	12.99	0.69	-18.30	8.23	A	A	A	2.91
26	15.4	0.5	-3.14	7.08	A	A	A	0.50
27	16.8	0.69	5.66	7.51	A	A	A	0.90
28	15.5	0.8	-2.52	8.14	A	A	A	0.40
29	15.74	0.65	-1.01	7.52	A	A	A	0.16
31	15.9	1.3	0.00	10.32	A	A	A	0.00
32	14.9	0.3	-6.29	6.60	A	A	A	1.00
33	13.66	0.72	-14.09	8.21	A	A	A	2.24
34	16.66	0.98	4.78	8.61	A	A	A	0.76
35	16.7	0.4	5.03	6.73	A	A	A	0.80
36	16	4	0.63	25.78	A	N	W	0.10
37	15.8	0.9	-0.63	8.49	A	A	A	0.10
38	16	1	0.63	8.87	A	A	A	0.10
39	16.0	0.9	0.63	8.44	A	A	A	0.10
40	15.7	0.3	-1.26	6.57	A	A	A	0.20
41	15.7	1.19	-1.26	9.85	A	A	A	0.20
42	7.76	0.30	-51.19	7.38	N	N	N	8.14
43	16.3	1.1	2.52	9.22	A	A	A	0.40
45	15.35	0.66	-3.46	7.62	A	A	A	0.55
46	15.3	0.6	-3.77	7.41	A	A	A	0.60
47	15.7	1.1	-1.26	9.42	A	A	A	0.20
49	17.3	0.7	8.81	7.48	A	A	A	1.40
50	17.1	0.2	7.55	6.40	A	A	A	1.20
51	16.3	0.3	2.52	6.55	A	A	A	0.40
52	13.05	1.14	-17.92	10.76	A	A	A	2.85
53	16.93	0.42	6.48	6.76	A	A	A	1.03
55	15.2	0.54	-4.40	7.22	A	A	A	0.70
56	14.6	0.9	-8.18	8.81	A	A	A	1.30
57	15.78	0.76	-0.75	7.92	A	A	A	0.12
58	15.99	0.24	0.57	6.47	A	A	A	0.09
59	16.0	0.9	0.63	8.44	A	A	A	0.10
60	16.1	1.4	1.26	10.73	A	A	A	0.20
61	16.2	0.61	1.89	7.33	A	A	A	0.30

TABLE 6. Evaluation Results for Cs-134 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
62	15.9	0.6	0.00	7.33	A	A	A	0.00
63	14.96	0.64	-5.91	7.61	A	A	A	0.94
64	15.6	0.8	-1.89	8.12	A	A	A	0.30
65	16.07	0.20	1.07	6.41	A	A	A	0.17
67	15.78	1.32	-0.75	10.47	A	A	A	0.12
68	16.1	0.8	1.26	8.02	A	A	A	0.20
69	14.66	0.13	-7.80	6.35	A	A	A	1.24
70	16.5	0.4	3.77	6.74	A	A	A	0.60
71	15.52	0.24	-2.39	6.48	A	A	A	0.38
72	14.8	0.9	-6.92	8.75	A	A	A	1.10
74	15.78	0.98	-0.75	8.84	A	A	A	0.12
75	15.6	0.6	-1.89	7.37	A	A	A	0.30
76	16.0	0.60	0.63	7.32	A	A	A	0.10
77	15.4	0.812	-3.14	8.21	A	A	A	0.50
78	15.6	0.5	-1.89	7.06	A	A	A	0.30
79	16.4	1.0	3.14	8.76	A	A	A	0.50
80	13.45	0.38	-15.41	6.89	A	A	A	2.45
81	16.0	0.8	0.63	8.03	A	A	A	0.10
83	15.70	0.65	-1.26	7.53	A	A	A	0.20
84	15.885	0.736	-0.09	7.81	A	A	A	0.02
85	15.7	0.52	-1.26	7.11	A	A	A	0.20
86	14.5	1.05	-8.81	9.59	A	A	A	1.40
87	15.6	0.55	-1.89	7.21	A	A	A	0.30
88	16.9	0.6	6.29	7.22	A	A	A	1.00
89	16.3	1.1	2.52	9.22	A	A	A	0.40
90	15.2	0.50	-4.40	7.10	A	A	A	0.70
91	15.3	1.1	-3.77	9.55	A	A	A	0.60
92	16.10	0.50	1.26	7.01	A	A	A	0.20
93	16.2	0.25	1.89	6.48	A	A	A	0.30
94	17.6	0.9	10.69	8.11	A	A	A	1.70
95	15.71	0.33	-1.19	6.63	A	A	A	0.19
96	15.83	0.21	-0.44	6.43	A	A	A	0.07
97	15.64	0.70	-1.64	7.72	A	A	A	0.26
98	16.0	0.8	0.63	8.03	A	A	A	0.10
99	15.58	1.87	-2.01	13.55	A	A	A	0.32
101	15.5	1.5	-2.52	11.54	A	A	A	0.40
102	15.8	0.765	-0.63	7.94	A	A	A	0.10
103	16.0	0.81	0.63	8.07	A	A	A	0.10
105	14.58	1.499	-8.30	12.05	A	A	A	1.32
106	16.00	0.596	0.63	7.31	A	A	A	0.10
107	16.7	0.62	5.03	7.30	A	A	A	0.80
109	13.90	0.53	-12.58	7.35	A	A	A	2.00
110	15.9	0.993	0.00	8.86	A	A	A	0.00
111	15.9	0.98	0.00	8.81	A	A	A	0.00
113	14.83	0.16	-6.73	6.38	A	A	A	1.07
114	16.3	0.7	2.52	7.62	A	A	A	0.40
115	15.90	0.86	0.00	8.30	A	A	A	0.00
117	15.6	0.6	-1.89	7.37	A	A	A	0.30
118	15	1	-5.66	9.17	A	A	A	0.90
119	16.442	1.726	3.41	12.24	A	A	A	0.54
120	15.5	0.7	-2.52	7.74	A	A	A	0.40
121	15.431	0.47	-2.95	6.99	A	A	A	0.47
123	16.4	0.9	3.14	8.35	A	A	A	0.50
124	12.41	0.25	-21.95	6.60	N	N	N	3.49

TABLE 6. Evaluation Results for Cs-134 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
125	19.96	1.40	25.53	9.42	N	N	N	4.06
126	16.6	1.2	4.40	9.58	A	A	A	0.70
127	15.660	1.25	-1.51	10.16	A	A	A	0.24
128	16.2	0.7	1.89	7.63	A	A	A	0.30
129	15.3	0.9	-3.77	8.61	A	A	A	0.60
130	15.5	0.15	-2.52	6.36	A	A	A	0.40
132	14.3	0.4	-10.06	6.88	A	A	A	1.60
133	16.4	1.7	3.14	12.12	A	A	A	0.50
134	14.4	0.6	-9.43	7.54	A	A	A	1.50
135	5.9	0.6	-62.89	11.96	N	N	N	10.00
136	16.1	1.3	1.26	10.23	A	A	A	0.20
137	13.10	0.29	-17.61	6.67	A	N	W	2.80
138	13.481	0.405	-15.21	6.97	A	A	A	2.42
139	14.8	0.26	-6.92	6.53	A	A	A	1.10
140	15.6	1.0	-1.89	8.98	A	A	A	0.30
143	14.9	0.67	-6.29	7.73	A	A	A	1.00
145	16.3	0.7	2.52	7.62	A	A	A	0.40
146	14.376	0.0359	-9.58	6.29	A	A	A	1.52
147	16.0	0.55	0.63	7.17	A	A	A	0.10
148	15.01	0.84	-5.60	8.42	A	A	A	0.89
149	13.5	0.5	-15.09	7.30	A	A	A	2.40
150	15.764	1.039	-0.86	9.11	A	A	A	0.14
152	15.0	0.8	-5.66	8.25	A	A	A	0.90
153	11.5	0.2	-27.67	6.53	N	N	N	4.40
154	15.4	0.66	-3.14	7.61	A	A	A	0.50
155	12.41	0.77	-21.95	8.83	N	A	N	3.49
156	15.9	1.2	0.00	9.82	A	A	A	0.00
157	15.9	1.2	0.00	9.82	A	A	A	0.00
158	15.876	0.953	-0.15	8.69	A	A	A	0.02
159	15.7	1.5	-1.26	11.44	A	A	A	0.20
160	13.2	0.3	-16.98	6.69	A	A	A	2.70
161	14.62	0.8	-8.05	8.34	A	A	A	1.28
162	14.3	0.9	-10.06	8.90	A	A	A	1.60
163	15.37	0.75	-3.33	7.96	A	A	A	0.53
164	14.95	0.75	-5.97	8.05	A	A	A	0.95
165	16.22	0.51	2.01	7.03	A	A	A	0.32
166	16.3	0.5	2.52	7.00	A	A	A	0.40
167	15.5	1.0	-2.52	9.01	A	A	A	0.40
168	17.08	0.38	7.42	6.67	A	A	A	1.18
169	14.76	0.44	-7.17	6.96	A	A	A	1.14
171	15.9	1.0	0.00	8.89	A	A	A	0.00
172	15.9	0.72	0.00	7.75	A	A	A	0.00
173	16.27	0.39	2.33	6.73	A	A	A	0.37
175	16.0	0.8	0.63	8.03	A	A	A	0.10
177	15.5	0.9	-2.52	8.56	A	A	A	0.40
178	15.8	0.7	-0.63	7.69	A	A	A	0.10
179	15.437	0.40893	-2.91	6.82	A	A	A	0.46
180	15.9	0.8	0.00	8.05	A	A	A	0.00
182	15.9	0.4	0.00	6.77	A	A	A	0.00
183	14.56	0.33	-8.43	6.69	A	A	A	1.34
184	15.8	0.6	-0.63	7.35	A	A	A	0.10
185	14.1	1.9	-11.32	14.87	A	A	A	1.80
187	15.9	0.8	0.00	8.05	A	A	A	0.00
188	15.1	1.05	-5.03	9.38	A	A	A	0.80

TABLE 6. Evaluation Results for Cs-134 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
189	16.22	1.33	2.01	10.33	A	A	A	0.32
191	16.2	0.9	1.89	8.39	A	A	A	0.30
192	15.34	1.48	-3.52	11.52	A	A	A	0.56
193	1.695	0.094	-89.34	8.39	N	N	N	14.20
194	44.9	5	182.39	12.79	N	N	N	29.00
195	13.7	0.6	-13.84	7.66	A	A	A	2.20
196	16.2	0.843	1.89	8.16	A	A	A	0.30
197	13.5	3.2	-15.09	24.52	A	N	W	2.40
198	15.91	0.89	0.06	8.42	A	A	A	0.01
201	16.1	0.6	1.26	7.31	A	A	A	0.20
202	15.6	0.8	-1.89	8.12	A	A	A	0.30
203	15.7	0.5	-1.26	7.05	A	A	A	0.20
204	15.5	1.0	-2.52	9.01	A	A	A	0.40
206	18.32	0.08	15.22	6.30	A	A	A	2.42
207	15.71	0.58	-1.19	7.29	A	A	A	0.19
209	14.88	3	-6.42	21.12	A	N	W	1.02
210	16.2	0.9	1.89	8.39	A	A	A	0.30
213	15.7	1.0	-1.26	8.95	A	A	A	0.20
214	16.0	0.3	0.63	6.56	A	A	A	0.10
215	14.8	0.85	-6.92	8.52	A	A	A	1.10
216	15.5	0.2	-2.52	6.42	A	A	A	0.40
217	15.0	1.0	-5.66	9.17	A	A	A	0.90
218	15.51	0.75	-2.45	7.93	A	A	A	0.39
219	14.4	0.35	-9.43	6.74	A	A	A	1.50
220	16.0	0.75	0.63	7.84	A	A	A	0.10
221	16.53	0.43	3.96	6.81	A	A	A	0.63
222	15.1	1.0	-5.03	9.13	A	A	A	0.80
223	15.33	0.09	-3.58	6.32	A	A	A	0.57
224	15.18	0.09	-4.53	6.32	A	A	A	0.72
226	13.36	0.585	-15.97	7.66	A	A	A	2.54
227	16.12	0.36	1.38	6.67	A	A	A	0.22
228	16.1	2.3	1.26	15.61	A	A	A	0.20
230	15.5	1.1	-2.52	9.48	A	A	A	0.40
231	10.6	1.1	-33.33	12.13	N	N	N	5.30
233	15.94	0.91	0.25	8.49	A	A	A	0.04
234	15.90	0.40	0.00	6.77	A	A	A	0.00
235	15.4	0.32	-3.14	6.62	A	A	A	0.50
237	16.06	0.22	1.01	6.44	A	A	A	0.16
238	15.7	0.5	-1.26	7.05	A	A	A	0.20
239	17.55	2.54	10.38	15.78	A	A	A	1.65
241	16.01	0.55	0.69	7.17	A	A	A	0.11
243	16.61	0.92	4.47	8.38	A	A	A	0.71
244	16.0	1.6	0.63	11.81	A	A	A	0.10
246	14.6	0.4	-8.18	6.86	A	A	A	1.30
247	16.00	0.82	0.63	8.11	A	A	A	0.10
248	16.54	0.56	4.03	7.14	A	A	A	0.64
249	8.2	7	-48.43	85.60	N	N	N	7.70
250	15.72	0.31	-1.13	6.59	A	A	A	0.18
251	15.68	0.65	-1.38	7.53	A	A	A	0.22
253	15.58	0.22	-2.01	6.45	A	A	A	0.32
255	17.0	0.5	6.92	6.94	A	A	A	1.10
256	16.1	0.48	1.26	6.96	A	A	A	0.20
258	15.6	0.296	-1.89	6.57	A	A	A	0.30
259	14.45	0.89	-9.12	8.80	A	A	A	1.45

TABLE 6. Evaluation Results for Cs-134 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
261	15.5	1.1	-2.52	9.48	A	A	A	0.40
262	16.4	2.13	3.14	14.43	A	A	A	0.50
263	15.421	1.058	-3.01	9.31	A	A	A	0.48
264	15.5	1.5	-2.52	11.54	A	A	A	0.40
265	13.21	0.33	-16.92	6.77	A	A	A	2.69
266	15.44	0.66	-2.89	7.60	A	A	A	0.46
269	14.9	0.87	-6.29	8.58	A	A	A	1.00
271	15.9	1.4	0.00	10.82	A	A	A	0.00
272	16.22	1.50	2.01	11.18	A	A	A	0.32
273	15.8	0.7	-0.63	7.69	A	A	A	0.10
274	18.24	2.98	14.72	17.51	A	A	A	2.34
276	14.15	0.96	-11.01	9.25	A	A	A	1.75
278	15.3	0.7	-3.77	7.78	A	A	A	0.60
281	15.1	0.8	-5.03	8.22	A	A	A	0.80
284	11.13	0.50	-30.00	7.73	N	N	N	4.77
285	13.71	0.94	-13.77	9.30	A	A	A	2.19
286	15.9	0.7	0.00	7.68	A	A	A	0.00
287	13.215	0.49	-16.89	7.30	A	A	A	2.69
288	15.6	1.6	-1.89	12.03	A	A	A	0.30
289	15.8	1.0	-0.63	8.92	A	A	A	0.10
290	16.76	0.80	5.41	7.90	A	A	A	0.86
292	16.0	2.2	0.63	15.12	A	A	A	0.10
295	15.06	0.43	-5.28	6.91	A	A	A	0.84
296	15.5	0.6	-2.52	7.39	A	A	A	0.40
297	1.27	0.28	-92.01	22.93	N	N	N	14.63
298	15.71	0.19	-1.19	6.40	A	A	A	0.19
299	15.60	0.79	-1.89	8.07	A	A	A	0.30
301	15.2	0.4	-4.40	6.82	A	A	A	0.70
303	14.24	0.26	-10.44	6.55	A	A	A	1.66
304	14.7	0.23	-7.55	6.48	A	A	A	1.20
305	13.213	1.739	-16.90	14.59	A	A	A	2.69
306	15.7	0.5	-1.26	7.05	A	A	A	0.20
307	12.50	0.23	-21.38	6.55	N	N	N	3.40
309	15.55	1.524	-2.20	11.65	A	A	A	0.35
310	15.9	1.7	0.00	12.40	A	A	A	0.00
312	16.52	0.68	3.90	7.52	A	A	A	0.62
313	16.0	1.0	0.63	8.87	A	A	A	0.10

Reported Results for Cs-137 Determination in Sample 1

TABLE 7. Target values

Parameter	Value
Target Value, Bq/kg	24.2
UNC, Bq/kg	1.5
MARB, %	20




FIGURE 7. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 8. Reported results and evaluation parameters for Cs-137. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 8. Evaluation Results for Cs-137 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
1	24.5	0.92	1.24	7.25	A	A	A	0.20
4	23.04	0.30	-4.79	6.33	A	A	A	0.77
5	23.5	0.72	-2.89	6.91	A	A	A	0.47
6	21.98	0.94	-9.17	7.53	A	A	A	1.48
7	23.7	1.1	-2.07	7.74	A	A	A	0.33
8	24.3	1.5	0.41	8.75	A	A	A	0.07
9	26.11	1.00	7.89	7.29	A	A	A	1.27
10	86.09	11.41	255.74	14.63	N	N	N	41.26
11	23.8	1.2	-1.65	7.99	A	A	A	0.27
12	25.5	1.2	5.37	7.78	A	A	A	0.87
13	23	1	-4.96	7.57	A	A	A	0.80
14	24.66	0.93	1.90	7.26	A	A	A	0.31
16	24.37	0.51	0.70	6.54	A	A	A	0.11
17	23.2	2.5	-4.13	12.43	A	A	A	0.67
18	23.57	1.41	-2.60	8.61	A	A	A	0.42
19	24.65	2.36	1.86	11.41	A	A	A	0.30
20	24.2	0.6	0.00	6.68	A	A	A	0.00
21	24.5	1.5	1.24	8.71	A	A	A	0.20
22	24.825	0.371	2.58	6.38	A	A	A	0.42
23	24.29	0.76	0.37	6.94	A	A	A	0.06
24	24.3	0.9	0.41	7.22	A	A	A	0.07
25	22.05	0.29	-8.88	6.34	A	A	A	1.43
26	24.9	1.5	2.89	8.64	A	A	A	0.47
27	24.3	0.95	0.41	7.33	A	A	A	0.07
28	24.3	1.2	0.41	7.93	A	A	A	0.07
29	24.27	0.61	0.29	6.69	A	A	A	0.05
31	24.7	1.7	2.07	9.26	A	A	A	0.33
32	25.2	0.5	4.13	6.51	A	A	A	0.67
33	24.25	1.60	0.21	9.05	A	A	A	0.03
34	24.04	1.44	-0.66	8.62	A	A	A	0.11
35	24.5	1.2	1.24	7.90	A	A	A	0.20
36	25	4	3.31	17.16	A	A	A	0.53
37	24.6	1.4	1.65	8.41	A	A	A	0.27
38	23	1.7	-4.96	9.65	A	A	A	0.80
39	24.0	1.2	-0.83	7.96	A	A	A	0.13
40	24.2	0.7	0.00	6.84	A	A	A	0.00
41	24.2	1.92	0.00	10.07	A	A	A	0.00
42	12.58	0.50	-48.02	7.36	N	N	N	7.75
43	24.4	0.9	0.83	7.21	A	A	A	0.13
45	23.50	0.36	-2.89	6.38	A	A	A	0.47
46	23.4	1.0	-3.31	7.53	A	A	A	0.53
47	23.5	1.3	-2.89	8.31	A	A	A	0.47
49	25.1	0.8	3.72	6.97	A	A	A	0.60
50	25.2	0.4	4.13	6.40	A	A	A	0.67
51	24.6	0.5	1.65	6.52	A	A	A	0.27
52	23.71	1.89	-2.02	10.10	A	A	A	0.33
53	29.08	0.70	20.17	6.65	N	N	N	3.25
55	23.5	0.73	-2.89	6.93	A	A	A	0.47
56	22.6	1.2	-6.61	8.16	A	A	A	1.07
57	24.32	1.52	0.50	8.80	A	A	A	0.08
58	23.99	0.34	-0.87	6.36	A	A	A	0.14
59	25.0	0.9	3.31	7.17	A	A	A	0.53
60	24.7	1.5	2.07	8.68	A	A	A	0.33

TABLE 8. Evaluation Results for Cs-137 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
61	23	1.1	-4.96	7.83	A	A	A	0.80
62	23.9	0.9	-1.24	7.25	A	A	A	0.20
63	25.26	1.58	4.38	8.81	A	A	A	0.71
64	24.4	1.8	0.83	9.64	A	A	A	0.13
65	24.93	0.32	3.02	6.33	A	A	A	0.49
67	24.80	1.55	2.48	8.80	A	A	A	0.40
68	24.3	1.0	0.41	7.44	A	A	A	0.07
69	22.59	0.23	-6.65	6.28	A	A	A	1.07
70	24.6	0.7	1.65	6.82	A	A	A	0.27
71	23.68	0.27	-2.15	6.30	A	A	A	0.35
72	23.4	0.6	-3.31	6.71	A	A	A	0.53
74	24.07	1.46	-0.54	8.67	A	A	A	0.09
75	24.2	0.5	0.00	6.53	A	A	A	0.00
76	24.0	0.60	-0.83	6.68	A	A	A	0.13
77	24.6	1.14	1.65	7.74	A	A	A	0.27
78	23.4	0.7	-3.31	6.88	A	A	A	0.53
79	26.0	1.2	7.44	7.73	A	A	A	1.20
80	23.75	0.78	-1.86	7.01	A	A	A	0.30
81	24.3	1.0	0.41	7.44	A	A	A	0.07
83	24.21	0.6	0.04	6.68	A	A	A	0.01
84	24.837	1.103	2.63	7.63	A	A	A	0.42
85	23.9	0.78	-1.24	7.01	A	A	A	0.20
86	23.4	1.71	-3.31	9.58	A	A	A	0.53
87	25.6	1.25	5.79	7.89	A	A	A	0.93
88	25.0	1.2	3.31	7.84	A	A	A	0.53
89	23.4	1.5	-3.31	8.92	A	A	A	0.53
90	23.5	1.44	-2.89	8.72	A	A	A	0.47
91	23.3	1.5	-3.72	8.94	A	A	A	0.60
92	24.22	0.79	0.08	7.00	A	A	A	0.01
93	24.0	0.57	-0.83	6.64	A	A	A	0.13
95	25.07	0.39	3.60	6.39	A	A	A	0.58
96	24.69	0.42	2.02	6.43	A	A	A	0.33
97	23.67	0.29	-2.19	6.32	A	A	A	0.35
98	25.0	1.3	3.31	8.09	A	A	A	0.53
99	24.10	2.41	-0.41	11.77	A	A	A	0.07
100	23.74	0.3	-1.90	6.33	A	A	A	0.31
101	25.0	2.5	3.31	11.77	A	A	A	0.53
102	24.5	0.896	1.24	7.20	A	A	A	0.20
103	24.7	1.23	2.07	7.95	A	A	A	0.33
105	23.17	1.53	-4.26	9.06	A	A	A	0.69
106	22.44	0.829	-7.27	7.22	A	A	A	1.17
107	24.8	1.44	2.48	8.49	A	A	A	0.40
109	23.70	0.90	-2.07	7.27	A	A	A	0.33
110	24.8	1.47	2.48	8.58	A	A	A	0.40
111	24.2	0.79	0.00	7.01	A	A	A	0.00
113	22.94	0.32	-5.21	6.35	A	A	A	0.84
114	25.2	1.0	4.13	7.36	A	A	A	0.67
115	23.8	1.2	-1.65	7.99	A	A	A	0.27
117	23.5	1.5	-2.89	8.90	A	A	A	0.47
118	26	3	7.44	13.10	A	A	A	1.20
119	23.803	2.499	-1.64	12.19	A	A	A	0.26
120	23.7	0.7	-2.07	6.87	A	A	A	0.33
121	23.844	0.55	-1.47	6.61	A	A	A	0.24
123	25.0	1.2	3.31	7.84	A	A	A	0.53

TABLE 8. Evaluation Results for Cs-137 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
124	23.03	0.42	-4.83	6.46	A	A	A	0.78
125	26.95	1.89	11.36	9.36	A	A	A	1.83
126	25.4	1.8	4.96	9.41	A	A	A	0.80
127	23.62	1.89	-2.40	10.12	A	A	A	0.39
128	24.5	0.8	1.24	7.01	A	A	A	0.20
129	23.1	1.2	-4.55	8.09	A	A	A	0.73
130	23.8	0.2	-1.65	6.26	A	A	A	0.27
132	22.5	1.4	-7.02	8.78	A	A	A	1.13
133	23.8	2.5	-1.65	12.20	A	A	A	0.27
134	24.9	0.7	2.89	6.81	A	A	A	0.47
135	21.2	3.3	-12.40	16.75	A	A	A	2.00
136	23.6	1.3	-2.48	8.29	A	A	A	0.40
137	23.10	0.26	-4.55	6.30	A	A	A	0.73
138	24.076	0.796	-0.51	7.02	A	A	A	0.08
139	23.8	0.32	-1.65	6.34	A	A	A	0.27
140	23.9	1.9	-1.24	10.08	A	A	A	0.20
143	25.6	1.16	5.79	7.68	A	A	A	0.93
145	24.7	0.8	2.07	6.99	A	A	A	0.33
146	24.85	0.0368	2.69	6.20	A	A	A	0.43
147	25.41	0.79	5.00	6.93	A	A	A	0.81
148	24.09	0.60	-0.45	6.68	A	A	A	0.07
149	21.8	1.1	-9.92	7.99	A	A	A	1.60
150	24.051	1.574	-0.62	9.01	A	A	A	0.10
152	22.6	1.2	-6.61	8.16	A	A	A	1.07
153	24.3	0.7	0.41	6.84	A	A	A	0.07
154	23.62	1.08	-2.40	7.70	A	A	A	0.39
155	21.86	1.38	-9.67	8.85	A	A	A	1.56
156	24.8	1.4	2.48	8.38	A	A	A	0.40
157	24.4	1.8	0.83	9.64	A	A	A	0.13
158	25.505	1.020	5.39	7.38	A	A	A	0.87
159	24.2	2.0	0.00	10.33	A	A	A	0.00
160	22.2	0.6	-8.26	6.76	A	A	A	1.33
161	23.35	1.1	-3.51	7.79	A	A	A	0.57
162	22.7	1.4	-6.20	8.74	A	A	A	1.00
163	23.91	1.22	-1.20	8.03	A	A	A	0.19
164	23.9	1.6	-1.24	9.12	A	A	A	0.20
165	24.19	0.73	-0.04	6.89	A	A	A	0.01
166	23.6	0.7	-2.48	6.87	A	A	A	0.40
167	24.1	1.4	-0.41	8.50	A	A	A	0.07
168	26.55	0.77	9.71	6.84	A	A	A	1.57
169	24.72	0.98	2.15	7.36	A	A	A	0.35
171	25.8	1.8	6.61	9.33	A	A	A	1.07
172	23.97	0.88	-0.95	7.20	A	A	A	0.15
173	24.67	0.52	1.94	6.55	A	A	A	0.31
175	24.4	1.2	0.83	7.91	A	A	A	0.13
177	23.6	1.0	-2.48	7.51	A	A	A	0.40
178	23.1	1.5	-4.55	8.98	A	A	A	0.73
179	25.710	0.66347	6.24	6.71	A	A	A	1.01
180	25.6	1.1	5.79	7.54	A	A	A	0.93
182	24.3	1.0	0.41	7.44	A	A	A	0.07
183	23.19	0.70	-4.17	6.89	A	A	A	0.67
184	24.2	0.9	0.00	7.23	A	A	A	0.00
185	21.5	3.4	-11.16	16.99	A	A	A	1.80
186	23.8	6.6	-1.65	28.42	A	N	W	0.27

TABLE 8. Evaluation Results for Cs-137 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
187	24.1	1.2	-0.41	7.95	A	A	A	0.07
188	23.8	1.48	-1.65	8.78	A	A	A	0.27
189	24.44	2.06	0.99	10.46	A	A	A	0.16
191	24.2	0.7	0.00	6.84	A	A	A	0.00
192	24.00	1.04	-0.83	7.56	A	A	A	0.13
193	2.798	0.184	-88.44	9.04	N	N	N	14.27
194	90	5	271.90	8.32	N	N	N	43.87
195	21.0	0.8	-13.22	7.28	A	A	A	2.13
196	24	0.813	-0.83	7.06	A	A	A	0.13
197	21.9	5.2	-9.50	24.54	A	N	W	1.53
198	23.43	1.34	-3.18	8.43	A	A	A	0.51
201	24.7	1.5	2.07	8.68	A	A	A	0.33
202	24.9	1.28	2.89	8.05	A	A	A	0.47
203	24.7	0.7	2.07	6.82	A	A	A	0.33
204	24.0	1.4	-0.83	8.51	A	A	A	0.13
206	24.06	0.14	-0.58	6.23	A	A	A	0.09
207	24.29	0.72	0.37	6.87	A	A	A	0.06
209	25.59	5	5.74	20.50	A	N	W	0.93
210	25.2	1.3	4.13	8.06	A	A	A	0.67
213	24.6	1.1	1.65	7.64	A	A	A	0.27
214	22.9	0.7	-5.37	6.91	A	A	A	0.87
215	24.3	1.1	0.41	7.68	A	A	A	0.07
216	24.1	0.2	-0.41	6.25	A	A	A	0.07
217	25.0	1.4	3.31	8.35	A	A	A	0.53
218	23.88	1.06	-1.32	7.62	A	A	A	0.21
219	25.2	0.36	4.13	6.36	A	A	A	0.67
220	24.8	0.9	2.48	7.18	A	A	A	0.40
221	24.00	0.62	-0.83	6.72	A	A	A	0.13
222	23.2	1.0	-4.13	7.55	A	A	A	0.67
223	23.5	0.17	-2.89	6.24	A	A	A	0.47
224	23.59	0.19	-2.52	6.25	A	A	A	0.41
226	23.0	0.945	-4.96	7.44	A	A	A	0.80
227	24.39	0.56	0.79	6.61	A	A	A	0.13
228	24.7	2.5	2.07	11.87	A	A	A	0.33
230	24.5	1.6	1.24	9.00	A	A	A	0.20
231	21.4	2.0	-11.57	11.21	A	A	A	1.87
233	25.7	1.1	6.20	7.53	A	A	A	1.00
234	24.20	0.50	0.00	6.53	A	A	A	0.00
235	24.0	0.46	-0.83	6.49	A	A	A	0.13
237	24.21	0.35	0.04	6.36	A	A	A	0.01
238	23.3	0.7	-3.72	6.89	A	A	A	0.60
239	22.5	3.26	-7.02	15.76	A	A	A	1.13
241	22.72	1.30	-6.12	8.44	A	A	A	0.99
243	24.57	1.13	1.53	7.72	A	A	A	0.25
244	23.9	1.6	-1.24	9.12	A	A	A	0.20
246	24.6	1.0	1.65	7.41	A	A	A	0.27
247	23.3	1.1	-3.72	7.79	A	A	A	0.60
248	24.37	0.64	0.70	6.73	A	A	A	0.11
249	16.2	4	-33.06	25.46	N	N	N	5.33
250	22.33	0.44	-7.73	6.50	A	A	A	1.25
251	23.61	1.01	-2.44	7.53	A	A	A	0.39
253	23.83	0.35	-1.53	6.37	A	A	A	0.25
255	24.4	0.7	0.83	6.83	A	A	A	0.13
256	25.34	0.78	4.71	6.92	A	A	A	0.76

TABLE 8. Evaluation Results for Cs-137 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
258	24.4	0.307	0.83	6.32	A	A	A	0.13
259	25.42	1.19	5.04	7.77	A	A	A	0.81
261	24.2	1.4	0.00	8.48	A	A	A	0.00
262	25.1	3.36	3.72	14.75	A	A	A	0.60
263	23.970	1.470	-0.95	8.72	A	A	A	0.15
264	23.7	2.2	-2.07	11.16	A	A	A	0.33
265	22.30	0.51	-7.85	6.61	A	A	A	1.27
266	23.87	1.51	-1.36	8.86	A	A	A	0.22
269	24.1	1.41	-0.41	8.52	A	A	A	0.07
271	24.1	1.1	-0.41	7.70	A	A	A	0.07
272	24.65	2.06	1.86	10.40	A	A	A	0.30
273	25.5	1.1	5.37	7.55	A	A	A	0.87
274	27.09	4.948	11.94	19.29	A	A	A	1.93
276	23.22	1.66	-4.05	9.46	A	A	A	0.65
278	24.2	1.3	0.00	8.20	A	A	A	0.00
281	24.4	1.3	0.83	8.17	A	A	A	0.13
284	2.24	0.99	-90.74	44.63	N	N	N	14.64
285	23.52	1.62	-2.81	9.27	A	A	A	0.45
286	24.9	1.0	2.89	7.39	A	A	A	0.47
287	22.52	0.92	-6.94	7.42	A	A	A	1.12
288	21.8	2.2	-9.92	11.84	A	A	A	1.60
289	23.4	1.5	-3.31	8.92	A	A	A	0.53
290	21.47	1.08	-11.28	7.98	A	A	A	1.82
292	25.3	3.5	4.55	15.16	A	A	A	0.73
295	24.26	1.25	0.25	8.06	A	A	A	0.04
296	23.1	0.8	-4.55	7.10	A	A	A	0.73
297	3.64	0.05	-84.96	6.35	N	N	N	13.71
298	24.16	0.29	-0.17	6.31	A	A	A	0.03
299	23.35	1.20	-3.51	8.05	A	A	A	0.57
301	26.4	0.6	9.09	6.60	A	A	A	1.47
303	24.47	0.34	1.12	6.35	A	A	A	0.18
304	24.9	5.9	2.89	24.49	A	N	W	0.47
305	25.587	2.003	5.73	9.98	A	A	A	0.92
306	24.4	0.8	0.83	7.01	A	A	A	0.13
307	23.60	0.44	-2.48	6.47	A	A	A	0.40
309	24.85	2.440	2.69	11.61	A	A	A	0.43
310	27.2	3.1	12.40	12.97	A	A	A	2.00
312	24.4	1.6	0.83	9.02	A	A	A	0.13
313	24.6	1.6	1.65	8.98	A	A	A	0.27

Reported Results for Po-210 Determination in Sample 1

TABLE 9. Target values

Parameter	Value
Target Value, Bq/kg	31.6
UNC, Bq/kg	1.9
MARB, %	30




FIGURE 9. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 10. Reported results and evaluation parameters for Po-210. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 10. Evaluation Results for Po-210 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
4	20.14	1.21	-36.27	8.50	N	N	N	6.03
8	30.9	1.8	-2.22	8.37	A	A	A	0.37
13	30.3	1.5	-4.11	7.79	A	A	A	0.68
14	32.34	1.07	2.34	6.86	A	A	A	0.39
16	33.35	0.8	5.54	6.47	A	A	A	0.92
17	35.1	4.5	11.08	14.16	A	A	A	1.84
24	30.4	1.7	-3.80	8.21	A	A	A	0.63
36	31	3	-1.90	11.39	A	A	A	0.32
42	99.59	10.72	215.16	12.33	N	N	N	35.78
52	27.01	0.0047	-14.53	6.01	A	A	A	2.42
57	32.87	2.75	4.02	10.30	A	A	A	0.67
58	30.9	0.7	-2.22	6.43	A	A	A	0.37
60	18.4	1.2	-41.77	8.87	N	N	N	6.95
62	32.7	1.0	3.48	6.75	A	A	A	0.58
68	32.5	2.27	2.85	9.22	A	A	A	0.47
71	103.5	2.62	227.53	6.52	N	N	N	37.84
78	86.83	5.54	174.78	8.77	N	N	N	29.07
81	45.4	1.4	43.67	6.76	N	N	N	7.26
83	41.0	2.4	29.75	8.39	A	N	W	4.95
84	30.001	1.124	-5.06	7.08	A	A	A	0.84
84	31.902	1.684	0.96	8.00	A	A	A	0.16
92	14.06	1.4	-55.51	11.63	N	N	N	9.23
97	36.2	1.3	14.56	7.00	A	A	A	2.42
99	82.46	15.67	160.95	19.93	N	N	N	26.77
118	28.6	3.6	-9.49	13.95	A	A	A	1.58
119	26.45	4.89	-16.30	19.44	A	A	A	2.71
123	31.98	3.42	1.20	12.27	A	A	A	0.20
128	24.4	3.0	-22.78	13.69	A	A	A	3.79
133	26.5	4.9	-16.14	19.44	A	A	A	2.68
135	105	6	232.28	8.29	N	N	N	38.63
155	90.82	4.99	187.41	8.14	N	N	N	31.17
156	28.0	2.3	-11.39	10.18	A	A	A	1.89
157	80.8	11.1	155.70	15.00	N	N	N	25.89
159	30.3	3.0	-4.11	11.58	A	A	A	0.68
161	100.7	0.7	218.67	6.05	N	N	N	36.37
163	118.6	10.10	275.32	10.42	N	N	N	45.79
163	118.60	10.10	275.32	10.42	N	N	N	45.79
164	27.2	1.6	-13.92	8.41	A	A	A	2.32
171	27.5	2.6	-12.97	11.20	A	A	A	2.16
177	111.4	10.4	252.53	11.10	N	N	N	42.00
186	33.6	7.2	6.33	22.26	A	A	A	1.05
203	17.9	1.1	-43.35	8.60	N	N	N	7.21
214	7.41	0.13	-76.55	6.26	N	N	N	12.73
217	24.8	3.0	-21.52	13.51	A	A	A	3.58
227	90.2	7	185.44	9.82	N	N	N	30.84
247	83.9	12.8	165.51	16.40	N	N	N	27.53
261	31.3	1.8	-0.95	8.32	A	A	A	0.16
272	42.42	2.8	34.24	8.93	N	N	N	5.69

Reported Results for Pb-210 Determination in Sample 1

TABLE 11. Target values

Parameter	Value
Target Value, Bq/kg	31.3
UNC, Bq/kg	1.9
MARB, %	30




FIGURE 11. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 12. Reported results and evaluation parameters for Pb-210. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 12. Evaluation Results for Pb-210 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
4	28.51	0.91	-8.91	6.86	A	A	A	1.47
8	31.6	2.0	0.96	8.77	A	A	A	0.16
8	32.3	3.2	3.19	11.62	A	A	A	0.53
9	62.42	12.77	99.42	21.34	N	N	N	16.38
9	62.42	12.77	99.42	21.34	N	N	N	16.38
13	29	4	-7.35	15.07	A	A	A	1.21
14	33.32	0.67	6.45	6.39	A	A	A	1.06
14	32.09	1.57	2.52	7.80	A	A	A	0.42

TABLE 12. Evaluation Results for Pb-210 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
16	36.01	3.95	15.05	12.54	A	A	A	2.48
17	35.1	4.5	12.14	14.18	A	A	A	2.00
17	34.0	4.4	8.63	14.29	A	A	A	1.42
18	32.71	3.17	4.50	11.44	A	A	A	0.74
20	30.3	1.8	-3.19	8.49	A	A	A	0.53
21	39.5	6.5	26.20	17.54	A	A	A	4.32
24	29.3	1.5	-6.39	7.94	A	A	A	1.05
24	36.5	4.0	16.61	12.53	A	A	A	2.74
28	37.8	1.9	20.77	7.88	A	N	W	3.42
31	31.2	2.9	-0.32	11.10	A	A	A	0.05
31	33.1	2.8	5.75	10.41	A	A	A	0.95
32	31.9	1.5	1.92	7.68	A	A	A	0.32
35	36.8	1.5	17.57	7.31	A	A	A	2.89
36	40	20	27.80	50.37	A	N	W	4.58
41	32.8	2	4.79	8.60	A	A	A	0.79
41	34.1	5.64	8.95	17.62	A	A	A	1.47
42	27.07	4.33	-13.51	17.11	A	A	A	2.23
43	30.0	3.0	-4.15	11.70	A	A	A	0.68
45	34.6	2.4	10.54	9.22	A	A	A	1.74
49	39	1	24.60	6.59	A	N	W	4.05
50	39	2	24.60	7.95	A	N	W	4.05
51	36.7	4.3	17.25	13.20	A	A	A	2.84
52	34.89	2.0	11.47	8.35	A	A	A	1.89
57	32.57	5.08	4.06	16.74	A	A	A	0.67
58	34.5	2.8	10.22	10.13	A	A	A	1.68
59	33.0	5.0	5.43	16.32	A	A	A	0.89
60	30.9	2.7	-1.28	10.64	A	A	A	0.21
60	36.2	3.6	15.65	11.65	A	A	A	2.58
62	34.4	1.0	9.90	6.73	A	A	A	1.63
67	30.59	3.22	-2.27	12.15	A	A	A	0.37
68	31.67	2.21	1.18	9.25	A	A	A	0.19
68	35.1	1.9	12.14	8.13	A	A	A	2.00
69	26.30	3.70	-15.97	15.32	A	A	A	2.63
71	32.89	0.69	5.08	6.42	A	A	A	0.84
74	33.05	5.48	5.59	17.66	A	A	A	0.92
76	45.2	3.40	44.41	9.67	N	N	N	7.32
78	33.7	2.7	7.67	10.05	A	A	A	1.26
81	34.1	3.2	8.95	11.18	A	A	A	1.47
83	37.0	2.0	18.21	8.13	A	A	A	3.00
83	36.12	1.55	15.40	7.43	A	A	A	2.54
84	31.378	1.656	0.25	8.04	A	A	A	0.04
87	31.2	2.7	-0.32	10.57	A	A	A	0.05
88	32	3.4	2.24	12.24	A	A	A	0.37
88	32.0	3.4	2.24	12.24	A	A	A	0.37
94	37.4	2.2	19.49	8.45	A	A	A	3.21
96	29.0	14.1	-7.35	49.00	A	N	W	1.21
98	39.8	2.4	27.16	8.56	A	N	W	4.47
101	40.0	4	27.80	11.70	A	A	A	4.58
104	27.36	2.54	-12.59	11.09	A	A	A	2.07
106	33.72	2.26	7.73	9.04	A	A	A	1.27
107	34.5	5.13	10.22	16.06	A	A	A	1.68
110	29.6	3.06	-5.43	11.99	A	A	A	0.89
111	31.7	2.87	1.28	10.90	A	A	A	0.21
113	30.28	3.54	-3.26	13.17	A	A	A	0.54

TABLE 12. Evaluation Results for Pb-210 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
114	56.7	8.34	81.15	15.91	N	N	N	13.37
115	35.1	2.0	12.14	8.33	A	A	A	2.00
117	30.6	3.4	-2.24	12.66	A	A	A	0.37
118	36	4	15.02	12.66	A	A	A	2.47
118	30	3	-4.15	11.70	A	A	A	0.68
119	32.050	4.487	2.40	15.26	A	A	A	0.39
119	32.643	4.570	4.29	15.26	A	A	A	0.71
120	33.3	4.4	6.39	14.54	A	A	A	1.05
121	34.152	1.828	9.11	8.09	A	A	A	1.50
123	30.0	1.5	-4.15	7.86	A	A	A	0.68
124	34.27	5.54	9.49	17.27	A	A	A	1.56
125	65.46	3.27	109.14	7.86	N	N	N	17.98
126	30	2.3	-4.15	9.78	A	A	A	0.68
127	30	5.4	-4.15	19.00	A	A	A	0.68
128	26.7	3.3	-14.70	13.77	A	A	A	2.42
128	25.5	3.5	-18.53	15.01	A	A	A	3.05
129	34.2	16.9	9.27	49.79	A	N	W	1.53
130	27.06	0.26	-13.55	6.15	A	A	A	2.23
132	30.6	4.5	-2.24	15.91	A	A	A	0.37
133	32.1	4.4	2.56	14.99	A	A	A	0.42
133	32.0	4.3	2.24	14.74	A	A	A	0.37
135	41	5	30.99	13.62	N	A	N	5.11
135	40.9	5.4	30.67	14.53	N	A	N	5.05
136	37.9	4.1	21.09	12.40	A	A	A	3.47
138	38.984	8.276	24.55	22.08	A	A	A	4.04
139	33.4	1.2	6.71	7.05	A	A	A	1.11
140	31.1	3.5	-0.64	12.79	A	A	A	0.11
140	32.7	5.2	4.47	17.02	A	A	A	0.74
143	30.3	2.95	-3.19	11.47	A	A	A	0.53
145	31	2	-0.96	8.86	A	A	A	0.16
148	32.9	3.2	5.11	11.47	A	A	A	0.84
149	48.4	7.2	54.63	16.07	N	N	N	9.00
150	30.036	2.187	-4.04	9.48	A	A	A	0.67
153	35.4	4.7	13.10	14.60	A	A	A	2.16
154	31.5	3.33	0.64	12.19	A	A	A	0.11
155	53.56	11.83	71.12	22.91	N	N	N	11.72
156	28.6	3.6	-8.63	13.97	A	A	A	1.42
158	40.080	4.810	28.05	13.45	A	A	A	4.62
159	37.0	5.0	18.21	14.81	A	A	A	3.00
160	42.2	8.0	34.82	19.91	N	A	N	5.74
163	54.57	3.36	74.35	8.65	N	N	N	12.25
163	54.57	3.36	74.35	8.65	N	N	N	12.25
164	29.6	1.2	-5.43	7.30	A	A	A	0.89
164	30.2	5.0	-3.51	17.63	A	A	A	0.58
165	33.26	1.46	6.26	7.49	A	A	A	1.03
166	35	8	11.82	23.65	A	A	A	1.95
169	32.50	4.00	3.83	13.72	A	A	A	0.63
169	32.50	4.00	3.83	13.72	A	A	A	0.63
171	29.3	0.8	-6.39	6.66	A	A	A	1.05
172	31	2.6	-0.96	10.35	A	A	A	0.16
173	36.77	2.66	17.48	9.44	A	A	A	2.88
177	35.5	4.0	13.42	12.80	A	A	A	2.21
178	32	5	2.24	16.76	A	A	A	0.37
180	34.8	4.1	11.18	13.25	A	A	A	1.84

TABLE 12. Evaluation Results for Pb-210 in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
182	56.6	4.0	80.83	9.32	N	N	N	13.32
186	27.9	5.8	-10.86	21.66	A	A	A	1.79
187	35.7	3.3	14.06	11.06	A	A	A	2.32
191	34.0	3.0	8.63	10.71	A	A	A	1.42
192	52.88	13.89	68.95	26.96	N	A	N	11.36
197	42	10	34.19	24.57	N	A	N	5.63
198	32.17	3.86	2.78	13.45	A	A	A	0.46
199	31.95	2.36	2.08	9.56	A	A	A	0.34
202	32.9	1.8	5.11	8.17	A	A	A	0.84
203	34.8	5.9	11.18	18.01	A	A	A	1.84
213	31	3	-0.96	11.42	A	A	A	0.16
214	32.6	3.2	4.15	11.54	A	A	A	0.68
216	36.3	1.3	15.97	7.05	A	A	A	2.63
216	36.3	1.3	15.97	7.05	A	A	A	2.63
217	29.9	1.1	-4.47	7.10	A	A	A	0.74
217	34.1	5.2	8.95	16.41	A	A	A	1.47
218	30.14	4.8	-3.71	17.04	A	A	A	0.61
220	29.51	3.5	-5.72	13.32	A	A	A	0.94
221	31.55	1.24	0.80	7.23	A	A	A	0.13
227	37.19	5.22	18.82	15.29	A	A	A	3.10
230	32.2	2.3	2.88	9.37	A	A	A	0.47
231	41.1	4.0	31.31	11.47	N	N	N	5.16
233	1.8	0.54	-94.25	30.61	N	N	N	15.53
235	31.36	2.4	0.19	9.77	A	A	A	0.03
238	37.7	3.1	20.45	10.22	A	A	A	3.37
243	50.2	4.6	60.38	10.99	N	N	N	9.95
244	34.2	4.8	9.27	15.29	A	A	A	1.53
247	29.1	2.1	-7.03	9.43	A	A	A	1.16
247	36.2	2.5	15.65	9.19	A	A	A	2.58
248	33.1	1.8	5.75	8.15	A	A	A	0.95
261	31.8	1.8	1.60	8.30	A	A	A	0.26
262	41.2	5.35	31.63	14.33	N	A	N	5.21
264	33.6	5.1	7.35	16.35	A	A	A	1.21
265	9.54	1.10	-69.52	13.03	N	N	N	11.45
271	25.388	2.096	-18.89	10.25	A	A	A	3.11
271	42.8	4.2	36.74	11.54	N	N	N	6.05
272	37.44	0.63	19.62	6.30	A	N	W	3.23
272	39.76	3.48	27.03	10.65	A	A	A	4.45
273	34.1	3.0	8.95	10.69	A	A	A	1.47
281	35.1	2.7	12.14	9.80	A	A	A	2.00
284	42.97	7.77	37.28	19.07	N	A	N	6.14
285	36.2	8.9	15.65	25.32	A	A	A	2.58
286	37.4	4.6	19.49	13.72	A	A	A	3.21
289	32.7	2.5	4.47	9.76	A	A	A	0.74
296	36.4	2.8	16.29	9.80	A	A	A	2.68
298	32.3	0.9	3.19	6.68	A	A	A	0.53
299	22.36	2.7	-28.56	13.52	A	A	A	4.71
301	37.3	1.7	19.17	7.59	A	A	A	3.16
303	54	16	72.52	30.25	N	N	N	11.95
304	37.0	2.49	18.21	9.06	A	A	A	3.00
309	38.63	7.21	23.42	19.63	A	A	A	3.86
312	32.6	5.0	4.15	16.49	A	A	A	0.68
313	31.7	2.7	1.28	10.46	A	A	A	0.21

Reported Results for H-3 Determination in Sample 2

TABLE 13. Target values

Parameter	Value
Target Value, Bq/kg	10.8
UNC, Bq/kg	0.6
MARB, %	30




FIGURE 13. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 14. Reported results and evaluation parameters for H-3. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 14. Evaluation Results for H-3 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
1	115.6	9.3	970.37	9.78	N	N	N	174.67
4	11.06	0.59	2.41	7.70	A	A	A	0.43
8	11.1	0.7	2.78	8.40	A	A	A	0.50
9	13.20	4.81	22.22	36.86	A	N	W	4.00
13	12.5	2.1	15.74	17.69	A	A	A	2.83
14	14.96	1.89	38.52	13.80	N	N	N	6.93
23	12.93	0.59	19.72	7.19	A	N	W	3.55
24	10.7	0.6	-0.93	7.89	A	A	A	0.17
25	160.00	3.32	1381.48	5.93	N	N	N	248.67
30	8.78	1.38	-18.70	16.67	A	A	A	3.37
36	11	2	1.85	19.01	A	A	A	0.33
38	11	4	1.85	36.79	A	N	W	0.33
41	7.31	1.46	-32.31	20.73	N	A	N	5.82
42	9.9	1.25	-8.33	13.79	A	A	A	1.50
47	10.0	0.54	-7.41	7.75	A	A	A	1.33
49	11.2	0.9	3.70	9.77	A	A	A	0.67
50	10.3	0.8	-4.63	9.55	A	A	A	0.83
51	10.8	1.6	0.00	15.82	A	A	A	0.00
53	14.01	1.96	29.72	15.05	A	A	A	5.35
57	10.45	0.65	-3.24	8.34	A	A	A	0.58
58	9.62	0.60	-10.93	8.35	A	A	A	1.97
60	12.09	0.91	11.94	9.36	A	A	A	2.15
62	10.9	0.5	0.93	7.20	A	A	A	0.17
68	11.03	0.77	2.13	8.92	A	A	A	0.38
69	13.91	1.29	28.80	10.81	A	N	W	5.18
74	12.2	1.3	12.96	12.02	A	A	A	2.33
76	11.7	2.27	8.33	20.18	A	A	A	1.50
78	10.74	0.24	-0.56	5.99	A	A	A	0.10
79	13.7	0.7	26.85	7.55	A	N	W	4.83
80	11.27	0.72	4.35	8.47	A	A	A	0.78
81	13.4	1.1	24.07	9.91	A	A	A	4.33
83	11.2	0.85	3.70	9.41	A	A	A	0.67
86	39.49	1.47	265.65	6.69	N	N	N	47.82
87	36	7.2	233.33	20.76	N	N	N	42.00
88	11.9	0.9	10.19	9.38	A	A	A	1.83
89	11.2	0.64	3.70	7.97	A	A	A	0.67
94	23.2	0.9	114.81	6.78	N	N	N	20.67
95	14.16	1.04	31.11	9.21	N	N	N	5.60
96	11.9	2.2	10.19	19.30	A	A	A	1.83
98	11.8	1.5	9.26	13.87	A	A	A	1.67
99	12.40	2.48	14.81	20.76	A	A	A	2.67
104	12.68	2.08	17.41	17.32	A	A	A	3.13
107	45.75	2.95	323.61	8.51	N	N	N	58.25
110	18.8	3.44	74.07	19.12	N	N	N	13.33
118	12	2	11.11	17.57	A	A	A	2.00
119	11.060	2.046	2.41	19.32	A	A	A	0.43
121	11.41	0.58	5.65	7.53	A	A	A	1.02
123	10.3	1.2	-4.63	12.91	A	A	A	0.83
128	7.17	0.29	-33.61	6.87	N	N	N	6.05
129	13.9	1.2	28.70	10.27	A	N	W	5.17
130	27.155	1.38	151.44	7.53	N	N	N	27.26
133	11.1	2.0	2.78	18.86	A	A	A	0.50
134	22.29	3.99	106.39	18.74	N	N	N	19.15

TABLE 14. Evaluation Results for H-3 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
137	11.30	0.66	4.63	8.06	A	A	A	0.83
146	19		75.93		N		N	13.67
155	11.21	2.80	3.80	25.59	A	A	A	0.68
156	11.2	0.8	3.70	9.05	A	A	A	0.67
157	11.2	0.8	3.70	9.05	A	A	A	0.67
159	10.4	1.6	-3.70	16.36	A	A	A	0.67
160	11.1	0.4	2.78	6.62	A	A	A	0.50
163	17.55	1.10	62.50	8.38	N	N	N	11.25
163	17.55	1.10	62.50	8.38	N	N	N	11.25
164	11.5	2.5	6.48	22.44	A	A	A	1.17
171	11.3	0.5	4.63	7.10	A	A	A	0.83
173	6.46	0.1	-40.19	5.77	N	N	N	7.23
175	6.9	1.0	-36.11	15.52	N	A	N	6.50
177	11.7	1.5	8.33	13.97	A	A	A	1.50
180	11.3	0.4	4.63	6.59	A	A	A	0.83
184	12.1	1.2	12.04	11.37	A	A	A	2.17
187	10.6	1.2	-1.85	12.61	A	A	A	0.33
188	11.0	0.88	1.85	9.74	A	A	A	0.33
189	8.5	1.8	-21.30	21.89	A	A	A	3.83
195	11.3	1.1	4.63	11.21	A	A	A	0.83
196	11.3	4.11	4.63	36.79	A	N	W	0.83
204	10.9	0.9	0.93	9.95	A	A	A	0.17
211	10.7	1.0	-0.93	10.87	A	A	A	0.17
214	8.58	0.5	-20.56	8.05	A	A	A	3.70
216	7.11	3.16	-34.17	44.79	N	N	N	6.15
217	10.8	0.55	0.00	7.54	A	A	A	0.00
221	12.69	1.27	17.50	11.45	A	A	A	3.15
226	8.69	1.5	-19.54	18.13	A	A	A	3.52
227	16.93	0.82	56.76	7.37	N	N	N	10.22
230	10.86	0.5	0.56	7.22	A	A	A	0.10
234	12.70	0.90	17.59	9.00	A	A	A	3.17
236	16.33	0.56	51.20	6.53	N	N	N	9.22
239	9.79	1.91	-9.35	20.29	A	A	A	1.68
240	10.9	3.48	0.93	32.41	A	N	W	0.17
241	13.60	0.30	25.93	5.98	A	N	W	4.67
246	12.0	1.0	11.11	10.02	A	A	A	2.00
247	13.2	1.2	22.22	10.65	A	A	A	4.00
248	14.1	0.56	30.56	6.83	N	N	N	5.50
250	12.10	1.00	12.04	9.96	A	A	A	2.17
254	11.61	1.40	7.50	13.28	A	A	A	1.35
261	11.0	0.66	1.85	8.18	A	A	A	0.33
262	14.37	1.88	33.06	14.21	N	A	N	5.95
263	12.00	0.56	11.11	7.26	A	A	A	2.00
264	11.1	0.8	2.78	9.10	A	A	A	0.50
272	11.1	0.8	2.78	9.10	A	A	A	0.50
273	11.63	0.67	7.69	8.00	A	A	A	1.38
278	9.0	3.2	-16.67	35.99	A	N	W	3.00
281	11.7	2.5	8.33	22.08	A	A	A	1.50
284	9.42	0.71	-12.78	9.36	A	A	A	2.30
289	11.1	0.7	2.78	8.40	A	A	A	0.50
296	10.86	0.56	0.56	7.58	A	A	A	0.10
309	11.8	2.33	9.26	20.51	A	A	A	1.67

Reported Results for Sr-90 Determination in Sample 2

TABLE 15. Target values

Parameter	Value
Target Value, Bq/kg	7.42
UNC, Bq/kg	0.45
MARB, %	30




FIGURE 15. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 16. Reported results and evaluation parameters for Sr-90. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 16. Evaluation Results for Sr-90 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
4	8.55	0.38	15.23	7.52	A	A	A	2.51
7	6.40	0.29	-13.75	7.57	A	A	A	2.27
8	7.35	0.4	-0.94	8.15	A	A	A	0.16
9	7.73	2.43	4.18	32.02	A	N	W	0.69
10	3.19	0.30	-57.01	11.19	N	N	N	9.40
13	7.3	0.6	-1.62	10.21	A	A	A	0.27
14	7.39	0.19	-0.40	6.59	A	A	A	0.07
16	7.83	0.46	5.53	8.44	A	A	A	0.91
17	7.11	0.92	-4.18	14.29	A	A	A	0.69
23	7.47	0.35	0.67	7.66	A	A	A	0.11
24	7.2	0.3	-2.96	7.36	A	A	A	0.49
25	4.3954	0.6	-40.76	14.94	N	N	N	6.72
27	7.64	0.76	2.96	11.65	A	A	A	0.49
31	7.42	0.5	0.00	9.07	A	A	A	0.00
36	7.2	0.8	-2.96	12.66	A	A	A	0.49
41	7.46	0.7	0.54	11.17	A	A	A	0.09
42	7.811	0.878	5.27	12.77	A	A	A	0.87
45	7.4	1.0	-0.27	14.81	A	A	A	0.04
47	7.5	0.6	1.08	10.04	A	A	A	0.18
49	7.40	0.42	-0.27	8.31	A	A	A	0.04
50	6.76	0.42	-8.89	8.68	A	A	A	1.47
51	7.13	0.50	-3.91	9.27	A	A	A	0.64
52	4.58	0.76	-38.27	17.67	N	A	N	6.31
53	6.92	0.22	-6.74	6.85	A	A	A	1.11
56	7.7	0.5	3.77	8.89	A	A	A	0.62
57	7.17	0.31	-3.37	7.45	A	A	A	0.56
60	7.76	0.69	4.58	10.76	A	A	A	0.76
61	10.2	1.2	37.47	13.24	N	N	N	6.18
62	8.27	0.45	11.46	8.15	A	A	A	1.89
65	8.23	0.5	10.92	8.58	A	A	A	1.80
68	7.48	0.46	0.81	8.64	A	A	A	0.13
69	8.06	0.14	8.63	6.31	A	A	A	1.42
74	7.2	0.4	-2.96	8.22	A	A	A	0.49
76	7.78	0.29	4.85	7.12	A	A	A	0.80
78	11.01	1.48	48.38	14.75	N	N	N	7.98
80	9.04	0.42	21.83	7.64	A	N	W	3.60
81	6.85	0.08	-7.68	6.18	A	A	A	1.27
83	6.5	0.3	-12.40	7.62	A	A	A	2.04
87	6.35	0.35	-14.42	8.20	A	A	A	2.38
88	6.93	0.81	-6.60	13.17	A	A	A	1.09
89	7.08	0.34	-4.58	7.74	A	A	A	0.76
90	7.41	0.57	-0.13	9.80	A	A	A	0.02
92	7.40	0.41	-0.27	8.21	A	A	A	0.04
93	7.23	0.38	-2.56	8.03	A	A	A	0.42
94	8.0	1.0	7.82	13.89	A	A	A	1.29
96	8.11	0.04	9.30	6.08	A	A	A	1.53
98	8.5	1.1	14.56	14.29	A	A	A	2.40
99	7.61	0.76	2.56	11.68	A	A	A	0.42
110	7.96	0.240	7.28	6.77	A	A	A	1.20
113	7.22	0.41	-2.70	8.31	A	A	A	0.44
118	6.6	0.4	-11.05	8.57	A	A	A	1.82
119	8.120	1.543	9.43	19.95	A	A	A	1.56
121	4.37	0.03	-41.11	6.10	N	N	N	6.78

TABLE 16. Evaluation Results for Sr-90 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
123	7.5	0.5	1.08	9.01	A	A	A	0.18
128	7.6	0.7	2.43	11.03	A	A	A	0.40
133	8.1	1.5	9.16	19.49	A	A	A	1.51
134	6.02	0.93	-18.87	16.60	A	A	A	3.11
137	7.26	0.44	-2.16	8.57	A	A	A	0.36
138	7.524	0.043	1.40	6.09	A	A	A	0.23
140	28.6	2.2	285.44	9.80	N	N	N	47.07
141	22	6	196.50	27.94	N	N	N	32.40
143	6.75	0.20	-9.03	6.75	A	A	A	1.49
146	8.04		8.36		A			1.38
149	7.8	0.3	5.12	7.18	A	A	A	0.84
156	7.45	0.52	0.40	9.25	A	A	A	0.07
157	7.3	0.7	-1.62	11.35	A	A	A	0.27
159	4.93	0.7	-33.56	15.44	N	A	N	5.53
164	6.37	0.38	-14.15	8.51	A	A	A	2.33
171	7.6	0.3	2.43	7.24	A	A	A	0.40
177	8.5	1.2	14.56	15.37	A	A	A	2.40
180	8.58	0.49	15.63	8.33	A	A	A	2.58
184	7.8	0.7	5.12	10.83	A	A	A	0.84
186	7.4	1.6	-0.27	22.46	A	A	A	0.04
187	7.5	0.7	1.08	11.13	A	A	A	0.18
188	7.42	0.32	0.00	7.44	A	A	A	0.00
192	7.88	0.58	6.20	9.54	A	A	A	1.02
196	9.07	1.31	22.24	15.66	A	A	A	3.67
199	7.64	0.45	2.96	8.45	A	A	A	0.49
203	7.4	0.6	-0.27	10.13	A	A	A	0.04
204	6.49	0.43	-12.53	8.98	A	A	A	2.07
211	7.3	0.3	-1.62	7.33	A	A	A	0.27
214	8.12	0.8	9.43	11.57	A	A	A	1.56
216	7.07	0.28	-4.72	7.24	A	A	A	0.78
217	7.54	0.43	1.62	8.32	A	A	A	0.27
226	7.79	0.5	4.99	8.83	A	A	A	0.82
228	6.95	1.15	-6.33	17.62	A	A	A	1.04
230	7.65	0.38	3.10	7.84	A	A	A	0.51
231	18.9	2	154.72	12.20	N	N	N	25.51
234	6.88	1.20	-7.28	18.47	A	A	A	1.20
239	6.3	1.23	-15.09	20.44	A	A	A	2.49
240	7.72	0.478	4.04	8.67	A	A	A	0.67
241	7.79	0.08	4.99	6.15	A	A	A	0.82
247	6.7	0.4	-9.70	8.51	A	A	A	1.60
248	15.0	1.92	102.16	14.16	N	N	N	16.84
250	6.44	0.6	-13.21	11.12	A	A	A	2.18
258	7.06	0.05	-4.85	6.11	A	A	A	0.80
262	6.60	0.86	-11.05	14.37	A	A	A	1.82
263	7.53	0.34	1.48	7.56	A	A	A	0.24
272	6.54	0.34	-11.86	7.99	A	A	A	1.96
273	7.5	0.4	1.08	8.08	A	A	A	0.18
276	5.63	0.33	-24.12	8.43	A	N	W	3.98
281	8.44	0.83	13.75	11.55	A	A	A	2.27
284	7.04	1.03	-5.12	15.84	A	A	A	0.84
289	7.7	0.3	3.77	7.21	A	A	A	0.62
296	7.56	0.58	1.89	9.78	A	A	A	0.31
303	8.31	0.30	11.99	7.06	A	A	A	1.98

TABLE 16. Evaluation Results for Sr-90 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
309	7.34	0.706	-1.08	11.37	A	A	A	0.18

Reported Results for Cs-137 Determination in Sample 2

TABLE 17. Target values

Parameter	Value
Target Value, Bq/kg	8.36
UNC, Bq/kg	0.5
MARB, %	20




FIGURE 17. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 18. Reported results and evaluation parameters for Cs-137. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 18. Evaluation Results for Cs-137 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
1	9.25	0.44	10.65	7.64	A	A	A	1.78
4	8.02	0.15	-4.07	6.27	A	A	A	0.68
5	7.97	0.29	-4.67	7.00	A	A	A	0.78
6	6.83	0.46	-18.30	9.01	A	A	A	3.06
7	8.26	0.45	-1.20	8.09	A	A	A	0.20
8	8.24	0.56	-1.44	9.05	A	A	A	0.24
9	8.64	0.49	3.35	8.24	A	A	A	0.56
10	28.75	11.00	243.90	38.73	N	N	N	40.78
11	9.05	0.88	8.25	11.42	A	A	A	1.38
12	8.85	0.4	5.86	7.50	A	A	A	0.98
13	8	1	-4.31	13.86	A	A	A	0.72
14	8.15	0.31	-2.51	7.09	A	A	A	0.42
16	8.7	0.23	4.07	6.54	A	A	A	0.68
17	8.40	1.1	0.48	14.40	A	A	A	0.08
18	8.20	0.50	-1.91	8.54	A	A	A	0.32
19	8.13	0.88	-2.75	12.37	A	A	A	0.46
20	8.7	0.4	4.07	7.54	A	A	A	0.68
21	8.1	0.5	-3.11	8.60	A	A	A	0.52
22	8.864	0.146	6.03	6.20	A	A	A	1.01
23	8.58	0.44	2.63	7.88	A	A	A	0.44
24	8.3	0.3	-0.72	6.99	A	A	A	0.12
25	7.88	0.71	-5.74	10.81	A	A	A	0.96
26	8.5	0.5	1.67	8.39	A	A	A	0.28
27	8.12	0.31	-2.87	7.10	A	A	A	0.48
28	8.1	0.4	-3.11	7.76	A	A	A	0.52
29	8.34	0.24	-0.24	6.64	A	A	A	0.04
31	8.44	0.8	0.96	11.21	A	A	A	0.16
32	8.67	0.17	3.71	6.29	A	A	A	0.62
33	8.38	0.66	0.24	9.89	A	A	A	0.04
34	8.56	0.61	2.39	9.30	A	A	A	0.40
35	8.9	0.3	6.46	6.87	A	A	A	1.08
36	8	2	-4.31	25.71	A	N	W	0.72
37	8.6	0.5	2.87	8.34	A	A	A	0.48
38	8	0.6	-4.31	9.59	A	A	A	0.72
39	8.3	0.4	-0.72	7.68	A	A	A	0.12
40	8.4	0.1	0.48	6.10	A	A	A	0.08
41	8.1	0.79	-3.11	11.44	A	A	A	0.52
42	4.29	0.21	-48.68	7.73	N	N	N	8.14
43	8.4	0.4	0.48	7.65	A	A	A	0.08
45	8.04	0.19	-3.83	6.43	A	A	A	0.64
46	8.5	0.4	1.67	7.61	A	A	A	0.28
47	8.32	0.47	-0.48	8.23	A	A	A	0.08
49	8.4	0.3	0.48	6.97	A	A	A	0.08
50	8.5	0.1	1.67	6.10	A	A	A	0.28
51	8.3	0.1	-0.72	6.10	A	A	A	0.12
52	8.04	0.67	-3.83	10.26	A	A	A	0.64
53	9.52	0.36	13.88	7.08	A	A	A	2.32
55	8.62	0.30	3.11	6.92	A	A	A	0.52
56	8.1	0.4	-3.11	7.76	A	A	A	0.52
57	8.43	0.60	0.84	9.30	A	A	A	0.14
58	8.53	0.18	2.03	6.34	A	A	A	0.34
59	8.5	0.4	1.67	7.61	A	A	A	0.28
60	8.33	0.53	-0.36	8.73	A	A	A	0.06

TABLE 18. Evaluation Results for Cs-137 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
61	8.18	0.58	-2.15	9.28	A	A	A	0.36
62	8.32	0.32	-0.48	7.11	A	A	A	0.08
63	8.50	0.56	1.67	8.90	A	A	A	0.28
64	8.4	0.6	0.48	9.32	A	A	A	0.08
65	8.55	0.15	2.27	6.23	A	A	A	0.38
67	8.58	0.34	2.63	7.17	A	A	A	0.44
68	8.29	0.31	-0.84	7.05	A	A	A	0.14
69	7.97	0.08	-4.67	6.06	A	A	A	0.78
70	8.2	0.4	-1.91	7.72	A	A	A	0.32
71	8.35	0.15	-0.12	6.24	A	A	A	0.02
72	8.4	0.4	0.48	7.65	A	A	A	0.08
74	8.38	0.55	0.24	8.88	A	A	A	0.04
75	8.5	0.6	1.67	9.25	A	A	A	0.28
76	8.10	0.22	-3.11	6.57	A	A	A	0.52
77	9.03	0.469	8.01	7.92	A	A	A	1.34
78	8.0	0.3	-4.31	7.06	A	A	A	0.72
79	8.91	0.41	6.58	7.55	A	A	A	1.10
80	8.12	0.267	-2.87	6.83	A	A	A	0.48
81	8.4	0.4	0.48	7.65	A	A	A	0.08
83	8.26	0.25	-1.20	6.70	A	A	A	0.20
84	8.578	0.395	2.61	7.55	A	A	A	0.44
85	8.42	0.28	0.72	6.84	A	A	A	0.12
86	8.0	0.807	-4.31	11.73	A	A	A	0.72
87	8.05	0.5	-3.71	8.62	A	A	A	0.62
88	8.82	0.57	5.50	8.81	A	A	A	0.92
89	8.97	0.68	7.30	9.66	A	A	A	1.22
90	7.99	0.52	-4.43	8.84	A	A	A	0.74
91	8.4	0.5	0.48	8.44	A	A	A	0.08
92	8.48	0.37	1.44	7.40	A	A	A	0.24
93	8.37	0.22	0.12	6.53	A	A	A	0.02
94	9.1	0.5	8.85	8.12	A	A	A	1.48
95	8.80	0.29	5.26	6.83	A	A	A	0.88
96	8.60	0.30	2.87	6.92	A	A	A	0.48
97	8.19	0.14	-2.03	6.22	A	A	A	0.34
98	8.7	0.5	4.07	8.29	A	A	A	0.68
99	8.34	0.83	-0.24	11.61	A	A	A	0.04
100	33000	300	394636.84	6.05	N	N	N	65983.28
101	8.9	0.9	6.46	11.75	A	A	A	1.08
102	8.41	0.434	0.60	7.90	A	A	A	0.10
103	8.47	0.43	1.32	7.84	A	A	A	0.22
105	8.22	0.64	-1.67	9.82	A	A	A	0.28
106	8.08	0.351	-3.35	7.39	A	A	A	0.56
107	8.76	0.55	4.78	8.67	A	A	A	0.80
109	7.70	0.35	-7.89	7.51	A	A	A	1.32
110	8.75	0.527	4.67	8.49	A	A	A	0.78
111	8.34	0.35	-0.24	7.31	A	A	A	0.04
113	7.93	0.16	-5.14	6.31	A	A	A	0.86
114	9.2	0.6	10.05	8.85	A	A	A	1.68
117	8.1	0.5	-3.11	8.60	A	A	A	0.52
118	9.1	0.7	8.85	9.74	A	A	A	1.48
119	8.339	0.876	-0.25	12.09	A	A	A	0.04
120	8.1	0.3	-3.11	7.03	A	A	A	0.52
121	8.383	0.273	0.28	6.81	A	A	A	0.05
123	8.2	0.6	-1.91	9.45	A	A	A	0.32

TABLE 18. Evaluation Results for Cs-137 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
124	7.88	0.38	-5.74	7.68	A	A	A	0.96
125	14.20	0.99	69.86	9.19	N	N	N	11.68
126	8.11	0.72	-2.99	10.70	A	A	A	0.50
127	8.508	0.681	1.77	9.99	A	A	A	0.30
128	9.07	0.28	8.49	6.73	A	A	A	1.42
129	8.1	0.5	-3.11	8.60	A	A	A	0.52
130	8.47	1.3	1.32	16.47	A	A	A	0.22
132	8.5	0.5	1.67	8.39	A	A	A	0.28
133	8.3	0.9	-0.72	12.38	A	A	A	0.12
134	9.17	0.77	9.69	10.31	A	A	A	1.62
135	6.8	0.5	-18.66	9.48	A	A	A	3.12
136	8.15	0.46	-2.51	8.22	A	A	A	0.42
137	7.79	0.17	-6.82	6.37	A	A	A	1.14
138	8.327	0.413	-0.39	7.77	A	A	A	0.07
139	8.35	0.11	-0.12	6.12	A	A	A	0.02
140	8.1	0.7	-3.11	10.51	A	A	A	0.52
143	8.30	0.54	-0.72	8.84	A	A	A	0.12
145	7.8	0.4	-6.70	7.88	A	A	A	1.12
146	7.2591	0.0871	-13.17	6.10	A	A	A	2.20
147	8.77	0.31	4.90	6.95	A	A	A	0.82
148	8.09	0.50	-3.23	8.60	A	A	A	0.54
149	7.6	0.5	-9.09	8.89	A	A	A	1.52
150	8.453	0.565	1.11	8.97	A	A	A	0.19
152	7.82	0.39	-6.46	7.79	A	A	A	1.08
153	8.6	0.3	2.87	6.92	A	A	A	0.48
154	8.29	0.5	-0.84	8.49	A	A	A	0.14
155	7.34	0.53	-12.20	9.38	A	A	A	2.04
156	8.42	0.58	0.72	9.12	A	A	A	0.12
157	8.4	0.8	0.48	11.25	A	A	A	0.08
158	8.187	0.327	-2.07	7.19	A	A	A	0.35
159	8.36	0.5	0.00	8.46	A	A	A	0.00
160	7.9	0.2	-5.50	6.49	A	A	A	0.92
161	8.22	0.5	-1.67	8.53	A	A	A	0.28
162	7.9	0.5	-5.50	8.71	A	A	A	0.92
163	8.46	0.45	1.20	8.00	A	A	A	0.20
164	8.07	0.60	-3.47	9.54	A	A	A	0.58
165	8.26	0.29	-1.20	6.94	A	A	A	0.20
166	8.6	0.4	2.87	7.58	A	A	A	0.48
167	8.1	0.5	-3.11	8.60	A	A	A	0.52
168	11.43	0.54	36.72	7.62	N	N	N	6.14
169	7.88	0.34	-5.74	7.37	A	A	A	0.96
171	9.35	0.4	11.84	7.35	A	A	A	1.98
172	7.98	0.48	-4.55	8.48	A	A	A	0.76
173	8.36	0.29	0.00	6.91	A	A	A	0.00
175	8.5	0.4	1.67	7.61	A	A	A	0.28
177	8.2	0.5	-1.91	8.54	A	A	A	0.32
178	8.0	0.6	-4.31	9.59	A	A	A	0.72
179	8.5733	0.45604	2.55	8.00	A	A	A	0.43
180	8.11	0.45	-2.99	8.16	A	A	A	0.50
182	8.5	0.4	1.67	7.61	A	A	A	0.28
183	8.37	0.39	0.12	7.58	A	A	A	0.02
184	8.37	0.36	0.12	7.37	A	A	A	0.02
185	7.1	1.3	-15.07	19.26	A	A	A	2.52
186	7.1	2.8	-15.07	39.89	A	N	W	2.52

TABLE 18. Evaluation Results for Cs-137 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
187	8.2	0.5	-1.91	8.54	A	A	A	0.32
188	7.91	0.58	-5.38	9.46	A	A	A	0.90
189	8.48	0.74	1.44	10.58	A	A	A	0.24
191	8.2	0.3	-1.91	7.01	A	A	A	0.32
192	8.25	0.67	-1.32	10.09	A	A	A	0.22
193	1.104	0.137	-86.79	13.78	N	N	N	14.51
194	30.2	5	261.24	17.60	N	N	N	43.68
195	8.6	0.42	2.87	7.72	A	A	A	0.48
196	8.21	0.364	-1.79	7.44	A	A	A	0.30
198	8.35	0.80	-0.12	11.29	A	A	A	0.02
201	8.5	0.5	1.67	8.39	A	A	A	0.28
202	8.6	0.5	2.87	8.34	A	A	A	0.48
203	8.2	0.5	-1.91	8.54	A	A	A	0.32
204	8.59	0.53	2.75	8.59	A	A	A	0.46
206	8.01	0.05	-4.19	6.01	A	A	A	0.70
207	8.35	0.36	-0.12	7.37	A	A	A	0.02
209	8.57	1.6	2.51	19.60	A	A	A	0.42
210	8.5	0.5	1.67	8.39	A	A	A	0.28
213	8.5	0.4	1.67	7.61	A	A	A	0.28
214	7.8	0.3	-6.70	7.11	A	A	A	1.12
215	8.22	0.74	-1.67	10.81	A	A	A	0.28
216	8.5	0.1	1.67	6.10	A	A	A	0.28
217	8.43	0.54	0.84	8.76	A	A	A	0.14
218	8.37	0.66	0.12	9.90	A	A	A	0.02
219	8.5	0.17	1.67	6.31	A	A	A	0.28
220	8.38	0.45	0.24	8.04	A	A	A	0.04
221	8.39	0.23	0.36	6.58	A	A	A	0.06
222	8.2	0.4	-1.91	7.72	A	A	A	0.32
223	8.38	0.1	0.24	6.10	A	A	A	0.04
224	8.53	0.11	2.03	6.12	A	A	A	0.34
226	8.69	0.56	3.95	8.79	A	A	A	0.66
227	8.65	0.23	3.47	6.55	A	A	A	0.58
228	9.45	1.28	13.04	14.81	A	A	A	2.18
230	8.2	0.6	-1.91	9.45	A	A	A	0.32
231	7.3	0.7	-12.68	11.30	A	A	A	2.12
233	9.8	0.23	17.22	6.42	A	N	W	2.88
234	8.40	0.40	0.48	7.65	A	A	A	0.08
235	8.4	0.23	0.48	6.58	A	A	A	0.08
237	8.21	0.20	-1.79	6.46	A	A	A	0.30
238	7.8	0.3	-6.70	7.11	A	A	A	1.12
239	7.61	1.1	-8.97	15.64	A	A	A	1.50
241	9.14	1.00	9.33	12.47	A	A	A	1.56
243	8.58	0.41	2.63	7.66	A	A	A	0.44
244	8.4	0.8	0.48	11.25	A	A	A	0.08
246	8.8	0.5	5.26	8.25	A	A	A	0.88
247	8.38	0.40	0.24	7.65	A	A	A	0.04
248	8.43	0.29	0.84	6.90	A	A	A	0.14
249	4.14	6	-50.48	145.05	N	N	N	8.44
250	7.70	0.17	-7.89	6.38	A	A	A	1.32
251	7.75	0.77	-7.30	11.60	A	A	A	1.22
253	8.08	0.15	-3.35	6.26	A	A	A	0.56
255	8.37	0.27	0.12	6.80	A	A	A	0.02
256	8.7	0.47	4.07	8.06	A	A	A	0.68
258	8.56	0.156	2.39	6.25	A	A	A	0.40

TABLE 18. Evaluation Results for Cs-137 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
259	8.86	0.31	5.98	6.93	A	A	A	1.00
261	8.21	0.6	-1.79	9.44	A	A	A	0.30
262	8.6	1.12	2.87	14.33	A	A	A	0.48
263	8.099	0.570	-3.12	9.24	A	A	A	0.52
264	8.21	0.81	-1.79	11.54	A	A	A	0.30
265	7.68	0.22	-8.13	6.63	A	A	A	1.36
266	8.67	0.59	3.71	9.06	A	A	A	0.62
269	7.95	0.46	-4.90	8.32	A	A	A	0.82
271	8.4	0.4	0.48	7.65	A	A	A	0.08
272	8.23	0.86	-1.56	12.04	A	A	A	0.26
273	8.7	0.5	4.07	8.29	A	A	A	0.68
274	10.10	1.93	20.81	20.02	N	N	N	3.48
276	7.20	0.86	-13.88	13.36	A	A	A	2.32
278	7.5	0.8	-10.29	12.23	A	A	A	1.72
281	8.1	0.5	-3.11	8.60	A	A	A	0.52
284	8.74	0.42	4.55	7.67	A	A	A	0.76
285	7.70	0.60	-7.89	9.82	A	A	A	1.32
286	8.7	0.4	4.07	7.54	A	A	A	0.68
287	7.771	0.50	-7.05	8.78	A	A	A	1.18
288	7.5	0.8	-10.29	12.23	A	A	A	1.72
289	8.0	0.5	-4.31	8.65	A	A	A	0.72
290	7.88	0.40	-5.74	7.84	A	A	A	0.96
292	8.48	1.19	1.44	15.25	A	A	A	0.24
295	8.40	0.43	0.48	7.87	A	A	A	0.08
296	7.97	0.3	-4.67	7.07	A	A	A	0.78
297	0.31	0.10	-96.29	32.81	N	N	N	16.10
298	8.40	0.12	0.48	6.15	A	A	A	0.08
299	8.05	0.15	-3.71	6.26	A	A	A	0.62
301	8.9	0.2	6.46	6.39	A	A	A	1.08
303	8.39	0.18	0.36	6.35	A	A	A	0.06
305	7.563	1.433	-9.53	19.87	A	A	A	1.59
306	8.4	0.3	0.48	6.97	A	A	A	0.08
307	7.91	0.24	-5.38	6.71	A	A	A	0.90
309	8.755	0.8686	4.72	11.58	A	A	A	0.79
310	8.49	0.8	1.56	11.16	A	A	A	0.26
312	8.57	0.32	2.51	7.05	A	A	A	0.42
313	8.4	0.5	0.48	8.44	A	A	A	0.08

Reported Results for Po-210 Determination in Sample 2

TABLE 19. Target values

Parameter	Value
Target Value, Bq/kg	5.29
UNC, Bq/kg	0.32
MARB, %	30




FIGURE 19. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 20. Reported results and evaluation parameters for Po-210. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 20. Evaluation Results for Po-210 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
4	3.56	0.22	-32.70	8.65	N	N	N	5.41
8	4.89	0.34	-7.56	9.22	A	A	A	1.25
13	4.9	0.3	-7.37	8.61	A	A	A	1.22
14	5.59	0.25	5.67	7.52	A	A	A	0.94
16	4.95	0.164	-6.43	6.90	A	A	A	1.06
17	4.66	0.61	-11.91	14.42	A	A	A	1.97
36	5.2	0.6	-1.70	13.03	A	A	A	0.28
42	16.83	1.81	218.15	12.34	N	N	N	36.06
52	4.831	0.0053	-8.68	6.05	A	A	A	1.43
58	5.2	0.3	-1.70	8.36	A	A	A	0.28
60	3.29	0.21	-37.81	8.79	N	N	N	6.25
68	5.18	0.36	-2.08	9.21	A	A	A	0.34
71	16.43	0.37	210.59	6.45	N	N	N	34.81
78	14.46	0.82	173.35	8.29	N	N	N	28.66
81	11.9	0.4	124.95	6.92	N	N	N	20.66
83	9.0	1.0	70.13	12.65	N	N	N	11.59
84	5.110	0.139	-3.40	6.63	A	A	A	0.56
84	5.035	0.486	-4.82	11.39	A	A	A	0.80
92	3.82	0.38	-27.79	11.64	A	A	A	4.59
97	5.60	0.20	5.86	7.02	A	A	A	0.97
99	15.82	3.01	199.05	19.97	N	N	N	32.91
118	5.1	0.7	-3.59	15.00	A	A	A	0.59
123	5.21	0.52	-1.51	11.67	A	A	A	0.25
128	4.9	0.7	-7.37	15.51	A	A	A	1.22
133	3.7	0.7	-30.06	19.86	N	A	N	4.97
135	16	1	202.46	8.70	N	N	N	33.47
155	18.05	0.96	241.21	8.05	N	N	N	39.88
156	5.25	0.41	-0.76	9.88	A	A	A	0.13
157	11.2	1.5	111.72	14.70	N	N	N	18.47
159	5.3	1.0	0.19	19.81	A	A	A	0.03
161	12.75	1.17	141.02	10.99	N	N	N	23.31
163	3.49	0.61	-34.03	18.50	N	A	N	5.62
163	3.49	0.61	-34.03	18.50	N	A	N	5.62
164	4.78	0.23	-9.64	7.73	A	A	A	1.59
171	6	0.3	13.42	7.85	A	A	A	2.22
177	4.1	0.4	-22.50	11.48	A	A	A	3.72
186	9.5	2.8	79.58	30.09	N	N	N	13.16
203	2.97	0.22	-43.86	9.56	N	N	N	7.25
214	1.42	0.04	-73.16	6.67	N	N	N	12.09
217	4.86	0.66	-8.13	14.87	A	A	A	1.34
227	14.8	1.1	179.77	9.58	N	N	N	29.72
247	15.3	2.4	189.22	16.81	N	N	N	31.28
261	5.20	0.41	-1.70	9.94	A	A	A	0.28
272	9.45	1.1	78.64	13.12	N	N	N	13.00

Reported Results for Pb-210 Determination in Sample 2

TABLE 21. Target values

Parameter	Value
Target Value, Bq/kg	5.2
UNC, Bq/kg	0.31
MARB, %	30




FIGURE 21. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 22. Reported results and evaluation parameters for Pb-210. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 22. Evaluation Results for Pb-210 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
4	4.93	0.21	-5.19	7.33	A	A	A	0.87
8	5.08	0.38	-2.31	9.57	A	A	A	0.39
9	36.98	9.93	611.15	27.51	N	N	N	102.52
9	36.98	9.93	611.15	27.51	N	N	N	102.52
14	5.47	0.11	5.19	6.29	A	A	A	0.87
14	4.23	0.26	-18.65	8.56	A	A	A	3.13
16	4.81	0.69	-7.50	15.53	A	A	A	1.26
17	4.69	0.61	-9.81	14.31	A	A	A	1.65

TABLE 22. Evaluation Results for Pb-210 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
18	4.69	0.56	-9.81	13.35	A	A	A	1.65
28	5.0	1.3	-3.85	26.67	A	A	A	0.65
31	5.08	0.6	-2.31	13.23	A	A	A	0.39
31	5.20	0.6	0.00	12.99	A	A	A	0.00
32	3.64	0.91	-30.00	25.70	A	A	A	5.03
41	5.54	1.5	6.54	27.72	A	A	A	1.10
41	5	1.33	-3.85	27.26	A	A	A	0.65
42	5.03	0.82	-3.27	17.36	A	A	A	0.55
45	4.05	0.43	-22.12	12.18	A	A	A	3.71
52	4.52	0.3	-13.08	8.92	A	A	A	2.19
58	5.8	0.7	11.54	13.46	A	A	A	1.94
60	5.05	0.46	-2.88	10.89	A	A	A	0.48
67	5.07	0.53	-2.50	12.03	A	A	A	0.42
68	5.01	0.35	-3.65	9.18	A	A	A	0.61
68	5.32	0.20	2.31	7.05	A	A	A	0.39
71	4.00	0.14	-23.08	6.91	A	N	W	3.87
74	6.14	1.59	18.08	26.57	A	A	A	3.03
78	4.8	0.6	-7.69	13.85	A	A	A	1.29
83	4.5	0.3	-13.46	8.94	A	A	A	2.26
83	4.97	0.30	-4.42	8.48	A	A	A	0.74
84	4.953	0.478	-4.75	11.34	A	A	A	0.80
87	4.33	0.785	-16.73	19.08	A	A	A	2.81
94	5.3	0.9	1.92	18.00	A	A	A	0.32
106	6.08	1.15	16.92	19.83	A	A	A	2.84
107	6.25	1.18	20.19	19.80	A	A	A	3.39
113	7.17	1.92	37.88	27.43	N	A	N	6.35
118	5.4	0.3	3.85	8.15	A	A	A	0.65
123	4.9	0.4	-5.77	10.11	A	A	A	0.97
125	76.33	3.82	1367.88	7.78	N	N	N	229.45
128	5.3	0.8	1.92	16.23	A	A	A	0.32
128	3.0	0.4	-42.31	14.61	N	N	N	7.10
133	4.5	1.0	-13.46	23.01	A	A	A	2.26
138	6.328	1.509	21.69	24.58	A	A	A	3.64
140	5.2	0.6	0.00	12.99	A	A	A	0.00
150	4.96	0.944	-4.62	19.94	A	A	A	0.77
153	6.5	0.9	25.00	15.08	A	A	A	4.19
154	4.94	0.67	-5.00	14.82	A	A	A	0.84
156	4.94	0.51	-5.00	11.92	A	A	A	0.84
159	4.1	1.6	-21.15	39.48	A	N	W	3.55
163	6.49	1.16	24.81	18.84	A	A	A	4.16
163	6.49	1.16	24.81	18.84	A	A	A	4.16
164	4.82	0.26	-7.31	8.04	A	A	A	1.23
164	6.1	1.1	17.31	18.99	A	A	A	2.90
165	5.23	0.82	0.58	16.77	A	A	A	0.10
169	5.31	1.76	2.12	33.68	A	N	W	0.35
169	5.31	1.76	2.12	33.68	A	N	W	0.35
171	5.0	0.2	-3.85	7.18	A	A	A	0.65
172	5.49	0.83	5.58	16.25	A	A	A	0.94
175	2.3	0.8	-55.77	35.29	N	N	N	9.35
186	7.6	2.3	46.15	30.84	N	N	N	7.74
199	6.50	0.50	25.00	9.73	A	A	A	4.19
202	5.37	0.35	3.27	8.83	A	A	A	0.55
214	4.5	1.2	-13.46	27.32	A	A	A	2.26
216	5.4	0.7	3.85	14.27	A	A	A	0.65

TABLE 22. Evaluation Results for Pb-210 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
216	5.4	0.7	3.85	14.27	A	A	A	0.65
217	4.82	0.26	-7.31	8.04	A	A	A	1.23
218	5.02	0.95	-3.46	19.84	A	A	A	0.58
220	4.91	0.9	-5.58	19.28	A	A	A	0.94
221	1.92	0.08	-63.08	7.27	N	N	N	10.58
231	23.2	2.3	346.15	11.57	N	N	N	58.06
235	5.33	1.0	2.50	19.69	A	A	A	0.42
238	6.5	0.8	25.00	13.68	A	A	A	4.19
247	5.39	0.40	3.65	9.52	A	A	A	0.61
247	5.7	1.0	9.62	18.53	A	A	A	1.61
262	8.95	1.16	72.12	14.27	N	N	N	12.10
271	5.524	0.472	6.23	10.42	A	A	A	1.05
272	4.4	0.41	-15.38	11.06	A	A	A	2.58
272	5.49	0.71	5.58	14.24	A	A	A	0.94
281	4.5	0.9	-13.46	20.87	A	A	A	2.26
286	4.8	1.8	-7.69	37.97	A	N	W	1.29
298	4.7	0.5	-9.62	12.19	A	A	A	1.61
301	5.3	0.7	1.92	14.49	A	A	A	0.32
304	4.0	1.5	-23.08	37.97	A	N	W	3.87
309	6.508	1.52	25.15	24.10	A	A	A	4.22
312	3.5	1.1	-32.69	31.99	N	N	N	5.48
313	3.4	0.7	-34.62	21.43	N	A	N	5.81

Reported Results for Am-241 Determination in Sample 2

TABLE 23. Target values

Parameter	Value
Target Value, Bq/kg	10.1
UNC, Bq/kg	0.6
MARB, %	30




FIGURE 23. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 24. Reported results and evaluation parameters for Am-241. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 24. Evaluation Results for Am-241 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
1	9.15	1.21	-9.41	14.50	A	A	A	1.58
4	9.94	0.17	-1.58	6.18	A	A	A	0.27
5	9.94	0.37	-1.58	7.01	A	A	A	0.27
7	9.5	0.8	-5.94	10.31	A	A	A	1.00
8	10.2	1.0	0.99	11.46	A	A	A	0.17
9	9.12	3.02	-9.70	33.64	A	N	W	1.63
9	9.46	0.90	-6.34	11.22	A	A	A	1.07
11	9.80	0.72	-2.97	9.45	A	A	A	0.50

TABLE 24. Evaluation Results for Am-241 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
12	10.7	1.0	5.94	11.07	A	A	A	1.00
13	10	1	-0.99	11.63	A	A	A	0.17
14	10.10	0.32	0.00	6.73	A	A	A	0.00
14	9.82	0.37	-2.77	7.03	A	A	A	0.47
16	10.23	0.25	1.29	6.42	A	A	A	0.22
18	10.47	0.91	3.66	10.53	A	A	A	0.62
19	10.72	1.42	6.14	14.52	A	A	A	1.03
20	11.3	1.2	11.88	12.17	A	A	A	2.00
21	10.6	1.1	4.95	11.96	A	A	A	0.83
22	10.020	0.729	-0.79	9.39	A	A	A	0.13
24	10.3	0.8	1.98	9.78	A	A	A	0.33
26	9.7	1.01	-3.96	11.99	A	A	A	0.67
26	9.7	1	-3.96	11.90	A	A	A	0.67
28	10.9	0.5	7.92	7.51	A	A	A	1.33
29	10.88	0.75	7.72	9.10	A	A	A	1.30
31	10.0	0.9	-0.99	10.78	A	A	A	0.17
31	10.0	0.9	-0.99	10.78	A	A	A	0.17
32	10.9	0.4	7.92	6.98	A	A	A	1.33
33	6.59	1.44	-34.75	22.64	N	A	N	5.85
34	10.25	0.71	1.49	9.13	A	A	A	0.25
35	9.7	0.3	-3.96	6.70	A	A	A	0.67
36	10	1	-0.99	11.63	A	A	A	0.17
36	10	2	-0.99	20.86	A	A	A	0.17
37	10.5	0.9	3.96	10.43	A	A	A	0.67
39	10.4	0.7	2.97	8.98	A	A	A	0.50
40	9.9	0.2	-1.98	6.27	A	A	A	0.33
41	10.4	1.23	2.97	13.24	A	A	A	0.50
42	5.92	0.39	-41.39	8.87	N	N	N	6.97
43	10.45	0.6	3.47	8.26	A	A	A	0.58
45	10.02	0.16	-0.79	6.15	A	A	A	0.13
46	10.5	0.5	3.96	7.61	A	A	A	0.67
47	10.2	1.7	0.99	17.69	A	A	A	0.17
49	10.1	0.3	0.00	6.64	A	A	A	0.00
50	10.1	0.2	0.00	6.26	A	A	A	0.00
51	10.2	0.2	0.99	6.26	A	A	A	0.17
52	10.96	0.87	8.51	9.91	A	A	A	1.43
53	15.56	1.44	54.06	11.00	N	N	N	9.10
55	10.33	0.49	2.28	7.60	A	A	A	0.38
56	9.4	0.6	-6.93	8.72	A	A	A	1.17
57	9.95	1.03	-1.49	11.94	A	A	A	0.25
58	9.80	0.79	-2.97	10.01	A	A	A	0.50
59	10.1	0.6	0.00	8.40	A	A	A	0.00
60	10.58	0.94	4.75	10.69	A	A	A	0.80
61	10.26	2.56	1.58	25.65	A	A	A	0.27
61	10.2	0.51	0.99	7.76	A	A	A	0.17
62	10.0	0.5	-0.99	7.76	A	A	A	0.17
63	10.50	0.86	3.96	10.12	A	A	A	0.67
64	10.2	1.1	0.99	12.31	A	A	A	0.17
67	9.97	0.63	-1.29	8.67	A	A	A	0.22
68	10.4	0.5	2.97	7.64	A	A	A	0.50
69	9.78	0.23	-3.17	6.39	A	A	A	0.53
70	11.0	1.6	8.91	15.71	A	A	A	1.50
71	10.08	0.12	-0.20	6.06	A	A	A	0.03
74	10.35	0.86	2.48	10.21	A	A	A	0.42

TABLE 24. Evaluation Results for Am-241 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
75	10.1	0.9	0.00	10.71	A	A	A	0.00
76	10.4	0.50	2.97	7.64	A	A	A	0.50
77	11.0	0.809	8.91	9.45	A	A	A	1.50
78	10.0	0.5	-0.99	7.76	A	A	A	0.17
79	6.99	1.90	-30.79	27.82	N	A	N	5.18
81	9.8	0.5	-2.97	7.83	A	A	A	0.50
81	9.8	0.5	-2.97	7.83	A	A	A	0.50
83	10.1	0.8	0.00	9.90	A	A	A	0.00
83	10.09	0.25	-0.10	6.44	A	A	A	0.02
84	10.184	0.447	0.83	7.39	A	A	A	0.14
84	10.878	0.503	7.70	7.53	A	A	A	1.30
87	6.036	0.1518	-40.24	6.45	N	N	N	6.77
87	10.6	0.6	4.95	8.21	A	A	A	0.83
88	10.3	0.7	1.98	9.03	A	A	A	0.33
88	10.3	0.7	1.98	9.03	A	A	A	0.33
90	9.68	1.30	-4.16	14.68	A	A	A	0.70
91	10.5	0.5	3.96	7.61	A	A	A	0.67
92	8.87	0.57	-12.18	8.75	A	A	A	2.05
93	9.87	0.36	-2.28	6.97	A	A	A	0.38
94	10.6	0.5	4.95	7.59	A	A	A	0.83
96	9.93	0.42	-1.68	7.29	A	A	A	0.28
97	11.23	0.23	11.19	6.28	A	A	A	1.88
98	10.3	0.6	1.98	8.32	A	A	A	0.33
101	10.1	1	0.00	11.55	A	A	A	0.00
106	8.84	0.354	-12.48	7.16	A	A	A	2.10
107	9.87	1.02	-2.28	11.92	A	A	A	0.38
109	10.60	0.79	4.95	9.53	A	A	A	0.83
110	10.3	0.755	1.98	9.44	A	A	A	0.33
111	10.1	0.41	0.00	7.20	A	A	A	0.00
113	9.33	0.22	-7.62	6.39	A	A	A	1.28
114	10.7	0.6	5.94	8.17	A	A	A	1.00
115	10.07	0.57	-0.30	8.21	A	A	A	0.05
117	9.8	1.0	-2.97	11.81	A	A	A	0.50
118	9.8	1	-2.97	11.81	A	A	A	0.50
119	9.742	1.218	-3.54	13.84	A	A	A	0.60
119	10.193	1.427	0.92	15.21	A	A	A	0.15
120	9.8	0.4	-2.97	7.21	A	A	A	0.50
121	10.954	0.304	8.46	6.56	A	A	A	1.42
123	9.5	1.0	-5.94	12.09	A	A	A	1.00
124	11.21	0.32	10.99	6.59	A	A	A	1.85
126	10.39	0.6	2.87	8.28	A	A	A	0.48
127	10.32	1.24	2.18	13.40	A	A	A	0.37
128	10.2	0.6	0.99	8.36	A	A	A	0.17
128	10.6	0.5	4.95	7.59	A	A	A	0.83
129	10.6	0.6	4.95	8.21	A	A	A	0.83
130	10.7	0.2	5.94	6.23	A	A	A	1.00
132	9.7	1.0	-3.96	11.90	A	A	A	0.67
133	9.7	1.2	-3.96	13.72	A	A	A	0.67
133	10.2	1.4	0.99	14.96	A	A	A	0.17
134	11.46	1.16	13.47	11.74	A	A	A	2.27
135	10	1	-0.99	11.63	A	A	A	0.17
135	9.8	1.2	-2.97	13.61	A	A	A	0.50
136	11.0	0.8	8.91	9.39	A	A	A	1.50
137	9.47	0.29	-6.24	6.68	A	A	A	1.05

TABLE 24. Evaluation Results for Am-241 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
138	10.218	1.049	1.17	11.86	A	A	A	0.20
139	9.69	0.9	-4.06	11.03	A	A	A	0.68
139	10.1	0.15	0.00	6.12	A	A	A	0.00
140	10.1	0.67	0.00	8.90	A	A	A	0.00
140	10.4	1.2	2.97	12.98	A	A	A	0.50
141	10.3	1.6	1.98	16.63	A	A	A	0.33
143	10.4	0.57	2.97	8.08	A	A	A	0.50
145	10.7	0.4	5.94	7.02	A	A	A	1.00
146	11.816	0.119	16.99	6.03	A	N	W	2.86
146	11.816	0.119	16.99	6.03	A	N	W	2.86
147	10.27	0.38	1.68	7.00	A	A	A	0.28
148	9.73	0.30	-3.66	6.69	A	A	A	0.62
149	15.1	1.1	49.50	9.40	N	N	N	8.33
150	10.070	0.655	-0.30	8.81	A	A	A	0.05
152	9.03	0.61	-10.59	9.00	A	A	A	1.78
153	11.3	0.9	11.88	9.94	A	A	A	2.00
154	10.07	0.61	-0.30	8.48	A	A	A	0.05
155	9.70	0.85	-3.96	10.59	A	A	A	0.67
156	10.4	0.9	2.97	10.50	A	A	A	0.50
157	10.7	1.1	5.94	11.87	A	A	A	1.00
158	10.229	0.716	1.28	9.18	A	A	A	0.21
159	10.5	0.5	3.96	7.61	A	A	A	0.67
160	10.5	0.6	3.96	8.24	A	A	A	0.67
161	11.5	1.3	13.86	12.77	A	A	A	2.33
162	10.1	1.1	0.00	12.41	A	A	A	0.00
163	8.32	0.76	-17.62	10.90	A	A	A	2.97
163	8.32	0.76	-17.62	10.90	A	A	A	2.97
164	10.6	1.3	4.95	13.63	A	A	A	0.83
165	10.77	0.32	6.63	6.64	A	A	A	1.12
166	10.7	0.6	5.94	8.17	A	A	A	1.00
167	10.1	0.7	0.00	9.13	A	A	A	0.00
169	9.58	0.87	-5.15	10.85	A	A	A	0.87
171	11.2	0.6	10.89	8.00	A	A	A	1.83
172	10.18	0.64	0.79	8.65	A	A	A	0.13
173	11.42	0.62	13.07	8.05	A	A	A	2.20
175	9.9	0.7	-1.98	9.24	A	A	A	0.33
177	10.8	0.6	6.93	8.13	A	A	A	1.17
178	11.0	1.2	8.91	12.42	A	A	A	1.50
179	9.2501	0.68596	-8.41	9.50	A	A	A	1.42
180	9.12	0.74	-9.70	10.06	A	A	A	1.63
182	9.9	0.5	-1.98	7.80	A	A	A	0.33
183	10.15	0.76	0.50	9.56	A	A	A	0.08
187	10.2	0.6	0.99	8.36	A	A	A	0.17
187	10.3	0.7	1.98	9.03	A	A	A	0.33
189	9.88	0.99	-2.18	11.65	A	A	A	0.37
191	10.2	0.6	0.99	8.36	A	A	A	0.17
192	6.96	1.35	-31.09	20.29	N	A	N	5.23
192	9.70	1.11	-3.96	12.89	A	A	A	0.67
194	4.5	25	-55.45	555.59	N	N	N	9.33
195	9.8	0.42	-2.97	7.33	A	A	A	0.50
196	10.2	1.42	0.99	15.14	A	A	A	0.17
198	9.93	0.64	-1.68	8.77	A	A	A	0.28
201	9.2	0.9	-8.91	11.45	A	A	A	1.50
202	9.71	0.52	-3.86	8.00	A	A	A	0.65

TABLE 24. Evaluation Results for Am-241 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
203	10.6	0.8	4.95	9.60	A	A	A	0.83
204	10.3	0.92	1.98	10.73	A	A	A	0.33
204	10.3	0.92	1.98	10.73	A	A	A	0.33
206	9.66	0.07	-4.36	5.98	A	A	A	0.73
207	9.99	2.17	-1.09	22.52	A	A	A	0.18
210	9.8	0.5	-2.97	7.83	A	A	A	0.50
213	9.9	0.5	-1.98	7.80	A	A	A	0.33
214	9.9	0.6	-1.98	8.49	A	A	A	0.33
215	9.91	0.88	-1.88	10.68	A	A	A	0.32
216	10.1	0.2	0.00	6.26	A	A	A	0.00
216	10.1	0.2	0.00	6.26	A	A	A	0.00
217	10.2	1.0	0.99	11.46	A	A	A	0.17
218	10.26	1.2	1.58	13.12	A	A	A	0.27
219	9.93	0.2	-1.68	6.27	A	A	A	0.28
220	9.99	1.5	-1.09	16.15	A	A	A	0.18
221	10.19	0.39	0.89	7.07	A	A	A	0.15
222	10.4	1.2	2.97	12.98	A	A	A	0.50
223	10.31	0.24	2.08	6.38	A	A	A	0.35
224	10.37	0.28	2.67	6.53	A	A	A	0.45
226	8.68	0.448	-14.06	7.87	A	A	A	2.37
226	8.38	0.78	-17.03	11.04	A	A	A	2.87
227	9.4	0.24	-6.93	6.47	A	A	A	1.17
228	8.57	1.27	-15.15	15.97	A	A	A	2.55
230	10.9	0.7	7.92	8.75	A	A	A	1.33
231	9.2	0.9	-8.91	11.45	A	A	A	1.50
233	4.7	0.47	-53.47	11.63	N	N	N	9.00
235	10.25	0.94	1.49	10.93	A	A	A	0.25
237	9.78	0.19	-3.17	6.25	A	A	A	0.53
238	9.8	0.5	-2.97	7.83	A	A	A	0.50
239	9.34	1.35	-7.52	15.63	A	A	A	1.27
243	10.03	0.51	-0.69	7.82	A	A	A	0.12
244	10.4	1.0	2.97	11.30	A	A	A	0.50
246	9.5	0.5	-5.94	7.94	A	A	A	1.00
247	9.86	0.52	-2.38	7.94	A	A	A	0.40
247	10.12	0.50	0.20	7.73	A	A	A	0.03
248	10.29	0.39	1.88	7.05	A	A	A	0.32
249	27	16	167.33	59.56	N	N	N	28.17
250	9.88	0.47	-2.18	7.61	A	A	A	0.37
251	10.8	0.62	6.93	8.26	A	A	A	1.17
253	10.65	0.50	5.45	7.57	A	A	A	0.92
255	10.0	0.8	-0.99	9.96	A	A	A	0.17
256	5.16	0.46	-48.91	10.71	N	N	N	8.23
258	9.67	0.215	-4.26	6.34	A	A	A	0.72
259	10.32	1.07	2.18	11.95	A	A	A	0.37
261	10.1	0.8	0.00	9.90	A	A	A	0.00
262	10.4	1.35	2.97	14.28	A	A	A	0.50
264	9.72	1.17	-3.76	13.42	A	A	A	0.63
265	11.84	0.36	17.23	6.67	A	N	W	2.90
266	10.51	1.23	4.06	13.12	A	A	A	0.68
269	10.49	0.63	3.86	8.45	A	A	A	0.65
271	10.2	0.6	0.99	8.36	A	A	A	0.17
272	10.17	0.95	0.69	11.07	A	A	A	0.12
272	10.56	0.93	4.55	10.62	A	A	A	0.77
273	10.8	0.5	6.93	7.53	A	A	A	1.17

TABLE 24. Evaluation Results for Am-241 in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
274	10.04	2.452	-0.59	25.13	A	A	A	0.10
276	11.91	1.25	17.92	12.06	A	A	A	3.02
278	10.7	1.0	5.94	11.07	A	A	A	1.00
281	10.3	0.7	1.98	9.03	A	A	A	0.33
284	9.84	0.52	-2.57	7.95	A	A	A	0.43
285	9.26	0.78	-8.32	10.31	A	A	A	1.40
286	9.9	0.5	-1.98	7.80	A	A	A	0.33
287	10.73	0.674	6.24	8.65	A	A	A	1.05
288	10.1	1.0	0.00	11.55	A	A	A	0.00
289	10.3	0.6	1.98	8.32	A	A	A	0.33
290	9.36	0.41	-7.33	7.38	A	A	A	1.23
292	10.3	1.4	1.98	14.83	A	A	A	0.33
295	10.99	1.69	8.81	16.49	A	A	A	1.48
296	10.7	0.4	5.94	7.02	A	A	A	1.00
298	10.28	0.16	1.78	6.14	A	A	A	0.30
299	7.83	0.83	-22.48	12.15	A	A	A	3.78
301	11.9	0.4	17.82	6.83	A	N	W	3.00
303	9.97	0.44	-1.29	7.40	A	A	A	0.22
304	11.8	0.43	16.83	6.97	A	A	A	2.83
305	10.0	2.566	-0.99	26.34	A	A	A	0.17
305	10.015	2.566	-0.84	26.30	A	A	A	0.14
306	9.3	0.7	-7.92	9.59	A	A	A	1.33
309	10.69	1.249	5.84	13.11	A	A	A	0.98
310	10.25	1.59	1.49	16.61	A	A	A	0.25
312	9.54	0.41	-5.54	7.33	A	A	A	0.93
313	10.6	0.7	4.95	8.88	A	A	A	0.83

Reported Results for Cs-134 Determination in Sample 3

TABLE 25. Target values

Parameter	Value
Target Value, Bq/kg	12.1
UNC, Bq/kg	0.7
MARB, %	20




FIGURE 25. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 26. Reported results and evaluation parameters for Cs-134. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 26. Evaluation Results for Cs-134 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
1	11.6	0.42	-4.13	6.82	A	A	A	0.71
4	12.29	0.23	1.57	6.08	A	A	A	0.27
5	10.45	0.32	-13.64	6.55	A	A	A	2.36
6	12.88	0.81	6.45	8.54	A	A	A	1.11
7	11.6	0.5	-4.13	7.21	A	A	A	0.71
8	12.0	1.0	-0.83	10.14	A	A	A	0.14
9	12.51	0.40	3.39	6.61	A	A	A	0.59
10	14.55	2.98	20.25	21.28	N	N	N	3.50
11	11.4	1.1	-5.79	11.25	A	A	A	1.00
12	13.1	0.5	8.26	6.93	A	A	A	1.43
13	12	1	-0.83	10.14	A	A	A	0.14
14	12.57	0.48	3.88	6.93	A	A	A	0.67
16	11.77	0.24	-2.73	6.13	A	A	A	0.47
17	12.9	1.7	6.61	14.39	A	A	A	1.14
18	11.90	0.97	-1.65	10.00	A	A	A	0.29
19	11.90	1.14	-1.65	11.19	A	A	A	0.29
20	11	0.5	-9.09	7.36	A	A	A	1.57
21	12.1	0.5	0.00	7.11	A	A	A	0.00
22	12.319	0.160	1.81	5.93	A	A	A	0.31
23	12.02	0.35	-0.66	6.48	A	A	A	0.11
24	12.3	0.5	1.65	7.07	A	A	A	0.29
25	10.59	0.73	-12.48	9.00	A	A	A	2.16
26	11.7	0.4	-3.31	6.72	A	A	A	0.57
27	12.4	0.52	2.48	7.15	A	A	A	0.43
28	11.7	0.6	-3.31	7.73	A	A	A	0.57
29	12.36	0.46	2.15	6.88	A	A	A	0.37
31	12.1	1.0	0.00	10.09	A	A	A	0.00
32	11.3	0.2	-6.61	6.05	A	A	A	1.14
33	10.37	0.57	-14.30	7.98	A	A	A	2.47
34	12.63	0.74	4.38	8.23	A	A	A	0.76
35	12.5	0.6	3.31	7.52	A	A	A	0.57
36	12	3	-0.83	25.66	A	N	W	0.14
37	12.0	0.7	-0.83	8.22	A	A	A	0.14
38	13	0.8	7.44	8.45	A	A	A	1.29
39	12.3	0.7	1.65	8.12	A	A	A	0.29
40	12.2	0.2	0.83	6.01	A	A	A	0.14
41	12.2	0.87	0.83	9.18	A	A	A	0.14
42	5.91	0.24	-51.16	7.07	N	N	N	8.84
43	12.3	0.6	1.65	7.57	A	A	A	0.29
45	12.18	0.55	0.66	7.34	A	A	A	0.11
46	12.1	0.6	0.00	7.62	A	A	A	0.00
47	11.8	1.0	-2.48	10.26	A	A	A	0.43
49	11.9	0.7	-1.65	8.25	A	A	A	0.29
50	12.1	0.2	0.00	6.02	A	A	A	0.00
51	12.7	0.2	4.96	6.00	A	A	A	0.86
52	10.73	0.92	-11.32	10.34	A	A	A	1.96
53	12.77	0.33	5.54	6.34	A	A	A	0.96
55	11.98	0.47	-0.99	6.99	A	A	A	0.17
56	11.5	0.7	-4.96	8.40	A	A	A	0.86
57	11.81	0.59	-2.40	7.64	A	A	A	0.41
58	12.15	0.20	0.41	6.01	A	A	A	0.07
59	11.5	0.8	-4.96	9.05	A	A	A	0.86
60	12.2	1.0	0.83	10.03	A	A	A	0.14

TABLE 26. Evaluation Results for Cs-134 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
61	11.8	0.54	-2.48	7.38	A	A	A	0.43
62	12.4	0.51	2.48	7.10	A	A	A	0.43
63	11.83	0.51	-2.23	7.21	A	A	A	0.39
64	11.9	0.6	-1.65	7.67	A	A	A	0.29
65	12.41	0.17	2.56	5.95	A	A	A	0.44
67	11.88	0.85	-1.82	9.20	A	A	A	0.31
68	12.05	0.54	-0.41	7.32	A	A	A	0.07
69	11.25	0.10	-7.02	5.85	A	A	A	1.21
70	13.1	0.3	8.26	6.22	A	A	A	1.43
71	11.78	0.26	-2.64	6.19	A	A	A	0.46
72	11.8	0.8	-2.48	8.91	A	A	A	0.43
74	11.85	0.77	-2.07	8.70	A	A	A	0.36
75	12.8	0.6	5.79	7.45	A	A	A	1.00
76	12.3	0.30	1.65	6.28	A	A	A	0.29
78	11.9	0.4	-1.65	6.69	A	A	A	0.29
79	12.7	0.9	4.96	9.15	A	A	A	0.86
80	10.55	0.30	-12.81	6.45	A	A	A	2.21
81	12.3	0.6	1.65	7.57	A	A	A	0.29
83	11.89	0.50	-1.74	7.15	A	A	A	0.30
84	11.974	0.561	-1.04	7.44	A	A	A	0.18
85	12.2	0.40	0.83	6.65	A	A	A	0.14
86	10.9	0.855	-9.92	9.75	A	A	A	1.71
87	11.7	0.45	-3.31	6.95	A	A	A	0.57
88	12.0	0.6	-0.83	7.65	A	A	A	0.14
90	11.6	0.34	-4.13	6.49	A	A	A	0.71
91	12.0	0.6	-0.83	7.65	A	A	A	0.14
92	12.36	0.41	2.15	6.67	A	A	A	0.37
93	12.3	0.17	1.65	5.95	A	A	A	0.29
94	13.4	0.7	10.74	7.79	A	A	A	1.86
95	11.72	0.26	-3.14	6.20	A	A	A	0.54
96	11.96	0.16	-1.16	5.94	A	A	A	0.20
97	11.91	0.82	-1.57	8.99	A	A	A	0.27
98	12.4	0.7	2.48	8.08	A	A	A	0.43
99	11.43	1.37	-5.54	13.31	A	A	A	0.96
101	11.8	1.2	-2.48	11.70	A	A	A	0.43
102	11.6	0.654	-4.13	8.08	A	A	A	0.71
103	12.4	0.56	2.48	7.34	A	A	A	0.43
105	10.50	1.20	-13.22	12.81	A	A	A	2.29
106	12.57	0.52	3.88	7.11	A	A	A	0.67
107	12.2	0.47	0.83	6.95	A	A	A	0.14
109	10.90	0.42	-9.92	6.95	A	A	A	1.71
110	11.9	0.761	-1.65	8.62	A	A	A	0.29
111	12.2	1.17	0.83	11.20	A	A	A	0.14
113	11.58	0.12	-4.30	5.88	A	A	A	0.74
114	11.4	0.6	-5.79	7.82	A	A	A	1.00
115	12.21	0.66	0.91	7.92	A	A	A	0.16
117	12.0	0.5	-0.83	7.13	A	A	A	0.14
118	10.6	0.2	-12.40	6.09	A	A	A	2.14
119	12.659	1.329	4.62	11.99	A	A	A	0.80
120	11.9	0.5	-1.65	7.15	A	A	A	0.29
121	12.083	0.479	-0.14	7.01	A	A	A	0.02
123	12.3	0.8	1.65	8.70	A	A	A	0.29
124	10.46	0.30	-13.55	6.46	A	A	A	2.34
126	11.77	0.86	-2.73	9.32	A	A	A	0.47

TABLE 26. Evaluation Results for Cs-134 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
127	11.65	0.93	-3.72	9.86	A	A	A	0.64
128	13.1	0.6	8.26	7.38	A	A	A	1.43
129	12.3	0.7	1.65	8.12	A	A	A	0.29
130	12.2	0.12	0.83	5.87	A	A	A	0.14
132	11.4	0.4	-5.79	6.77	A	A	A	1.00
133	12.6	1.3	4.13	11.83	A	A	A	0.71
134	10.4	0.5	-14.05	7.52	A	A	A	2.43
135	3.6	0.6	-70.25	17.64	N	N	N	12.14
136	12.2	1.0	0.83	10.03	A	A	A	0.14
137	9.49	0.20	-21.57	6.16	N	N	N	3.73
138	10.5251	0.3968	-13.02	6.91	A	A	A	2.25
139	11.4	0.17	-5.79	5.97	A	A	A	1.00
140	12.1	0.8	0.00	8.79	A	A	A	0.00
143	11.8	0.54	-2.48	7.38	A	A	A	0.43
145	12.0	0.4	-0.83	6.68	A	A	A	0.14
146	10.77	0.0394	-10.99	5.80	A	A	A	1.90
147	12.80	0.36	5.79	6.43	A	A	A	1.00
148	11.22	0.82	-7.27	9.32	A	A	A	1.26
149	10.1	0.3	-16.53	6.50	A	A	A	2.86
150	12.185	0.807	0.70	8.79	A	A	A	0.12
152	11.9	0.6	-1.65	7.67	A	A	A	0.29
153	7.7	0.2	-36.36	6.34	N	N	N	6.29
154	11.97	0.51	-1.07	7.18	A	A	A	0.19
155	9.58	0.60	-20.83	8.53	N	A	N	3.60
156	12.3	0.9	1.65	9.33	A	A	A	0.29
157	12.2	0.9	0.83	9.37	A	A	A	0.14
158	11.932	0.716	-1.39	8.34	A	A	A	0.24
159	12.0	0.5	-0.83	7.13	A	A	A	0.14
160	10.2	0.2	-15.70	6.11	A	A	A	2.71
161	11.82	0.8	-2.31	8.90	A	A	A	0.40
162	10.8	0.7	-10.74	8.69	A	A	A	1.86
163	11.32	0.6	-6.45	7.85	A	A	A	1.11
164	11.00	0.78	-9.09	9.15	A	A	A	1.57
165	12.22	0.40	0.99	6.65	A	A	A	0.17
166	12.6	0.4	4.13	6.60	A	A	A	0.71
167	11.5	0.7	-4.96	8.40	A	A	A	0.86
168	12.87	0.32	6.36	6.30	A	A	A	1.10
169	11.34	0.32	-6.28	6.44	A	A	A	1.09
171	13.0	1.6	7.44	13.60	A	A	A	1.29
172	12.37	0.62	2.23	7.65	A	A	A	0.39
173	12.59	0.34	4.05	6.38	A	A	A	0.70
175	12.2	0.6	0.83	7.59	A	A	A	0.14
176	9.9	0.4	-18.18	7.06	A	A	A	3.14
177	11.3	0.9	-6.61	9.84	A	A	A	1.14
178	11.5	0.5	-4.96	7.24	A	A	A	0.86
179	12.109	0.35063	0.07	6.47	A	A	A	0.01
180	11.5	0.6	-4.96	7.79	A	A	A	0.86
181	11.66	0.80	-3.64	8.97	A	A	A	0.63
182	12.1	0.3	0.00	6.29	A	A	A	0.00
183	11.49	0.28	-5.04	6.28	A	A	A	0.87
184	12.3	0.5	1.65	7.07	A	A	A	0.29
185	11.7	1.5	-3.31	14.07	A	A	A	0.57
187	11.9	0.6	-1.65	7.67	A	A	A	0.29
188	11.9	0.88	-1.65	9.39	A	A	A	0.29

TABLE 26. Evaluation Results for Cs-134 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
189	12.14	1.00	0.33	10.07	A	A	A	0.06
191	12.0	0.7	-0.83	8.22	A	A	A	0.14
192	11.75	0.49	-2.89	7.13	A	A	A	0.50
193	1.351	0.084	-88.83	8.49	N	N	N	15.36
194	33.1	5	173.55	16.18	N	N	N	30.00
195	12.6	0.56	4.13	7.30	A	A	A	0.71
196	12.5	0.673	3.31	7.90	A	A	A	0.57
197	10.7	2.5	-11.57	24.07	A	N	W	2.00
198	12.83	0.76	6.03	8.28	A	A	A	1.04
201	12.4	0.5	2.48	7.05	A	A	A	0.43
202	11.5	0.6	-4.96	7.79	A	A	A	0.86
203	12.6	0.6	4.13	7.49	A	A	A	0.71
204	12.3	0.81	1.65	8.77	A	A	A	0.29
206	13.58	0.06	12.23	5.80	A	A	A	2.11
207	12.02	0.42	-0.66	6.76	A	A	A	0.11
209	12.3	2.4	1.65	20.35	A	N	W	0.29
210	12.1	0.4	0.00	6.66	A	A	A	0.00
213	12.0	0.6	-0.83	7.65	A	A	A	0.14
214	10.4	0.3	-14.05	6.46	A	A	A	2.43
215	11.55	0.95	-4.55	10.06	A	A	A	0.79
216	12.7	0.2	4.96	6.00	A	A	A	0.86
217	11.5	0.84	-4.96	9.32	A	A	A	0.86
218	12.1	0.56	0.00	7.41	A	A	A	0.00
219	11.1	0.3	-8.26	6.39	A	A	A	1.43
220	12.27	0.5	1.40	7.08	A	A	A	0.24
221	12.53	0.33	3.55	6.36	A	A	A	0.61
222	11.9	0.8	-1.65	8.87	A	A	A	0.29
223	11.85	0.07	-2.07	5.82	A	A	A	0.36
224	11.8	0.08	-2.48	5.82	A	A	A	0.43
226	10.40	1.19	-14.05	12.82	A	A	A	2.43
227	12.65	0.3	4.55	6.25	A	A	A	0.79
228	12.6	1.6	4.13	13.95	A	A	A	0.71
230	11.8	0.8	-2.48	8.91	A	A	A	0.43
231	8.1	0.8	-33.06	11.45	N	N	N	5.71
233	13.1	0.64	8.26	7.57	A	A	A	1.43
234	12.20	0.40	0.83	6.65	A	A	A	0.14
235	11.9	0.27	-1.65	6.21	A	A	A	0.29
237	12.47	0.24	3.06	6.10	A	A	A	0.53
238	12.2	0.4	0.83	6.65	A	A	A	0.14
239	13.04	1.89	7.77	15.61	A	A	A	1.34
241	9.87	0.50	-18.43	7.69	A	A	A	3.19
244	12.1	1.3	0.00	12.20	A	A	A	0.00
246	10.9	0.3	-9.92	6.41	A	A	A	1.71
247	12.65	0.50	4.55	7.01	A	A	A	0.79
248	12.39	0.36	2.40	6.47	A	A	A	0.41
249	3.93	5	-67.52	127.36	N	N	N	11.67
250	11.79	0.24	-2.56	6.13	A	A	A	0.44
251	12.01	0.93	-0.74	9.67	A	A	A	0.13
253	12.34	0.21	1.98	6.03	A	A	A	0.34
255	13.1	0.4	8.26	6.54	A	A	A	1.43
256	11.3	0.27	-6.61	6.26	A	A	A	1.14
258	11.7	0.282	-3.31	6.27	A	A	A	0.57
259	11.2	0.47	-7.44	7.15	A	A	A	1.29
261	12.3	0.9	1.65	9.33	A	A	A	0.29

TABLE 26. Evaluation Results for Cs-134 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
262	12.6	1.63	4.13	14.17	A	A	A	0.71
263	12.033	0.852	-0.55	9.14	A	A	A	0.10
264	11.8	1.16	-2.48	11.41	A	A	A	0.43
266	11.97	0.53	-1.07	7.29	A	A	A	0.19
269	11.07	0.64	-8.51	8.18	A	A	A	1.47
271	12.4	1.2	2.48	11.27	A	A	A	0.43
272	9.78	0.12	-19.17	5.91	A	N	W	3.31
274	14.03	2.309	15.95	17.44	A	A	A	2.76
276	8.88	0.65	-26.61	9.33	N	N	N	4.60
278	11.4	0.5	-5.79	7.26	A	A	A	1.00
281	11.9	0.7	-1.65	8.25	A	A	A	0.29
284	8.45	0.38	-30.17	7.33	N	N	N	5.21
285	10.68	0.74	-11.74	9.03	A	A	A	2.03
286	12.2	0.5	0.83	7.09	A	A	A	0.14
287	10.455	0.43	-13.60	7.10	A	A	A	2.35
289	12.2	0.8	0.83	8.74	A	A	A	0.14
290	10.29	0.50	-14.96	7.56	A	A	A	2.59
292	11.2	1.6	-7.44	15.41	A	A	A	1.29
295	12.90	0.36	6.61	6.42	A	A	A	1.14
296	12.6	0.5	4.13	7.02	A	A	A	0.71
297	0.39	0.04	-96.78	11.78	N	N	N	16.73
298	12.13	0.15	0.25	5.92	A	A	A	0.04
299	9.69	0.49	-19.92	7.68	A	N	W	3.44
301	11.8	0.3	-2.48	6.32	A	A	A	0.43
303	10.88	0.20	-10.08	6.07	A	A	A	1.74
304	22.8	3.1	88.43	14.78	N	N	N	15.29
305	9.4	1.687	-22.31	18.86	N	A	N	3.86
306	12.2	0.4	0.83	6.65	A	A	A	0.14
307	10.00	0.20	-17.36	6.12	A	N	W	3.00
309	11.88	1.167	-1.82	11.40	A	A	A	0.31
310	11.8	1.3	-2.48	12.44	A	A	A	0.43
312	12.48	0.42	3.14	6.69	A	A	A	0.54
313	12.3	0.8	1.65	8.70	A	A	A	0.29

Reported Results for Cs-137 Determination in Sample 3

TABLE 27. Target values

Parameter	Value
Target Value, Bq/kg	22.6
UNC, Bq/kg	1.4
MARB, %	20




FIGURE 27. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 28. Reported results and evaluation parameters for Cs-137. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 28. Evaluation Results for Cs-137 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
1	23.7	0.81	4.87	7.07	A	A	A	0.79
4	21.65	0.32	-4.20	6.37	A	A	A	0.68
5	22.24	0.70	-1.59	6.95	A	A	A	0.26
6	21.34	0.89	-5.58	7.47	A	A	A	0.90
7	22.7	1.1	0.44	7.86	A	A	A	0.07
8	23.2	1.2	2.65	8.07	A	A	A	0.43
9	24.48	1.21	8.32	7.92	A	A	A	1.34
10	22.99	4.47	1.73	20.41	A	N	W	0.28
11	23.3	1.3	3.10	8.34	A	A	A	0.50
12	24.4	1.2	7.96	7.91	A	A	A	1.29
13	22	1	-2.65	7.68	A	A	A	0.43
14	23.51	0.89	4.03	7.26	A	A	A	0.65
16	22.39	0.48	-0.93	6.56	A	A	A	0.15
17	23.1	3.0	2.21	14.39	A	A	A	0.36
18	22.08	1.32	-2.30	8.61	A	A	A	0.37
19	22.32	2.12	-1.24	11.34	A	A	A	0.20
20	20.8	1.1	-7.96	8.15	A	A	A	1.29
21	23.3	1.43	3.10	8.72	A	A	A	0.50
22	23.287	0.338	3.04	6.36	A	A	A	0.49
23	22.99	0.71	1.73	6.92	A	A	A	0.28
24	22.9	0.8	1.33	7.11	A	A	A	0.21
25	23.87	0.98	5.62	7.43	A	A	A	0.91
26	23.1	1.4	2.21	8.67	A	A	A	0.36
27	21.9	0.86	-3.10	7.33	A	A	A	0.50
28	22.4	1.1	-0.88	7.91	A	A	A	0.14
29	23.04	0.55	1.95	6.64	A	A	A	0.31
31	23.2	2.2	2.65	11.33	A	A	A	0.43
32	23.7	0.5	4.87	6.54	A	A	A	0.79
33	23.45	1.51	3.76	8.94	A	A	A	0.61
34	22.88	1.36	1.24	8.59	A	A	A	0.20
35	28.3	1.5	25.22	8.15	N	N	N	4.07
36	22	3	-2.65	14.98	A	A	A	0.43
37	23.4	1.3	3.54	8.32	A	A	A	0.57
38	22	1.6	-2.65	9.55	A	A	A	0.43
39	22.8	1.1	0.88	7.85	A	A	A	0.14
40	23.0	0.2	1.77	6.26	A	A	A	0.29
41	23.3	1.74	3.10	9.70	A	A	A	0.50
42	11.82	0.47	-47.70	7.36	N	N	N	7.70
43	23.0	1.0	1.77	7.57	A	A	A	0.29
45	22.79	0.47	0.84	6.53	A	A	A	0.14
46	22.6	1.0	0.00	7.61	A	A	A	0.00
47	22.8	1.4	0.88	8.72	A	A	A	0.14
49	22.7	0.7	0.44	6.92	A	A	A	0.07
50	22.8	0.4	0.88	6.44	A	A	A	0.14
51	23.5	0.3	3.98	6.32	A	A	A	0.64
52	22.34	1.78	-1.15	10.09	A	A	A	0.19
53	26.62	0.62	17.79	6.62	A	N	W	2.87
55	23.0	0.78	1.77	7.06	A	A	A	0.29
56	21.7	1.1	-3.98	8.00	A	A	A	0.64
57	22.33	1.41	-1.19	8.85	A	A	A	0.19
58	23.07	0.32	2.08	6.35	A	A	A	0.34
59	23.4	0.8	3.54	7.08	A	A	A	0.57
60	22.6	1.4	0.00	8.76	A	A	A	0.00

TABLE 28. Evaluation Results for Cs-137 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
61	22.2	1.0	-1.77	7.66	A	A	A	0.29
62	23.6	0.9	4.42	7.27	A	A	A	0.71
63	22.95	1.43	1.55	8.79	A	A	A	0.25
64	22.7	1.7	0.44	9.72	A	A	A	0.07
65	23.43	0.31	3.67	6.33	A	A	A	0.59
67	23.30	1.00	3.10	7.54	A	A	A	0.50
68	22.9	0.8	1.33	7.11	A	A	A	0.21
69	21.73	0.23	-3.85	6.28	A	A	A	0.62
70	22.9	0.8	1.33	7.11	A	A	A	0.21
71	22.32	0.32	-1.24	6.36	A	A	A	0.20
72	23.3	0.8	3.10	7.08	A	A	A	0.50
74	23.09	1.42	2.17	8.73	A	A	A	0.35
75	23.2	0.7	2.65	6.89	A	A	A	0.43
76	23.0	0.60	1.77	6.72	A	A	A	0.29
78	22.6	0.7	0.00	6.93	A	A	A	0.00
79	23.4	1.1	3.54	7.78	A	A	A	0.57
80	21.76	0.72	-3.72	7.02	A	A	A	0.60
81	23.1	1.0	2.21	7.56	A	A	A	0.36
83	22.81	0.60	0.93	6.73	A	A	A	0.15
84	23.361	1.025	3.37	7.59	A	A	A	0.54
85	23.1	0.76	2.21	7.01	A	A	A	0.36
86	20.7	1.72	-8.41	10.36	A	A	A	1.36
87	22.9	1.1	1.33	7.84	A	A	A	0.21
88	23.5	1.1	3.98	7.76	A	A	A	0.64
90	21.9	1.34	-3.10	8.71	A	A	A	0.50
91	22.6	1.3	0.00	8.45	A	A	A	0.00
92	23.50	0.94	3.98	7.37	A	A	A	0.64
93	22.8	0.55	0.88	6.65	A	A	A	0.14
94	24.5	1.2	8.41	7.90	A	A	A	1.36
95	23.76	0.35	5.13	6.37	A	A	A	0.83
96	22.92	0.32	1.42	6.35	A	A	A	0.23
97	22.32	0.28	-1.24	6.32	A	A	A	0.20
98	22.9	1.3	1.33	8.40	A	A	A	0.21
99	22.67	2.27	0.31	11.77	A	A	A	0.05
100	23.0	0.3	1.77	6.33	A	A	A	0.29
101	24.1	2.4	6.64	11.73	A	A	A	1.07
102	23.2	0.851	2.65	7.20	A	A	A	0.43
103	23.2	1.14	2.65	7.91	A	A	A	0.43
105	21.6	1.43	-4.42	9.07	A	A	A	0.71
106	22.46	0.823	-0.62	7.20	A	A	A	0.10
107	23.1	1.34	2.21	8.49	A	A	A	0.36
109	22.48	0.85	-0.53	7.26	A	A	A	0.09
110	23.2	1.39	2.65	8.62	A	A	A	0.43
111	23.0	1.10	1.77	7.83	A	A	A	0.29
113	22.01	0.30	-2.61	6.34	A	A	A	0.42
114	24.1	0.9	6.64	7.23	A	A	A	1.07
115	22.1	1.1	-2.21	7.95	A	A	A	0.36
117	22.2	1.4	-1.77	8.84	A	A	A	0.29
118	22.8	0.9	0.88	7.35	A	A	A	0.14
119	22.910	2.406	1.37	12.19	A	A	A	0.22
120	22.3	0.7	-1.33	6.94	A	A	A	0.21
121	22.901	0.541	1.33	6.63	A	A	A	0.21
123	22.7	1.1	0.44	7.86	A	A	A	0.07
124	22.83	0.55	1.02	6.65	A	A	A	0.16

TABLE 28. Evaluation Results for Cs-137 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
125	25.61	1.79	13.32	9.34	A	A	A	2.15
126	23.3	1.5	3.10	8.93	A	A	A	0.50
127	22.04	1.76	-2.48	10.11	A	A	A	0.40
128	23.7	0.8	4.87	7.05	A	A	A	0.79
129	22.2	1.2	-1.77	8.22	A	A	A	0.29
130	23.03	0.19	1.90	6.25	A	A	A	0.31
132	24.5	1.5	8.41	8.71	A	A	A	1.36
133	22.9	2.4	1.33	12.17	A	A	A	0.21
134	21.9	0.6	-3.10	6.77	A	A	A	0.50
135	20	2	-11.50	11.76	A	A	A	1.86
136	22.3	1.2	-1.33	8.21	A	A	A	0.21
137	20.90	0.20	-7.52	6.27	A	A	A	1.21
138	22.6639	0.7498	0.28	7.02	A	A	A	0.05
139	23	0.26	1.77	6.30	A	A	A	0.29
140	22.4	1.8	-0.88	10.15	A	A	A	0.14
143	22.4	1.04	-0.88	7.74	A	A	A	0.14
145	23.4	0.8	3.54	7.08	A	A	A	0.57
146	22.53	0.0392	-0.31	6.20	A	A	A	0.05
147	25.17	0.74	11.37	6.86	A	A	A	1.84
148	22.95	0.58	1.55	6.69	A	A	A	0.25
149	22.4	1.1	-0.88	7.91	A	A	A	0.14
150	22.855	1.496	1.13	9.01	A	A	A	0.18
152	21.3	1.1	-5.75	8.07	A	A	A	0.93
153	22.8	0.7	0.88	6.91	A	A	A	0.14
154	22.5	1.03	-0.44	7.70	A	A	A	0.07
155	22.17	1.40	-1.90	8.85	A	A	A	0.31
156	23.0	1.1	1.77	7.83	A	A	A	0.29
157	23.2	1.7	2.65	9.60	A	A	A	0.43
158	22.997	0.920	1.76	7.37	A	A	A	0.28
159	23.1	1.0	2.21	7.56	A	A	A	0.36
160	21.9	0.6	-3.10	6.77	A	A	A	0.50
161	24.53	1.0	8.54	7.42	A	A	A	1.38
162	20.9	1.3	-7.52	8.78	A	A	A	1.21
163	22.91	1.17	1.37	8.03	A	A	A	0.22
164	20.9	1.8	-7.52	10.61	A	A	A	1.21
165	23.22	0.69	2.74	6.87	A	A	A	0.44
166	22.3	0.7	-1.33	6.94	A	A	A	0.21
167	22.2	1.3	-1.77	8.52	A	A	A	0.29
168	25.41	0.66	12.43	6.72	A	A	A	2.01
169	22.90	0.86	1.33	7.24	A	A	A	0.21
171	25.2	3.3	11.50	14.49	A	A	A	1.86
172	22.35	0.84	-1.11	7.25	A	A	A	0.18
173	24.17	0.48	6.95	6.51	A	A	A	1.12
175	22.7	1.1	0.44	7.86	A	A	A	0.07
176	22.9	0.9	1.33	7.34	A	A	A	0.21
177	21.6	1.1	-4.42	8.02	A	A	A	0.71
178	22.4	1.4	-0.88	8.80	A	A	A	0.14
179	22.677	0.64953	0.34	6.82	A	A	A	0.05
180	22.8	1.1	0.88	7.85	A	A	A	0.14
181	23.65	0.99	4.65	7.48	A	A	A	0.75
182	22.8	0.9	0.88	7.35	A	A	A	0.14
183	20.37	0.64	-9.87	6.95	A	A	A	1.59
184	23.0	0.9	1.77	7.33	A	A	A	0.29
185	24.7	3.6	9.29	15.84	A	A	A	1.50

TABLE 28. Evaluation Results for Cs-137 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
186	25.1	6.3	11.06	25.85	A	N	W	1.79
187	22.8	1.2	0.88	8.13	A	A	A	0.14
188	22.3	1.39	-1.33	8.79	A	A	A	0.21
189	23.34	1.94	3.27	10.37	A	A	A	0.53
191	23.0	0.7	1.77	6.90	A	A	A	0.29
192	23.77	0.8672	5.18	7.19	A	A	A	0.84
193	3.340	0.194	-85.22	8.49	N	N	N	13.76
194	84	5	271.68	8.59	N	N	N	43.86
195	22.2	1.0	-1.77	7.66	A	A	A	0.29
196	22.9	0.785	1.33	7.08	A	A	A	0.21
197	21.6	5.1	-4.42	24.41	A	N	W	0.71
198	23.12	1.20	2.30	8.08	A	A	A	0.37
201	23.4	1.4	3.54	8.61	A	A	A	0.57
202	22.5	1.2	-0.44	8.17	A	A	A	0.07
203	23.2	1.0	2.65	7.55	A	A	A	0.43
204	22.1	1.4	-2.21	8.86	A	A	A	0.36
206	64.38	0.36	184.87	6.22	N	N	N	29.84
207	23.13	0.75	2.35	6.99	A	A	A	0.38
209	24.3	5	7.52	21.49	A	N	W	1.21
210	23.2	1.2	2.65	8.07	A	A	A	0.43
213	22.8	1.0	0.88	7.59	A	A	A	0.14
214	22.4	0.7	-0.88	6.94	A	A	A	0.14
215	23.0	1.3	1.77	8.39	A	A	A	0.29
216	23.7	0.2	4.87	6.25	A	A	A	0.79
217	23.0	1.3	1.77	8.39	A	A	A	0.29
218	22.84	0.96	1.06	7.49	A	A	A	0.17
219	23.5	0.5	3.98	6.55	A	A	A	0.64
220	23.3	0.8	3.10	7.08	A	A	A	0.50
221	23.04	0.59	1.95	6.70	A	A	A	0.31
222	22.5	0.9	-0.44	7.37	A	A	A	0.07
223	22.31	0.17	-1.28	6.24	A	A	A	0.21
224	22.33	0.16	-1.19	6.24	A	A	A	0.19
226	23.7	1.4	4.87	8.56	A	A	A	0.79
227	24	0.47	6.19	6.50	A	A	A	1.00
228	23.8	2.8	5.31	13.30	A	A	A	0.86
230	22.9	1.5	1.33	9.02	A	A	A	0.21
231	20.1	2.0	-11.06	11.72	A	A	A	1.79
233	25.1	0.75	11.06	6.88	A	A	A	1.79
234	23.30	0.60	3.10	6.71	A	A	A	0.50
235	22.48	0.44	-0.53	6.50	A	A	A	0.09
237	22.76	0.33	0.71	6.36	A	A	A	0.11
238	21.7	0.7	-3.98	6.98	A	A	A	0.64
239	21.13	3.06	-6.50	15.75	A	A	A	1.05
241	22.69	1.29	0.40	8.41	A	A	A	0.06
244	23.2	1.6	2.65	9.27	A	A	A	0.43
246	23.2	1.0	2.65	7.55	A	A	A	0.43
247	23.5	1.1	3.98	7.76	A	A	A	0.64
248	22.80	0.53	0.88	6.62	A	A	A	0.14
249	9.5	4	-57.96	42.56	N	N	N	9.36
250	21.34	0.41	-5.58	6.49	A	A	A	0.90
251	21.55	0.98	-4.65	7.68	A	A	A	0.75
253	22.68	0.35	0.35	6.38	A	A	A	0.06
255	23.7	0.7	4.87	6.86	A	A	A	0.79
256	23.92	0.78	5.84	7.00	A	A	A	0.94

TABLE 28. Evaluation Results for Cs-137 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
258	22.8	0.299	0.88	6.33	A	A	A	0.14
259	24.26	0.83	7.35	7.08	A	A	A	1.19
261	23.1	1.4	2.21	8.67	A	A	A	0.36
262	23.3	3.02	3.10	14.37	A	A	A	0.50
263	24.102	1.472	6.65	8.70	A	A	A	1.07
264	22.3	2.12	-1.33	11.35	A	A	A	0.21
266	22.01	1.40	-2.61	8.88	A	A	A	0.42
269	22.17	1.27	-1.90	8.44	A	A	A	0.31
271	23.1	1.1	2.21	7.81	A	A	A	0.36
272	23.18	0.18	2.57	6.24	A	A	A	0.41
274	25.45	4.662	12.61	19.34	A	A	A	2.04
276	23.82	1.62	5.40	9.20	A	A	A	0.87
278	23.7	1.0	4.87	7.50	A	A	A	0.79
281	22.4	1.3	-0.88	8.49	A	A	A	0.14
284	21.13	0.94	-6.50	7.63	A	A	A	1.05
285	22.00	1.57	-2.65	9.45	A	A	A	0.43
286	23.2	0.8	2.65	7.09	A	A	A	0.43
287	22.685	0.92	0.38	7.40	A	A	A	0.06
289	22.8	1.3	0.88	8.42	A	A	A	0.14
290	22.51	1.10	-0.40	7.89	A	A	A	0.06
292	22.3	3.1	-1.33	15.22	A	A	A	0.21
295	22.84	1.18	1.06	8.07	A	A	A	0.17
296	22.0	0.7	-2.65	6.96	A	A	A	0.43
298	22.79	0.26	0.84	6.30	A	A	A	0.14
299	22.10	1.13	-2.21	8.03	A	A	A	0.36
301	24.4	0.6	7.96	6.66	A	A	A	1.29
303	23.17	0.30	2.52	6.33	A	A	A	0.41
304	56.2	1.3	148.67	6.61	N	N	N	24.00
305	23.047	1.95	1.98	10.49	A	A	A	0.32
306	22.7	0.7	0.44	6.92	A	A	A	0.07
307	22.50	0.42	-0.44	6.47	A	A	A	0.07
309	23.93	2.353	5.88	11.62	A	A	A	0.95
310	25.1	2.8	11.06	12.76	A	A	A	1.79
312	23.49	0.76	3.94	6.99	A	A	A	0.64
313	23.1	1.5	2.21	8.97	A	A	A	0.36

Reported Results for Pu-239 Determination in Sample 3

TABLE 29. Target values

Parameter	Value
Target Value, Bq/kg	5.16
UNC, Bq/kg	0.31
MARB, %	30




FIGURE 29. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 30. Reported results and evaluation parameters for Pu-239. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 30. Evaluation Results for Pu-239 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
7	5.99	0.35	16.09	8.38	A	A	A	2.68
9	6.06	1.17	17.44	20.22	A	A	A	2.90
13	6.1	0.4	18.22	8.89	A	A	A	3.03
14	4.97	0.16	-3.68	6.82	A	A	A	0.61
16	4.79	0.5	-7.17	12.04	A	A	A	1.19
24	3.11	0.28	-39.73	10.82	N	N	N	6.61
36	4.2	0.4	-18.60	11.26	A	A	A	3.10
41	5.05	0.5	-2.13	11.58	A	A	A	0.35
42	5.5464	0.277	7.49	7.81	A	A	A	1.25
52	3.906	0.078	-24.30	6.33	A	N	W	4.05
57	4.34	0.2	-15.89	7.57	A	A	A	2.65
60	4.00	0.60	-22.48	16.16	A	A	A	3.74
61	1.29	0.1	-75.00	9.81	N	N	N	12.48
68	2.40	0.16	-53.49	8.97	N	N	N	8.90
71	10.02	0.84	94.19	10.31	N	N	N	15.68
83	2.7	0.3	-47.67	12.63	N	N	N	7.94
87	6.039	0.2373	17.03	7.18	A	A	A	2.84
92	3.62	0.21	-29.84	8.35	A	N	W	4.97
94	3.2	0.2	-37.98	8.67	N	N	N	6.32
96	4.39	0.37	-14.92	10.35	A	A	A	2.48
97	5.51	0.07	6.78	6.14	A	A	A	1.13
99	5.39	0.81	4.46	16.18	A	A	A	0.74
110	4.48	0.47	-13.18	12.09	A	A	A	2.19
118	5.7	0.8	10.47	15.27	A	A	A	1.74
119	6.311	1.136	22.31	18.98	A	A	A	3.71
133	6.3	1.1	22.09	18.46	A	A	A	3.68
134	10.5	2	103.49	19.97	N	N	N	17.23
140	3.9	0.95	-24.42	25.09	A	A	A	4.06
141	4.69	0.45	-9.11	11.32	A	A	A	1.52
143	4.43	2.29	-14.15	52.04	A	N	W	2.35
146	5.13		-0.58		A			0.10
157	4.2	0.4	-18.60	11.26	A	A	A	3.10
163	2.62	0.25	-49.22	11.28	N	N	N	8.19
163	2.62	0.25	-49.22	11.28	N	N	N	8.19
164	4.03	0.15	-21.90	7.07	A	N	W	3.65
168	4.27	0.32	-17.25	9.60	A	A	A	2.87
171	4.4	0.3	-14.73	9.09	A	A	A	2.45
177	3.4	0.4	-34.11	13.21	N	N	N	5.68
184	4.51	0.28	-12.60	8.64	A	A	A	2.10
187	3.7	0.4	-28.29	12.37	A	A	A	4.71
196	1.48	0.148	-71.32	11.67	N	N	N	11.87
199	4.3	0.3	-16.67	9.21	A	A	A	2.77
204	3.96	0.14	-23.26	6.97	A	N	W	3.87
207	5.65	0.29	9.50	7.90	A	A	A	1.58
214	2.2	0.3	-57.36	14.90	N	N	N	9.55
216	4.66	0.13	-9.69	6.62	A	A	A	1.61
217	5.52	0.22	6.98	7.21	A	A	A	1.16
226	4.67	0.378	-9.50	10.08	A	A	A	1.58
240	5.16	0.38	0.00	9.50	A	A	A	0.00
247	5.00	0.35	-3.10	9.22	A	A	A	0.52
253	4.03	0.13	-21.90	6.82	A	N	W	3.65
272	2.69	0.31	-47.87	13.00	N	N	N	7.97
276	2.91	0.24	-43.60	10.20	N	N	N	7.26

TABLE 30. Evaluation Results for Pu-239 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
281	3.34	0.35	-35.27	12.08	N	N	N	5.87

Reported Results for Am-241 Determination in Sample 3

TABLE 31. Target values

Parameter	Value
Target Value, Bq/kg	0.902
UNC, Bq/kg	0.054
MARB, %	30




FIGURE 31. Percentage of Accepted vs. Warnings vs. Not Accepted vs. Not Reported results




FIGURE 32. Reported results and evaluation parameters for Am-241. Blue points represent accepted scores, whereas yellow and red represent warning and not accepted scores respectively.

TABLE 32. Evaluation Results for Am-241 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
4	0.816	0.087	-9.53	12.23	A	A	A	1.59
5	7.81	0.21	765.85	6.56	N	N	N	127.93
8	0.96	0.15	6.43	16.73	A	A	A	1.07
9	0.73	0.25	-19.07	34.77	A	N	W	3.19
14	0.937	0.044	3.88	7.61	A	A	A	0.65
16	0.91	0.1	0.89	12.51	A	A	A	0.15
18	0.919	0.135	1.88	15.86	A	A	A	0.31
20	1.1	0.2	21.95	19.14	A	A	A	3.67
21	1.24	0.27	37.47	22.58	N	A	N	6.26
24	1.03	0.10	14.19	11.41	A	A	A	2.37
26	1.04	0.2	15.30	20.14	A	A	A	2.56
28	0.88	0.07	-2.44	9.96	A	A	A	0.41
31	0.89	0.1	-1.33	12.73	A	A	A	0.22
32	1.03	0.06	14.19	8.35	A	A	A	2.37
37	0.5	0.2	-44.57	40.45	N	N	N	7.44
40	0.9	0.1	-0.22	12.62	A	A	A	0.04
42	0.51	0.19	-43.46	37.73	N	N	N	7.26
45	0.96	0.15	6.43	16.73	A	A	A	1.07
49	0.78	0.07	-13.53	10.79	A	A	A	2.26
51	0.8	0.2	-11.31	25.71	A	A	A	1.89
52	0.9	0.07	-0.22	9.82	A	A	A	0.04
58	0.79	0.14	-12.42	18.71	A	A	A	2.07
59	0.9	0.15	-0.22	17.71	A	A	A	0.04
60	0.901	0.090	-0.11	11.65	A	A	A	0.02
62	0.92	0.06	2.00	8.85	A	A	A	0.33
64	1.1	0.1	21.95	10.89	A	A	A	3.67
67	0.83	0.10	-7.98	13.45	A	A	A	1.33
68	0.931	0.089	3.22	11.28	A	A	A	0.54
69	0.91	0.15	0.89	17.54	A	A	A	0.15
71	0.77	0.04	-14.63	7.93	A	A	A	2.44
75	1.2	0.4	33.04	33.87	N	N	N	5.52
76	0.85	0.18	-5.76	22.01	A	A	A	0.96
78	0.89	0.07	-1.33	9.88	A	A	A	0.22
83	0.84	0.05	-6.87	8.44	A	A	A	1.15
84	0.905	0.0655	0.33	9.39	A	A	A	0.06
87	0.908	0.116	0.67	14.11	A	A	A	0.11
88	1.15	0.24	27.49	21.71	A	A	A	4.59
93	0.845	0.053	-6.32	8.67	A	A	A	1.06
94	1.0	0.1	10.86	11.66	A	A	A	1.81
96	0.96	0.08	6.43	10.26	A	A	A	1.07
97	1.24	0.33	37.47	27.28	N	A	N	6.26
101	1.2	0.4	33.04	33.87	N	N	N	5.52
106	0.671	0.122	-25.61	19.14	A	A	A	4.28
107	0.87	0.12	-3.55	15.04	A	A	A	0.59
111	0.94	0.16	4.21	18.04	A	A	A	0.70
113	0.70	0.08	-22.39	12.90	A	A	A	3.74
115	0.92	0.17	2.00	19.42	A	A	A	0.33
117	0.8	0.1	-11.31	13.86	A	A	A	1.89
118	0.8	0.1	-11.31	13.86	A	A	A	1.89
119	0.850	0.208	-5.76	25.19	A	A	A	0.96
121	1.172	0.175	29.93	16.09	A	A	A	5.00
123	1.0	0.05	10.86	7.80	A	A	A	1.81
127	0.930	0.226	3.10	25.03	A	A	A	0.52

TABLE 32. Evaluation Results for Am-241 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
129	0.69	0.16	-23.50	23.95	A	A	A	3.93
133	0.9	0.2	-0.22	23.01	A	A	A	0.04
136	1.01	0.09	11.97	10.74	A	A	A	2.00
138	0.9204	0.2421	2.04	26.98	A	A	A	0.34
140	0.9	0.2	-0.22	23.01	A	A	A	0.04
143	1.27	0.23	40.80	19.07	N	A	N	6.81
148	0.71	0.21	-21.29	30.18	A	N	W	3.56
150	0.885	0.1	-1.88	12.79	A	A	A	0.31
154	0.944	0.077	4.66	10.12	A	A	A	0.78
156	1.02	0.12	13.08	13.20	A	A	A	2.19
159	1.19	0.24	31.93	21.04	N	A	N	5.33
160	0.9	0.2	-0.22	23.01	A	A	A	0.04
162	0.77	0.11	-14.63	15.49	A	A	A	2.44
164	0.75	0.15	-16.85	20.88	A	A	A	2.81
165	0.89	0.10	-1.33	12.73	A	A	A	0.22
169	0.92	0.12	2.00	14.35	A	A	A	0.33
171	1.05	0.18	16.41	18.16	A	A	A	2.74
172	0.81	0.21	-10.20	26.61	A	A	A	1.70
173	1.27	0.49	40.80	39.04	N	N	N	6.81
176	0.70	0.08	-22.39	12.90	A	A	A	3.74
177	1.1	0.2	21.95	19.14	A	A	A	3.67
187	0.87	0.07	-3.55	10.03	A	A	A	0.59
191	1.0	0.1	10.86	11.66	A	A	A	1.81
195	0.86	0.14	-4.66	17.34	A	A	A	0.78
197	0.92	0.29	2.00	32.09	A	N	W	0.33
201	0.6	0.2	-33.48	33.87	N	N	N	5.59
202	0.81	0.05	-10.20	8.60	A	A	A	1.70
204	1.00	0.15	10.86	16.15	A	A	A	1.81
206	93.4	0.71	10254.77	6.03	N	N	N	1712.93
213	0.9	0.2	-0.22	23.01	A	A	A	0.04
214	2.4	0.4	166.08	17.71	N	N	N	27.74
216	1.01	0.05	11.97	7.77	A	A	A	2.00
218	1.1	0.19	21.95	18.28	A	A	A	3.67
219	1.32	0.02	46.34	6.18	N	N	N	7.74
220	0.95	0.2	5.32	21.89	A	A	A	0.89
221	1.02	0.07	13.08	9.11	A	A	A	2.19
227	0.98	0.17	8.65	18.35	A	A	A	1.44
230	0.982	0.090	8.87	10.95	A	A	A	1.48
231	0.78	0.08	-13.53	11.88	A	A	A	2.26
235	0.67	0.11	-25.72	17.48	A	A	A	4.30
237	0.84	0.08	-6.87	11.25	A	A	A	1.15
238	0.82	0.07	-9.09	10.43	A	A	A	1.52
247	1.22	0.15	35.25	13.68	N	A	N	5.89
248	1.02	0.12	13.08	13.20	A	A	A	2.19
250	1.53	0.44	69.62	29.37	N	A	N	11.63
255	1.01	0.12	11.97	13.30	A	A	A	2.00
261	0.94	0.11	4.21	13.14	A	A	A	0.70
262	1.2	0.16	33.04	14.62	N	A	N	5.52
264	0.91	0.23	0.89	25.97	A	A	A	0.15
265	0.98	0.20	8.65	21.27	A	A	A	1.44
266	2.27	0.49	151.66	22.40	N	N	N	25.33
269	1.11	0.14	23.06	13.96	A	A	A	3.85
271	1.0	0.2	10.86	20.88	A	A	A	1.81
272	0.55	0.03	-39.02	8.10	N	N	N	6.52

TABLE 32. Evaluation Results for Am-241 in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Rel. Bias	P Test	Accuracy	Precision	Final Score	Z-Score
281	0.84	0.13	-6.87	16.59	A	A	A	1.15
284	1.03	0.32	14.19	31.64	A	N	W	2.37
286	0.92	0.15	2.00	17.37	A	A	A	0.33
289	1.1	0.3	21.95	27.92	A	A	A	3.67
296	1.00	0.1	10.86	11.66	A	A	A	1.81
298	0.96	0.06	6.43	8.65	A	A	A	1.07
301	1.0	0.1	10.86	11.66	A	A	A	1.81
303	1.18	0.25	30.82	22.02	N	A	N	5.15
309	0.9625	0.153	6.71	16.99	A	A	A	1.12
312	0.758	0.071	-15.96	11.12	A	A	A	2.67
313	0.9	0.1	-0.22	12.62	A	A	A	0.04

4 Intercomparison Parameters

Reported Results for Gross alpha Determination in Sample 1

TABLE 33. Target values

Parameter	Value
Robust Mean, Bq/kg	23.98
Robust SD, Bq/kg	7.78
N	211




FIGURE 33. Reported results and evaluation parameters for Gross alpha in Sample 1.

TABLE 34. IC Evaluation Results for in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
4	18.55	1.93	0.70	A
5	3.61	0.96	2.62	W
7	38	8	1.80	A
8	27.7	1.4	0.48	A
9	28.83	2.22	0.62	A
10	13.52	0.51	1.34	A
12	16.6	0.6	0.95	A
13	22.3	1.1	0.22	A
14	27.13	5.43	0.40	A
17	38.7	7.7	1.89	A
24	23.0	1.2	0.13	A
25	18.0639	0.2926	0.76	A
26	31.8	0.4	1.01	A
28	29.0	2.9	0.65	A
30	22.18	11.1	0.23	A
31	28.1	3.9	0.53	A
36	26	5	0.26	A
38	27.5	0.3	0.45	A
42	31.49	7.9	0.97	A
45	26.1	4.2	0.27	A

TABLE 34. IC Evaluation Results for in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
47	13.8	1.3	1.31	A
49	26.4	0.8	0.31	A
50	26.8	0.8	0.36	A
51	23.033	0.309	0.12	A
52	9.52	0.643	1.86	A
56	35.7	4.9	1.51	A
58	29.1	5.1	0.66	A
60	25.1	1.9	0.14	A
62	27.1	1.4	0.40	A
64	23.6	1.18	0.05	A
67	25	1.4	0.13	A
68	24.8	1.2	0.11	A
69	12.24	0.22	1.51	A
74	32.2	3.6	1.06	A
75	28.2	1.6	0.54	A
76	27.1	4.68	0.40	A
78	10.41	0.44	1.74	A
79	8.54	0.53	1.98	A
80	22.98	1.90	0.13	A
81	18.3	0.3	0.73	A
83	19.5	2.1	0.58	A
87	25.59	5	0.21	A
88	37.8	3.9	1.78	A
90	37.2	0.79	1.70	A
94	20.0	1.3	0.51	A
95	4.6712	0.1939	2.48	W
96	17.94	0.70	0.78	A
98	26.0	2.8	0.26	A
99	31.79	6.36	1.00	A
100	6.744	0.119	2.22	W
101	33.4	2	1.21	A
103	19.1	1.43	0.63	A
104	27.77	2.36	0.49	A
105	22.08	1.75	0.24	A
107	7.43	0.3	2.13	W
109	63.27	0.14	5.05	N
110	16.5	1.49	0.96	A
113	24.21	2.22	0.03	A
118	30	3	0.77	A
119	26.080	3.651	0.27	A
120	24.5	2.5	0.07	A
123	27.49	1.44	0.45	A
124	27.28	0.86	0.42	A
126	32.7	2.5	1.12	A
127	7.52	1.1	2.12	W
128	25.0	3.0	0.13	A
129	29	5	0.65	A
130	0.8	0.28	2.98	W
133	26.1	3.7	0.27	A
135	24.6	0.3	0.08	A
137	16.80	0.77	0.92	A
138	21.17	0.20	0.36	A
139	31.5	0.7	0.97	A
143	30.2	4.8	0.80	A

TABLE 34. IC Evaluation Results for in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
146	30.31		0.81	A
149	22.6	2	0.18	A
150	5.70	2.5	2.35	W
155	14.56	0.82	1.21	A
156	29.4	2.8	0.70	A
157	29.0	2.4	0.65	A
158	26.176	5.111	0.28	A
159	22.5	4.0	0.19	A
160	16.10	0.85	1.01	A
161	25.9	2.3	0.25	A
163	27.13	3.45	0.40	A
171	29.2	0.9	0.67	A
173	30.67	3.16	0.86	A
175	24.2	3.1	0.03	A
177	22.0	3.0	0.25	A
183	32.98	0.41	1.16	A
186	26.4	4.7	0.31	A
192	22.53	8.89	0.19	A
194	1.91	0.22	2.84	W
196	30.6	2.81	0.85	A
199	29.15	1.75	0.66	A
202	38.4	4.0	1.85	A
203	25.7	1.4	0.22	A
204	25.4	3.6	0.18	A
205	37.606	2.969	1.75	A
208	10.1666	0.00063	1.78	A
211	27.2	1.9	0.41	A
213	30	3	0.77	A
214	21.44	1.07	0.33	A
216	9.8	0.4	1.82	A
217	26.6	1.4	0.34	A
220	5.74	2.2	2.34	W
221	29.3	4.39	0.68	A
222	34.7	2.0	1.38	A
223	6.9	1.8	2.20	W
224	18.5	5	0.70	A
225	26.0	1.1	0.26	A
226	25.43	1.91	0.19	A
227	22.04	1.39	0.25	A
230	26	3.3	0.26	A
231	5.8	0.5	2.34	W
239	20.41	6.02	0.46	A
240	28.5	6.7	0.58	A
241	28.43	0.56	0.57	A
243	17.33	1.52	0.85	A
246	29.1	0.3	0.66	A
247	29.0	2.4	0.65	A
248	22.6	0.95	0.18	A
250	35.25	4.23	1.45	A
253	19.79	0.39	0.54	A
261	27.9	3.9	0.50	A
262	12.0	1.5	1.54	A
266	39.5	1.8	1.99	A
272	19.0	2.61	0.64	A

TABLE 34. IC Evaluation Results for in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
276	24.4	2.8	0.05	A
277	15.35	1.49	1.11	A
278	0.36	0.01	3.04	N
281	16.2	3.3	1.00	A
284	16.88	0.08	0.91	A
289	30.0	2.4	0.77	A
290	27.25	0.50	0.42	A
296	28.1	3.1	0.53	A
299	20.06	1.64	0.50	A
303	13.59	1.1	1.34	A
305	21.92	3.0262	0.26	A
306	12.5	0.2	1.48	A
312	29.8	1.6	0.75	A

Reported Results for Gross beta Determination in Sample 1

TABLE 35. Target values

Parameter	Value
Robust Mean, Bq/kg	124.75
Robust SD, Bq/kg	29.46
N	213




FIGURE 34. Reported results and evaluation parameters for Gross beta in Sample 1.

TABLE 36. IC Evaluation Results for in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
4	135.3	8.4	0.36	A
5	109.64	17.05	0.51	A
7	174	36	1.67	A
8	126	5	0.04	A
9	174.03	12.79	1.67	A
10	82.49	2.18	1.43	A
12	123	4	0.06	A
13	139.3	6.3	0.49	A
14	143.63	34.48	0.64	A
17	133	27	0.28	A
23	106.85	9.14	0.61	A
24	120	3	0.16	A
25	80.1893	2.0084	1.51	A
26	186.6	1.2	2.10	W
28	130.0	13.0	0.18	A
30	87.28	43.69	1.27	A
31	132	18	0.25	A
34	104.67	2.43	0.68	A
36	120	10	0.16	A
38	145.7	0.7	0.71	A
41	178.3	71.3	1.82	A

TABLE 36. IC Evaluation Results for in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
42	187.723	8.137	2.14	W
45	140	15	0.52	A
47	158	14	1.13	A
49	87.2	1.7	1.27	A
50	89.4	1.7	1.20	A
51	94.077	0.493	1.04	A
52	105	6.53	0.67	A
56	133.6	14.3	0.30	A
57	107.35	2.81	0.59	A
58	152	9	0.92	A
60	132.9	9.6	0.28	A
62	136	6	0.38	A
64	167.5	8.45	1.45	A
65	143.61	1.39	0.64	A
67	116	4	0.30	A
68	122.3	5.6	0.08	A
69	111.71	0.44	0.44	A
74	133.5	4.0	0.30	A
75	142	7	0.59	A
76	125.9	9.15	0.04	A
78	98.43	1.68	0.89	A
79	90.9	2.4	1.15	A
80	43.88	1.58	2.75	W
81	90.4	0.2	1.17	A
83	117.0	12.5	0.26	A
87	156.24	31	1.07	A
88	104	10	0.70	A
90	144.3	1.72	0.66	A
94	115.7	7.5	0.31	A
95	87.2228	1.4728	1.27	A
96	133.4	4.1	0.29	A
98	113.8	10.2	0.37	A
99	117.0	17.55	0.26	A
100	250.9	0.502	4.28	N
101	71.5	1.7	1.81	A
103	173	12.6	1.64	A
104	121.89	1.74	0.10	A
105	143.83	40.0	0.65	A
107	158.3	11.3	1.14	A
109	172.00	0.12	1.60	A
110	88.6	5.28	1.23	A
113	130.20	6.82	0.18	A
118	119	11	0.20	A
119	107.320	15.025	0.59	A
120	148	15	0.79	A
123	123.88	5.96	0.03	A
124	109.65	0.69	0.51	A
126	120	10	0.16	A
127	89.34	8.9	1.20	A
128	134.2	10.8	0.32	A
129	167	23	1.43	A
130	28.1	1.68	3.28	N
133	107.3	15.0	0.59	A
135	31.7	0.4	3.16	N

TABLE 36. IC Evaluation Results for in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
137	119.00	3.86	0.20	A
138	120.20	0.37	0.15	A
139	120	1.4	0.16	A
143	128	7.3	0.11	A
149	123.0	1.1	0.06	A
155	98.98	1.89	0.87	A
156	93.1	13.3	1.07	A
157	108.9	8.7	0.54	A
158	182.662	16.324	1.97	A
159	82.5	10.0	1.43	A
160	127.04	5.68	0.08	A
161	123.3	11	0.05	A
163	186.00	6.20	2.08	W
171	131	2	0.21	A
173	161.87	16.29	1.26	A
175	170.8	4.2	1.56	A
177	109.0	15.0	0.53	A
180	145	9	0.69	A
183	111.17	1.20	0.46	A
186	82.2	12.3	1.44	A
192	137.7	8.1	0.44	A
194	91.18	1.78	1.14	A
196	150	12	0.86	A
199	127.1	5.1	0.08	A
202	131	14	0.21	A
203	112.0	11.0	0.43	A
204	121	17	0.13	A
205	128.798	8.427	0.14	A
208	80.1639	0.0169	1.51	A
211	80	6	1.52	A
213	130	15	0.18	A
214	120.01	1.41	0.16	A
216	188	3	2.15	W
217	140	7.5	0.52	A
220	168.96	40.2	1.50	A
221	123.05	18.46	0.06	A
222	126.0	8.0	0.04	A
223	117.8	21	0.24	A
224	134.7	24	0.34	A
225	118.7	0.4	0.21	A
226	121	6.05	0.13	A
227	136.25	3.14	0.39	A
230	108	14	0.57	A
231	57.44	2.5	2.28	W
234	102.90	5.30	0.74	A
235	110.4	2.43	0.49	A
239	106.5	31.4	0.62	A
240	183	56.1	1.98	A
241	181.21	1.80	1.92	A
243	134.1	8.2	0.32	A
246	114.7	2.2	0.34	A
247	119.3	6.3	0.18	A
248	102	4.2	0.77	A
250	137.08	16.45	0.42	A

TABLE 36. IC Evaluation Results for in Sample 1, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
253	103.26	0.81	0.73	A
261	129.8	11.8	0.17	A
262	108.2	14.1	0.56	A
272	116.35	18.03	0.29	A
276	122.5	8.1	0.08	A
277	111.88	12.26	0.44	A
284	165.55	0.32	1.38	A
289	88.9	3.4	1.22	A
290	98.60	3.52	0.89	A
296	145.5	13.2	0.70	A
299	160.04	11.66	1.20	A
303	117.9	27.2	0.23	A
306	63.8	0.8	2.07	W
312	153.9	11.4	0.99	A

Reported Results for Gross alpha Determination in Sample 2

TABLE 37. Target values

Parameter	Value
Robust Mean, Bq/kg	12.72
Robust SD, Bq/kg	3.73
N	207




FIGURE 35. Reported results and evaluation parameters for Gross alpha in Sample 2.

TABLE 38. IC Evaluation Results for in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
4	14.93	0.24	0.59	A
5	5.17	0.45	2.02	W
7	18.1	4.4	1.44	A
8	14.8	0.9	0.56	A
9	15.37	1.94	0.71	A
10	6.53	0.36	1.66	A
12	9.06	0.4	0.98	A
13	12.5	0.7	0.06	A
14	16.61	3.33	1.04	A
17	14.2	2.8	0.40	A
24	14.0	0.8	0.34	A
25	11.8046	0.1914	0.25	A
26	14.2	0.3	0.40	A
28	14.0	1.4	0.34	A
30	21.88	12.63	2.46	W
31	14.3	2.0	0.42	A
36	14	3	0.34	A
38	13.8	0.2	0.29	A
42	16.959	4.255	1.14	A
45	13.6	2.1	0.24	A
47	13.0	0.89	0.08	A

TABLE 38. IC Evaluation Results for in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
49	10.7	0.5	0.54	A
50	11.2	0.5	0.41	A
52	6.04	0.387	1.79	A
56	18.4	2.7	1.52	A
58	15.1	2.6	0.64	A
60	14.2	1.1	0.40	A
62	13.8	0.8	0.29	A
64	9.38	0.46	0.90	A
67	15.6	1.0	0.77	A
68	14.5	0.9	0.48	A
69	8.20	0.16	1.21	A
74	15.5	2.2	0.75	A
75	13.3	1.1	0.16	A
76	13.7	4.07	0.26	A
78	4.78	0.44	2.13	W
79	4.07	0.37	2.32	W
80	6.87	1.23	1.57	A
81	17.6	0.3	1.31	A
83	15.3	1.5	0.69	A
87	11.86	2.36	0.23	A
88	13.8	1.4	0.29	A
90	16.5	0.55	1.01	A
94	17.9	1.2	1.39	A
95	2.2578	0.1378	2.80	W
96	13.09	0.62	0.10	A
98	15.4	1.2	0.72	A
99	12.23	2.45	0.13	A
100	3.622	0.2	2.44	W
101	18.5	3.1	1.55	A
103	14.1	1.07	0.37	A
104	14.10	1.13	0.37	A
105	11.33	1.25	0.37	A
107	11.62	0.4	0.29	A
109	30.00	0.07	4.63	N
110	6.70	0.945	1.61	A
113	12.71	1.19	0.00	A
118	9	0.9	1.00	A
119	14.050	1.967	0.36	A
120	15.9	1.6	0.85	A
123	12.46	1.32	0.07	A
124	14.24	0.45	0.41	A
126	17.0	1.4	1.15	A
127	6.75	1.0	1.60	A
128	15.5	1.9	0.75	A
129	15	3	0.61	A
130	0.3	0.17	3.33	N
133	14.1	1.9	0.37	A
134	0.45	0.068	3.29	N
135	13.3	0.2	0.16	A
137	11.50	0.54	0.33	A
138	11.87	0.13	0.23	A
143	18.1	2.89	1.44	A
146	7.29		1.46	A
149	13.2	1.2	0.13	A

TABLE 38. IC Evaluation Results for in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
150	5.9	2.45	1.83	A
155	5.69	0.36	1.88	A
156	19.9	1.9	1.92	A
157	12.2	1.1	0.14	A
158	12.365	2.424	0.10	A
159	12.2	1.9	0.14	A
160	10.99	0.57	0.46	A
161	17.25	1.6	1.21	A
163	14.43	1.75	0.46	A
171	14.2	0.5	0.40	A
173	14.92	1.58	0.59	A
175	14.0	1.1	0.34	A
177	11.0	2.0	0.46	A
183	16.79	0.29	1.09	A
186	12.1	2.1	0.17	A
192	13.0	4.7	0.08	A
194	1.33	0.07	3.05	N
196	15.4	2.24	0.72	A
199	13.26	0.92	0.14	A
202	19.2	2.0	1.74	A
203	15.0	1.4	0.61	A
204	8.42	1.18	1.15	A
205	17.187	1.396	1.20	A
208	5.1974	0.00063	2.02	W
211	13.0	0.9	0.08	A
213	15.2	1.6	0.66	A
214	14.14	0.87	0.38	A
216	7.2	0.3	1.48	A
217	14.3	0.82	0.42	A
220	5.88	2.5	1.83	A
221	15.39	2.31	0.72	A
222	16.3	1.0	0.96	A
223	15.7	4	0.80	A
224	9.8	2.3	0.78	A
225	12.4	0.5	0.09	A
227	10.46	0.97	0.61	A
230	16	2.3	0.88	A
231	2.37	0.2	2.77	W
239	16.0	4.72	0.88	A
240	17.4	4.36	1.25	A
241	13.76	0.11	0.28	A
243	8.81	0.72	1.05	A
246	10.2	0.3	0.68	A
247	14.0	1.1	0.34	A
248	18.0	0.72	1.42	A
250	14.00	1.68	0.34	A
253	9.26	0.25	0.93	A
262	7.5	0.94	1.40	A
266	19.3	0.9	1.76	A
272	15.44	1.67	0.73	A
276	12.2	1.84	0.14	A
277	9.521	0.89	0.86	A
281	8.41	1.75	1.16	A
284	9.57	0.06	0.84	A

TABLE 38. IC Evaluation Results for in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
289	14.7	1.5	0.53	A
290	3.81	0.08	2.39	W
296	15.5	1.6	0.75	A
299	11.38	0.91	0.36	A
303	10.19	1.1	0.68	A
305	12.33	3.4569	0.10	A
306	6.8	0.2	1.59	A
312	15.3	0.8	0.69	A

Reported Results for Gross beta Determination in Sample 2

TABLE 39. Target values

Parameter	Value
Robust Mean, Bq/kg	28.94
Robust SD, Bq/kg	6.35
N	209




FIGURE 36. Reported results and evaluation parameters for Gross beta in Sample 2.

TABLE 40. IC Evaluation Results for in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
4	28.52	1.77	0.07	A
5	22.58	7.3	1.00	A
7	43	9	2.21	W
8	27.9	1.4	0.16	A
9	41.83	5.04	2.03	W
10	22.44	1.34	1.02	A
12	27.1	0.9	0.29	A
13	31.3	1.4	0.37	A
14	31.75	7.62	0.44	A
17	33.2	6.6	0.67	A
23	24.62	2.11	0.68	A
24	27.6	1.1	0.21	A
25	18.9754	0.4753	1.57	A
26	53	1	3.79	N
28	28.6	2.9	0.05	A
30	9.48	5.43	3.06	N
31	32.0	4.4	0.48	A
34	25.26	0.75	0.58	A
36	30	4	0.17	A
38	29.7	0.3	0.12	A
42	43.065	1.916	2.22	W

TABLE 40. IC Evaluation Results for in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
45	28.2	5.2	0.12	A
47	35.3	3.2	1.00	A
49	21.4	0.7	1.19	A
50	22.2	0.7	1.06	A
52	38.6	2.52	1.52	A
56	28.5	3.3	0.07	A
57	24.78	0.74	0.66	A
58	39.8	3.6	1.71	A
60	30.7	2.3	0.28	A
62	30.9	1.7	0.31	A
64	31.05	1.55	0.33	A
67	26.2	1.2	0.43	A
68	28.8	1.8	0.02	A
69	24.66	0.21	0.67	A
74	30.5	1.7	0.25	A
75	31.2	1.2	0.36	A
76	29.7	4.88	0.12	A
78	22.02	0.71	1.09	A
79	21.1	0.8	1.23	A
80	20.06	0.86	1.40	A
81	23.8	0.1	0.81	A
83	30.0	2.5	0.17	A
87	37.13	7.4	1.29	A
88	23.4	2.3	0.87	A
90	33.8	1.11	0.77	A
94	30.5	2.1	0.25	A
95	13.6937	1.0917	2.40	W
96	32.7	1.9	0.59	A
98	27.1	2.3	0.29	A
99	25.54	3.83	0.54	A
100	37.31	0.8672	1.32	A
101	15.6	2.1	2.10	W
103	37.4	2.73	1.33	A
104	27.05	0.41	0.30	A
105	34.42	2.0	0.86	A
107	34.83	4.3	0.93	A
109	28.76	0.05	0.03	A
110	23.0	2.05	0.94	A
113	27.80	1.49	0.18	A
118	27	3	0.31	A
119	23.460	3.284	0.86	A
120	32.6	3.3	0.58	A
123	28.41	1.50	0.08	A
124	27.88	0.20	0.17	A
126	20.2	2.8	1.38	A
127	22.05	2.2	1.09	A
128	28.6	2.3	0.05	A
129	34	5	0.80	A
130	18.7	1.37	1.61	A
133	23.5	3.3	0.86	A
134	13.69	1.45	2.40	W
135	17.0	0.3	1.88	A
137	26.20	0.87	0.43	A
138	26.50	0.13	0.38	A

TABLE 40. IC Evaluation Results for in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
143	27.3	1.56	0.26	A
149	24.2	0.5	0.75	A
155	21.18	0.62	1.22	A
156	21.4	2.1	1.19	A
157	24.3	2.1	0.73	A
158	48.024	4.355	3.01	N
159	24.0	4.6	0.78	A
160	31.07	1.41	0.34	A
161	59.2	5.1	4.77	N
163	34.56	1.35	0.89	A
171	28.4	0.7	0.09	A
173	36.45	3.7	1.18	A
175	28.2	1.8	0.12	A
177	25.0	5.0	0.62	A
180	38.5	3.8	1.51	A
183	25.97	0.57	0.47	A
186	18.5	3.3	1.64	A
192	29.9	1.2	0.15	A
194	22.35	0.48	1.04	A
196	33.1	6.83	0.66	A
199	27.60	1.90	0.21	A
202	30.6	3.0	0.26	A
203	26.0	3.0	0.46	A
204	28.9	4.1	0.01	A
205	27.658	1.830	0.20	A
208	17.8436	0.0169	1.75	A
211	17.8	1.3	1.75	A
213	31	5	0.32	A
214	27.74	0.68	0.19	A
216	46.0	0.6	2.69	W
217	37.3	0.12	1.32	A
220	40.6	10.1	1.84	A
221	29.85	4.48	0.14	A
222	28.1	1.8	0.13	A
223	31	5.5	0.32	A
224	32.6	6.2	0.58	A
225	26.0	0.2	0.46	A
227	31.77	1.68	0.45	A
230	26	3.4	0.46	A
231	13.6	1.3	2.42	W
234	25.60	1.30	0.53	A
235	20.51	1.17	1.33	A
239	31.9	9.41	0.47	A
240	44.2	13.6	2.40	W
241	34.06	0.20	0.81	A
243	31.32	2.25	0.37	A
246	25.9	0.9	0.48	A
247	27.7	1.5	0.20	A
248	54.4	3.4	4.01	N
250	29.43	3.53	0.08	A
253	25.77	0.39	0.50	A
262	31.5	4.1	0.40	A
272	29.46	3.0	0.08	A
276	23.8	2.56	0.81	A

TABLE 40. IC Evaluation Results for in Sample 2, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
277	27.89	2.68	0.17	A
284	35.48	0.11	1.03	A
289	20.2	0.9	1.38	A
290	8.11	0.31	3.28	N
296	38.1	3.1	1.44	A
299	37.11	2.73	1.29	A
303	25.27	5.8	0.58	A
306	13.4	0.2	2.45	W
312	35.6	4.7	1.05	A

Reported Results for Gross alpha Determination in Sample 3

TABLE 41. Target values

Parameter	Value
Robust Mean, Bq/kg	8.93
Robust SD, Bq/kg	3.04
N	206




FIGURE 37. Reported results and evaluation parameters for Gross alpha in Sample 3.

TABLE 42. IC Evaluation Results for in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
4	9.42	0.33	0.16	A
5	7.64	1.8	0.42	A
7	10.5	2.8	0.52	A
8	8.89	0.53	0.01	A
9	10.80	1.67	0.62	A
10	5.11	0.32	1.26	A
12	5.71	0.3	1.06	A
13	9.2	0.6	0.09	A
14	11.94	2.39	0.99	A
17	11.5	2.3	0.85	A
24	10.7	0.9	0.58	A
25	7.9536	0.1293	0.32	A
26	10.8	0.3	0.62	A
28	12.4	1.2	1.14	A
30	12.15	7.54	1.06	A
36	9	3	0.02	A
38	9.7	0.1	0.25	A
42	11.972	3.004	1.00	A
45	9.9	1.6	0.32	A
47	7.38	0.52	0.51	A
49	9.7	0.5	0.25	A

TABLE 42. IC Evaluation Results for in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
50	10.1	0.5	0.38	A
51	10.295	0.198	0.45	A
52	3.86	0.238	1.67	A
56	13.1	2.0	1.37	A
60	9.69	0.71	0.25	A
62	9.01	0.63	0.03	A
64	7.46	0.37	0.48	A
67	11.0	0.9	0.68	A
68	10.2	0.8	0.42	A
69	5.51	0.12	1.12	A
74	10.9	1.4	0.65	A
75	9.4	0.8	0.15	A
76	11.6	3.82	0.88	A
78	3.18	0.37	1.89	A
79	3.30	0.32	1.85	A
80	5.09	1.24	1.26	A
81	9.06	0.21	0.04	A
83	6.5	0.7	0.80	A
87	8.42	1.6	0.17	A
88	11.4	1.2	0.81	A
90	11.7	0.48	0.91	A
94	7.1	0.5	0.60	A
95	2.510	0.1447	2.11	W
96	8.52	0.54	0.13	A
98	11.7	0.9	0.91	A
99	9.83	1.97	0.30	A
100	3.281	0.92	1.86	A
101	10.1	0.4	0.38	A
103	42.4	3.12	11.01	N
104	10.59	0.84	0.55	A
105	9.33	1.08	0.13	A
107	7.75	0.3	0.39	A
109	24.70	0.05	5.19	N
110	2.89	0.649	1.99	A
113	9.14	0.87	0.07	A
118	14.4	1.5	1.80	A
119	10.573	1.480	0.54	A
120	8.8	0.8	0.04	A
123	9.27	1.07	0.11	A
124	9.66	0.31	0.24	A
126	12.9	1.1	1.31	A
127	5.73	0.86	1.05	A
128	16.4	2.0	2.46	W
129	10.2	1.8	0.42	A
130	0.3	0.17	2.84	W
133	10.6	1.5	0.55	A
135	8.5	0.4	0.14	A
137	7.25	0.34	0.55	A
138	8.878	0.110	0.02	A
143	12.3	1.95	1.11	A
146	3.12		1.91	A
149	10.3	1.0	0.45	A
150	1.43	0.45	2.47	W
155	3.47	0.28	1.80	A

TABLE 42. IC Evaluation Results for in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
156	13.3	1.3	1.44	A
157	6.4	0.6	0.83	A
158	4.321	0.864	1.52	A
159	13.9	2.2	1.63	A
160	9.29	0.49	0.12	A
161	7.38	1.1	0.51	A
163	10.13	1.52	0.39	A
171	10.6	0.3	0.55	A
173	11.39	1.23	0.81	A
175	9.1	0.9	0.06	A
177	6.3	1.2	0.87	A
183	16.14	0.28	2.37	W
186	7.6	1.5	0.44	A
192	11.0	2.2	0.68	A
194	1.00	0.19	2.61	W
196	6.47	1.04	0.81	A
199	10.50	1.10	0.52	A
202	13.3	1.4	1.44	A
203	12.9	1.0	1.31	A
204	6.21	0.87	0.89	A
205	13.151	1.078	1.39	A
208	4.6890	0.00063	1.40	A
211	10.0	0.7	0.35	A
213	11.2	1.2	0.75	A
214	9.19	0.66	0.09	A
216	2.88	0.12	1.99	A
217	8.75	0.52	0.06	A
220	1.44	0.5	2.46	W
221	11.14	1.67	0.73	A
222	12.2	0.8	1.08	A
223	8.1	2	0.27	A
224	8.7	2	0.08	A
225	8.8	0.3	0.04	A
226	10.06	0.75	0.37	A
227	6.41	0.8	0.83	A
230	12	2.1	1.01	A
231	2.75	0.3	2.03	W
239	6.81	2.0	0.70	A
240	14.5	3.78	1.83	A
241	8.32	0.37	0.20	A
243	7.57	0.41	0.45	A
246	7.95	0.13	0.32	A
247	9.76	0.81	0.27	A
248	9.10	0.43	0.06	A
250	7.20	0.86	0.57	A
253	6.32	0.21	0.86	A
262	5.40	0.67	1.16	A
266	15.5	0.8	2.16	W
272	6.44	1.64	0.82	A
276	10.00	1.7	0.35	A
277	1.021	0.95	2.60	W
281	5.79	1.2	1.03	A
284	5.37	0.04	1.17	A
289	11.5	1.2	0.85	A

TABLE 42. IC Evaluation Results for in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
290	9.45	0.18	0.17	A
296	8.8	0.9	0.04	A
299	5.40	0.53	1.16	A
303	5.63	0.73	1.09	A
305	5.98	1.828	0.97	A
306	5.3	0.2	1.19	A
312	9.5	0.6	0.19	A

Reported Results for Gross beta Determination in Sample 3

TABLE 43. Target values

Parameter	Value
Robust Mean, Bq/kg	27.63
Robust SD, Bq/kg	6.78
N	210




FIGURE 38. Reported results and evaluation parameters for Gross beta in Sample 3.

TABLE 44. IC Evaluation Results for in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
4	30.40	1.88	0.41	A
5	22.21	6.6	0.80	A
7	39	9	1.68	A
8	27.0	1.1	0.09	A
9	34.57	2.49	1.02	A
10	13.57	1.16	2.07	W
12	27.8	1.0	0.03	A
13	25.6	1.2	0.30	A
14	31.73	7.62	0.60	A
17	40.4	8.1	1.88	A
23	23.72	2.03	0.58	A
24	27.2	1.0	0.06	A
25	19.0960	0.4784	1.26	A
26	47.7	1	2.96	W
28	30.0	3.0	0.35	A
30	11.08	6.20	2.44	W
31	27.2	3.7	0.06	A
34	24.05	0.72	0.53	A
36	22	3	0.83	A
38	31	0.3	0.50	A
42	49.682	2.182	3.25	N

TABLE 44. IC Evaluation Results for in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
45	33.3	6.7	0.84	A
47	32.9	2.8	0.78	A
49	20.2	0.6	1.10	A
50	19.3	0.6	1.23	A
51	18.977	0.252	1.28	A
52	17.9	1.29	1.44	A
56	27.8	3.3	0.03	A
57	22.93	0.74	0.69	A
58	32.2	2.9	0.67	A
60	28.1	2.1	0.07	A
62	26.1	1.5	0.23	A
64	29.8	1.49	0.32	A
67	24.2	1.1	0.51	A
68	27.5	1.8	0.02	A
69	30.32	0.25	0.40	A
74	25.6	1.7	0.30	A
75	30.8	1.6	0.47	A
76	30.0	4.94	0.35	A
78	20.05	0.71	1.12	A
79	20.2	0.7	1.10	A
80	12.35	0.65	2.25	W
81	15.9	0.1	1.73	A
83	38.8	3.1	1.65	A
87	27.66	5.5	0.00	A
88	22.3	2.2	0.79	A
90	29.2	1.08	0.23	A
94	22.7	1.5	0.73	A
95	18.9302	1.1583	1.28	A
96	34.8	1.8	1.06	A
98	24.4	3.1	0.48	A
99	26.62	3.99	0.15	A
100	44.85	0.87	2.54	W
101	8.3	1	2.85	W
103	34.5	2.51	1.01	A
104	26.66	0.40	0.14	A
105	35.67	2.0	1.19	A
107	36.26	3.2	1.27	A
109	35.54	0.05	1.17	A
110	82.7	4.92	8.12	N
113	27.64	1.48	0.00	A
118	26.3	2.4	0.20	A
119	26.200	3.668	0.21	A
120	31.0	3.1	0.50	A
123	29.44	1.55	0.27	A
124	24.53	0.17	0.46	A
126	26.0	3.1	0.24	A
127	19.52	1.95	1.20	A
128	31.2	1.6	0.53	A
129	33	4	0.79	A
130	21.8	1.48	0.86	A
133	26.2	3.7	0.21	A
135	10.7	0.2	2.50	W
137	25.60	0.84	0.30	A
138	26.34	0.13	0.19	A

TABLE 44. IC Evaluation Results for in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
143	26.4	1.51	0.18	A
149	25.7	0.5	0.28	A
155	19.50	0.59	1.20	A
156	18.8	1.7	1.30	A
157	25.8	2.2	0.27	A
158	29.736	2.797	0.31	A
159	16.6	4.0	1.63	A
160	28.56	0.90	0.14	A
161	15.71	2	1.76	A
163	36.48	1.18	1.31	A
171	27.8	0.7	0.03	A
173	35.75	3.63	1.20	A
175	40.6	1.7	1.91	A
177	16.0	3.0	1.72	A
180	36.8	3.7	1.35	A
183	26.18	0.56	0.21	A
186	14.9	2.5	1.88	A
192	28.6	1.6	0.14	A
194	20.08	1.18	1.11	A
196	34.3	5.79	0.98	A
199	28.90	1.65	0.19	A
202	28.7	3.0	0.16	A
203	23.0	2.0	0.68	A
204	22.4	3.2	0.77	A
205	27.121	1.795	0.08	A
208	17.4929	0.0169	1.50	A
211	16.2	1.2	1.69	A
213	29	5	0.20	A
214	27.06	0.67	0.08	A
216	44.2	0.6	2.44	W
217	31.8	2.0	0.62	A
220	23.24	8.0	0.65	A
221	25.08	3.76	0.38	A
222	28.6	1.9	0.14	A
223	23.3	4	0.64	A
224	26.8	5	0.12	A
225	26.3	0.2	0.20	A
226	27.98	1.4	0.05	A
227	29.04	1.63	0.21	A
230	29.5	4.1	0.28	A
231	13.2	1.3	2.13	W
234	21.20	2.20	0.95	A
235	22.07	1.2	0.82	A
239	20	5.9	1.13	A
240	42.3	13.0	2.16	W
241	31.66	1.29	0.59	A
243	30.73	1.71	0.46	A
246	25.7	0.9	0.28	A
247	27.0	1.5	0.09	A
248	50.5	4.2	3.37	N
250	28.99	3.48	0.20	A
253	22.60	0.37	0.74	A
262	27.2	3.54	0.06	A
272	38.75	3.32	1.64	A

TABLE 44. IC Evaluation Results for in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
276	30.20	2.92	0.38	A
277	27.142	2.83	0.07	A
284	32.65	0.11	0.74	A
289	13.9	0.6	2.03	W
290	21.22	0.78	0.95	A
296	28.7	2.2	0.16	A
299	31.72	2.34	0.60	A
303	25.84	6.0	0.26	A
306	12.4	0.2	2.25	W
312	37.4	4.2	1.44	A

Reported Results for Cm-244 Determination in Sample 3

TABLE 45. Target values

Parameter	Value
Robust Mean, Bq/kg	4.51
Robust SD, Bq/kg	0.66
N	38




FIGURE 39. Reported results and evaluation parameters for Cm-244 in Sample 3.

TABLE 46. IC Evaluation Results for in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
9	4.04	1.35	0.71	A
14	4.78	0.16	0.41	A
36	4.4	0.4	0.17	A
41	5.2	0.5	1.05	A
60	5.41	0.26	1.36	A
61	4.94	1.20	0.65	A
71	5.39	0.36	1.33	A
83	0.8	0.1	5.62	N
84	4.568	0.235	0.09	A
87	4.745	0.1753	0.36	A
94	4.0	0.2	0.77	A
96	5.04	0.43	0.80	A
119	3.860	0.695	0.98	A
133	3.9	0.7	0.92	A
137	4.327	0.326	0.28	A
140	5.19	0.64	1.03	A
146	4.23	0	0.42	A
157	3.5	0.4	1.53	A
187	3.3	0.2	1.83	A
207	5.76	0.25	1.89	A
217	4.80	0.19	0.44	A

TABLE 46. IC Evaluation Results for in Sample 3, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
226	4.88	0.6	0.56	A
247	4.02	0.22	0.74	A
253	4.22	0.18	0.44	A
272	0.8	0.13	5.62	N

Reported Results for Gross alpha Determination in Sample 5

TABLE 47. Target values

Parameter	Value
Robust Mean, $\text{counts}/\text{cm}^2/\text{second}$	0.081
Robust SD, $\text{counts}/\text{cm}^2/\text{second}$	0.052
N	182




FIGURE 40. Reported results and evaluation parameters for Gross alpha in Sample 5.

TABLE 48. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in $\text{counts}/\text{cm}^2/\text{second}$

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
4	0.323	0.070	4.65	N
6	0.000005		1.56	A
8	0.10	0.01	0.37	A
9	0.232	0.013	2.90	W
10	0.2246	0.007296	2.76	W
13	0.095	0.011	0.27	A
14	0.059	0.007	0.42	A
16	0.046	0.006	0.67	A
17	0.09	0.02	0.17	A
18	0.0361	0.0021	0.86	A
19	0.0423	0.0030	0.74	A
21	0.084	0.004	0.06	A
22	0.381	0.148	5.77	N
24	0.097	0.008	0.31	A
26	0.11	0.04	0.56	A
27	0.10	0.01	0.37	A
28	0.22	0.02	2.67	W
31	0.10	0.03	0.37	A
32	2.0	0.5	36.90	N
36	0.031	0.002	0.96	A
39	0.089	0.036	0.15	A

TABLE 48. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in *counts/cm²/second*

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
41	0.036	0.01	0.87	A
42	0.052	0.005	0.56	A
44	0.056	0.008	0.48	A
45	0.090	0.018	0.17	A
47	0.124	0.010	0.83	A
49	0.105	0.033	0.46	A
50	0.11	0.03	0.56	A
51	0.112	0.017	0.60	A
52	0.213	0.0102	2.54	W
56	0.29	0.04	4.02	N
58	0.036	0.001	0.87	A
60	0.041	0.004	0.77	A
62	0.097	0.005	0.31	A
64	0.011	0.0002	1.35	A
67	0.0654	0.0082	0.30	A
68	0.095	0.0050	0.27	A
69	0.017	0.002	1.23	A
72	0.036	0.004	0.87	A
74	0.037	0.004	0.85	A
76	0.066	0.0264	0.29	A
81	0.076	0.004	0.10	A
83	0.025	0.0004	1.08	A
86	0.065	0.00975	0.31	A
88	0.0869	0.0094	0.11	A
92	0.0853	0.0128	0.08	A
93	0.0088	0.0022	1.39	A
94	0.043	0.001	0.73	A
95	0.0424	0.0025	0.74	A
96	0.0408	0.0004	0.77	A
98	0.077	0.016	0.08	A
99	0.101	0.003	0.38	A
101	0.083	0.007	0.04	A
107	0.042	0.011	0.75	A
109	0.057	0.006	0.46	A
110	0.219	0.00759	2.65	W
114	0.0860	0.0094	0.10	A
117	0.025	0.07	1.08	A
120	0.053	0.001	0.54	A
121	0.21	0.02	2.48	W
122	0.110	0.007	0.56	A
123	0.052	0.005	0.56	A
125	0.226	0.007	2.79	W
126	0.098	0.014	0.33	A
127	0.0893	0.0037	0.16	A
128	0.015	0.003	1.27	A
130	0.0642	0.0025	0.32	A
134	0.109	0.012	0.54	A
138	0.0732		0.15	A
140	0.014	0.002	1.29	A
143	0.12	0.01	0.75	A
147	0.026	0.003	1.06	A
149	0.04	0.01	0.79	A
153	0.12	0.03	0.75	A
156	0.080	0.016	0.02	A

TABLE 48. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in *counts/cm²/second*

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
157	0.058	0.002	0.44	A
158	0.061	0.0025	0.38	A
159	0.056	0.006	0.48	A
161	0.08	0.01	0.02	A
162	0.115	0.007	0.65	A
164	0.0403	0.0026	0.78	A
171	0.103	0.006	0.42	A
172	0.041	0.005	0.77	A
173	0.054	0.017	0.52	A
183	0.26	0.01	3.44	N
186	0.07	0.01	0.21	A
191	0.105	0.011	0.46	A
192	0.095	0.010	0.27	A
195	0.16	0.032	1.52	A
198	0.0407	0.0081	0.78	A
199	0.044	0.005	0.71	A
203	0.08	0.01	0.02	A
204	0.10	0.012	0.37	A
213	0.070	0.008	0.21	A
214	0.277	0.1	3.77	N
217	0.104	0.0033	0.44	A
220	0.07	0.03	0.21	A
221	0.05	0.005	0.60	A
227	0.0833	0.0002	0.04	A
230	0.095	0.010	0.27	A
238	0.0371	0.0017	0.84	A
239	0.05	0.02	0.60	A
240	0.112	0.0125	0.60	A
247	0.076	0.007	0.10	A
248	0.11	0.007	0.56	A
250	14.82	2.15	283.44	N
260	0.0307	0.0012	0.97	A
261	0.10	0.01	0.37	A
262	0.0745	0.0089	0.13	A
264	0.12	0.03	0.75	A
272	0.025	0.0002	1.08	A
276	0.04	0.004	0.79	A
277	0.015	0.002	1.27	A
280	0.1025	0.0087	0.41	A
284	0.282	0.065	3.87	N
286	0.0346	0.004	0.89	A
289	0.012	0.001	1.33	A
295	0.68	0.20	11.52	N
296	0.481	0.037	7.69	N
303	0.569	0.026	9.38	N
305	0.05920	0.01	0.42	A
306	0.0407	0.0003	0.78	A
308	0.091	0.028	0.19	A
310	0.10	0.03	0.37	A
312	0.100	0.021	0.37	A

Reported Results for Gross beta Determination in Sample 5

TABLE 49. Target values

Parameter	Value
Robust Mean, $\text{counts}/\text{cm}^2/\text{second}$	0.158
Robust SD, $\text{counts}/\text{cm}^2/\text{second}$	0.084
N	189




FIGURE 41. Reported results and evaluation parameters for Gross beta in Sample 5.

TABLE 50. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in $\text{counts}/\text{cm}^2/\text{second}$

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
4	0.249	0.059	1.08	A
6	0.0101		1.76	A
8	0.17	0.02	0.14	A
9	0.365	0.054	2.46	W
10	0.3425	0.01788	2.20	W
13	0.212	0.030	0.64	A
14	0.123	0.010	0.42	A
16	0.07	0.025	1.05	A
17	0.14	0.03	0.21	A
18	0.1199	0.0061	0.45	A
19	0.0866	0.0060	0.85	A
21	0.149	0.021	0.11	A
22	0.301	0.026	1.70	A
24	0.26	0.03	1.21	A
25	52	2.6	617.17	N
26	0.11	0.01	0.57	A
27	0.19	0.03	0.38	A
28	0.38	0.04	2.64	W
32	10	0.5	117.17	N
36	0.057	0.005	1.20	A
39	0.11	0.04	0.57	A

TABLE 50. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in *counts/cm²/second*

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
40	0.194	0.020	0.43	A
41	0.097	0.03	0.73	A
42	0.198	0.025	0.48	A
43	0.187	0.022	0.35	A
44	0.141	0.008	0.20	A
45	0.26	0.04	1.21	A
47	0.163	0.014	0.06	A
49	0.17	0.03	0.14	A
50	0.17	0.03	0.14	A
51	0.183	0.021	0.30	A
52	0.237	0.017	0.94	A
56	0.32	0.04	1.93	A
58	0.090	0.013	0.81	A
59	0.19	0.02	0.38	A
60	0.115	0.016	0.51	A
62	0.235	0.012	0.92	A
64	0.066	0.001	1.10	A
67	0.174	0.035	0.19	A
68	0.23	0.014	0.86	A
69	0.05	0.01	1.29	A
72	0.34	0.034	2.17	W
74	0.06	0.007	1.17	A
76	0.061	0.0244	1.15	A
81	0.18	0.01	0.26	A
83	0.025	0.0004	1.58	A
86	0.116	0.0174	0.50	A
88	0.142	0.015	0.19	A
92	0.0959	0.0144	0.74	A
93	0.0822	0.0084	0.90	A
94	0.114	0.017	0.52	A
95	0.1406	0.0086	0.21	A
96	0.107	0.001	0.61	A
98	0.168	0.026	0.12	A
99	0.130	0.006	0.33	A
101	0.049	0.001	1.30	A
107	0.15	0.068	0.10	A
109	0.084	0.008	0.88	A
110	0.232	0.0123	0.88	A
114	0.1411	0.0116	0.20	A
117	0.124	0.04	0.40	A
120	0.114	0.002	0.52	A
121	0.24	0.03	0.98	A
122	0.167	0.015	0.11	A
123	0.190	0.042	0.38	A
125	0.299	0.003	1.68	A
126	0.141	0.017	0.20	A
127	0.1535	0.0056	0.05	A
128	0.063	0.008	1.13	A
130	0.1478	0.0038	0.12	A
134	0.145	0.016	0.15	A
138	0.1484		0.11	A
140	0.13	0.01	0.33	A
143	0.65	0.10	5.86	N
147	0.042	0.004	1.38	A

TABLE 50. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in *counts/cm²/second*

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
149	0.12	0.2	0.45	A
153	0.21	0.04	0.62	A
156	0.14	0.028	0.21	A
157	0.120	0.002	0.45	A
158	0.097	0.0039	0.73	A
159	0.166	0.017	0.10	A
160	0.25	0.02	1.10	A
161	0.118	0.01	0.48	A
162	0.184	0.008	0.31	A
164	0.0634	0.0036	1.13	A
171	0.060	0.003	1.17	A
172	0.1	0.005	0.69	A
173	0.030	0.013	1.52	A
183	0.14	0.01	0.21	A
186	0.16	0.02	0.02	A
191	0.179	0.018	0.25	A
192	0.197	0.02	0.46	A
195	0.11	0.031	0.57	A
198	0.1347	0.0269	0.28	A
199	0.107	0.016	0.61	A
203	0.10	0.01	0.69	A
204	0.21	0.025	0.62	A
213	0.217	0.018	0.70	A
214	0.362	0.1	2.43	W
217	0.156	0.017	0.02	A
220	0.1	0.05	0.69	A
221	0.167	0.017	0.11	A
227	0.1253	0.0002	0.39	A
230	0.22	0.02	0.74	A
235	0.10		0.69	A
238	0.1421	0.0037	0.19	A
239	0.06	0.02	1.17	A
240	0.156	0.0175	0.02	A
247	0.172	0.010	0.17	A
248	0.27	0.02	1.33	A
250	33.54	4.86	397.40	N
260	0.1611	0.0026	0.04	A
262	0.083	0.010	0.89	A
264	0.22	0.05	0.74	A
272	0.024	0.0002	1.60	A
276	0.12	0.02	0.45	A
277	0.062	0.007	1.14	A
280	0.1828	0.0125	0.30	A
284	0.284	0.042	1.50	A
286	0.0672	0.007	1.08	A
289	0.105	0.004	0.63	A
295	0.412	0.120	3.02	N
296	0.041	0.004	1.39	A
303	0.3246	0.0091	1.98	A
305	0.02023	0.01	1.64	A
306	0.077	0.001	0.96	A
308	0.20	0.051	0.50	A
310	0.21	0.06	0.62	A

TABLE 50. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in *counts/cm²/second*

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
312	0.211	0.015	0.63	A

Reported Results for Cs-137 Determination in Sample 5

TABLE 51. Target values

Parameter	Value
Robust Mean, Bq/cm^2	0.332
Robust SD, Bq/cm^2	0.048
N	196




FIGURE 42. Reported results and evaluation parameters for Cs-137 in Sample 5.

TABLE 52. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in Bq/cm^2

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
4	0.329	0.006	0.06	A
6	0.0848	0.0057	5.15	N
7	0.324	0.016	0.17	A
9	0.344	0.058	0.25	A
10	0.3683	0.0399	0.76	A
12	0.34	0.04	0.17	A
13	0.294	0.047	0.79	A
14	0.3872	0.016	1.15	A
16	0.28	0.13	1.08	A
17	0.38	0.06	1.00	A
18	0.3278	0.0195	0.09	A
20	0.301	0.015	0.65	A
21	0.316	0.019	0.33	A
22	0.302	0.018	0.63	A
24	0.335	0.019	0.06	A
25	0.22	0.01	2.33	W
28	0.34	0.02	0.17	A
33	0.31	0.02	0.46	A
34	0.36	0.01	0.58	A
37	0.31	0.02	0.46	A
39	0.36	0.05	0.58	A

TABLE 52. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in Bq/cm^2

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
40	0.315	0.020	0.35	A
42	0.299	0.018	0.69	A
43	0.331	0.019	0.02	A
44	0.359	0.02	0.56	A
45	0.248	0.013	1.75	A
47	0.493	0.041	3.35	N
52	0.31	0.02	0.46	A
56	0.31	0.02	0.46	A
57	0.339	0.02	0.15	A
58	0.38	0.05	1.00	A
59	0.345	0.025	0.27	A
60	0.331	0.020	0.02	A
62	0.345	0.014	0.27	A
64	0.36	0.02	0.58	A
68	0.336	0.010	0.08	A
69	0.32	0.02	0.25	A
70	0.44	0.02	2.25	W
72	0.35	0.03	0.37	A
74	0.32	0.03	0.25	A
76	0.32	0.04	0.25	A
83	0.325	0.0137	0.15	A
84	0.3292	0.01443	0.06	A
86	0.433	0.0247	2.10	W
87	0.2293	0.009	2.14	W
88	0.254	0.027	1.63	A
94	0.34	0.01	0.17	A
95	0.3668	0.0092	0.72	A
96	0.336	0.016	0.08	A
110	0.304	0.0175	0.58	A
114	0.35	0.01	0.37	A
117	0.21	0.06	2.54	W
118	0.34	0.03	0.17	A
119	0.327	0.036	0.10	A
120	0.4	0.1	1.42	A
121	0.219	0.004	2.35	W
123	0.34	0.02	0.17	A
125	197.1	9.86	4099.33	N
128	0.275	0.033	1.19	A
130	0.48	0.009	3.08	N
133	0.327	0.036	0.10	A
134	0.38	0.024	1.00	A
136	0.33	0.02	0.04	A
138	0.2498		1.71	A
140	0.34	0.03	0.17	A
143	0.307	0.021	0.52	A
147	0.03	0.001	6.29	N
149	0.4	0.1	1.42	A
150	0.320	0.022	0.25	A
152	0.332	0.017	0.00	A
153	0.34	0.08	0.17	A
154	0.339	0.015	0.15	A
157	0.26	0.05	1.50	A
158	0.286	0.011	0.96	A
159	0.341	0.010	0.19	A

TABLE 52. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in Bq/cm^2

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
160	0.35	0.02	0.37	A
161	0.32	0.02	0.25	A
162	0.30	0.02	0.67	A
164	0.562	0.039	4.79	N
166	0.35	0.02	0.37	A
167	0.35	0.02	0.37	A
169	0.36	0.02	0.58	A
172	0.345	0.01	0.27	A
173	0.51	0.01	3.71	N
178	0.32	0.02	0.25	A
179	0.13161	0.00061095	4.17	N
180	0.354	0.014	0.46	A
183	0.14	0.01	4.00	N
191	0.356	0.015	0.50	A
192	0.58	0.02	5.17	N
194	0.202	0.04	2.71	W
196	0.385	0.0251	1.10	A
198	0.39	0.08	1.21	A
201	0.23	0.01	2.12	W
203	0.35	0.01	0.37	A
204	0.363	0.025	0.65	A
207	0.33	0.03	0.04	A
210	0.31	0.02	0.46	A
214	0.32	0.006	0.25	A
217	0.337	0.017	0.10	A
218	0.34	0.03	0.17	A
220	0.35	0.02	0.37	A
221	0.335	0.033	0.06	A
227	0.33	0.01	0.04	A
233	0.7	0.023	7.67	N
234	0.35	0.10	0.37	A
235	0.2	0.05	2.75	W
239	0.31	0.1	0.46	A
240	0.332	0.0212	0.00	A
247	0.345	0.023	0.27	A
248	0.324	0.008	0.17	A
251	0.36	0.026	0.58	A
253	0.22	0.02	2.33	W
258	0.346	0.000438	0.29	A
259	0.34	0.02	0.17	A
262	0.34	0.04	0.17	A
264	0.36	0.09	0.58	A
265	0.21	0.004	2.54	W
271	0.340	0.016	0.17	A
272	0.29	0.02	0.88	A
274	0.32	0.057	0.25	A
276	0.27	0.01	1.29	A
284	0.38	0.019	1.00	A
285	0.38	0.073	1.00	A
295	0.76	0.46	8.92	N
303	0.5684	0.0064	4.92	N
305	0.3544	0.0512	0.47	A
306	0.0021	0.0002	6.87	N
309	0.3181	0.03718	0.29	A

TABLE 52. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in Bq/cm^2

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
310	0.29	0.02	0.88	A
312	0.338	0.025	0.13	A

Reported Results for Pu-239 Determination in Sample 5

TABLE 53. Target values

Parameter	Value
Robust Mean, Bq/cm^2	0.236
Robust SD, Bq/cm^2	0.073
N	79




FIGURE 43. Reported results and evaluation parameters for Pu-239 in Sample 5.

TABLE 54. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in Bq/cm^2

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
7	0.367	0.021	1.79	A
9	0.212	0.078	0.33	A
14	0.250	0.008	0.19	A
16	0.26	0.1	0.33	A
24	0.268	0.020	0.44	A
42	0.259	0.0125	0.32	A
44	0.221	0.02	0.21	A
45	0.075	0.007	2.21	W
47	0.230	0.065	0.08	A
60	0.440	0.042	2.79	W
68	0.275	0.019	0.53	A
76	0.31	0.007	1.01	A
83	0.12	0.002	1.59	A
86	0.2705	0.02813	0.47	A
88	0.280	0.030	0.60	A
94	0.24	0.01	0.05	A
95	0.2414	0.0014	0.07	A
96	0.279	0.012	0.59	A
97	0.264	0.002	0.38	A
110	0.261	0.0271	0.34	A
117	0.23	0.07	0.08	A

TABLE 54. IC Evaluation Results for in Sample 5, Rep. Value and Rep. Unc. in Bq/cm^2

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
118	0.23	0.02	0.08	A
119	0.125	0.045	1.52	A
128	0.198	0.030	0.52	A
133	0.125	0.045	1.52	A
141	0.24	0.02	0.05	A
143	0.266	0.011	0.41	A
149	1.0	0.2	10.47	N
153	0.23	0.05	0.08	A
157	0.25	0.03	0.19	A
160	0.84	0.06	8.27	N
168	0.237	0.015	0.01	A
187	0.27	0.03	0.47	A
204	0.255	0.009	0.26	A
207	0.22	0.02	0.22	A
214	0.266	0.09	0.41	A
217	0.248	0.0094	0.16	A
221	0.365	0.037	1.77	A
227	0.26	0.01	0.33	A
239	0.23	0.7	0.08	A
240	0.230	0.0167	0.08	A
247	0.153	0.021	1.14	A
253	0.16	0.02	1.04	A
264	0.23	0.05	0.08	A
272	0.12	0.001	1.59	A
295	2.70	1.96	33.75	N
306	0.0027	0.0001	3.20	N
310	0.20	0.06	0.49	A
312	0.2175	0.046	0.25	A

5 Gamma spectrum analysis exercise results (Sample 7)

For Sample 07 (gamma-ray spectrum analysis exercise), the Z-Score evaluation is only calculated for those analytes which were reported by > 30 participants.

Seven analytes have an overall single reporting occurrence; for the WorldWide network, Y-88 was reported as present in the spectrum by one laboratory.

Reported Results for K-40 Determination in Sample 7

TABLE 55. Target values

Parameter	Value
Robust Mean, Bq/kg	449.74
Robust SD, Bq/kg	31.25
N	193




FIGURE 44. Reported results and evaluation parameters for K-40 in Sample 7.

TABLE 56. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	454.18	17.51	0.14	A
7	400	20	1.59	A
8	443	32	0.22	A
13	4.5	0.3	14.25	N
14	446	66	0.12	A
16	436	21	0.44	A
18	443.0	26.6	0.22	A
20	518	27	2.18	W
21	411	17.1	1.24	A
22	407.582	17.478	1.35	A
24	435	26	0.47	A
29	433.0	38.0	0.54	A
31	437	38	0.41	A
32	474	9	0.78	A

TABLE 56. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
33	431.32	17.37	0.59	A
38	559	41	3.50	N
39	449	16	0.02	A
40	468	20	0.58	A
47	412	21	1.21	A
49	490	20	1.29	A
50	485	15	1.13	A
51	422	23	0.89	A
52	384.8	13.46	2.08	W
57	439.58	19.87	0.33	A
58	459	7	0.30	A
59	428	25	0.70	A
60	443	27	0.22	A
65	439.7	18.8	0.32	A
66	434	26	0.50	A
68	448.3	24.2	0.05	A
69	435.96	18.65	0.44	A
70	437.2	16.0	0.40	A
71	517.4	13.66	2.17	W
72	405	22	1.43	A
74	438	19	0.38	A
76	452	44.0	0.07	A
83	406.8	13.05	1.37	A
84	452.132	19.935	0.08	A
88	445	19	0.15	A
92	441	19	0.28	A
93	437	19	0.41	A
94	469.0	25.1	0.62	A
95	432.63	2.00	0.55	A
96	465	7.1	0.49	A
101	36	1.3	13.24	N
106	453.3	7.1	0.11	A
107	443	18.3	0.22	A
110	400	35.2	1.59	A
113	441.3	18.9	0.27	A
115	455	23	0.17	A
117	439	19	0.34	A
118	500	20	1.61	A
120	443	14	0.22	A
121	442.379	8.374	0.24	A
122	382	24	2.17	W
123	392.3	10.0	1.84	A
126	444	20	0.18	A
128	333	40	3.74	N
130	360	15.3	2.87	W
132	470	20.0	0.65	A
136	436	36	0.44	A
140	440	23	0.31	A
143	469.00	7.35	0.62	A
147	455.87	11.92	0.20	A
148	470.1	1.7	0.65	A
153	405	11	1.43	A
155	463.36	7.15	0.44	A
156	438	23	0.38	A

TABLE 56. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
157	405.6	21.9	1.41	A
158	437	18.700	0.41	A
160	590.1	25.2	4.49	N
162	438.7	18.8	0.35	A
163	362.61	18.24	2.79	W
168	465.28	7.29	0.50	A
169	46488.00	721.00	1473.22	N
172	447.73	10.14	0.06	A
175	391.6	19.7	1.86	A
178	444	19	0.18	A
183	565.10	10.31	3.69	N
187	436	28	0.44	A
191	430	26	0.63	A
196	469.43	13.31	0.63	A
198	483.94	9.97	1.09	A
201	461.8	19.8	0.39	A
207	453.10	26.52	0.11	A
210	439.4	22.9	0.33	A
213	440	20	0.31	A
214	555.2	17.2	3.37	N
216	512	36	1.99	A
217	446	13	0.12	A
219	736.1	23.5	9.16	N
220	468.49	23.4	0.60	A
221	428.19	42.87	0.69	A
228	436	18.7	0.44	A
230	439	25	0.34	A
231	472	15	0.71	A
233	2766	69	74.12	N
235	448.5	10.45	0.04	A
238	438	19	0.38	A
239	516	74.8	2.12	W
242	445	32	0.15	A
247	446	14	0.12	A
250	2444.9	35.85	63.85	N
251	462.01	12.31	0.39	A
253	443.9	15.2	0.19	A
255	446	99	0.12	A
259	521.79	23.05	2.31	W
261	433	26	0.54	A
264	438	44	0.38	A
265	459.56	6.43	0.31	A
266	438.60	19.20	0.36	A
269	494.27	7.44	1.42	A
271	450.6	20.6	0.03	A
276	443	23.6	0.22	A
284	734.12	10.59	9.10	N
290	491.2	33.0	1.33	A
295	13.53	0.70	13.96	N
297	507.97	45.76	1.86	A
298	456	6	0.20	A
299	686	23.7	7.56	N
303	452	13	0.07	A
305	527.7234	7.2868	2.50	W

TABLE 56. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
306	530	16	2.57	W
309	436.2	28.85	0.43	A
310	447.1	13.9	0.08	A
312	450	23	0.01	A
500	1	2	14.36	N

Reported Results for Mn-54 Determination in Sample 7

TABLE 57. Target values

Parameter	Value
Robust Mean, Bq/kg	0.59
Robust SD, Bq/kg	0.08
N	17




FIGURE 45. Reported results and evaluation parameters for Mn-54 in Sample 7.

TABLE 58. IC Evaluation Results for Mn-54 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
29	0.56	0.1	0.37
33	0.58	0.11	0.13
69	0.56	0.11	0.37
83	0.76	0.05	2.13
143	0.52	0.10	0.87
148	0.2688	0.0042	4.02
158	0.561	0.110	0.36
169	491.90	5.15	6141.38
214	0.63	0.12	0.50
233	5.05	0.12	55.75
299	0.539	0.113	0.64

Reported Results for Co-60 Determination in Sample 7

TABLE 59. Target values

Parameter	Value
Robust Mean, Bq/kg	123.96
Robust SD, Bq/kg	8.12
N	175




FIGURE 46. Reported results and evaluation parameters for Co-60 in Sample 7.

TABLE 60. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	127.73	4.72	0.46	A
7	123.5	3.7	0.06	A
8	127	11	0.37	A
13	1.26	0.07	15.11	N
14	125	16	0.13	A
18	125.9	10.1	0.24	A
20	129.1	6.7	0.63	A
21	120	3.6	0.49	A
22	109.623	3.296	1.77	A
24	126	7.6	0.25	A
29	114.56	4.5	1.16	A
32	113	2	1.35	A
33	110.80	3.16	1.62	A
37	121	1	0.36	A
38	129	10	0.62	A
39	127	4	0.37	A
40	117	2	0.86	A
47	126	6.3	0.25	A
49	135	6	1.36	A
50	131	5	0.87	A
51	117	5	0.86	A

TABLE 60. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
52	96.10	3.08	3.43	N
57	126.66	5.24	0.33	A
58	114.44	0.86	1.17	A
59	129	7	0.62	A
60	127	11	0.37	A
65	126.1	5.4	0.26	A
66	126	7.6	0.25	A
68	131.8	4.9	0.97	A
69	123.37	3.71	0.07	A
70	123.6	3.0	0.04	A
71	129.1	2.2	0.63	A
72	127	7	0.37	A
74	126.6	3.8	0.33	A
76	113	1.60	1.35	A
83	113.3	2.1	1.31	A
84	127.316	5.388	0.41	A
88	123	4	0.12	A
92	128	5	0.50	A
93	123	4	0.12	A
94	135.3	7.3	1.40	A
95	112.75	0.28	1.38	A
96	112	1.2	1.47	A
107	126	3.79	0.25	A
110	113	9.94	1.35	A
113	126.5	3.8	0.31	A
115	134.3	6.7	1.27	A
117	110	4	1.72	A
118	125	5	0.13	A
120	129	4	0.62	A
121	110.440	2.045	1.67	A
122	115.1	7.4	1.09	A
123	113.1	1.7	1.34	A
126	126.6	4.1	0.33	A
128	114.5	3.5	1.17	A
132	134	4	1.24	A
136	121.8	10.1	0.27	A
140	126	7	0.25	A
143	112.00	1.17	1.47	A
147	116.61	3.01	0.91	A
148	116.3	1.0	0.94	A
153	107	2	2.09	W
155	110.59	1.15	1.65	A
156	124	5.6	0.00	A
157	118.5	6.4	0.67	A
158	123.68	3.720	0.03	A
162	125.5	5.2	0.19	A
168	111.7	1.173	1.51	A
169	11168.00	116.60	1360.10	N
172	124.79	3.29	0.10	A
175	112.3	5.6	1.44	A
178	126	4	0.25	A
187	126	7	0.25	A
191	127	6	0.37	A
196	112.88	3.14	1.36	A

TABLE 60. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
198	125.74	1.26	0.22	A
201	105.6	3.4	2.26	W
207	128.25	7.48	0.53	A
210	126.3	4.6	0.29	A
213	127	7	0.37	A
214	134.7	2.2	1.32	A
216	147	10	2.84	W
217	127	3.6	0.37	A
220	126.77	4.4	0.35	A
221	118.79	11.88	0.64	A
230	127	5.9	0.37	A
231	116	5	0.98	A
233	740	17.7	75.87	N
235	125.8	2.34	0.23	A
238	126	4	0.25	A
239	124.5	18	0.07	A
242	131	8.5	0.87	A
247	129.9	2.8	0.73	A
250	633.4	8.63	62.74	N
251	120.88	2.37	0.38	A
253	123.4	3.0	0.07	A
255	126	45	0.25	A
259	143.63	4.40	2.42	W
261	125.7	6.0	0.21	A
264	126	11	0.25	A
265	113.44	1.47	1.30	A
266	126.60	3.83	0.33	A
269	122.6	3.5	0.17	A
271	128.3	6.1	0.53	A
290	181.15	7.7	7.04	N
295	3.36	0.17	14.85	N
297	18.63	1.2	12.97	N
298	128.2	1.1	0.52	A
299	173	3.53	6.04	N
303	107.1	3.0	2.08	W
306	124	4	0.00	A
309	109.8	7.487	1.74	A
310	127.2	2.9	0.40	A
312	125.7	9.0	0.21	A

Reported Results for Cd-109 Determination in Sample 7

TABLE 61. Target values

Parameter	Value
Robust Mean, Bq/kg	55.5
Robust SD, Bq/kg	39.4
N	8




FIGURE 47. Reported results and evaluation parameters for Cd-109 in Sample 7.

TABLE 62. IC Evaluation Results for Cd-109 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
29	4.7	0.89	1.29
32	63.8	1.4	0.21
76	57.2	2.50	0.04
83	62.1	1.8	0.17
175	6.9	1.0	1.23
299	74.5	11.3	0.48

Reported Results for Sn-113 Determination in Sample 7

TABLE 63. Target values

Parameter	Value
Robust Mean, Bq/kg	1.84
Robust SD, Bq/kg	0.26
N	27




FIGURE 48. Reported results and evaluation parameters for Sn-113 in Sample 7.

TABLE 64. IC Evaluation Results for Sn-113 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
29	0.87	0.15	3.73
33	1.90	0.18	0.23
40	1.7	0.2	0.54
51	1.6	0.2	0.92
59	1.7	0.3	0.54
69	1.99	0.17	0.58
70	2.0	0.1	0.62
76	1.46	0.18	1.46
83	1.88	0.13	0.15
96	1.65	0.08	0.73
148	1.492	0.017	1.34
158	2.009	0.169	0.65
169	166.79	8.10	634.42
187	2.0	0.2	0.62
191	1.8	0.3	0.15
214	2.1	0.2	1.00
259	1.70	0.17	0.54
271	2.0	0.1	0.62
299	1.38	0.112	1.77
309	1.902	0.2226	0.24

Reported Results for Ba-133 Determination in Sample 7

TABLE 65. Target values

Parameter	Value
Robust Mean, Bq/kg	62.13
Robust SD, Bq/kg	8.76
N	165




FIGURE 49. Reported results and evaluation parameters for Ba-133 in Sample 7.

TABLE 66. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	68.03	2.47	0.67	A
7	58.1	2.3	0.46	A
8	66.7	6.4	0.52	A
13	0.66	0.05	7.02	N
14	73.3	8.6	1.28	A
18	68.65	5.37	0.74	A
20	54	2.9	0.93	A
21	64.5	3.14	0.27	A
22	54.036	4.258	0.92	A
24	66.2	4.0	0.46	A
29	43.51	5.6	2.13	W
32	39.3	0.6	2.61	W
33	48.87	1.28	1.51	A
37	59.2	0.4	0.33	A
39	67.6	3.2	0.62	A
40	63	4	0.10	A
47	68.2	3.6	0.69	A
49	63	3	0.10	A
50	67	3	0.56	A
51	65	7	0.33	A
52	46.37	3.09	1.80	A

TABLE 66. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
57	65.54	5.30	0.39	A
58	42.78	0.35	2.21	W
60	65.6	5.4	0.40	A
65	73.2	8.1	1.26	A
66	65.4	3.9	0.37	A
68	67.9	3.4	0.66	A
69	61.81	2.19	0.04	A
70	59.5	2.3	0.30	A
72	66	4	0.44	A
74	65.2	5.3	0.35	A
76	44.6	2.50	2.00	A
83	61.8	0.7	0.04	A
84	67.802	2.961	0.65	A
88	60.3	2.1	0.21	A
93	58.1	1.7	0.46	A
94	55.0	3.4	0.81	A
95	42.75	0.14	2.21	W
96	41.0	0.4	2.41	W
107	68.1	2.88	0.68	A
110	48	4.22	1.61	A
113	64.9	2.7	0.32	A
115	62.4	3.1	0.03	A
117	53.7	2.2	0.96	A
118	63	3	0.10	A
120	67.8	2.1	0.65	A
121	47.656	0.896	1.65	A
122	57.8	4.4	0.49	A
123	56.1	0.85	0.69	A
126	66.6	2.6	0.51	A
128	56.4	7.9	0.65	A
132	69.9	4	0.89	A
136	58.9	5.8	0.37	A
140	67	4	0.56	A
143	41.40	0.47	2.37	W
147	46.77	1.04	1.75	A
148	44.4	2.7	2.02	W
153	55.4	1.06	0.77	A
155	37.97	0.41	2.76	W
156	61.6	4.7	0.06	A
157	58.7	3.7	0.39	A
158	60.822	2.602	0.15	A
168	40.95	0.43	2.42	W
169	4407.57	49.59	496.05	N
172	56.79	36.16	0.61	A
175	59.8	3.0	0.27	A
178	64	5	0.21	A
187	65.0	3.9	0.33	A
191	71	8	1.01	A
198	56.72	0.82	0.62	A
201	52.0	4.2	1.16	A
207	66.76	3.90	0.53	A
210	64.5	3.3	0.27	A
213	65	4	0.33	A
214	49.5	1.2	1.44	A

TABLE 66. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
216	68.0	1.5	0.67	A
217	67.1	1.9	0.57	A
220	60.88	4.5	0.14	A
221	60.57	6.06	0.18	A
235	61.42	2.38	0.08	A
238	65.6	1.2	0.40	A
239	52.32	7.6	1.12	A
242	67.3	4.5	0.59	A
247	68.1	2.1	0.68	A
250	263.17	3.91	22.95	N
251	67.52	1.12	0.62	A
253	53.8	1.3	0.95	A
255	67.0	18.6	0.56	A
259	59.93	3.33	0.25	A
261	67.6	5.0	0.62	A
264	68.7	6.9	0.75	A
265	41.57	0.58	2.35	W
266	64.60	1.86	0.28	A
269	68.0	1.96	0.67	A
271	70.6	3.9	0.97	A
290	62.25	3.84	0.01	A
295	1.24	0.07	6.95	N
297	130.48	7.46	7.80	N
298	64.1	1.2	0.22	A
299	54.5	3.40	0.87	A
303	45.9	1.3	1.85	A
309	47.44	4.603	1.68	A
310	69.1	2.7	0.80	A
312	62.9	2.7	0.09	A

Reported Results for Cs-134 Determination in Sample 7

TABLE 67. Target values

Parameter	Value
Robust Mean, Bq/kg	39.8
Robust SD, Bq/kg	3.85
N	175




FIGURE 50. Reported results and evaluation parameters for Cs-134 in Sample 7.

TABLE 68. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	40.54	1.50	0.19	A
7	38.9	0.8	0.23	A
8	42.8	3.6	0.78	A
13	0.42	0.03	10.23	N
14	42.3	5.0	0.65	A
18	41.27	3.25	0.38	A
20	35.6	1.9	1.09	A
21	40.6	1.79	0.21	A
22	33.111	0.692	1.74	A
24	42.6	2.5	0.73	A
29	31.82	2.83	2.07	W
32	29.1	0.5	2.78	W
33	32.85	2.49	1.81	A
37	40.6	0.2	0.21	A
38	35	3	1.25	A
39	41.2	1.8	0.36	A
40	44	3	1.09	A
47	40.5	2.4	0.18	A
49	44	2	1.09	A
50	41	2	0.31	A
51	43	3	0.83	A

TABLE 68. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
52	29.6	1.19	2.65	W
57	40.72	1.83	0.24	A
58	30.66	0.16	2.37	W
59	43	3	0.83	A
60	41.5	3.4	0.44	A
65	41.8	2.1	0.52	A
66	42.4	2.6	0.68	A
68	42.7	1.4	0.75	A
69	39.17	1.00	0.16	A
70	39.5	0.8	0.08	A
71	40.9	0.78	0.29	A
72	50	3	2.65	W
74	41.4	2.5	0.42	A
76	30.0	0.60	2.55	W
83	35.4	0.3	1.14	A
84	43.204	1.886	0.88	A
88	38.4	1.1	0.36	A
92	42.8	1.6	0.78	A
93	39.6	1.2	0.05	A
94	40.9	2.4	0.29	A
95	29.12	0.07	2.77	W
96	29.8	0.3	2.60	W
107	42.3	1.48	0.65	A
110	34.6	3.04	1.35	A
113	39.3	1.7	0.13	A
115	38.4	2.0	0.36	A
117	32.2	1.2	1.97	A
118	38	2	0.47	A
120	42.7	1.3	0.75	A
121	32.028	0.622	2.02	W
122	39.4	2.7	0.10	A
123	37.2	0.5	0.68	A
126	41.5	1.6	0.44	A
128	35.8	4.7	1.04	A
132	43.4	1.3	0.94	A
136	39.3	3.3	0.13	A
140	41	3	0.31	A
143	30.00	0.33	2.55	W
147	28.65	0.75	2.90	W
148	30.54	0.48	2.41	W
153	32.05	0.5	2.01	W
155	29.79	0.34	2.60	W
156	40.1	2.3	0.08	A
157	40.3	2.2	0.13	A
158	39.167	1.030	0.16	A
162	41.3	2.5	0.39	A
169	2982.79	32.69	764.41	N
172	42.37	2.76	0.67	A
175	37.1	1.9	0.70	A
178	41	2	0.31	A
187	42.5	2.4	0.70	A
191	44	4	1.09	A
198	30.16	0.36	2.50	W
201	33.2	1.5	1.71	A

TABLE 68. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
207	41.42	2.42	0.42	A
210	41.4	1.1	0.42	A
213	41	2	0.31	A
214	36.4	0.8	0.88	A
216	43.7	2.2	1.01	A
217	41.3	1.2	0.39	A
220	39.42	3.5	0.10	A
221	38.81	3.88	0.26	A
230	40.5	2.2	0.18	A
231	28.5	7	2.94	W
233	282	7.05	62.91	N
235	42.1	0.78	0.60	A
238	40.5	0.8	0.18	A
239	39	5.6	0.21	A
242	41.7	3.9	0.49	A
247	42.8	0.9	0.78	A
250	145.81	2.08	27.54	N
251	40.21	0.85	0.11	A
253	37.6	1.0	0.57	A
255	41.6	9.5	0.47	A
259	41.48	1.12	0.44	A
261	42.0	3.1	0.57	A
264	41.3	4.2	0.39	A
265	31.91	0.45	2.05	W
266	41.20	0.98	0.36	A
269	41.74	1.20	0.50	A
271	41.3	5.8	0.39	A
290	49.59	2.3	2.54	W
295	1.00	0.04	10.08	N
297	45.99	3.85	1.61	A
298	41.8	0.7	0.52	A
299	36.4	1.16	0.88	A
303	29.48	0.89	2.68	W
306	30.1	0.9	2.52	W
309	34.29	2.701	1.43	A
310	41.5	0.7	0.44	A
312	41.3	1.4	0.39	A

Reported Results for Cs-137 Determination in Sample 7

TABLE 69. Target values

Parameter	Value
Robust Mean, Bq/kg	83.2
Robust SD, Bq/kg	4.79
N	175




FIGURE 51. Reported results and evaluation parameters for Cs-137 in Sample 7.

TABLE 70. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	83.26	2.96	0.01	A
7	84.4	3.5	0.25	A
8	84.8	6.1	0.33	A
13	0.85	0.05	17.19	N
14	87	13	0.79	A
18	84.74	5.03	0.32	A
20	85.3	4.5	0.44	A
21	81.6	4.97	0.33	A
22	84.382	3.493	0.25	A
24	85.0	5.1	0.38	A
29	72.7	2.69	2.19	W
32	72.1	1.4	2.32	W
33	86.05	5.17	0.59	A
38	89	7	1.21	A
39	84.6	2.8	0.29	A
40	74	4	1.92	A
47	83.2	4.4	0.00	A
49	87	4	0.79	A
50	85	3	0.38	A
51	79	6	0.88	A
52	74.27	2.85	1.86	A

TABLE 70. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
57	83.81	5.12	0.13	A
58	75.79	1.18	1.55	A
59	87	7	0.79	A
60	84.8	5.1	0.33	A
65	85.2	4.4	0.42	A
66	84.4	5.1	0.25	A
68	88.6	4.5	1.13	A
69	84.77	5.17	0.33	A
70	84.7	3.2	0.31	A
71	100.0	2.36	3.51	N
72	75.9	1.4	1.52	A
74	84.4	5.2	0.25	A
83	75.4	0.5	1.63	A
84	84.162	3.628	0.20	A
88	84.4	5.2	0.25	A
92	85.3	4.4	0.44	A
93	85.0	4.4	0.38	A
94	78.6	4.5	0.96	A
95	72.76	0.21	2.18	W
96	74.8	0.8	1.75	A
107	82.5	4.68	0.15	A
110	74.7	6.56	1.77	A
113	84.9	5.2	0.35	A
115	82.1	4.1	0.23	A
117	84.7	5.2	0.31	A
118	97	4	2.88	W
120	83.9	2.5	0.15	A
121	77.754	1.436	1.14	A
122	78.1	5.3	1.06	A
123	73.6	1.0	2.00	A
126	85.1	3.6	0.40	A
128	74.6	2.3	1.80	A
132	89.2	5.4	1.25	A
136	83.4	6.9	0.04	A
140	85	5	0.38	A
143	74.20	0.86	1.88	A
147	74.15	1.92	1.89	A
148	76.5	1.1	1.40	A
153	82.1	2.5	0.23	A
155	73.98	0.86	1.92	A
156	84.3	5.1	0.23	A
157	78.5	4.6	0.98	A
158	84.45	5.146	0.26	A
162	84.5	5.21	0.27	A
168	75.55	0.938	1.60	A
169	7475.26	91.24	1543.23	N
172	86.50	2.48	0.69	A
175	73.8	3.7	1.96	A
178	84	5	0.17	A
187	85.3	6.7	0.44	A
191	86	7	0.58	A
196	78.07	2.22	1.07	A
198	75.40	1.12	1.63	A
201	87.8	5.3	0.96	A

TABLE 70. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
207	85.26	4.98	0.43	A
210	84.1	4.3	0.19	A
213	85	4	0.38	A
214	91	2.8	1.63	A
216	87.6	6.1	0.92	A
217	84.8	2.4	0.33	A
220	85.09	3.0	0.39	A
221	82.53	8.26	0.14	A
230	84.7	5.2	0.31	A
231	75	3.5	1.71	A
233	797	7.3	149.02	N
235	86.51	1.55	0.69	A
238	84.7	3.4	0.31	A
239	80.13	11.62	0.64	A
242	86.8	7.3	0.75	A
247	85.7	2.6	0.52	A
250	307.45	4.2	46.82	N
251	84.56	1.61	0.28	A
253	84.3	4.3	0.23	A
255	85.6	31.3	0.50	A
259	97.43	6.01	2.97	W
261	84.3	5.1	0.23	A
264	84.3	7.6	0.23	A
265	73.82	0.96	1.96	A
266	84.63	3.52	0.30	A
269	83.36	2.38	0.03	A
271	83.0	3.8	0.04	A
290	74.01	3.6	1.92	A
295	2.19	0.11	16.91	N
297	1.67	0.09	17.02	N
298	85.9	1.4	0.56	A
299	96.60	6.69	2.80	W
303	74.3	2.0	1.86	A
306	85	3	0.38	A
309	84.67	6.69	0.31	A
310	84	2.7	0.17	A
312	85.7	4.4	0.52	A

Reported Results for Tl-208 Determination in Sample 7

TABLE 71. Target values

Parameter	Value
Robust Mean, Bq/kg	13.25
Robust SD, Bq/kg	1.92
N	90




FIGURE 52. Reported results and evaluation parameters for Tl-208 in Sample 7.

TABLE 72. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	13.73	0.55	0.25	A
7	14.7	0.5	0.76	A
13	0.14	0.01	6.83	N
14	10.2	1.3	1.59	A
18	13.47	1.13	0.11	A
20	10.8	0.6	1.28	A
21	13.3	0.81	0.03	A
22	10.638	0.486	1.36	A
33	12.21	1.66	0.54	A
39	14.0	0.9	0.39	A
40	12.1	0.8	0.60	A
59	12.1	1.1	0.60	A
60	13.3	1.2	0.03	A
68	13.8	0.7	0.29	A
70	12.9	0.4	0.18	A
72	12	1	0.65	A
76	9.96	0.40	1.71	A
84	13.818	0.737	0.30	A
88	13.0	0.5	0.13	A
94	14.5	0.9	0.65	A
96	9.25	0.14	2.08	W

TABLE 72. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
110	11.7	1.15	0.81	A
113	14.7	0.7	0.76	A
115	12.65	0.80	0.31	A
120	14.7	0.5	0.76	A
121	10.221	0.230	1.58	A
122	11.7	1.1	0.81	A
123	11.8	0.2	0.76	A
126	13.62	0.57	0.19	A
128	10.2	1.2	1.59	A
132	43.9	1.9	15.96	N
136	12.7	1.1	0.29	A
140	13	1	0.13	A
148	10.09	0.14	1.65	A
155	9.46	0.14	1.97	A
157	12.9	0.8	0.18	A
158	13.03	0.651	0.11	A
162	13.4	0.83	0.08	A
175	12.1	0.6	0.60	A
178	13.3	0.7	0.03	A
191	12.2	1.0	0.55	A
198	12.01	0.42	0.65	A
210	13.3	0.6	0.03	A
217	13.5	0.40	0.13	A
221	10.7	1.08	1.33	A
231	26	2.6	6.64	N
233	104	0.73	47.27	N
247	11.14	0.35	1.10	A
251	35.12	1.11	11.39	N
253	39.4	1.4	13.62	N
255	13.9	1.4	0.34	A
264	13.4	1.4	0.08	A
271	12.8	0.5	0.23	A
303	10.34	0.33	1.52	A
312	13.65	0.73	0.21	A

Reported Results for Pb-210 Determination in Sample 7

TABLE 73. Target values

Parameter	Value
Robust Mean, Bq/kg	621.79
Robust SD, Bq/kg	80.1
N	126




FIGURE 53. Reported results and evaluation parameters for Pb-210 in Sample 7.

TABLE 74. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	712.80	27.13	1.14	A
13	6.3	0.7	7.68	N
14	647	97	0.31	A
18	636.6	55.6	0.18	A
20	754	39	1.65	A
21	618	88.7	0.05	A
24	667	40	0.56	A
29	409.2	60.56	2.65	W
32	865	26	3.04	N
33	640.86	64.11	0.24	A
39	685	42	0.79	A
40	636	40	0.18	A
49	980	40	4.47	N
51	606	47	0.20	A
52	545.79	66.7	0.95	A
57	634.15	92.59	0.15	A
58	520	6	1.27	A
59	629	53	0.09	A
60	641	56	0.24	A
66	663	40	0.51	A
68	668.8	34.2	0.59	A

TABLE 74. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
69	625.54	90.93	0.05	A
70	645.8	81.6	0.30	A
71	755.2	67.3	1.67	A
74	678	68	0.70	A
76	387	6.70	2.93	W
83	514.3	18.5	1.34	A
84	644.053	97.092	0.28	A
88	638	93	0.20	A
94	636.9	67.1	0.19	A
96	416	5.9	2.57	W
107	636	93.9	0.18	A
110	394	34.8	2.84	W
113	637.9	92.9	0.20	A
115	646	33	0.30	A
118	1750	150	14.09	N
120	659	50	0.46	A
121	429.409	1.458	2.40	W
123	416.0	22.0	2.57	W
126	646	31	0.30	A
128	372	53	3.12	N
132	665	97	0.54	A
136	652	57	0.38	A
148	560.4	2.9	0.77	A
153	625.4	84.1	0.05	A
155	435.33	8.39	2.33	W
156	695	95	0.91	A
158	643.38	92.584	0.27	A
162	647	94	0.31	A
169	47801.17	1784.20	589.01	N
172	634.30	43.26	0.16	A
178	644	93	0.28	A
187	646	99	0.30	A
191	631	50	0.11	A
198	488.70	23.23	1.66	A
207	569.00	87.06	0.66	A
210	642.8	33.4	0.26	A
213	620	40	0.02	A
214	548.8	44.8	0.91	A
216	577	40	0.56	A
217	619	17	0.03	A
220	643.1	103.1	0.27	A
221	554.00	55.42	0.85	A
231	333	16	3.61	N
233	508	12.7	1.42	A
235	662.4	44.9	0.51	A
238	647	27	0.31	A
239	586	84.97	0.45	A
242	641	56	0.24	A
247	635	26	0.16	A
253	646.7	33.8	0.31	A
264	633	79	0.14	A
271	646.6	29.5	0.31	A
298	648	12	0.33	A
299	432.60	20.45	2.36	W

TABLE 74. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
303	408.4	9.2	2.66	W
309	668.7	76.85	0.59	A
310	652	74	0.38	A
312	643.2	33.4	0.27	A

Reported Results for Pb-212 Determination in Sample 7

TABLE 75. Target values

Parameter	Value
Robust Mean, Bq/kg	38.24
Robust SD, Bq/kg	3.99
N	96




FIGURE 54. Reported results and evaluation parameters for Pb-212 in Sample 7.

TABLE 76. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	42.28	1.59	1.01	A
7	42.5	3.3	1.07	A
13	0.4	0.02	9.48	N
14	44.0	5.6	1.44	A
18	36.20	2.04	0.51	A
20	38.5	2.1	0.07	A
21	35.95	2.90	0.57	A
22	37.523	2.288	0.18	A
33	34.89	3.92	0.84	A
39	39.6	1.8	0.34	A
50	38	3	0.06	A
60	39.0	2.5	0.19	A
68	37.6	1.9	0.16	A
69	34.42	1.68	0.96	A
70	36.3	1.4	0.49	A
72	33	1	1.31	A
74	38.4	3.1	0.04	A
76	33.7	0.80	1.14	A
84	41.804	1.919	0.89	A
88	36.5	2.2	0.44	A
94	42.5	2.7	1.07	A

TABLE 76. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
96	30.7	0.3	1.89	A
110	34.1	3.00	1.04	A
113	43.4	3.6	1.29	A
115	38.5	2.1	0.07	A
120	42.2	1.4	0.99	A
121	36.696	0.687	0.39	A
122	38.7	3.0	0.12	A
123	32.3	0.8	1.49	A
126	38.2	2.1	0.01	A
128	34.9	4.2	0.84	A
132	46.6	3.8	2.10	W
136	38.8	2.9	0.14	A
140	39	3	0.19	A
148	34.60	0.24	0.91	A
157	38.6	2.6	0.09	A
158	39.842	3.165	0.40	A
162	38.3	3.1	0.02	A
169	2791.38	36.50	690.01	N
175	32.3	1.6	1.49	A
178	36	3	0.56	A
198	35.44	0.69	0.70	A
210	34.9	1.8	0.84	A
213	36	2	0.56	A
214	54.4	1.3	4.05	N
217	38.4	1.1	0.04	A
221	40.64	4.07	0.60	A
231	26.5	7	2.94	W
233	214	1.49	44.05	N
247	39.1	1.2	0.22	A
253	39.0	2.1	0.19	A
255	37.4	8.9	0.21	A
264	38.3	3.9	0.02	A
271	36.4	1.6	0.46	A
299	53800.00	4860.00	13474.13	N
303	35.7	1.0	0.64	A
312	38.2	2.0	0.01	A

Reported Results for Bi-212 Determination in Sample 7

TABLE 77. Target values

Parameter	Value
Robust Mean, Bq/kg	39.39
Robust SD, Bq/kg	5.02
N	81




FIGURE 55. Reported results and evaluation parameters for Bi-212 in Sample 7.

TABLE 78. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	43.60	2.34	0.84	A
7	38.4	1.6	0.20	A
13	0.43	0.07	7.76	N
14	33.0	4.6	1.27	A
18	39.58	3.60	0.04	A
21	41.3	2.5	0.38	A
22	22.692	1.273	3.33	N
33	23.25	1.54	3.22	N
39	40.0	6.8	0.12	A
50	41	3	0.32	A
60	38.7	3.8	0.14	A
68	37.6	1.9	0.36	A
69	21.62	1.27	3.54	N
70	38.4	1.7	0.20	A
76	33.6	1.60	1.15	A
84	40.431	3.467	0.21	A
94	47.2	3.2	1.56	A
96	19.9	0.7	3.88	N
110	37.2	4.69	0.44	A
113	40.8	2.3	0.28	A
115	37.9	2.2	0.30	A

TABLE 78. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
120	40.5	4.5	0.22	A
121	23.744	0.976	3.12	N
122	36.4	4.4	0.60	A
128	38.2	4.6	0.24	A
132	46.2	2.3	1.36	A
136	38.1	3.3	0.26	A
140	42	3	0.52	A
148	38.42	0.28	0.19	A
157	39.4	3.1	0.00	A
158	39.778	6.775	0.08	A
162	41.6	2.8	0.44	A
169	2038.03	42.77	398.14	N
175	38.2	2.3	0.24	A
178	42	3	0.52	A
198	41.39	1.51	0.40	A
210	38.6	2.3	0.16	A
217	37.9	1.6	0.30	A
231	39.6	9	0.04	A
247	40.0	1.6	0.12	A
253	35.0	2.0	0.87	A
255	45.2	3.0	1.16	A
264	37.3	4.4	0.42	A
271	41.9	2.4	0.50	A
297	47.88	3.78	1.69	A
299	7070000.00	428000.00	1408358.69	N
303	23.7	1.3	3.13	N
312	42.0	2.5	0.52	A

Reported Results for Bi-214 Determination in Sample 7

TABLE 79. Target values

Parameter	Value
Robust Mean, Bq/kg	31.49
Robust SD, Bq/kg	3.39
N	100




FIGURE 56. Reported results and evaluation parameters for Bi-214 in Sample 7.

TABLE 80. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	32.38	1.24	0.26	A
13	0.33	0.02	9.19	N
14	22.1	2.5	2.77	W
20	27.6	1.5	1.15	A
21	31.1	1.9	0.12	A
22	26.639	1.118	1.43	A
32	25.3	0.4	1.83	A
33	31.28	0.99	0.06	A
39	33.1	1.5	0.47	A
40	32	2	0.15	A
49	33.7	1.5	0.65	A
50	32	1	0.15	A
57	31.68	0.80	0.06	A
58	27.01	2.01	1.32	A
59	34	4	0.74	A
68	33.7	1.7	0.65	A
69	29.43	0.65	0.61	A
70	30.8	0.6	0.20	A
72	28	2	1.03	A
76	23.1	0.60	2.47	W
83	35.4	1.1	1.15	A

TABLE 80. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
84	32.320	1.418	0.24	A
94	31.6	1.8	0.03	A
96	24.6	0.3	2.03	W
110	30.5	2.72	0.29	A
113	30.2	0.8	0.38	A
115	31.6	2.0	0.03	A
120	33.6	1.1	0.62	A
121	25.515	0.531	1.76	A
122	30.9	2.3	0.17	A
123	28.0	0.6	1.03	A
128	23.7	3.4	2.30	W
132	33.3	0.8	0.53	A
140	31	2	0.14	A
148	26.4	2.6	1.50	A
157	31.1	1.9	0.12	A
158	31.360	2.509	0.04	A
162	31.8	1.9	0.09	A
169	2588.92	30.97	754.40	N
175	29.2	1.5	0.68	A
178	32.8	1.1	0.39	A
191	35	3	1.04	A
198	24.74	2.22	1.99	A
207	35.44	2.20	1.17	A
210	29.0	0.8	0.73	A
213	32.5	1.9	0.30	A
216	35.1	1.9	1.06	A
217	32.6	0.95	0.33	A
221	37.53	3.77	1.78	A
231	24	5	2.21	W
233	286	6.6	75.08	N
239	27.66	4.01	1.13	A
247	29.20	0.59	0.68	A
250	118.27	1.83	25.60	N
251	32.98	1.01	0.44	A
253	30.3	0.7	0.35	A
255	33.0	4.4	0.45	A
264	31.9	3.8	0.12	A
271	32.6	1.1	0.33	A
297	35.55	2.58	1.20	A
303	24.70	0.88	2.00	A
310	31.4	0.7	0.03	A
312	31.8	1.2	0.09	A

Reported Results for Pb-214 Determination in Sample 7

TABLE 81. Target values

Parameter	Value
Robust Mean, Bq/kg	32.86
Robust SD, Bq/kg	2.5
N	94




FIGURE 57. Reported results and evaluation parameters for Pb-214 in Sample 7.

TABLE 82. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	34.83	1.31	0.79	A
14	34.8	4.4	0.78	A
20	31.7	1.7	0.46	A
21	33.1	2.5	0.10	A
22	32.228	1.971	0.25	A
32	30.0	0.6	1.14	A
33	31.18	0.22	0.67	A
39	34.2	1.8	0.54	A
40	34	2	0.46	A
49	34.1	1.5	0.50	A
50	33	1	0.06	A
58	27.43	0.56	2.17	W
59	35	4	0.86	A
68	33.7	1.7	0.34	A
69	30.72	1.47	0.86	A
70	33.4	1	0.22	A
72	29	2	1.54	A
74	34.1	2.8	0.50	A
76	25.8	0.15	2.82	W
83	34.9	0.3	0.82	A
84	34.003	2.031	0.46	A

TABLE 82. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
94	30.3	1.9	1.02	A
96	25.0	0.3	3.14	N
110	28.1	2.48	1.90	A
113	33.7	2.6	0.34	A
115	32.9	1.7	0.02	A
120	34.8	1.1	0.78	A
121	29.981	0.616	1.15	A
122	31.4	2.4	0.58	A
123	27.9	0.6	1.98	A
128	29.7	4.2	1.26	A
132	35.4	2.1	1.02	A
140	34	2	0.46	A
148	31.65	0.59	0.48	A
157	32.0	2.0	0.34	A
158	33.69	2.695	0.33	A
162	32.5	2.5	0.14	A
169	2769.54	34.55	1094.67	N
175	28.2	1.4	1.86	A
178	33.0	2.0	0.06	A
191	36	3	1.26	A
198	30.53	2.74	0.93	A
207	32.02	1.89	0.34	A
210	31.2	1.1	0.66	A
213	32	2	0.34	A
216	31.6	2.2	0.50	A
217	34.4	1.0	0.62	A
221	42.95	4.30	4.04	N
231	19.5	2	5.34	N
247	32.47	0.72	0.16	A
251	34.48	1.11	0.65	A
253	35.1	0.9	0.90	A
255	33.9	4.8	0.42	A
264	32.6	3.9	0.10	A
271	33.9	1.1	0.42	A
297	44.81	3.12	4.78	N
303	29.1	1.0	1.50	A
312	33.7	1.2	0.34	A

Reported Results for Ra-224 Determination in Sample 7

TABLE 83. Target values

Parameter	Value
Robust Mean, Bq/kg	39.91
Robust SD, Bq/kg	5.74
N	27




FIGURE 58. Reported results and evaluation parameters for Ra-224 in Sample 7.

TABLE 84. IC Evaluation Results for Ra-224 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
14	39.0	5.8	0.16
84	42.393	3.344	0.43
94	44.8	3.2	0.85
110	32.1	3.17	1.36
115	37.1	2.0	0.49
126	38	1.3	0.33
158	39.962	6.458	0.01
175	32.0	1.9	1.38
214	17.6	3.4	3.89
217	35.2	1.7	0.82
251	52.20	1.58	2.14
255	36.0	2.5	0.68
299	2430.00	257.00	416.39
312	37.8	2.1	0.37

Reported Results for Ra-226 Determination in Sample 7

TABLE 85. Target values

Parameter	Value
Robust Mean, Bq/kg	38
Robust SD, Bq/kg	10.91
N	146




FIGURE 59. Reported results and evaluation parameters for Ra-226 in Sample 7.

TABLE 86. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	42.25	2.28	0.39	A
7	42	7	0.37	A
14	12.8	5.1	2.31	W
16	32.43	1.9	0.51	A
22	47.852	3.519	0.90	A
24	32.8	1.2	0.48	A
29	48.45	3.92	0.96	A
31	32.2	4.0	0.53	A
32	33.7	0.8	0.39	A
33	37.46	3.08	0.05	A
39	39.6	2.9	0.15	A
40	56	6	1.65	A
47	36.4	2.4	0.15	A
49	34.4	1.5	0.33	A
50	33	1	0.46	A
51	61	6	2.11	W
52	26.47	0.98	1.06	A
58	35.57	1.83	0.22	A
59	60	8	2.02	W
66	32.6	2.1	0.49	A
68	33.7	1.7	0.39	A

TABLE 86. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
69	36.31	8.60	0.15	A
71	35.16	1.69	0.26	A
72	28	6	0.92	A
74	32.3	2.0	0.52	A
76	24.5	0.40	1.24	A
83	32.5	2.4	0.50	A
84	39.179	4.958	0.11	A
88	41.5	6.8	0.32	A
93	27.2	0.6	0.99	A
94	36.2	2.8	0.16	A
96	24.8	0.3	1.21	A
101	94.5	4	5.18	N
106	61.90	1.35	2.19	W
110	35.9	5.46	0.19	A
113	49.3	8.0	1.04	A
115	30.0	2.2	0.73	A
117	29.9	0.8	0.74	A
118	36	2	0.18	A
120	35.1	1.8	0.27	A
122	34.9	3.1	0.28	A
123	27.9	1.7	0.93	A
126	37.9	5	0.01	A
128	55.3	6.7	1.59	A
130	51.94	3.9	1.28	A
132	64.4	5.30	2.42	W
143	30.30	2.51	0.71	A
153	29.7	3.06	0.76	A
155	22.09	0.36	1.46	A
156	31.0	1.3	0.64	A
157	34.9	3.5	0.28	A
158	29.870	5.970	0.75	A
160	38.1	0.9	0.01	A
163	24.61	1.19	1.23	A
168	26.11	0.33	1.09	A
169	2769.54	34.55	250.37	N
175	28.2	1.7	0.90	A
178	33	3	0.46	A
183	29.67	0.41	0.76	A
191	61	6	2.11	W
198	31.70	2.85	0.58	A
201	43.4	3.9	0.49	A
210	51.4	7.9	1.23	A
213	33	2	0.46	A
214	34.9	0.8	0.28	A
217	31.5	1.8	0.60	A
219	81.8	2.6	4.01	N
220	53.18	7.29	1.39	A
221	40.24	4.03	0.21	A
228	28.3	2.9	0.89	A
231	24.3	2.4	1.26	A
233	17.8	0.12	1.85	A
238	44.2	6.3	0.57	A
239	34.13	4.95	0.35	A
242	34.5	4.4	0.32	A

TABLE 86. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
247	36.0	1.7	0.18	A
251	33.45	1.11	0.42	A
253	51.3	7.2	1.22	A
255	56.2	3.8	1.67	A
259	45.45	3.66	0.68	A
261	32.7	3.0	0.49	A
264	48.1	5.8	0.93	A
265	27.61	0.61	0.95	A
266	29.30	0.52	0.80	A
269	34.66	1.76	0.31	A
271	33.3	0.8	0.43	A
276	29.3	0.81	0.80	A
284	92.827	3.765	5.03	N
297	45.86	3.48	0.72	A
305	35.3891	1.1988	0.24	A
306	32	1	0.55	A
309	32.74	2.386	0.48	A
310	33.9	1.1	0.38	A
312	36.1	4.1	0.17	A

Reported Results for Ac-227 Determination in Sample 7

TABLE 87. Target values

Parameter	Value
Robust Mean, Bq/kg	1.56
Robust SD, Bq/kg	0.21
N	5




FIGURE 60. Reported results and evaluation parameters for Ac-227 in Sample 7.

TABLE 88. IC Evaluation Results for Ac-227 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
84	1.484	0.245	0.36
132	1.79	0.28	1.10
298	1.67	0.26	0.52
309	1.561	0.274	0.00

Reported Results for Th-227 Determination in Sample 7

TABLE 89. Target values

Parameter	Value
Robust Mean, Bq/kg	1.65
Robust SD, Bq/kg	0.41
N	15




FIGURE 61. Reported results and evaluation parameters for Th-227 in Sample 7.

TABLE 90. IC Evaluation Results for Th-227 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
7	1.46	0.19	0.46
69	2.16	0.23	1.24
84	1.464	0.242	0.45
110	1.29	0.322	0.88
169	196.73	12.82	475.80
217	1.49	0.38	0.39
253	0.8	0.2	2.07
312	1.60	0.22	0.12

Reported Results for Ra-228 Determination in Sample 7

TABLE 91. Target values

Parameter	Value
Robust Mean, Bq/kg	37.77
Robust SD, Bq/kg	2.2
N	29




FIGURE 62. Reported results and evaluation parameters for Ra-228 in Sample 7.

TABLE 92. IC Evaluation Results for Ra-228 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
21	36.05	1.73	0.78
24	37.8	1.3	0.01
66	37.8	2.4	0.01
68	37.6	1.9	0.08
71	39.1	1.03	0.60
74	35.3	1.8	1.12
84	38.613	2.289	0.38
126	36.5	1.3	0.58
169	3071.99	25.21	1379.19
207	38.21	2.61	0.20
214	42.8	1	2.29
216	40.6	1.2	1.29
230	37.2	1.8	0.26
238	35.3	0.7	1.12
271	38.5	1.1	0.33
298	38.6	1.0	0.38
309	34.54	2.575	1.47
310	35.7	0.8	0.94

Reported Results for Th-228 Determination in Sample 7

TABLE 93. Target values

Parameter	Value
Robust Mean, Bq/kg	38.98
Robust SD, Bq/kg	4.11
N	27




FIGURE 63. Reported results and evaluation parameters for Th-228 in Sample 7.

TABLE 94. IC Evaluation Results for Th-228 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
7	31.5	4.7	1.82
14	29.7	8.3	2.26
40	36	2	0.73
59	36	4	0.73
83	35.0	0.7	0.97
84	42.393	3.344	0.83
126	37.9	1.3	0.26
143	81.90	3.42	10.44
191	36	3	0.73
207	38.30	2.3	0.17
216	39.5	3.1	0.13
238	36.1	0.8	0.70
255	47.6	3.7	2.10
271	37.0	1.0	0.48
298	42.3	0.9	0.81
309	43.07	4.102	1.00
310	34.2	0.9	1.16

Reported Results for Ac-228 Determination in Sample 7

TABLE 95. Target values

Parameter	Value
Robust Mean, Bq/kg	36.27
Robust SD, Bq/kg	3.03
N	118




FIGURE 64. Reported results and evaluation parameters for Ac-228 in Sample 7.

TABLE 96. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	39.77	2.09	1.16	A
7	35.7	1.0	0.19	A
13	0.38	0.03	11.84	N
14	36.8	4.2	0.17	A
20	41.4	2.2	1.69	A
21	36.05	1.73	0.07	A
22	32.977	1.411	1.09	A
33	35.17	1.25	0.36	A
39	39.9	2	1.20	A
40	35	2	0.42	A
49	36.8	1.7	0.17	A
50	39	1	0.90	A
51	31	2	1.74	A
57	36.99	1.93	0.24	A
58	33.51	1.1	0.91	A
59	32	3	1.41	A
60	34.9	2.9	0.45	A
68	37.6	1.9	0.44	A
69	32.68	0.72	1.18	A
70	33.5	0.6	0.91	A
72	35	2	0.42	A

TABLE 96. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
76	32.8	1.00	1.15	A
83	31.9	0.24	1.44	A
84	38.613	2.289	0.77	A
88	34.8	0.9	0.49	A
92	33.4	0.7	0.95	A
93	33.2	0.8	1.01	A
94	39.8	2.3	1.17	A
96	33.1	0.4	1.05	A
110	32.0	2.87	1.41	A
113	37.2	1.3	0.31	A
115	36.5	1.9	0.08	A
118	37	2	0.24	A
120	38.6	1.3	0.77	A
121	31.095	0.789	1.71	A
122	35.7	2.6	0.19	A
123	34.0	0.5	0.75	A
126	36.5	1.3	0.08	A
128	33.0	3.7	1.08	A
132	40.2	1.2	1.30	A
136	35.6	2.8	0.22	A
140	35	2	0.42	A
148	36.61	0.42	0.11	A
156	34.3	1.9	0.65	A
157	36.3	2.1	0.01	A
158	35.76	0.784	0.17	A
162	36.4	1.7	0.04	A
169	3071.99	25.21	1001.89	N
175	33.3	1.7	0.98	A
178	32.1	1.4	1.38	A
191	34	2	0.75	A
198	35.26	0.41	0.33	A
210	36.3	1.4	0.01	A
213	37	2	0.24	A
217	38.0	1.2	0.57	A
221	36.03	3.64	0.08	A
231	35.2	3.5	0.35	A
233	328	8.09	96.28	N
239	35.25	5.11	0.34	A
247	36.39	0.71	0.04	A
250	147.28	3.22	36.64	N
251	36.53	1.20	0.09	A
253	34.4	0.7	0.62	A
255	38.7	3.0	0.80	A
261	37.3	1.8	0.34	A
264	35.6	4.3	0.22	A
271	38.5	1.1	0.74	A
297	40.36	3.45	1.35	A
299	187000.00	10300.00	61704.20	N
303	32.5	1.1	1.24	A
312	36.3	1.9	0.01	A

Reported Results for Th-230 Determination in Sample 7

TABLE 97. Target values

Parameter	Value
Robust Mean, Bq/kg	39.65
Robust SD, Bq/kg	4.58
N	4




FIGURE 65. Reported results and evaluation parameters for Th-230 in Sample 7.

TABLE 98. IC Evaluation Results for Th-230 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
255	36.7	8.9	0.64

Reported Results for Th-231 Determination in Sample 7

TABLE 99. Target values

Parameter	Value
Robust Mean, Bq/kg	7.3
Robust SD, Bq/kg	6.82
N	5



FIGURE 66. Reported results and evaluation parameters for Th-231 in Sample 7.

TABLE 100. IC Evaluation Results for Th-231 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
69	4.79	0.52	0.37
169	1116.88	24.84	162.70

Reported Results for Th-232 Determination in Sample 7

TABLE 101. Target values

Parameter	Value
Robust Mean, Bq/kg	36.55
Robust SD, Bq/kg	4.41
N	11




FIGURE 67. Reported results and evaluation parameters for Th-232 in Sample 7.

TABLE 102. IC Evaluation Results for Th-232 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
29	96.65	14.5	13.63
38	36	3	0.12
49	36.8	1.7	0.06
50	39	1	0.56
52	31.70	1.54	1.10
69	32.68	0.72	0.88
264	37.1	4.4	0.12
266	30.40	0.52	1.39
306	41	1	1.01

Reported Results for Th-234 Determination in Sample 7

TABLE 103. Target values

Parameter	Value
Robust Mean, Bq/kg	28.6
Robust SD, Bq/kg	6.56
N	66




FIGURE 68. Reported results and evaluation parameters for Th-234 in Sample 7.

TABLE 104. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
7	27.1	2.2	0.23	A
14	34.9	4.9	0.96	A
39	27.2	1.8	0.21	A
40	17	2	1.77	A
52	992.96	99.2	147.01	N
59	18	3	1.62	A
68	28.6	2.9	0.00	A
69	25.27	2.59	0.51	A
76	11.3	1.80	2.64	W
84	27.382	1.928	0.19	A
88	24.2	1.8	0.67	A
92	29.7	1.7	0.17	A
94	29.0	2.6	0.06	A
110	7.98	1.18	3.14	N
113	26.0	2.7	0.40	A
118	60	3	4.79	N
122	44.5	4.8	2.42	W
126	27.5	1.9	0.17	A
128	19.5	2.4	1.39	A
132	34.8	2.8	0.95	A
143	478.00	11.20	68.51	N

TABLE 104. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
148	40.12	0.13	1.76	A
157	37.3	4.0	1.33	A
158	28.81	2.753	0.03	A
175	19.9	1.3	1.33	A
178	23	3	0.85	A
191	19	2	1.46	A
210	19.3	4.1	1.42	A
213	29	6	0.06	A
217	34.4	1.5	0.88	A
247	27.6	1.2	0.15	A
251	31.55	1.36	0.45	A
255	22.0	1.6	1.01	A
269	591.82	23.06	85.86	N
271	27.8	6.8	0.12	A
312	26.2	1.6	0.37	A

Reported Results for Pa-234m Determination in Sample 7

TABLE 105. Target values

Parameter	Value
Robust Mean, Bq/kg	26.57
Robust SD, Bq/kg	10.19
N	35




FIGURE 69. Reported results and evaluation parameters for Pa-234m in Sample 7.

TABLE 106. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
7	28	5	0.14	A
14	29.2	9.0	0.26	A
16	31.2	2.2	0.45	A
47	18.1	10	0.83	A
71	18.54	3.83	0.79	A
83	21.2	0.74	0.53	A
84	27.382	1.928	0.08	A
96	0		2.61	W
115	29.2	8.8	0.26	A
130	40.32	2.94	1.35	A
155	17.29	0.99	0.91	A
160	0.35	0.12	2.57	W
163	20.19	2.81	0.63	A
168	26.11	0.33	0.05	A
214	-9.2	59	3.51	N
216	25.0	1.8	0.15	A
228	28.3	15.5	0.17	A
231	7.75	1	1.85	A
233	57.9	1.44	3.07	N
253	22.0	1.4	0.45	A
297	33.99	2.86	0.73	A

TABLE 106. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
500	5	2.5	2.12	W

Reported Results for U-235 Determination in Sample 7

TABLE 107. Target values

Parameter	Value
Robust Mean, Bq/kg	1.46
Robust SD, Bq/kg	0.55
N	117




FIGURE 70. Reported results and evaluation parameters for U-235 in Sample 7.

TABLE 108. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	1.62	0.40	0.29	A
7	0.96	0.33	0.91	A
14	2.56	0.44	2.00	A
16	1.44	0.28	0.04	A
29	1.1	0.33	0.65	A
32	1.50	0.05	0.07	A
33	1.62	0.13	0.29	A
39	1.3	0.4	0.29	A
40	0.79	0.15	1.22	A
47	2.36	0.31	1.64	A
49	1.1	0.1	0.65	A
50	1.3	0.1	0.29	A
52	1.67	0.1	0.38	A
59	0.86	0.19	1.09	A
68	1.506	0.298	0.08	A
69	1.75	0.39	0.53	A
72	1.7	0.3	0.44	A
76	0.58	0.33	1.60	A
83	1.4	0.1	0.11	A
84	1.449	0.264	0.02	A
88	1.30	0.16	0.29	A

TABLE 108. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
93	1.97	0.11	0.93	A
94	1.7	0.4	0.44	A
96	0.64	0.35	1.49	A
101	1.1	0.04	0.65	A
110	1.09	0.285	0.67	A
115	1.44	0.28	0.04	A
117	1.9	0.3	0.80	A
118	2.3	0.4	1.53	A
120	1.5	0.4	0.07	A
121	3.523	0.139	3.75	N
122	1.60	0.41	0.25	A
123	1.32	0.08	0.25	A
126	1.27	0.1	0.35	A
130	1.92	0.14	0.84	A
143	1.18	0.14	0.51	A
147	3.82	0.16	4.29	N
148	1.440	0.018	0.04	A
155	3.64	0.09	3.96	N
157	1.6	0.6	0.25	A
158	1.630	0.650	0.31	A
160	3.2	0.6	3.16	N
163	1.41	0.25	0.09	A
168	4.14	0.101	4.87	N
169	365.08	6.95	661.13	N
175	1.2	0.1	0.47	A
178	1.5	0.2	0.07	A
191	0.89	0.14	1.04	A
198	1.23	0.09	0.42	A
210	0.7	0.4	1.38	A
213	1.43	0.11	0.05	A
214	0.8	0.43	1.20	A
217	1.22	0.30	0.44	A
219	0.6	0.02	1.56	A
220	1.0	0.4	0.84	A
221	1.68	0.28	0.40	A
228	2.14	0.37	1.24	A
233	0.51	0.003	1.73	A
238	1.2	0.4	0.47	A
239	1.6	0.31	0.25	A
247	1.27	0.1	0.35	A
251	1.84	0.46	0.69	A
253	0.7	0.4	1.38	A
259	1.20	0.12	0.47	A
265	1.26	0.16	0.36	A
266	1.80	0.16	0.62	A
269	1.65	0.14	0.35	A
271	1.5	0.1	0.07	A
276	2.61	0.25	2.09	W
284	16.54	1.50	27.42	N
290	4.75	0.28	5.98	N
295	0.07	0.01	2.53	W
297	1.91	0.26	0.82	A
303	1.89	0.61	0.78	A
305	1.6548	0.4201	0.35	A

TABLE 108. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
306	2.0	0.3	0.98	A
309	1.708	0.2548	0.45	A
310	1.5	0.2	0.07	A
312	1.34	0.17	0.22	A

Reported Results for U-238 Determination in Sample 7

TABLE 109. Target values

Parameter	Value
Robust Mean, Bq/kg	28
Robust SD, Bq/kg	6.45
N	16




FIGURE 71. Reported results and evaluation parameters for U-238 in Sample 7.

TABLE 110. IC Evaluation Results for U-238 in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score
14	525	98	77.05
29	343.7	103.1	48.95
38	29	2	0.16
49	20	1	1.24
50	25	2	0.47
84	27.382	1.928	0.10
107	25.6	2.60	0.37
115	29.9	1.9	0.29
126	27.5	1.9	0.08
214	28.28	0.66	0.04

Reported Results for Am-241 Determination in Sample 7

TABLE 111. Target values

Parameter	Value
Robust Mean, Bq/kg	78.89
Robust SD, Bq/kg	7.59
N	169




FIGURE 72. Reported results and evaluation parameters for Am-241 in Sample 7.

TABLE 112. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
6	83.40	2.96	0.59	A
7	80	8	0.15	A
8	81.5	8.2	0.34	A
13	0.83	0.09	10.28	N
18	82.22	7.13	0.44	A
20	86.7	4.5	1.03	A
21	77.8	7.91	0.14	A
22	80.148	5.871	0.17	A
24	82.9	5.0	0.53	A
29	56.9	2.85	2.90	W
33	78.12	7.81	0.10	A
37	84.5	1.1	0.74	A
39	84	5	0.67	A
40	58	4	2.75	W
47	83.4	4.5	0.59	A
49	95	4	2.12	W
50	75	5	0.51	A
51	74	8	0.64	A
52	85.81	5.88	0.91	A
57	78.36	8.15	0.07	A
58	66.76	1.14	1.60	A

TABLE 112. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
59	81	10	0.28	A
60	80.5	7.1	0.21	A
66	82.7	5.0	0.50	A
68	83.4	4.3	0.59	A
69	80.27	8.20	0.18	A
70	81.2	6.6	0.30	A
71	94.6	6.9	2.07	W
72	61	3	2.36	W
74	80.2	8.1	0.17	A
76	59.1	1.50	2.61	W
83	67.1	0.8	1.55	A
84	82.359	5.153	0.46	A
88	81.0	8.3	0.28	A
92	81.6	4.2	0.36	A
93	80.9	4.2	0.26	A
94	75.2	6.3	0.49	A
96	58.4	0.6	2.70	W
110	57.8	5.08	2.78	W
113	81.3	8.3	0.32	A
115	72.5	3.6	0.84	A
117	86.1	8.7	0.95	A
118	204	8	16.48	N
120	82.8	3.1	0.52	A
121	58.239	1.079	2.72	W
122	85.9	8.9	0.92	A
123	58.7	0.91	2.66	W
126	80.9	3.7	0.26	A
128	59.3	4.8	2.58	W
132	85.8	8.8	0.91	A
136	83.1	6.9	0.55	A
140	80	7	0.15	A
143	57.60	0.74	2.81	W
147	61.17	1.15	2.33	W
148	63.79	0.27	1.99	A
153	79.1	6.3	0.03	A
156	83.8	8.5	0.65	A
157	84.7	8.4	0.77	A
158	80.57	8.491	0.22	A
162	78.8	8	0.01	A
169	5864.51	76.62	762.27	N
172	80.29	7.69	0.18	A
175	61.0	3.1	2.36	W
178	80	8	0.15	A
187	80.8	9.2	0.25	A
191	81	8	0.28	A
196	60.65	1.74	2.40	W
198	60.26	1.06	2.45	W
201	83.5	8.5	0.61	A
207	83.08	8.68	0.55	A
210	80.0	4.5	0.15	A
213	81	4	0.28	A
214	89.27	2	1.37	A
216	78.8	5.5	0.01	A
217	80.6	2.3	0.23	A

TABLE 112. IC Evaluation Results for in Sample 7, Rep. Value and Rep. Unc. in Bq/kg

Labcode	Rep. Value	Rep. Unc.	Z-Score	Z-Score Evaluation
220	80.2	14.2	0.17	A
221	73.44	7.35	0.72	A
230	79.3	4.9	0.05	A
231	56.4	2.3	2.96	W
233	734	18.3	86.31	N
235	81.4	7.36	0.33	A
238	80.8	3.3	0.25	A
239	70.28	10.19	1.13	A
242	81.4	6.2	0.33	A
247	76.1	2.3	0.37	A
251	100.12	1.35	2.80	W
253	80.4	4.1	0.20	A
255	79.6	31.4	0.09	A
259	99.27	33.47	2.69	W
261	80.6	8.0	0.23	A
264	79.7	9.6	0.11	A
265	59.31	0.77	2.58	W
266	82.80	6.08	0.52	A
269	72.51	2.28	0.84	A
271	78.9	4.1	0.00	A
290	59.55	2.6	2.55	W
295	2.01	0.09	10.13	N
297	114.12	9.45	4.64	N
298	81.0	1.4	0.28	A
299	65	3.16	1.83	A
303	61.3	1.4	2.32	W
306	85	3	0.81	A
309	81.58	9.124	0.35	A
310	79.3	5.1	0.05	A
312	79.3	4.1	0.05	A
500	87	65	1.07	A

6 References

References

- [1] International Organization for Standardization (ISO). (2010). Conformity assessment - General requirements for proficiency testing, ISO/IEC 17043:2010. Geneva: Switzerland.
- [2] International Organization for Standardization (ISO). (2015). Statistical methods for use in proficiency testing by interlaboratory comparison, ISO 13528:2015. Geneva: Switzerland.

7 Contributors to evaluation and report

B. Nadalut, S. Patterson, B. Seslak, A. Trinkl: Terrestrial Environmental Radiochemistry Laboratory, Division of Physical and Chemical Sciences, Department of Nuclear Sciences and Applications

Contributors to sample preparation and parcel distribution:

C. Abraham-Ponti, A. Kazmierska, B. Massinger, L. Rován, Z. Sera: Terrestrial Environmental Radiochemistry Laboratory, Division of Physical and Chemical Sciences, Department of Nuclear Sciences and Applications