Effect of NaCl concentration on target-ion intensity and repeatability at m/z ≈ 590.2 (CD = 100 μM)



High NaCl (10 μM): marked signal attenuation and poor repeatability.

Peak window: m/z 590.10–590.32. No normalization or significance testing.