


Filename: 2024-05-14 - 15.28.20.HSRNA




Default image (Contrast 114%)

Sample Info

Well	RIN ^e	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
A1	-	-	2160	Ladder		Caution! Expired ScreenTape device; Ladder
B1	-	-		LCM_3		Marker(s) not detected; Caution! Expired ScreenTape device; The lower ribosomal fragment is missing
C1	-	-		LCM_10		Marker(s) not detected; Caution! Expired ScreenTape device; The lower ribosomal fragment is missing
D1	-	-		LCM_2		Marker(s) not detected; Caution! Expired ScreenTape device; The lower ribosomal fragment is missing
E1	-	-		LCM_6		Marker(s) not detected; Caution! Expired ScreenTape device; The lower ribosomal fragment is missing

A1: Ladder




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
A1	-	-	2160	Ladder		Caution! Expired ScreenTape device; Ladder

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
200	130	-	1910	7.78		
500	476	-	2800	28.55		
1000	360	-	1060	21.58		
2000	390	-	573	23.39		
4000	226	-	166	13.54		
6000	86.1	-	42.2	5.17		

B1: LCM_3




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
B1	-	-		LCM_3		Marker(s) not detected; Caution! Expired ScreenTape device; The lower ribosomal fragment is missing

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	100.00		

C1: LCM_10




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
C1	-	-		LCM_10		Marker(s) not detected; Caution! Expired ScreenTape device; The lower ribosomal fragment is missing

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	100.00		

D1: LCM_2




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
D1	-	-		LCM_2		Marker(s) not detected; Caution! Expired ScreenTape device; The lower ribosomal fragment is missing

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	100.00		

E1: LCM_6



Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
E1	-	-		LCM_6		Marker(s) not detected; Caution! Expired ScreenTape device; The lower ribosomal fragment is missing

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
-	-	-	-	100.00		