



Filename: 2026-09-07 - 15.54.15.HSRNA



Default image (Contrast 50%), Image is Scaled to Sample

EL1: Electronic Ladder




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
EL1	-	-	3750	Electronic Ladder		Ladder

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
200	189	-	2770	5.98		
500	638	-	3760	20.25		
1000	557	-	1640	17.66		
2000	580	-	854	18.42		
4000	634	-	466	20.11		
6000	554	-	272	17.58		

A1: POC_120_C_clean




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
A1	-	-	181	POC_120_C_clean	▲	Sample concentration outside functional range for RINe and the assay


Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
933	32.1	-	101	72.68		18S
1817	12.1	-	19.5	27.32		28S



Region Table


From [nt]	To [nt]	Average Size [nt]	Conc. [pg/ul]	Region Molarity [pmol/l]	% of Total	Region Comment	Color
200	10000	2436	146	176	80.58	dv200	■

B1: POC_120_H_clean**Sample Table**

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
B1	-	-	162	POC_120_H_clean	▲	Sample concentration outside functional range for RINe and the assay


Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
378	17.5	-	136	22.85		
584	25.4	-	128	33.21		
939	24.7	-	77.4	32.35		18S
1808	8.86	-	14.4	11.59		28S

**Region Table**

From [nt]	To [nt]	Average Size [nt]	Conc. [pg/ul]	Region Molarity [pmol/l]	% of Total	Region Comment	Color
200	10000	2243	129	170	80.08	dv200	

C1: POR_120_C_clean




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
C1	-	-	124	POR_120_C_clean		Sample concentration outside functional range for RINe and the assay


Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
279	7.95	-	84.0	9.78		
413	15.2	-	108	18.72		
572	16.2	-	83.1	19.88		
695	12.0	-	50.6	14.71		
900	23.9	-	78.1	29.40		18S
1764	6.11	-	10.2	7.51		



Region Table


From [nt]	To [nt]	Average Size [nt]	Conc. [pg/ul]	Region Molarity [pmol/l]	% of Total	Region Comment	Color
200	10000	1988	96.1	142	77.47	dv200	

D1: POR_120_H_clean**Sample Table**

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
D1	-	-	127	POR_120_H_clean	▲	Sample concentration outside functional range for RINe and the assay


Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
713	12.7	-	52.2	19.74		
885	31.9	-	106	49.78		18S
1731	19.5	-	33.2	30.48		

**Region Table**

From [nt]	To [nt]	Average Size [nt]	Conc. [pg/ul]	Region Molarity [pmol/l]	% of Total	Region Comment	Color
200	10000	2220	103	137	81.36	dv200	■

E1: MON_120_C_clean




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
E1	-	-	214	MON_120_C_clean	▲	Sample concentration outside functional range for RINe and the assay


Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
658	16.7	-	74.4	17.79		
775	24.9	-	94.5	26.61		
1037	35.3	-	100	37.73		18S
1928	16.7	-	25.5	17.88		28S



Region Table


From [nt]	To [nt]	Average Size [nt]	Conc. [pg/ul]	Region Molarity [pmol/l]	% of Total	Region Comment	Color
200	10000	2578	180	205	84.19	dv200	

F1: MON_120_H_clean**Sample Table**

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
F1	4.6	0.9	446	MON_120_H_clean	⚠	Sample concentration outside functional range for RINe and the assay


Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
928	78.8	-	250	52.82		18S
1699	70.4	-	122	47.18		28S edited

**Region Table**

From [nt]	To [nt]	Average Size [nt]	Conc. [pg/ul]	Region Molarity [pmol/l]	% of Total	Region Comment	Color
200	10000	1902	404	624	90.41	dv200	■

G1: POR_3_C_RS_clean




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
G1	-	-	143	POR_3_C_RS_clean	▲	Sample concentration outside functional range for RINe and the assay

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
389	4.63	-	35.0	4.28		
572	10.4	-	53.3	9.59		
866	58.2	-	198	53.81		18S
1328	9.97	-	22.1	9.22		
1679	25.0	-	43.7	23.10		




Region Table

From [nt]	To [nt]	Average Size [nt]	Conc. [pg/ul]	Region Molarity [pmol/l]	% of Total	Region Comment	Color
-----------	---------	-------------------	---------------	--------------------------	------------	----------------	-------

200	10000	1893	124	193	86.98	dv200	
-----	-------	------	-----	-----	-------	-------	---

H1: POR_3_H_RS_clean




Sample Table

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
H1	-	-	62.6	POR_3_H_RS_clean	▲	Sample concentration outside functional range for RINe and the assay

Peak Table

Size [nt]	Calibrated Conc. [pg/ul]	Assigned Conc. [pg/ul]	Peak Molarity [pmol/l]	% Integrated Area	Peak Comment	Observations
25	700	700	82400	-		Lower Marker
706	5.20	-	21.7	23.76		
890	14.0	-	46.4	64.16		18S
1773	2.64	-	4.38	12.08		28S edited



Region Table

From [nt]	To [nt]	Average Size [nt]	Conc. [pg/ul]	Region Molarity [pmol/l]	% of Total	Region Comment	Color
200	10000	2556	42.2	48.6	67.53	dv200	

Sample Info

Well	RINe	28S/18S (Area)	Conc. [pg/ul]	Sample Description	Alert	Observations
EL1	-	-	3750	Electronic Ladder		Ladder
A1	-	-	181	POC_120_C_clean	▲	Sample concentration outside functional range for RINe and the assay
B1	-	-	162	POC_120_H_clean	▲	Sample concentration outside functional range for RINe and the assay
C1	-	-	124	POR_120_C_clean	▲	Sample concentration outside functional range for RINe and the assay
D1	-	-	127	POR_120_H_clean	▲	Sample concentration outside functional range for RINe and the assay
E1	-	-	214	MON_120_C_clean	▲	Sample concentration outside functional range for RINe and the assay
F1	4.6	0.9	446	MON_120_H_clean	⚠	Sample concentration outside functional range for RINe and the assay
G1	-	-	143	POR_3_C_RS_clean	▲	Sample concentration outside functional range for RINe and the assay
H1	-	-	62.6	POR_3_H_RS_clean	▲	Sample concentration outside functional range for RINe and the assay