CleanNGS-R

Catalog Nos. CNGS-R-0,5L
Manual revision v1.00

Contents

Introduction and Principle 2
Kit Contents and Materials 3
CleanNGS-R - 96-well Plate Protocol 4
CleanNGS-R - 384-well Plate Protocol 5
CleanNGS-R – Single Tube Protocol 6
Trouble Shooting Guide 7
Ordering Information 7

For Research Use Only. Not for use in diagnostic procedures.

Information in this document is subject to change without notice.

CLEANNA IS A REGISTERED TRADE MARK OF GC BIOTECH BV.
GC BIOTECH DISCLAIMS ALL WARRANTIES WITH RESPECT TO THIS DOCUMENT, EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO THOSE OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. TO THE FULLEST EXTEND ALLOWED BY LAW, IN NO EVENT SHALL GC BIOTECH BE LIABLE, WHETER CONTRACT, TORT, WARRANTY, OR UNDER ANY STATUTE OR ON ANY OTHER BASIS FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE OR CONSEQUENTAL DAMAGES IN CONNECTION WITH OR ARRISING FROM THIS DOCUMENT, INCLUDING, BUT NOT LIMITED TO THE USE THEREOF, WHETER OR NOT FORSEEABLE AND WHETER OR NOT GC BIOTEH IS ADVISED OF THE POSSIBILITY OF SUCH DAMAGES.

TRADEMARKS
The Trademarks mentioned herein are the property of GC biotech, or their respective owners.

Introduction and Principle

The CleanNGS-R kit is an efficient PCR and Next Gen library prep clean up system based on paramagnetic beads technology, providing an efficient purification of PCR amplicons.
With its simple, three-step protocol, CleanNGS-R removes salts, primers, primer-dimers, dNTPs, while DNA fragments are selectively bound to the magnetic particles; and highly purified DNA is eluted with low salt elution buffer or water and can be used directly for downstream applications. The protocol can be adapted to your current liquid handling workstation (e.g. Beckman, Hamilton, Tecan, Caliper, Perkin Elmer, Agilent and Eppendorf) utilizing your current protocol as well as it can be performed manually.

Features:

- Designed for both DNA purification
- Ideal for (double-sided) size selection for Next-Generation Sequencing
- High recovery of amplicons greater than 100 bp
- Efficiently removes unincorporated dNTPs, primers, primer dimers and other contaminants
- No centrifugation or filtration

Amplicons purified with CleanNGS-R system are ready to be used in the following applications:

- PCR[1)] and RT-PCR
- Mutation detection and Genotyping
- Sequencing (Sanger and Next Generation)
- Fragment Analysis
- Restriction enzyme clean up
- Cloning

[1)] The PCR Process is covered by patents owned by Roche Molecular Systems, Inc., and F. Hoffmann-La Roche, Ltd.

Phone: +31 182 22 33 00 | Web: www.cleanna.com | E-mail: info@cleanna.com

Kit Contents and Materials

Kit Contents:

Product Number	Description	Number of	Storage Conditions
CNGS-R-0,5L	CleanNGS-R – 500 mL	27.777 *	

* Number of reactions is based on a typical 10 µL PCR reaction volume.
Volume of CleanNGS-R to be used per reaction = 1.8x the sample volume.

Materials Supplied in the CleanNGS-R kit:
CleanNGS-R magnetic particle solution

Materials and Equipment to be supplied by User:

- 96-well PCR plate containing PCR samples (up to 50 µL/well)
- Magnetic separation device, recommended Clean Magnet Plate 96-Well RN50 (Part# CMAG-RN50)
- Multichannel pipettor
- Multichannel Disposable Reservoirs
- 96-well microplate for elution
- 70% ethanol (freshly prepared from non-denatured alcohol)
- Reagent grade water, nuclease free or Elution Buffer (10mM TRIS-HCL pH 8.0)

Phone: +31 182 22 33 00 | Web: www.cleanna.com | E-mail: info@cleanna.com

CleanNGS-R - 96-well Plate Protocol

1. Shake the CleanNGS-R reagent thoroughly too fully resuspend the magnetic particles prior to usage.

2. Measure the sample(s) reaction volume in the wells of the 96-well plate. Determine if transferring the sample(s) to a processing plate is required. If necessary, transfer the reactions to a 96-well microplate.

   

   Note: If the reaction volume * 2.8 exceeds the volume of the PCR plate, a transfer to a 300 µl round bottom plate is required.

3. Add 1.8x the reaction volume of CleanNGS-R to each well.

PCR Reaction Volume (µL)	CleanNGS-R (µL)
10	18
20	36
50	90

4. Pipet up and down 5-10 times or vortex for 30 seconds.

5. Incubate at room temperature for 5 minutes.

6. Place the plate on a magnetic separation device to magnetize the CleanNGS-R particles. Incubate at room temperature until the CleanNGS-R is completely cleared from solution.

7. Aspirate and discard the cleared supernatant. Do not disturb the CleanNGS-R particles.

8. Add 200 µL 70% ethanol to each well.

9. Incubate at room temperature for 1 minute. It is not necessary to resuspend the CleanNGS-R particles.

10. Aspirate and discard the cleared supernatant. Do not disturb the CleanNGS-R.

11. Repeat Steps 8-10 for a second 70% ethanol wash step.

12. Leave the plate on the magnetic separation device for 10-15 minutes to air dry the CleanNGS-R particles. Remove any residue liquid with a pipet.

    

    Note: It is important to dry the CleanNGS-R particles before elution. Residual ethanol may interfere with downstream applications.

13. Remove the plate from magnetic separation device.

14. Add 30-40 µL Elution Buffer (not provided) to each well.

15. Pipet up and down 20 times or vortex for 30 seconds.

16. Incubate at room temperature for 2-3 minutes.

17. Place the plate on a magnetic separation device to magnetize the CleanNGS-R particles. Incubate at room temperature until the CleanNGS-R particles are completely cleared from solution.

18. Transfer the cleared supernatant containing purified DNA to a new 96-well microplate and seal with non-permeable sealing film.

19. Store the plate at 2-8°C if storage is only for a few days. For long-term storage, samples should be kept at -20°C.

Phone: +31 182 22 33 00 | Web: www.cleanna.com | E-mail: info@cleanna.com

CleanNGS-R - 384-well Plate Protocol

1. Shake the CleanNGS-R reagent thoroughly too fully resuspend the magnetic particles prior to usage.

2. Place the 384-well PCR plate on the bench and measure the volume of the reaction. Transfer the sample to a skirted 384-well PCR plate.

3. Add 1.8x the sample volume of CleanNGS-R reagent to each well.

PCR Reaction Volume (µL)	CleanNGS-R (µL)
5	9
7	12,6
10	18

4. Pipet up and down 5-10 times or vortex for 30 seconds.

5. Incubate at room temperature for 5 minutes.

6. Place the plate on a magnetic separation device to magnetize the CleanNGS-R particles. Incubate at room temperature until the CleanNGS-R particles are completely cleared from solution.

7. Aspirate and discard the cleared supernatant. Do not disturb the CleanNGS-R particles.

8. Add 30 µL 70% ethanol to each well.

9. Incubate at room temperature for 1 minute. It is not necessary to resuspend the CleanNGS-R particles.

10. Aspirate and discard the cleared supernatant. Do not disturb the CleanNGS-R particles.

11. Repeat Steps 8-10 for a second 70% ethanol wash step.

12. Leave the plate on the magnetic separation device for 10-15 minutes to air dry the CleanNGS-R particles. Remove any residue liquid with a pipet.



Note: It is important to dry the CleanNGS-R particles before elution. Residual ethanol may interfere with downstream applications.

13. Remove the plate from magnetic separation device.

14. Add 20 µL Elution Buffer (not provided) to each well.

15. Pipet up and down 20 times or vortex for 30 seconds.

16. Incubate at room temperature for 2-3 minutes.

17. Place the plate on a magnetic separation device to magnetize the CleanNGS-R. Incubate at room temperature until the CleanNGS-R is completely cleared from solution.

18. Transfer the cleared supernatant containing purified DNA to a new 384-well microplate and seal with non-permeable sealing film.

19. Store the plate at 2-8°C if storage is only for a few days. For long-term storage, samples should be kept at -20°C.

Phone: +31 182 22 33 00 | Web: www.cleanna.com | E-mail: info@cleanna.com

CleanNGS-R – Single Tube Protocol

1. Shake the CleanNGS-R reagent thoroughly too fully resuspend the magnetic particles prior to usage.

2. Measure the volume of the PCR reaction and transfer the sample to a single tube, for example an 1,5 mL single tube.

3. Add 1.8x the sample volume of CleanNGS-R reagent to each well.

PCR Reaction Volume (µL)	CleanNGS-R (µL)
50	90
100	180
150	270

4. Pipet up and down 5-10 times or vortex for 30 seconds.

5. Incubate at room temperature for 5 minutes.

6. Place the tube in the magnetic separation stand to magnetize the CleanNGS-R particles. Incubate at room temperature until the CleanNGS-R particles are completely cleared from solution.

7. Aspirate and discard the cleared supernatant. Do not disturb the CleanNGS-R particles.

8. Add 500 - 1000 µL 70% ethanol to each well.

9. Incubate at room temperature for 1 minute. It is not necessary to resuspend the CleanNGS-R particles.

10. Aspirate and discard the cleared supernatant. Do not disturb the CleanNGS-R particles.

11. Repeat Steps 8-10 for a total of three 70% ethanol wash step.

12. Leave the tube in the magnetic separation stand for 10-15 minutes to air dry the CleanNGS-R particles. Remove any residue liquid with a pipet.



Note: It is important to dry the CleanNGS-R particles before elution. Residual ethanol may interfere with downstream applications.

13. Remove the tube from the magnetic separation stand.

14. Add a minimum of 30 µL nuclease free water or Elution Buffer (not provided) to each well.

15. Pipet up and down 20 times or vortex for 30 seconds.

16. Incubate at room temperature for 2-3 minutes.

17. Place the tube in the magnetic separation stand to magnetize the CleanNGS-R. Incubate at room temperature until the CleanNGS-R is completely cleared from solution.

18. Transfer the cleared supernatant containing purified DNA to a new tube or microplate and seal with non-permeable sealing film.

19. Store the plate at 2-8°C if storage is only for a few days. For long-term storage, samples should be kept at -20°C.

Phone: +31 182 22 33 00 | Web: www.cleanna.com | E-mail: info@cleanna.com

Trouble Shooting Guide

Please use this guide to troubleshoot any problems that may arise. For further assistance, please contact your local distributor.

Possible Problems and Suggestions

Problem	Cause	Solution
Low yield	Low yield	Increase the number amplification cycles for PCR
	Smaller product size	Small DNA fragments normally give lower yield.
	Ethanol residue	During the drying step, remove any liquid from bottom of the well
	Particle loss during the procedure	Increase magnetization time. Aspirate slowly.
	DNA remains bound to beads	Increase elution volume
	Incomplete resuspension of the particles during elution	Vortex or pipet up and down to fully resuspend the particles.
Primer carryover	Insufficient wash of the particles	Wash the beads one more time with 70% ethanol.
Non-specific amplification products were not removed	The size of the non-specific amplification products are larger than 100 bp	Non-specific amplification products larger than 100 bp are not efficiently removed from PCR products.
Problems in downstream applications	Salt carryover	70% ethanol must be stored at room temperature.
	Ethanol carryover	Ensure the beads are completely dried before elution.

Ordering Information

Contact your local distributor to order.

Product	Part Number
CleanNGS-R (500 mL)	CNGS-R-0,5L

Product	Part Number
Clean Magnet Plate 96-Well RN50	CMAG-RN50

Phone: +31 182 22 33 00 | Web: www.cleanna.com | E-mail: info@cleanna.com



CleanNA is a registered trade mark of GC biotech BV