Orthopedics Power System_Oscillating Saw
(XY-G02)

I. Overview

The power system adopts rechargeable battery power supply. The whole machine can be sterilized with high temperature and pressure (except battery). The clinical use of space is not limited by cables. It is easy to carry with 14.4V battery. It works with strong power, fast charging, manual stepless speed regulation, built-in imported motor, high speed, low noise and stable performance.

This combination is suitable for general orthopedics. The oscillating saw can perform a variety of procedures such as joint replacement without the limitation of the patient's body shape and position.

II. Technical Parameters

1	Battery voltage	12V
2	Charger input voltage	220V 50Hz
3	Output power	200W
4	Torque	1000g. cm
5	Hardness	HRC58
6	Saw swing	0～20000 times/min
7	Noise is less than	75dba

Safety performance indicators comply with the *"GB9706.1 Medical Electrical Equipment Part I, General Safety Requirements"* .

III. Operating Environment

Meets the requirements of the national standard GB/T14710

- Environmental temperature 5°C～40°C
- Relative humidity ≤80%

IV. Instructions

1. Before use, the working head and tools etc. should be sterilized together with the sterilized cloth under high temperature and pressure. (except battery)

2. The saw blade has a quick mounting interface. Press the interface of the hand-held motor handle downwards to expose the saw blade interface, then insert the saw blade to the blade interface and insert it. After inserting it, release the hand. When the blade is removed, press down on the saw blade interface and pull out the piece.

3. When operating, hold it steady, use even force, and gradually press the button on the hand-held motor handle to gradually increase the speed from slow to fast. Familiarize yourself with the performance and simulate operation before using it.

V. Cautions

1. Before use, check whether all parts are normal, whether there is noise or abnormal vibration, and whether the rotating part is smooth.

2. Check if the work head is sharp enough and the installation is firm and proper.

3. Check if the battery is fully charged and the charge does not exceed two hours per charge. Unplug the plug immediately after the indicator light goes out.

4. Keep the handheld motor handle and the battery from water splash, the battery can not be sterilized under high temperature and pressure.

5. The system should be operated by trained physicians to prevent injury.

VI. Maintenance

1. After use, the blood should be cleaned and wiped, oiled and maintained in time to avoid damage.

2. The machine should be stored in a cool, dry and ventilated place.

3. The machine must avoid strong vibrations, collisions, and rough handling during transportation.

4. Contact the manufacturer or maintenance personnel.that authorized by the manufacturer in case of malfunction.

VII. Packing List

1	motor handle	x1
2	rechargeable battery	x1
3	charger	x1
4	saw blade	x2
5	disinfection channel	x1
6	instruction manual	x1

VIII. Quality Track Card

The entire system has one year warranty and life-long maintenance guarantee. Please fill in this card carefully and cut it along the dotted line to send it back or fax it to our factory so that we can enter your data into the customer database to better provide you after-sales service.

Please cut along the dotted line and send it back

Product Number：No　　Product Name：

Registration Number：　　Model：

Factory time：　　QC Number：

Manufacturer：Tianjin Jindong Xiyi Medical Instruments Factory

Add.：Xinmao Industrial Park C2-1-D, Rongyuan Road 16, Tianjin 300384, PR China

Tel.：+86 (0)22-83710960　　Fax.：+86 (0)22-83710950

Customer：　　Contact Person：

Add.：　　Postal Code：

Tel.：　　Purchase Time：

Distributor：　　Purchase Location：

Signature：