Please read the manual before using

OWNER'S MANUAL

Shoulder Joint CPM

Type **YTK-E**

This publication is printed

On 100% recycled paper

The sketch map of shoulder joint CPM

Technique feature of product：

The Shoulder joint CPM is the newly-designed medical equipment product, the main technique feature of product are as follows:

1. Special designed motor drive, run steadily, low noise, and long span life.
2. Microcomputer digital control design and it also adopt imported integrated parts, high volume imported CPU and intelligent program design, the machine also has the function of over-load protection.
3. Big screen liquid display, all the parameters can be read clearly.
4. Considering the need of medical curing, the motion range is wide.
5. Optimized structure design.

Main structure

This machine is made up of host computer, supporting pole, leather cushion, hand controller, and so on.

The main capability parameters are as follows:

Angle motion range of joint：0°—150°；

Angle motion speed range： 0.5°—3.5°/s；

Power supply： 220Vac， 50Hz

Power： ≤60W

The adaptively of unit

This unit is used for the convalescence of Shoulder joint fracture.

This unit can adjust the motion angle of joint, it can promote the convalescence of Shoulder joint.

Operation method

1. Install the main body head according to patient's need. Adjust the "height adjusting wheel". Attention that be more careful for loosening the "movable fasten bolt of main body" when moving the main body head.

2. Turn on the power, open the power switch, then the machine runs to the initial position. The display shows "YATAI", and it stays in setting state, the right-hand corner shows"SETUP"。

3. Install the arm placing part (at the initial position) according to patient's need. Don't install the front arm shelf parts!

4.Set **Extension angle**、**Flex angle**、**Speed**、**Time** according to the s↓ of p↑nt. press the correspond key, it shows the parameter on the machine（the character shows as"EXT"、"FLEX"、"S"、"T"）the label"."is blinking. And you can press or to adjust the parameter.

5. Confirm the parameter on the display is correct. you can press **start/pause** to start the machine, it displays“RUN” and immediate motion angle, watching the machine running for more than two times, if it runs in normal state, press the **start/pause** to pause it, then fix the patient’s arm and palm.

After the above operation，the machine can be started，it is suggested that run in low speed and range is good to patient, then increase the speed and motion range if the patient can hold.

6. The main function of hand controller is for safe protection. One key can stop machine, and the other key is extension key, press any of them can protect the patient from machine malfunction.

Attention item:

a. For safety, install the arm placing part two men together; notice that protecting the balance of the main body head.

b. It is a little difficult to loosening the fixing bolt, turn a little of the main body head can loosen it easily.

c. Fasten the bolt when curing.

d. When the time is end, it is strictly forbidden for patient ending the curing and standing up, avoid the unbalance of machine.

Simple failure elimination and maintenance

If the machine can’t run, first check the power supply, check the power switch light is normal or not, then check the screen display. If you are sure the power supply is off, then check inner fuse of power outlet. If it’s fine, but there is no power supply, please contact with us.

If the machine has power, but it has no display, maybe the inner connection has some problem. This is small failure, but it is suggested that you contact with us.

If the screen can display, but the key on the screen can’t work, it should be the problem of printing line of control panel. If you can’t eliminate the problem, please contact with us.

The shell of the unit should be cleaned using dry detergent and soft cloth.

Commitment for after service

If the product has any quality problem in one year, we provide the service of change and repair, and we also provide long-term repairing service. And only charge the material expenses in guarantee period.

ZHENGDA Medical Instrument Co.,Ltd.of Hangzhou

Producing the article protects to fix single

Ltd. model:

_____

Number of the machine：

_____

Purchasing date：

_____

Invoice number：

_____

User's name: _____

User's address _____

Telephone number _____

Maintain the recond						
Date	The reason of Maintain	The content of Maintain	The cost of Maintain	Maintain signs	User's signs	Remark

Precautions

1. The free guarantee period of products is half a year from the day when purchasing
2. If the products that you bought have any doubtful point while using, we provide guidance and service.
3. If that's users cause the products' problem or it exceeds guarantee period, our company, we provide repairing service, and only charge material expenses
4. Must return this card while maintaining, so please keep this singly.