五功能高级电动床规格表

Specification List for High-level Five-function Electric Bed

床面尺寸 Bed Base Size	
● 床面长度 Length	2220mm
● 床面宽度 Width	1050mm
● 床面离地高 hi-Lo from floor	450-790mm
升降功能 Technical Date	
● 床面上升高度 Vertical travel	340mm
● 背部上升角度 Back rest upward	0°-75°
● 脚部上升角度 Foot rest upward	0°-45°
● 床体倾斜角度 Tilt headward and footward respectively	±10°

Shanghai Rongshun Medical Techology Co.,Ltd

地址：上海市宝山区罗店镇苗圃路607号
电话：021-56309894　56636868
传真：021-56728791

Add: NO.607,MiaoPu Road,LuoDian Town,BaoShan District,ShangHai,201908 China.
Tel:0086-021-56309894　0086-021-56636868
Fax:0086-021-56728791

http://www.shrongshun.cn
E-mail:shrs@shrongshun.cn

产品使用说明书

Instruction Manual

上海荣顺医疗科技有限公司

高级电动床

High-level Electric Bed

Hospital Bed 医用病床系列

护栏

○ 两段内缩式PP护栏，可分别升降便于患者上下床。

○ PP材质制作，外表滑顺触感温润，擦拭容易。

Side Rail

○ 2-Section collapsible PP side rails, which can be Raised and Lowered resepectively. It's very conveniently for the patient to get on or get off the bed.

○ PP material . It's smooth, tender in out-side profile, and easy to clean.

PP塑料床头尾板

PP材质制作，钢制夹紧后，床头尾板不会晃动及震动异声的情形发生。方便推送，急救医疗进行时可快速取下。

PP Head and Foot Board

PP material, Head and foot board will never vibrate and noise after the safe steel clamp tightly to make push the bed conveniently. Detachable head and foot board can be quickly removed for emergenly position.

背板CPR拉把机构

配合医疗急救的需要，床体两侧装设CPR拉把机构，床背板于任何角度均可快速放平，以利CPR的进行。

CPR Control Handle

CPR Control Handle Located under both sides of head section to meet medical emergency use. Continually pull and hold the handle, hold the handle until the head section come to a stop in the flat postion .

输液架
Drip stand

升降控制器
Side Rail Control Panel

钢制冲压床面板
Pressed Steel Bed Base

中控脚轮
Centrally Cotrolled

中控刹车脚踏板

Foot centrally controlled pedal

床脚装设两个脚踏板，可同时刹住四个脚轮

Two centrally controlled foot tradele. While can brake four castors at the same.

○ 红色 Red (刹车) Brake

○ 蓝色 Blue (启动) Release Brake

控制器位于两侧护栏内，具有床面、背部、脚部和倾斜的升降图形，识别容易且易于操作。

Side Rail Control Panel located inside and outside of head end side rails, which is easy to recognised and operated.

护士控制器

控制器位于床尾，具有床面、背部、腿部和倾斜的升降图形，并有锁定按键功能，识别容易、易于操作。

Controller for Nurse

Controller located foot end. There are function indication figures for bed base, back and foot section and some function lock button, which is easy to recognised and operated.



升降控制器
Side Rail Control Panel

1、背部上升按钮 Back up
2、背部下降按钮 Back down
3、腿部上升按钮 Leg up
4、腿部下降按钮 Leg down
5、背部和腿部同时上升按钮 Back and leg simuteneously rise
6、背部和腿部同时下降按钮 Back and leg simuteneously down
7、倾斜控制按钮 Tilt headward and footward
8、整体上升按钮 Bed rise
9、整体下降按钮 Bed down
10、护栏锁定按钮 Automatic side rail lock

护士控制器
Control for Nurse

12、电源开关按钮 Power on
13、背部上升按钮 Back up
14、背部下降按钮 Back down
15、背部锁定按钮 Back function lock
16、腿部锁定按钮 Leg function lock
17、腿部上升按钮 Leg up
18、腿部下降按钮 Leg down
19、升降、倾斜锁定按钮 Lock for vitical tralling and titl function
20、整体上升、下降按钮 Vitical tralling
21、倾斜控制按钮 Tilt headward and footward

Hospital Bed 医用病床系列

病床的使用及注意事项：

- 病床最大静态载重不得大于250KG。
- 操作部分的使用（由医生、护士操作）。
- 因医疗需要卸床头尾板时人应站稳重心，打开床头保险，两手分别握在床架主管两端，用力向上拔即可。
- 病人及家属不能利用床的任何部件做各种锻炼恢复活动，以免意外事故发生。
- 病床的室内活动：需要先松开脚轮刹车，方可推动，但不能在高低不平的地面上强行推行。到目标位置后踩脚轮的刹车（使床固定）
- 病床不能使用含氯消毒剂进行消毒。
- 病床如有异常情况，请致电本公司，公司负责修配服务，超过保修期，则适当收费。

Using instruction and notes:

- Safe working load-Maximum 250KG
- All controlled parts must be operated by medical personnel.
- When need to remove head and foot boards to meet medical requirement,Medical personnel should stand well,open the clamp,Clutch both side ends of head and foot boards and pull it vertically.
- Patient and family numbers are not allowded to use any parts of bed to take exercise,get well exercise,or personal injury or equipment damage could occur.
- Before moving the bed indoor,have to release the brake.When reach the desired postition,set the brake.It's not allowered to move the bed in force on uneven floor.
- Do not use chlorine disinfeetor to disinfect hospital bed.
- If there are some problems existing,please call for company for assistance. The company will offer mainantenance service,Out of guarantee will charge on the circumstance.