








lys_instr: A Python Package for Automating Scientific Measurements

Ziqian Wang^{1,2}[¶], Hidenori Tsuji², Toshiya Shiratori³³, and Asuka Nakamura^{2,3}

¹ Research Institute for Quantum and Chemical Innovation, Institutes of Innovation for Future Society, Nagoya University, Japan^{ROR} ² RIKEN Center for Emergent Matter Science, Japan^{ROR} ³ Department of Applied Physics, The University of Tokyo, Japan^{ROR} [¶] Corresponding author

DOI: [10.xxxxxx/draft](https://doi.org/10.xxxxxx/draft)

Software

- [Review](#) 
- [Repository](#) 
- [Archive](#) 

Editor: [Open Journals](#) 

Reviewers:

- [@openjournals](#)

Submitted: 01 January 1970

Published: unpublished

License

Authors of papers retain copyright and release the work under a Creative Commons Attribution 4.0 International License ([CC BY 4.0](#)).

Summary

Modern experiments increasingly demand automation frameworks that coordinate diverse scientific instruments while remaining flexible and customizable. Existing solutions, however, often require explicit management of low-level communication and concurrency, resulting in substantial development overhead. We present `lys_instr`, a Python package that addresses these challenges through an object-oriented, multi-layered architecture for instrument control, workflow coordination, and GUI construction. It enables researchers to rapidly build responsive, asynchronous measurement systems with minimal coding effort. Seamlessly integrated with the `lys` platform (Nakamura, 2023), `lys_instr` unifies experiment control, data acquisition, and visualization, providing an efficient foundation for next-generation automation-driven experimental research.

Statement of need

Modern scientific research increasingly relies on comprehensive measurements across wide parameter spaces to elucidate physical phenomena. As experiments grow in complexity—with longer measurement times and greater instrument diversity—efficient automation has become essential. Measurement automation is evolving beyond simple parameter scans toward informatics-driven, condition-based optimization, paving the way for AI-assisted experimental workflow management. This shift demands robust software infrastructure with high integrability and flexible logic control.

However, building such a measurement system remains time-consuming for researchers. At the low level, instruments are often tightly coupled to diverse communication protocols (e.g., TCP/IP, VISA, serial), limiting interchangeability and cross-system flexibility. At the high level, coordinating workflows that combine conditional logic, iterative processes, and advanced algorithms across multiple libraries frequently leads to redundant implementations, reducing development efficiency. For example, capturing temperature-dependent images with a camera or acquiring temperature-dependent spectra are routine tasks. Although the underlying workflow—iterative parameter adjustment followed by data acquisition—is conceptually identical, such logic is often reimplemented independently across experiments. Moreover, implementing graphical user interfaces (GUIs) for these low- and high-level functionalities typically involves complex multithreading, requiring familiarity with GUI frameworks and operating system (OS) event-handling mechanisms. These challenges impose substantial development overhead and highlight the need for a control framework that balances architectural flexibility with reduced implementation complexity.

State of the field

In principle, the challenges in measurement system development described above can be addressed by frameworks that adopt well-known object-oriented GoF design patterns (Gamma et al., 1994). Encapsulating low-level communication protocols behind standardized interfaces enables the development of high-level workflows that are independent of specific instruments, while still allowing device-specific customization through inheritance. Such loose coupling improves reusability of high-level logic across different experimental setups. Furthermore, the *Template Method* design pattern allows complex multithreading-related functionalities to be implemented within superclass definitions, enabling users to write measurement logic without explicitly handling thread management. Reusable GUI components can likewise be constructed on top of these abstract interfaces, significantly reducing implementation effort. This design philosophy also enhances stability, as most components can be developed and tested independently of physical hardware.

However, existing software platforms do not explicitly adopt this interface-centered design philosophy. Commercial platforms such as LabVIEW (National Instruments Corporation, 2024) and MATLAB's Instrument Control Toolbox (The MathWorks, Inc., 2024) provide mature environments for instrument communication, workflow execution, and GUI development. Python-based frameworks including QCoDeS (Nielsen et al., 2025), PyMeasure (PyMeasure Developers, 2024), and PyOpticon (Randall & Majumdar, 2025) likewise offer instrument drivers, experiment routines, and graphical components. While these tools provide powerful capabilities, flexible workflow orchestration—particularly for conditional logic, nested procedures, and multithreaded execution—often requires substantial user-defined implementation. In many cases, these platforms function either as general-purpose programming environments or as collections of concrete drivers and predefined workflows, rather than as a framework that defines a unified set of abstract interfaces for measurement systems. Therefore, development of an interface-driven architecture grounded in object-oriented design patterns is essential for building reusable, low-code, flexible, and stable measurement systems.

Software design

To address this gap, `lys_instr` introduces a layered architecture that spans low-level instrument interfaces to high-level workflow and GUI integration. Most importantly, it defines a unified set of abstract interfaces in the lowest layer that capture the common functionalities shared by many scientific instruments. This design enables the reuse of high-level GUIs and workflow logic, ranging from simple parameter sweeps to informatics-driven adaptive experiments, as discussed above. `lys_instr` adopts a three-layer architecture organized by functional separation (Figure 1). Each layer applies object-oriented design patterns described by Gamma et al. (Gamma et al., 1994) according to its responsibilities, thereby enhancing flexibility, modularity, and usability.




Figure 1: Schematic of the code architecture of `lys_instr`.

1. Base Layer: Device Controller Abstraction

This layer defines abstract interfaces that standardize core instrument controllers. The interfaces encapsulate the concrete implementations, following the *Template Method* design pattern. Typically, most measurement systems include two types of components: *controllers*, which adjust experimental parameters such as external fields, temperature, or physical positions, and *detectors*, which record experimental data, e.g., cameras, spectrometers. Accordingly, `lys_instr` provides standardized *controller* and *detector* interfaces that unify instrument behavior, allowing higher layers to operate on different devices uniformly through common interfaces. Users only need to provide device-specific subclasses that inherit from these interfaces to handle communication with their respective hardware devices. Moreover, each interface manages its own thread(s), ensuring responsiveness and asynchronous operation without blocking other controllers or the GUIs in higher layers. This structure enables users to create controller objects that can be readily integrated into higher-level workflows with minimal device-specific coding.

2. Top Layer: Workflow Coordination

This layer implements workflows common across many setups. Most measurements share similar procedural structures, such as a *scan* process in which data are sequentially recorded while parameters like fields, temperature, or positions are varied. These workflows are standardized using the abstract interfaces defined in the Base Layer, independent of any specific hardware devices, following the *Bridge* and *Composite* design patterns. For example, `lys_instr` provides a standardized *scan* routine that calls *controller* and *detector* interface methods without requiring knowledge of the underlying concrete implementations. This abstraction allows such workflows to be reused across different hardware configurations, greatly improving coding efficiency. In addition, `lys_instr` includes prebuilt GUI components corresponding to each Base Layer component, enabling direct GUI-based control through the same abstract methods. This design cleanly separates workflow logic from device-specific details, simplifying extension to complex measurement systems. Moreover, the GUI communicates with Base Layer interfaces via event-driven messaging, following the *Observer* design pattern to ensure low coupling and high extensibility. With this layer, users can design measurement workflows from scratch without manually creating GUI components.

3. Connection Layer: Control-System Assembly

This layer enables flexible assembly of components from the Base and Top Layers into a complete control system by managing connections within and across layers. Following the *Mediator* design pattern, it connects abstract Base Layer interfaces (and the corresponding hardware devices) to enable automatic data flow, and links GUI components to their respective interfaces, fully hiding device-specific implementations from this layer and above. It also organizes the GUI components into a cohesive application for user interaction. This design grants users maximum freedom to construct tailored control systems without handling low-level tasks such as inter-device communication or multi-threading. Several prebuilt GUI templates for common scenarios are provided for quick hands-on use.

Overall, `lys_instr` provides prebuilt support for standard device controllers, common experimental workflows, and GUI components and assemblies, so users generally need to implement only device-specific subclasses to handle communication with their hardware. This enables rapid integration of new instruments into automated measurement workflows with minimal coding and design effort. A potential limitation of this architecture is that highly unconventional or non-standard measurement workflows may require customization beyond the predefined abstractions. However, the layered interface design covers the vast majority of multi-parameter experimental scenarios encountered in typical laboratory environments.

Example of constructed GUI

With `lys_instr`, users can easily construct a GUI like the one shown in Figure 2. In this example, the `lys_instr` window is embedded in the `lys` platform, with Sector A for storage, Sector B for detector, and Sector C for controllers. Multi-dimensional, nested scan sequences can be defined via the visual interface in the Scan tab in Sector C. `lys` tools in the outer window tabs allow customization of data display, enabling advanced, on-the-fly customization of data visualization.




Figure 2: Example GUI of `lys_instr`. The main window, embedded in the `lys` window, contains three sectors: Storage panel (A), Detector panel (B), and controller panel (C). The Scan tab in (C) enables dynamic configuration of multi-dimensional, nested experimental workflows.

Research impact statement

lys_instr has been deployed in complex, real-world scientific experiments and has supported multiple peer-reviewed publications. It automates ultrafast electron diffraction (UED) and ultrafast transmission electron microscopy (UTEM) systems, coordinating ultrafast laser excitation and pulsed electron beam detection in pump-probe experiments (Koga et al., 2024; Nakamura et al., 2020, 2021, 2022, 2023; Shimojima et al., 2021, 2023a, 2023b). It enables precise control of electromagnetic lenses and electron deflectors for advanced microscopy involving electron-beam precession, a capability that would be difficult to implement without lys_instr (Hayashi et al., 2026; Shiratori et al., 2024).

The software has demonstrated seamless control of transmission electron microscopes from multiple manufacturers across different institutes, including RIKEN Center for Emergent Matter Science and Nagoya University, illustrating reproducible performance and hardware-independent workflow management. Through integration with sister packages in the lys family, including lys_em (lys_em Developers, 2026) and lys_fem (lys_fem Developers, 2026), lys_instr supports complex multi-instrument automation within a research-driven ecosystem, enabling efficient deployment of advanced workflows while preserving modularity and extensibility.

AI usage disclosure

Generative AI tools were used to provide debugging suggestions during the final stages of software development. All code was implemented, reviewed, and verified on real hardware, with functionality confirmed through unit tests and experimental validation.

Acknowledgements

We acknowledge valuable comments from Takahiro Shimojima and Kyoko Ishizaka. This work was partially supported by Grant-in-Aid for Scientific Research (KAKENHI) Grants No. 21K13889 and No. 25K00057, and JST PRESTO Grant No. JPMJPR24JA.

References

- Gamma, E., Helm, R., Johnson, R., & Vlissides, J. (1994). *Design patterns: Elements of reusable object-oriented software*. Addison-Wesley. ISBN: 978-0201633610
- Hayashi, S., Han, D., Tsuji, H., Ishizaka, K., & Nakamura, A. (2026). Development of precession lorentz transmission electron microscopy. *Ultramicroscopy*, 280, 114276. <https://doi.org/10.1016/j.ultramic.2025.114276>
- Koga, J., Chiashi, Y., Nakamura, A., Akiba, T., Takahashi, H., Shimojima, T., Ishiwata, S., & Ishizaka, K. (2024). Unusual photoinduced crystal structure dynamics in TaTe₂ with double zigzag chain superstructure. *Applied Physics Express*, 17(4), 042007. <https://doi.org/10.35848/1882-0786/ad3b61>
- lys_em Developers. (2026). *Lys_em: Python package for electron microscopy control and data analysis*. https://github.com/a-tock/lys_em
- lys_fem Developers. (2026). *Lys_fem: Python package for finite-element modeling and experiment integration*. https://github.com/lys-devel/lys_fem
- Nakamura, A. (2023). lys: Interactive Multi-Dimensional Data Analysis and Visualization Platform. *Journal of Open Source Software*, 8(92), 5869. <https://doi.org/10.21105/joss.05869>

- 174 Nakamura, A., Shimojima, T., Chiashi, Y., Kamitani, M., Sakai, H., Ishiwata, S., Li, H., &
175 Ishizaka, K. (2020). Nanoscale Imaging of Unusual Photoacoustic Waves in Thin Flake
176 VTe₂. *Nano Letters*, 20(7), 4932–4938. <https://doi.org/10.1021/acs.nanolett.0c01006>
- 177 Nakamura, A., Shimojima, T., & Ishizaka, K. (2021). Finite-element Simulation of
178 Photoinduced Strain Dynamics in Silicon Thin Plates. *Structural Dynamics*, 8(2), 024103.
179 <https://doi.org/10.1063/4.0000059>
- 180 Nakamura, A., Shimojima, T., & Ishizaka, K. (2022). Visualizing Optically-Induced Strains
181 by Five-Dimensional Ultrafast Electron Microscopy. *Faraday Discussions*, 237, 27–39.
182 <https://doi.org/10.1039/D2FD00062H>
- 183 Nakamura, A., Shimojima, T., & Ishizaka, K. (2023). Characterizing an Optically Induced
184 Sub-micrometer Gigahertz Acoustic Wave in a Silicon Thin Plate. *Nano Letters*, 23(7),
185 2490–2495. <https://doi.org/10.1021/acs.nanolett.2c03938>
- 186 National Instruments Corporation. (2024). *LabVIEW*. National Instruments Corporation.
187 <https://www.ni.com/en-us/shop/labview.html>
- 188 Nielsen, J. H., Astafev, M., Nielsen, W. H. P., & others. (2025). *QCoDeS: Modular data*
189 *acquisition framework* (Version v0.54.2). <https://doi.org/10.5281/zenodo.17459861>
- 190 PyMeasure Developers. (2024). *PyMeasure: Scientific measurement library for instruments,*
191 *experiments, and live-plotting*. (Version 0.14.0). Zenodo. <https://doi.org/10.5281/zenodo.11241567>
- 192
- 193 Randall, R., & Majumdar, A. (2025). PyOpticon: An open-source python package for laboratory
194 control, automation, and visualization. *Chemistry of Materials*, 37(13), 4585–4592. <https://doi.org/10.1021/acs.chemmater.5c00644>
- 195
- 196 Shimojima, T., Nakamura, A., & Ishizaka, K. (2023a). Development of Five-Dimensional
197 Scanning Transmission Electron Microscopy. *Review of Scientific Instruments*, 94(2),
198 023705. <https://doi.org/10.1063/5.0106517>
- 199 Shimojima, T., Nakamura, A., & Ishizaka, K. (2023b). Development and Applications
200 of Ultrafast Transmission Electron Microscopy. *Microscopy*, 72(4), 287–298. <https://doi.org/10.1093/jmicro/dfad021>
- 201
- 202 Shimojima, T., Nakamura, A., Yu, X., Karube, K., Taguchi, Y., Tokura, Y., & Ishizaka, K.
203 (2021). Nano-to-Micro Spatiotemporal Imaging of Magnetic Skyrmion's Life Cycle. *Science*
204 *Advances*, 7(25), eabg1322. <https://doi.org/10.1126/sciadv.abg1322>
- 205 Shiratori, T., Koga, J., Shimojima, T., Ishizaka, K., & Nakamura, A. (2024). Development of
206 ultrafast four-dimensional precession electron diffraction. *Ultramicroscopy*, 267, 114064.
207 <https://doi.org/10.1016/j.ultramic.2024.114064>
- 208 The MathWorks, Inc. (2024). *MATLAB instrument control toolbox*. The MathWorks, Inc.
209 <https://www.mathworks.com/products/instrument.html>